

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM011121\
 Data File : BM028368.D
 Acq On : 11 Jan 2021 12:46
 Operator : CG/JU
 Sample : SSTD3.204
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.2004

Quant Time: Jan 11 13:16:24 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM011121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 11 11:35:38 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.13	152	543	0.40	ng/ul	0.00
4) Naphthalene-d8	10.95	136	2111	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.76	164	1146	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.48	188	2192	0.40	ng/ul	0.00
17) Chrysene-d12	21.60	240	1808	0.40	ng/ul	0.00
23) Perylene-d12	23.92	264	1386	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.37	96	1842	2.99	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.53	152	9912	3.07	ng/ul	0.00
18) Fluoranthene-d10	19.48	212	17894	3.34	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.40	88	1915	2.787	ng/ul#	90
5) Naphthalene	11.00	128	18980	3.185	ng/ul	94
7) 2-Methylnaphthalene	12.60	142	12635	3.214	ng/ul	99
8) 1-Methylnaphthalene	12.82	142	11975	3.190	ng/ul	99
10) Acenaphthylene	14.48	152	16725	3.613	ng/ul#	93
11) Acenaphthene	14.82	153	13429	3.302	ng/ul	97
12) Fluorene	15.80	166	15182	3.284	ng/ul#	98
14) Pentachlorophenol	17.13	266	1923	3.457	ng/ul	97
15) Phenanthrene	17.52	178	22646	3.367	ng/ul	94
16) Anthracene	17.61	178	21585	3.658	ng/ul	93
19) Fluoranthene	19.51	202	25039	3.523	ng/ul#	94
20) Pyrene	19.87	202	24490	3.452	ng/ul	95
21) Benzo(a)anthracene	21.58	228	19756	3.314	ng/ul	95
22) Chrysene	21.63	228	20414	3.180	ng/ul	96
24) Benzo(b)fluoranthene	23.22	252	18719	3.358	ng/ul	85
25) Benzo(k)fluoranthene	23.26	252	18315	3.419	ng/ul#	83
26) Benzo(a)pyrene	23.82	252	15883	3.448	ng/ul#	79
27) Indeno(1,2,3-cd)pyrene	26.30	276	19231	3.441	ng/ul#	98
28) Dibenzo(a,h)anthracene	26.30	278	15990	3.414	ng/ul#	85
29) Benzo(g,h,i)perylene	27.03	276	16444	3.376	ng/ul	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM011121\
Data File : BM028368.D
Acq On : 11 Jan 2021 12:46
Operator : CG/JU
Sample : SSTD3.204
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD3.2004

Quant Time: Jan 11 13:16:24 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM011121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 11 11:35:38 2021
Response via : Initial Calibration

