

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM011121\
 Data File : BM028367.D
 Acq On : 11 Jan 2021 12:09
 Operator : CG/JU
 Sample : SSTD1.603
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD1.6003

Quant Time: Jan 11 12:40:50 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM011121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 11 11:35:38 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.13	152	535	0.40	ng/ul	0.00
4) Naphthalene-d8	10.95	136	2050	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.76	164	1078	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.48	188	2232	0.40	ng/ul	0.00
17) Chrysene-d12	21.60	240	1719	0.40	ng/ul	0.00
23) Perylene-d12	23.92	264	1380	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.37	96	899	1.48	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.52	152	4637	1.48	ng/ul	0.00
18) Fluoranthene-d10	19.48	212	7733	1.52	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.41	88	947	1.399	ng/ul#	91
5) Naphthalene	11.00	128	8987	1.553	ng/ul	95
7) 2-Methylnaphthalene	12.60	142	5881	1.540	ng/ul	99
8) 1-Methylnaphthalene	12.82	142	5612	1.540	ng/ul	100
10) Acenaphthylene	14.48	152	7440	1.708	ng/ul	94
11) Acenaphthene	14.82	153	6188	1.618	ng/ul	97
12) Fluorene	15.80	166	6949	1.598	ng/ul	99
14) Pentachlorophenol	17.13	266	797	1.407	ng/ul	98
15) Phenanthrene	17.52	178	10534	1.538	ng/ul	95
16) Anthracene	17.61	178	9653	1.607	ng/ul	94
19) Fluoranthene	19.51	202	11112	1.644	ng/ul#	95
20) Pyrene	19.87	202	10772	1.597	ng/ul	95
21) Benzo(a)anthracene	21.58	228	8540	1.507	ng/ul	95
22) Chrysene	21.63	228	9069	1.486	ng/ul	97
24) Benzo(b)fluoranthene	23.22	252	8379	1.510	ng/ul	87
25) Benzo(k)fluoranthene	23.26	252	8425	1.580	ng/ul#	85
26) Benzo(a)pyrene	23.82	252	7157	1.561	ng/ul#	81
27) Indeno(1,2,3-cd)pyrene	26.29	276	9105	1.636	ng/ul#	99
28) Dibenzo(a,h)anthracene	26.30	278	7575	1.624	ng/ul#	87
29) Benzo(g,h,i)perylene	27.03	276	7940	1.637	ng/ul	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM011121\
Data File : BM028367.D
Acq On : 11 Jan 2021 12:09
Operator : CG/JU
Sample : SSTD1.603
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD1.6003

Quant Time: Jan 11 12:40:50 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM011121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 11 11:35:38 2021
Response via : Initial Calibration

