

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM011123\  
 Data File : BM038426.D  
 Acq On : 11 Jan 2023 10:01  
 Operator : CG/JU  
 Sample : PB150059BL  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :

Quant Time: Jan 12 01:11:20 2023  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\SFAM-EPA-SIM-BM123022.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Jan 12 01:09:22 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc  | Units   | Dev(Min) |
|-----------------------------|--------|------|----------|-------|---------|----------|
| -----                       |        |      |          |       |         |          |
| Internal Standards          |        |      |          |       |         |          |
| 1) 1,4-Dichlorobenzene-d4   | 8.005  | 152  | 1531     | 0.400 | ng/ul   | 0.00     |
| 4) Naphthalene-d8           | 10.812 | 136  | 3759     | 0.400 | ng/ul # | 0.00     |
| 9) Acenaphthene-d10         | 14.634 | 164  | 2026     | 0.400 | ng/ul   | 0.00     |
| 13) Phenanthrene-d10        | 17.375 | 188  | 3995     | 0.400 | ng/ul   | 0.00     |
| 17) Chrysene-d12            | 21.545 | 240  | 4475     | 0.400 | ng/ul   | 0.00     |
| 23) Perylene-d12            | 23.974 | 264  | 5101     | 0.400 | ng/ul # | 0.00     |
| System Monitoring Compounds |        |      |          |       |         |          |
| 3) 1,4-Dioxane-d8           | 3.368  | 96   | 1268     | 0.556 | ng/ul   | 0.00     |
| 6) 2-Methylnaphthalene-d10  | 12.401 | 152  | 1463     | 0.259 | ng/ul   | 0.00     |
| 18) Fluoranthene-d10        | 19.394 | 212  | 3559     | 0.281 | ng/ul   | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM011123\  
Data File : BM038426.D  
Acq On : 11 Jan 2023 10:01  
Operator : CG/JU  
Sample : PB150059BL  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :

Quant Time: Jan 12 01:11:20 2023  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\SFAM-EPA-SIM-BM123022.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Jan 12 01:09:22 2023  
Response via : Initial Calibration

