


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011123\
Data File : BM038424.D
Acq On : 11 Jan 2023 08:47
Operator : CG/JU
Sample : DFTPP102
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP102

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM123022.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Jan 12 01:15:30 2023



AutoFind: Scans 2481, 2482, 2483; Background Corrected with Scan 2475

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	56.9	62464	PASS
68	69	0.00	2	1.8	1097	PASS
69	198	0.00	100	54.2	59539	PASS
70	69	0.00	2	0.5	270	PASS
127	198	10	80	51.3	56323	PASS
197	198	0.00	2	0.4	457	PASS
198	198	100	100	100.0	109867	PASS
199	198	5	9	6.8	7514	PASS
275	198	10	60	25.5	28011	PASS
365	198	1	100	4.1	4551	PASS
441	198	0.01	100	14.0	15354	PASS
442	198	50	100	79.2	87019	PASS
443	442	15	24	19.9	17299	PASS

m/z	Abundance
39.90	193.0
44.00	276.0
49.00	235.0
73.00	327.0
76.80	152.0
133.10	163.0
190.90	152.0
207.00	1392.0
207.90	252.0
281.00	178.0

Instrument :
BNA_M
ClientSampleId :
DFTPP102

m/z	Abundance
36.90	377.0
38.00	712.0
39.00	4422.0
39.90	542.0
41.00	207.0
43.90	352.0
44.90	223.0
48.90	438.0
50.00	16362.0
51.00	57936.0
52.00	3070.0
53.10	156.0
55.00	336.0
55.90	1883.0
57.00	3648.0
57.90	180.0
60.90	705.0
61.90	642.0
63.00	2317.0
64.00	333.0
65.00	999.0
68.00	1204.0
69.00	53768.0
70.00	163.0
73.00	886.0
74.00	5707.0
75.00	9206.0
76.00	3152.0
77.00	60992.0
78.00	4157.0
79.00	3894.0
80.00	3179.0
81.00	4162.0
82.00	1195.0
82.90	993.0
85.00	656.0
85.90	1585.0
86.90	662.0
87.90	187.0
91.00	973.0
91.90	1158.0
93.00	7781.0
93.90	641.0
96.00	582.0
98.00	5394.0
98.90	4754.0
99.90	340.0
100.90	2468.0
102.90	904.0
103.90	1345.0
104.90	1152.0
105.90	551.0
107.00	18576.0
108.00	2403.0
108.90	440.0
110.00	25104.0
111.00	4429.0
112.00	613.0
112.90	236.0
115.90	1010.0
117.00	18232.0
117.90	1253.0
119.00	272.0
119.90	266.0
121.80	1105.0
123.00	1961.0
123.80	871.0
124.90	966.0
125.90	213.0
127.00	47192.0
128.00	3341.0
129.00	21344.0
129.90	1840.0
131.00	305.0
133.10	214.0
134.00	472.0
134.90	1539.0
135.90	830.0
137.00	893.0
140.00	316.0
141.00	2781.0
141.90	793.0
142.90	611.0
143.80	151.0
146.00	438.0
146.90	1746.0
148.00	4546.0
148.90	724.0
149.90	290.0
151.00	323.0
151.90	184.0
152.90	903.0

Instrument :
BNA_M
ClientSampleId :
DFTPP102

154.00	659.0
154.90	1799.0
156.00	2145.0
156.80	424.0
157.90	391.0
159.00	308.0
159.90	846.0
161.00	1330.0
161.90	505.0
162.80	179.0
163.90	155.0
165.00	1017.0
165.90	597.0
167.00	6783.0
168.00	4035.0
168.90	587.0
170.90	306.0
172.00	448.0
172.90	547.0
173.90	1183.0
175.00	2325.0
175.90	556.0
176.90	1100.0
178.00	284.0
178.90	4367.0
179.90	2837.0
180.90	1291.0
182.00	305.0
183.90	456.0
185.00	2097.0
186.00	14783.0
187.00	4148.0
187.90	601.0
188.90	1109.0
190.90	615.0
191.90	1330.0
193.00	1894.0
193.90	379.0
194.90	188.0
195.90	2219.0
197.90	85904.0
198.90	5713.0
199.90	434.0
200.80	165.0
201.50	254.0
202.00	187.0
202.90	893.0
203.90	4227.0
205.00	6919.0
206.00	24928.0
207.00	4779.0
207.90	1451.0
209.00	759.0
210.00	349.0
210.90	1044.0
211.80	244.0
214.90	301.0
216.00	662.0
216.90	7520.0
217.90	1018.0
221.00	2637.0
221.90	942.0
223.00	1654.0
224.00	15219.0
225.00	3703.0
226.00	501.0
227.00	6661.0
227.90	1167.0
229.00	1455.0
231.00	563.0
233.90	427.0
234.80	571.0
235.90	391.0
236.90	609.0
238.90	287.0
240.00	304.0
240.90	402.0
242.00	768.0
243.00	773.0
244.00	11927.0
245.00	1501.0
245.90	2474.0
246.90	528.0
249.00	432.0
249.80	169.0
252.70	150.0
253.80	324.0
255.00	54152.0
256.00	8306.0
256.90	695.0
257.90	3743.0
258.90	837.0
260.00	186.0
265.00	1181.0
266.00	214.0
270.90	156.0

Instrument :
BNA_M
ClientSampleId :
DFTPP102

273.00	1671.0
273.90	3982.0
275.00	19376.0
276.00	2562.0
276.90	1787.0
278.00	489.0
280.90	289.0
282.90	380.0
284.00	211.0
284.90	347.0
292.80	512.0
293.80	162.0
296.00	6929.0
296.90	893.0
302.90	782.0
303.90	303.0
313.90	295.0
314.90	867.0
315.90	386.0
321.00	224.0
323.00	1848.0
323.90	452.0
326.90	536.0
328.00	150.0
332.00	161.0
333.90	1208.0
334.90	328.0
341.00	228.0
346.00	357.0
351.90	355.0
352.90	407.0
354.00	401.0
355.00	182.0
364.90	2883.0
365.80	393.0
372.00	822.0
373.00	312.0
382.90	343.0
389.80	181.0
401.90	385.0
402.80	454.0
404.00	237.0
420.90	348.0
421.90	362.0
423.00	2459.0
424.00	1015.0
441.00	7902.0
442.00	48160.0
443.00	9490.0
444.00	829.0

Instrument :
BNA_M
ClientSampleId :
DFTPP102

m/z	Abundance
37.10	384.0
38.00	943.0
39.00	5627.0
39.90	526.0
40.90	212.0
43.90	254.0
44.80	174.0
49.00	518.0
50.00	21720.0
51.00	77200.0
52.00	3797.0
55.00	340.0
56.00	2311.0
57.00	4785.0
57.90	336.0
60.90	859.0
62.00	1165.0
63.00	2735.0
64.00	455.0
65.00	1207.0
67.90	1157.0
69.00	74232.0
70.00	402.0
73.00	952.0
74.00	9007.0
75.00	13256.0
76.00	3925.0
77.00	85224.0
78.00	5811.0
79.00	6365.0
80.00	4146.0
81.00	6024.0
82.00	1505.0
83.00	1292.0
83.80	308.0
84.90	1057.0
86.00	1766.0
86.90	861.0
87.90	278.0
91.00	1244.0
91.90	1495.0
93.00	10607.0
93.90	752.0
95.90	395.0
96.90	216.0
97.90	8168.0
99.00	6333.0
100.00	642.0
101.00	3313.0
101.90	213.0
102.90	665.0
104.00	1995.0
104.90	1729.0
106.00	578.0
107.00	24984.0
107.90	3483.0
108.90	687.0
110.00	36528.0
111.00	7301.0
111.90	884.0
112.90	185.0
115.90	1290.0
117.00	26952.0
118.00	1717.0
118.90	319.0
120.00	219.0
121.90	1344.0
122.90	2176.0
124.00	1256.0
125.00	1187.0
126.00	230.0
127.00	70752.0
128.00	5194.0
129.00	30752.0
129.90	2373.0
130.90	446.0
131.90	179.0
132.90	313.0
133.90	1031.0
135.00	2595.0
136.00	1005.0
136.90	1405.0
137.90	271.0
138.80	203.0
139.90	516.0
140.90	4486.0
141.90	1114.0
142.90	960.0
144.00	235.0
145.00	186.0
145.90	820.0
147.00	2129.0

Instrument :
BNA_M
ClientSampleId :
DFTPP102

147.90	5735.0
149.00	947.0
149.90	353.0
150.90	595.0
151.80	206.0
152.90	1264.0
153.90	1074.0
154.90	2174.0
156.00	3322.0
157.00	572.0
157.90	822.0
158.90	669.0
159.90	1228.0
160.90	1982.0
161.90	536.0
164.00	222.0
165.00	1480.0
166.00	1303.0
167.00	9221.0
168.00	5644.0
169.00	1047.0
169.90	164.0
170.70	402.0
171.90	701.0
173.00	988.0
173.90	1794.0
175.00	3196.0
175.90	608.0
176.90	1354.0
178.00	729.0
178.90	7529.0
180.00	4714.0
181.00	1907.0
182.00	278.0
183.90	365.0
185.00	3490.0
186.00	23720.0
187.00	6520.0
188.00	726.0
188.90	1521.0
189.90	181.0
191.00	982.0
191.90	2257.0
193.00	2263.0
193.90	458.0
194.90	235.0
195.90	3787.0
196.90	696.0
197.90	139136.0
198.90	9537.0
200.00	939.0
201.50	601.0
202.90	1487.0
204.00	6612.0
205.00	10217.0
206.00	44864.0
207.00	7239.0
208.00	1552.0
208.90	822.0
209.90	779.0
211.00	1724.0
212.00	357.0
213.00	154.0
214.90	423.0
215.90	1271.0
216.90	13009.0
218.00	1700.0
218.90	229.0
219.90	210.0
221.00	5271.0
222.00	1527.0
223.00	2721.0
224.00	25992.0
225.00	6453.0
226.00	872.0
227.00	11929.0
228.00	1605.0
228.90	2524.0
229.90	444.0
230.90	1081.0
232.90	206.0
233.80	801.0
234.90	875.0
235.90	603.0
237.00	1140.0
237.90	236.0
239.00	368.0
239.80	273.0
240.90	610.0
242.00	1275.0
243.00	1671.0
244.00	18944.0
245.00	2327.0
245.90	4634.0
247.00	797.0
247.90	211.0

Instrument :
BNA_M
ClientSampleId :
DFTPP102

248.90	600.0
251.80	246.0
252.90	446.0
254.00	596.0
255.00	92976.0
256.00	14632.0
256.90	1158.0
258.00	6007.0
258.90	1085.0
264.90	2281.0
265.90	211.0
266.80	173.0
270.90	284.0
271.90	202.0
273.00	2734.0
274.00	7090.0
275.00	35072.0
275.90	4596.0
277.00	3551.0
277.90	961.0
281.00	505.0
282.90	519.0
284.00	275.0
285.00	680.0
286.00	155.0
288.90	151.0
292.90	971.0
294.00	199.0
294.90	193.0
295.90	12256.0
296.90	1747.0
301.90	239.0
303.00	1505.0
303.90	391.0
307.90	169.0
310.00	170.0
313.90	499.0
315.00	1460.0
316.00	867.0
320.90	611.0
323.00	3716.0
323.90	748.0
326.90	757.0
327.90	403.0
331.90	227.0
332.90	207.0
334.00	2520.0
335.00	513.0
341.00	405.0
342.00	174.0
345.90	879.0
347.10	160.0
351.90	1147.0
353.00	670.0
354.00	863.0
354.90	204.0
365.00	5460.0
366.00	847.0
371.00	360.0
372.00	2021.0
373.00	358.0
383.00	380.0
389.90	295.0
390.80	184.0
402.00	746.0
402.90	866.0
404.10	276.0
420.90	801.0
421.90	671.0
423.00	5122.0
424.00	1770.0
424.90	157.0
441.00	17320.0
442.00	101064.0
443.00	20120.0
444.00	2016.0

Instrument :
BNA_M
ClientSampleId :
DFTPP102

m/z	Abundance
37.00	210.0
37.90	618.0
39.00	3595.0
39.90	406.0
43.90	327.0
44.90	155.0
49.00	577.0
50.00	13778.0
51.00	52256.0
52.00	2466.0
54.90	396.0
55.90	1463.0
57.00	3030.0
57.90	151.0
60.90	652.0
61.90	720.0
63.00	2169.0
63.90	247.0
65.00	1131.0
68.00	929.0
69.00	50616.0
69.90	244.0
73.00	902.0
74.00	5780.0
75.00	8151.0
76.00	3041.0
77.00	59504.0
78.00	3862.0
79.00	3857.0
80.00	3114.0
80.90	4344.0
82.00	998.0
83.00	1098.0
84.00	185.0
84.80	570.0
86.00	1057.0
86.90	654.0
87.90	219.0
90.90	818.0
91.90	1068.0
93.00	8024.0
94.00	536.0
96.00	490.0
98.00	5424.0
99.00	4461.0
99.90	288.0
101.00	2074.0
103.00	736.0
103.90	1632.0
105.00	1384.0
106.00	464.0
107.00	17608.0
108.00	2180.0
108.80	332.0
110.00	24904.0
111.00	4722.0
112.00	486.0
112.90	225.0
115.90	899.0
117.00	18912.0
118.00	1146.0
118.90	272.0
122.00	1140.0
123.00	1623.0
123.90	852.0
124.80	605.0
126.10	241.0
127.00	51024.0
127.90	3650.0
129.00	20592.0
129.90	1570.0
131.00	346.0
133.00	244.0
133.90	698.0
135.00	1649.0
135.90	718.0
137.00	965.0
139.90	258.0
141.00	3268.0
142.00	922.0
142.90	670.0
143.90	308.0
144.80	212.0
146.00	503.0
147.00	1476.0
148.00	4210.0
149.00	776.0
150.00	231.0
151.00	403.0
152.10	176.0
152.90	1289.0
153.90	750.0

Instrument :
BNA_M
ClientSampleId :
DFTPP102

155.00	1499.0
156.00	2293.0
156.90	532.0
157.90	605.0
158.90	467.0
159.90	964.0
161.00	1440.0
161.80	239.0
162.90	278.0
164.90	1312.0
165.90	702.0
167.00	7238.0
168.00	4058.0
169.00	648.0
170.80	349.0
171.90	714.0
173.00	750.0
173.90	1209.0
175.00	2753.0
175.90	403.0
177.00	1200.0
177.80	362.0
178.90	4712.0
180.00	3511.0
180.90	1589.0
182.00	353.0
183.80	251.0
185.00	2502.0
186.00	17336.0
187.00	5277.0
187.90	550.0
188.90	1509.0
189.90	307.0
190.90	871.0
191.90	1881.0
193.00	2148.0
194.00	318.0
195.90	3088.0
196.90	675.0
197.90	104560.0
198.90	7293.0
199.90	641.0
201.40	497.0
203.00	1006.0
204.00	5417.0
205.00	8217.0
206.00	32936.0
207.00	5461.0
208.00	1575.0
208.90	684.0
209.90	702.0
211.00	1395.0
211.90	242.0
214.90	536.0
216.00	842.0
216.90	11083.0
217.90	1508.0
219.00	223.0
221.00	3503.0
222.00	1172.0
223.00	2296.0
224.00	21072.0
225.00	5242.0
226.00	479.0
227.00	9466.0
227.90	1313.0
228.90	2049.0
229.90	219.0
230.90	482.0
233.90	743.0
235.00	842.0
235.90	421.0
236.90	973.0
238.90	357.0
240.10	319.0
240.90	561.0
242.00	1083.0
242.90	1287.0
244.00	15889.0
245.00	2007.0
245.90	3490.0
246.90	469.0
248.90	619.0
252.00	191.0
253.00	295.0
253.90	320.0
255.00	79184.0
256.00	12474.0
256.90	785.0
257.90	5369.0
258.90	837.0
260.00	205.0
261.00	200.0
263.90	166.0
265.00	2327.0
265.90	301.0

Instrument :
BNA_M
ClientSampleId :
DFTPP102

270.90	273.0
272.00	183.0
273.00	2474.0
274.00	5613.0
275.00	29584.0
275.90	3461.0
277.00	3021.0
278.00	437.0
278.90	221.0
281.00	305.0
281.90	223.0
282.90	621.0
284.00	396.0
284.90	633.0
292.90	841.0
293.90	213.0
294.90	171.0
296.00	11513.0
297.00	1582.0
302.90	1148.0
303.90	356.0
307.90	201.0
313.90	528.0
314.90	1565.0
315.90	683.0
321.00	329.0
322.00	298.0
323.00	2925.0
324.00	598.0
326.90	707.0
327.90	264.0
332.00	309.0
332.90	407.0
333.90	2074.0
335.00	416.0
341.00	581.0
345.90	582.0
352.00	1193.0
352.90	722.0
354.00	871.0
354.90	218.0
365.00	5310.0
366.00	609.0
372.00	1655.0
372.90	377.0
382.90	455.0
389.90	302.0
392.00	178.0
402.00	819.0
403.00	991.0
403.90	436.0
421.00	969.0
421.90	723.0
423.00	4903.0
424.00	1560.0
441.00	20840.0
442.00	111832.0
443.00	22288.0
444.00	1923.0

Instrument :
BNA_M
ClientSampleId :
DFTPP102