


Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_M\DATA\BM011218\
Data File : BM013566.D
Acq On : 12 Jan 2018 14:49
Operator : SJ/JU
Sample : DFTPP73
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP73

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM011018.M
Title : SVOA CALIBRATION
Last Update : Sat Jan 13 01:50:56 2018



AutoFind: Scans 2445, 2446, 2447; Background Corrected with Scan 2440

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.2	102610	PASS
68	69	0.00	2	2.0	2067	PASS
69	198	0.00	100	44.2	105005	PASS
70	69	0.00	2	0.6	620	PASS
127	198	10	80	47.7	113429	PASS
197	198	0.00	2	0.3	782	PASS
198	198	100	100	100.0	237781	PASS
199	198	5	9	6.8	16084	PASS
275	198	10	60	28.1	66704	PASS
365	198	1	100	3.9	9265	PASS
441	443	0.01	100	80.9	28409	PASS
442	198	50	100	75.4	179245	PASS
443	442	15	24	19.6	35130	PASS

m/z	Abundance
38.10	1279.0
39.10	6727.0
39.90	586.0
41.10	522.0
44.00	1947.0
44.90	569.0
48.90	1178.0
50.10	24904.0
51.00	82696.0
52.10	3917.0
54.80	594.0
56.00	2515.0
57.00	4934.0
61.00	1473.0
62.00	1198.0
63.10	3865.0
65.00	2115.0
68.10	1078.0
69.00	79752.0
69.90	382.0
72.90	971.0
74.00	10088.0
75.00	15476.0
76.10	4420.0
77.00	86480.0
78.10	6209.0
79.00	5916.0
80.00	6433.0
81.00	5962.0
81.90	2046.0
82.90	1592.0
84.00	412.0
85.00	1154.0
86.00	2091.0
87.00	1576.0
88.10	462.0
90.90	1900.0
92.00	1499.0
93.00	13218.0
93.90	838.0
95.00	527.0
96.10	322.0
98.00	11175.0
99.10	7649.0
100.10	456.0
101.00	4403.0
102.90	1753.0
104.00	2678.0
105.00	3470.0
106.00	999.0
107.00	28280.0
108.00	4739.0
108.90	1046.0
110.00	43936.0
111.00	6941.0
111.90	1057.0
112.90	697.0
116.10	1429.0
117.00	34824.0
118.00	2303.0
119.00	391.0
119.90	645.0
121.90	2696.0
123.00	2519.0
123.90	2033.0
125.00	1277.0
125.80	860.0
127.00	87304.0
128.00	5967.0
129.00	33184.0
130.00	3183.0
133.80	1418.0
135.00	2540.0
135.80	1191.0
136.90	2230.0
137.70	303.0
139.80	787.0
141.00	4825.0
141.80	1385.0
143.00	1383.0
146.00	904.0
147.00	3431.0
148.00	6303.0
149.20	694.0
152.80	1770.0
153.90	1500.0
155.00	2997.0
156.00	4367.0
157.00	1079.0
157.90	914.0
158.90	758.0

Instrument :
BNA_M
ClientSampleId :
DFTPP73

160.00	2381.0
161.10	2401.0
161.90	999.0
163.70	522.0
165.00	2231.0
166.10	1382.0
167.00	10791.0
168.00	5800.0
169.00	1055.0
169.90	359.0
170.80	419.0
172.00	1022.0
172.80	1263.0
173.30	639.0
173.90	1484.0
175.00	3727.0
176.00	915.0
176.80	1598.0
178.10	834.0
179.00	6569.0
180.00	5619.0
181.00	2568.0
181.80	351.0
183.90	731.0
185.00	3990.0
186.10	25736.0
187.00	7459.0
187.90	1265.0
188.90	1936.0
190.00	771.0
191.10	877.0
192.10	2373.0
193.00	3498.0
194.10	813.0
196.00	3382.0
197.10	1079.0
198.00	168704.0
198.90	10700.0
199.90	1832.0
201.20	785.0
201.80	598.0
202.90	1588.0
203.90	6080.0
205.10	9682.0
206.00	49944.0
207.00	9367.0
208.00	1742.0
208.90	1006.0
210.00	803.0
210.80	2207.0
214.90	1025.0
215.80	741.0
217.00	13849.0
217.90	1169.0
221.00	5711.0
222.00	1989.0
222.90	3668.0
224.00	24808.0
225.10	6234.0
226.10	683.0
227.00	12510.0
228.00	1292.0
228.90	1958.0
230.10	511.0
230.90	935.0
232.10	468.0
234.00	758.0
234.80	316.0
235.80	551.0
237.00	1176.0
239.00	574.0
239.90	346.0
240.80	305.0
241.20	370.0
242.00	1401.0
243.00	1209.0
244.00	18440.0
245.10	2432.0
246.00	4739.0
247.00	921.0
249.10	729.0
254.00	1287.0
255.00	97960.0
256.00	14678.0
256.90	1006.0
258.00	7309.0
258.90	1218.0
260.00	444.0
265.00	3672.0
265.90	514.0
271.90	357.0
273.10	3088.0
274.00	7927.0
275.00	39368.0
276.00	5582.0

Instrument :
BNA_M
ClientSampleId :
DFTPP73

277.00	5307.0
278.00	652.0
281.00	852.0
281.70	393.0
283.00	938.0
284.10	304.0
285.10	465.0
293.00	761.0
294.00	666.0
295.00	443.0
296.00	13655.0
297.00	2291.0
302.00	535.0
303.00	1452.0
313.90	558.0
315.00	1017.0
316.00	958.0
320.90	381.0
323.10	3268.0
324.00	615.0
326.80	544.0
334.00	2126.0
334.80	449.0
335.10	391.0
346.10	758.0
352.10	1135.0
353.00	645.0
354.10	1159.0
354.90	337.0
365.00	5383.0
365.90	714.0
370.80	373.0
372.00	1734.0
382.80	599.0
401.90	526.0
403.10	1117.0
404.20	384.0
421.10	967.0
422.10	658.0
423.00	3946.0
424.00	1188.0
441.10	10868.0
442.10	83976.0
443.10	14751.0
444.00	2547.0

Instrument :
BNA_M
ClientSampleId :
DFTPP73

m/z	Abundance
37.10	877.0
38.00	1918.0
39.00	8165.0
40.00	838.0
43.90	1573.0
45.00	491.0
49.10	1364.0
50.00	35112.0
51.00	126584.0
52.10	7468.0
55.00	644.0
56.00	3629.0
57.00	8622.0
57.90	388.0
61.00	1300.0
62.00	2573.0
63.00	4295.0
64.00	814.0
64.90	2030.0
66.90	488.0
68.10	3248.0
69.00	125928.0
69.90	465.0
73.00	2079.0
74.10	15231.0
75.00	21312.0
76.10	7351.0
77.00	137344.0
78.00	9527.0
79.00	11648.0
80.00	8442.0
81.00	11006.0
82.10	2459.0
83.10	2854.0
84.90	1333.0
86.00	3469.0
86.90	2104.0
87.90	599.0
91.00	3123.0
92.00	2854.0
93.00	21904.0
93.90	1267.0
96.00	664.0
97.10	437.0
98.00	16480.0
99.00	13698.0
100.00	1380.0
101.00	5930.0
102.00	428.0
103.00	1711.0
104.10	4047.0
104.90	3434.0
105.90	1133.0
107.00	46000.0
108.00	5483.0
109.00	1630.0
110.00	70240.0
111.00	10363.0
112.00	2265.0
112.90	584.0
114.80	323.0
116.00	2935.0
117.00	51200.0
117.90	3792.0
119.10	880.0
120.00	345.0
122.00	2576.0
123.00	4316.0
123.90	1539.0
124.90	2021.0
127.00	133184.0
127.90	11690.0
129.00	59800.0
130.00	4731.0
131.00	857.0
132.10	326.0
132.90	466.0
133.90	2183.0
135.00	5172.0
136.00	1751.0
137.00	3420.0
137.90	460.0
140.00	856.0
140.90	8057.0
142.00	2103.0
142.90	1511.0
143.90	633.0
144.90	411.0
145.90	1920.0
147.00	4419.0
148.00	10412.0

Instrument :
BNA_M
ClientSampleId :
DFTPP73

149.00	2148.0
150.00	501.0
150.90	1166.0
151.90	557.0
152.90	2195.0
154.00	2370.0
155.00	4000.0
156.10	7395.0
157.10	1160.0
158.00	1513.0
159.10	1475.0
160.00	2213.0
161.00	4082.0
162.00	970.0
164.00	999.0
165.00	2564.0
166.10	2313.0
167.00	15729.0
168.00	9329.0
169.10	1862.0
169.90	546.0
171.10	879.0
172.00	1549.0
173.00	1675.0
174.00	3336.0
175.00	5882.0
175.90	1508.0
177.00	2895.0
177.80	1550.0
179.00	11456.0
179.90	7896.0
181.00	4711.0
182.00	1020.0
182.70	543.0
184.10	1343.0
185.00	4864.0
186.00	44752.0
187.00	12619.0
188.10	1694.0
188.90	3070.0
190.10	842.0
190.90	1437.0
192.00	5213.0
193.00	5032.0
194.00	1143.0
194.80	486.0
195.90	7141.0
197.90	272896.0
199.00	17752.0
200.00	1451.0
200.90	803.0
201.40	1660.0
203.00	2363.0
204.00	13024.0
205.00	19424.0
206.00	85152.0
207.00	12476.0
207.90	3064.0
208.90	1757.0
209.90	1171.0
210.20	980.0
211.00	4252.0
211.70	968.0
212.80	333.0
215.00	1366.0
216.10	2331.0
217.00	21232.0
217.90	3718.0
219.10	373.0
220.20	546.0
221.00	9580.0
222.00	3333.0
223.00	4513.0
224.00	47448.0
225.00	12268.0
226.10	1120.0
227.00	22672.0
227.90	3399.0
229.00	4964.0
229.90	927.0
230.90	1918.0
232.00	497.0
232.90	642.0
234.00	2006.0
235.00	1334.0
236.10	1466.0
237.10	1671.0
238.90	1381.0
239.80	701.0
241.00	1032.0
242.00	2865.0
242.90	2443.0
244.00	28992.0
245.00	3781.0
245.90	6587.0

Instrument :
BNA_M
ClientSampleId :
DFTPP73

246.90	2300.0
248.90	2138.0
252.10	359.0
253.00	1271.0
253.80	910.0
255.00	184512.0
256.00	28312.0
256.90	2327.0
258.00	13593.0
258.90	2086.0
261.00	410.0
265.00	4848.0
265.90	791.0
271.10	944.0
272.00	789.0
273.10	4785.0
274.00	12475.0
275.00	81392.0
276.10	8628.0
277.00	7789.0
277.80	1325.0
280.90	553.0
281.70	345.0
283.10	1060.0
283.90	412.0
285.10	1053.0
285.90	358.0
289.00	694.0
291.00	403.0
292.00	516.0
293.00	2283.0
294.10	582.0
295.00	489.0
296.00	27992.0
296.90	3234.0
302.10	482.0
303.10	2417.0
304.10	1549.0
308.00	417.0
314.10	693.0
314.90	2405.0
316.00	1241.0
321.30	743.0
322.10	575.0
323.00	6983.0
324.00	1454.0
327.10	1262.0
328.00	1079.0
331.90	1160.0
332.90	1382.0
334.00	3420.0
334.90	1779.0
341.10	1510.0
346.00	1430.0
347.00	337.0
352.00	2236.0
352.90	1551.0
354.10	2688.0
355.10	784.0
365.00	9450.0
366.00	1066.0
371.00	498.0
372.00	2792.0
373.00	637.0
382.90	1068.0
390.00	446.0
401.10	303.0
401.90	1228.0
403.00	2793.0
404.00	943.0
421.00	1401.0
422.00	1556.0
423.00	10917.0
424.10	2623.0
425.20	326.0
440.20	355.0
441.10	33920.0
442.10	196416.0
443.10	39232.0
444.10	4299.0

Instrument :
BNA_M
ClientSampleId :
DFTPP73

m/z	Abundance
37.00	814.0
38.00	1278.0
39.00	8255.0
40.00	1185.0
44.00	1621.0
49.00	960.0
50.10	28400.0
51.10	98552.0
52.10	5176.0
54.80	803.0
56.00	3342.0
57.00	5482.0
61.00	1168.0
61.90	976.0
63.10	4687.0
64.00	685.0
65.10	1595.0
67.90	1875.0
69.00	109336.0
69.90	1014.0
73.10	1536.0
74.00	11371.0
75.00	18688.0
76.10	6057.0
77.10	115248.0
78.10	9187.0
79.00	10491.0
80.00	8168.0
81.00	9621.0
81.90	2712.0
83.00	2099.0
85.00	1599.0
86.00	2820.0
87.00	971.0
90.90	1674.0
91.90	2424.0
93.00	18320.0
93.90	1527.0
94.70	359.0
96.00	950.0
97.00	739.0
98.00	15207.0
99.00	9303.0
100.00	671.0
101.00	5488.0
101.90	346.0
103.00	1165.0
104.00	3400.0
105.00	3633.0
106.20	2071.0
107.00	40344.0
108.00	7241.0
108.90	1318.0
110.00	59288.0
111.00	10541.0
111.90	1175.0
116.00	2256.0
117.00	45112.0
118.00	3573.0
119.00	561.0
119.90	382.0
122.10	3197.0
123.10	4026.0
124.00	1774.0
125.00	1533.0
126.10	602.0
127.00	119800.0
128.00	8841.0
129.00	53056.0
130.00	3407.0
130.80	760.0
132.10	487.0
133.00	382.0
134.00	1539.0
135.00	3785.0
135.90	1576.0
137.00	2400.0
138.00	316.0
138.90	636.0
140.10	959.0
141.00	6561.0
142.10	2414.0
143.00	1985.0
144.10	438.0
146.00	1593.0
147.10	4676.0
148.00	11295.0
149.00	2229.0
151.00	1315.0
152.00	536.0
152.90	2885.0

Instrument :
BNA_M
ClientSampleId :
DFTPP73

154.10	2153.0
155.00	5411.0
156.00	7090.0
157.00	717.0
157.90	949.0
158.80	1281.0
160.00	2595.0
161.00	4122.0
161.90	1007.0
165.10	2622.0
166.00	3324.0
167.00	16009.0
168.00	9322.0
168.90	1445.0
170.90	1313.0
171.90	1501.0
173.10	2219.0
174.00	3610.0
175.00	5241.0
175.90	1016.0
177.10	3395.0
178.10	671.0
179.00	11429.0
180.00	6966.0
181.10	3384.0
181.90	791.0
182.70	322.0
183.90	1214.0
185.00	5080.0
186.00	39248.0
187.00	11014.0
187.90	1272.0
189.00	4548.0
190.00	702.0
190.90	1269.0
192.00	3936.0
193.00	4197.0
194.00	1216.0
194.80	399.0
196.00	5743.0
197.00	1268.0
197.90	271744.0
198.90	19800.0
200.00	2427.0
200.70	672.0
201.60	1273.0
202.90	2616.0
204.00	10945.0
205.10	16359.0
206.10	74688.0
207.10	11632.0
208.00	2825.0
208.90	1111.0
210.10	1962.0
211.00	3952.0
211.80	1165.0
214.90	1132.0
216.00	2451.0
217.00	22544.0
218.00	3650.0
219.00	328.0
219.70	435.0
221.10	8960.0
221.90	2712.0
223.00	4490.0
224.00	49040.0
225.10	11045.0
226.00	1183.0
227.00	22304.0
227.90	2863.0
229.00	4405.0
229.90	581.0
230.90	938.0
232.80	457.0
233.90	1332.0
234.90	1482.0
236.00	1533.0
237.00	2455.0
239.00	1161.0
240.10	727.0
240.90	1528.0
241.90	2107.0
242.90	2578.0
244.00	33768.0
245.10	3601.0
245.90	6816.0
247.10	1265.0
247.80	377.0
249.00	942.0
250.00	363.0
250.80	547.0
253.10	898.0
253.90	1628.0
255.00	190848.0
256.00	28608.0

Instrument :
BNA_M
ClientSampleId :
DFTPP73

257.00	1865.0
258.00	13458.0
258.90	1459.0
260.10	638.0
261.00	405.0
263.90	465.0
264.90	4579.0
265.90	1037.0
271.00	873.0
271.90	484.0
273.00	5885.0
274.00	14676.0
275.00	79352.0
276.00	9679.0
277.00	7480.0
278.00	1354.0
281.00	538.0
281.90	658.0
282.80	505.0
283.90	626.0
285.10	1306.0
290.80	511.0
293.00	2279.0
294.00	340.0
294.90	687.0
296.00	27960.0
297.00	4070.0
300.90	323.0
302.00	704.0
303.10	3767.0
304.00	1608.0
310.10	398.0
314.10	1811.0
314.90	4285.0
316.10	1068.0
321.00	1123.0
321.90	640.0
323.10	8524.0
324.10	1689.0
327.00	2476.0
328.00	964.0
332.10	345.0
332.70	571.0
334.00	5790.0
335.00	1765.0
341.10	849.0
342.10	311.0
346.00	2136.0
347.10	888.0
352.00	2726.0
353.00	1761.0
354.00	1943.0
365.00	12962.0
366.10	1465.0
371.00	631.0
372.00	3772.0
372.90	805.0
382.80	511.0
383.80	524.0
390.00	404.0
401.00	317.0
402.00	1094.0
403.10	1998.0
404.00	976.0
421.10	1900.0
421.90	1967.0
423.00	12101.0
424.00	2769.0
425.10	527.0
441.10	40440.0
442.10	257344.0
443.00	51408.0
444.10	4079.0

Instrument :
BNA_M
ClientSampleId :
DFTPP73

Scan 2440 (17.439 min): BM013566.D
DFTPP73

m/z	Abundance
39.90	728.0
44.00	1393.0
72.90	436.0
207.00	715.0

Instrument :
BNA_M
ClientSampleId :
DFTPP73