

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011222\
 Data File : BM033704.D
 Acq On : 12 Jan 2022 18:23
 Operator : CG/JU
 Sample : N1099-05
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 VACLOT-BACKFILL-TILCON-05

Quant Time: Jan 13 07:51:38 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM010522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 13 07:49:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.972	152	184570	20.000	ng	0.00
21) Naphthalene-d8	10.784	136	750276	20.000	ng	0.00
39) Acenaphthene-d10	14.607	164	424410	20.000	ng	0.00
64) Phenanthrene-d10	17.342	188	672728	20.000	ng	0.00
76) Chrysene-d12	21.501	240	657022	20.000	ng	0.00
86) Perylene-d12	23.924	264	717856	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.513	112	1226003	104.761	ng	0.00
7) Phenol-d6	7.125	99	1578498	100.078	ng	0.00
23) Nitrobenzene-d5	9.131	82	1058287	65.142	ng	0.00
42) 2,4,6-Tribromophenol	16.083	330	451867	101.936	ng	0.00
45) 2-Fluorobiphenyl	13.236	172	2176880	67.181	ng	0.00
79) Terphenyl-d14	19.954	244	2066443	56.613	ng	0.00
Target Compounds						
9) Benzaldehyde	7.131	77	4392	Below Cal	Qvalue #	10

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011222\
Data File : BM033704.D
Acq On : 12 Jan 2022 18:23
Operator : CG/JU
Sample : N1099-05
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
VACLOT-BACKFILL-TILCON-05

Quant Time: Jan 13 07:51:38 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM010522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 13 07:49:01 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.972 min Scan# 807
Delta R.T. 0.000 min
Lab File: BM033704.D
Acq: 12 Jan 2022 18:23




Instrument :
BNA_M
ClientSampleId :
VACLOT-BACKFILL-TILCON-05


Tgt Ion:152 Resp: 184570
Ion Ratio Lower Upper
152 100
150 154.1 128.9 193.3
115 63.6 53.0 79.6



#5
2-Fluorophenol
Concen: 104.761 ng
RT: 5.513 min Scan# 387
Delta R.T. 0.000 min
Lab File: BM033704.D
Acq: 12 Jan 2022 18:23

Tgt Ion:112 Resp: 1226003
Ion Ratio Lower Upper
112 100
64 61.7 55.5 83.3
63 33.4 31.3 46.9









#7
Phenol-d6
Concen: 100.078 ng
RT: 7.125 min Scan# 661
Delta R.T. 0.000 min
Lab File: BM033704.D
Acq: 12 Jan 2022 18:23




Instrument :
BNA_M
ClientSampleId :
VACLOT-BACKFILL-TILCON-05


Tgt Ion: 99 Resp: 1578498
Ion Ratio Lower Upper
99 100
42 19.4 20.3 30.5#
71 39.0 34.5 51.7



#9
Benzaldehyde
Concen: Below Cal
RT: 7.131 min Scan# 662
Delta R.T. 0.036 min
Lab File: BM033704.D
Acq: 12 Jan 2022 18:23

Tgt Ion: 77 Resp: 4392
Ion Ratio Lower Upper
77 100
105 6.9 67.1 107.1#
106 0.0 64.3 104.3#









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.784 min Scan# 1002
Delta R.T. -0.006 min
Lab File: BM033704.D
Acq: 12 Jan 2022 18:23




Instrument :
BNA_M
ClientSampleId :
VACLOT-BACKFILL-TILCON-05


Tgt Ion:136 Resp: 750276
Ion Ratio Lower Upper
136 100
137 10.9 8.6 13.0
54 9.0 8.3 12.5
68 5.3 5.0 7.6



#23
Nitrobenzene-d5
Concen: 65.142 ng
RT: 9.131 min Scan# 1002
Delta R.T. -0.006 min
Lab File: BM033704.D
Acq: 12 Jan 2022 18:23

Tgt Ion: 82 Resp: 1058287
Ion Ratio Lower Upper
82 100
128 41.9 28.1 42.1
54 54.6 46.1 69.1









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.607 min Scan# 1936
Delta R.T. 0.000 min
Lab File: BM033704.D
Acq: 12 Jan 2022 18:23




Instrument :
BNA_M
ClientSampleId :
VACLOT-BACKFILL-TILCON-05


Tgt Ion:164 Resp: 424410
Ion Ratio Lower Upper
164 100
162 99.7 79.7 119.5
160 43.6 34.2 51.4



#42
2,4,6-Tribromophenol
Concen: 101.936 ng
RT: 16.083 min Scan# 2184
Delta R.T. -0.006 min
Lab File: BM033704.D
Acq: 12 Jan 2022 18:23

Tgt Ion:330 Resp: 451867
Ion Ratio Lower Upper
330 100
332 96.5 76.5 114.7
141 40.1 35.9 53.9









#45
2-Fluorobiphenyl
Concen: 67.181 ng
RT: 13.236 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM033704.D
Acq: 12 Jan 2022 18:23




Instrument :
BNA_M
ClientSampleId :
VACLOT-BACKFILL-TILCON-05


Tgt Ion:172 Resp: 2176880
Ion Ratio Lower Upper
172 100
171 35.9 27.2 40.8
170 23.6 18.7 28.1



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.342 min Scan# 2398
Delta R.T. 0.000 min
Lab File: BM033704.D
Acq: 12 Jan 2022 18:23

Tgt Ion:188 Resp: 672728
Ion Ratio Lower Upper
188 100
94 8.6 7.5 11.3
80 10.0 8.9 13.3









#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.501 min Scan# 3109
Delta R.T. -0.006 min
Lab File: BM033704.D
Acq: 12 Jan 2022 18:23




Instrument :
BNA_M
ClientSampleId :
VACLOT-BACKFILL-TILCON-05


Tgt Ion:240 Resp: 657022
Ion Ratio Lower Upper
240 100
120 10.7 7.8 11.6
236 25.9 20.3 30.5



#79
Terphenyl-d14
Concen: 56.613 ng
RT: 19.954 min Scan# 2842
Delta R.T. -0.006 min
Lab File: BM033704.D
Acq: 12 Jan 2022 18:23

Tgt Ion:244 Resp: 2066443
Ion Ratio Lower Upper
244 100
212 7.7 5.9 8.9
122 10.4 8.4 12.6





#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 23.924 min Scan# 31
 Delta R.T. -0.006 min
 Lab File: BM033704.D
 Acq: 12 Jan 2022 18:23

Instrument :
 BNA_M
 ClientSampleId :
 VACLOT-BACKFILL-TILCON-05

Tgt Ion:	264	Resp:	717856
Ion	Ratio	Lower	Upper
264	100		
260	24.3	17.7	26.5
265	22.2	17.7	26.5

