

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011222\
 Data File : BM033709.D
 Acq On : 12 Jan 2022 21:21
 Operator : CG/JU
 Sample : N1099-10
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 VACLOT-BACKFILL-TILCON-10

Quant Time: Jan 13 07:52:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM010522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 13 07:49:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.966	152	177925	20.000	ng	0.00
21) Naphthalene-d8	10.783	136	668258	20.000	ng	# 0.00
39) Acenaphthene-d10	14.607	164	332566	20.000	ng	0.00
64) Phenanthrene-d10	17.342	188	582312	20.000	ng	0.00
76) Chrysene-d12	21.500	240	672289	20.000	ng	0.00
86) Perylene-d12	23.924	264	732649	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.513	112	1088115	96.451	ng	0.00
7) Phenol-d6	7.125	99	1340845	88.186	ng	0.00
23) Nitrobenzene-d5	9.130	82	888709	61.418	ng	0.00
42) 2,4,6-Tribromophenol	16.083	330	358350	103.165	ng	0.00
45) 2-Fluorobiphenyl	13.236	172	1674423	65.946	ng	0.00
79) Terphenyl-d14	19.953	244	1971710	52.791	ng	0.00
Target Compounds						
9) Benzaldehyde	7.125	77	3223	Below Cal	#	Qvalue 14

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011222\
Data File : BM033709.D
Acq On : 12 Jan 2022 21:21
Operator : CG/JU
Sample : N1099-10
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
VACLOT-BACKFILL-TILCON-10

Quant Time: Jan 13 07:52:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM010522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 13 07:49:01 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.966 min Scan# 807
Delta R.T. -0.006 min
Lab File: BM033709.D
Acq: 12 Jan 2022 21:21




Instrument :
BNA_M
ClientSampleId :
VACLOT-BACKFILL-TILCON-10


Tgt Ion:152 Resp: 177925
Ion Ratio Lower Upper
152 100
150 155.3 128.9 193.3
115 65.5 53.0 79.6



#5
2-Fluorophenol
Concen: 96.451 ng
RT: 5.513 min Scan# 387
Delta R.T. -0.000 min
Lab File: BM033709.D
Acq: 12 Jan 2022 21:21

Tgt Ion:112 Resp: 1088115
Ion Ratio Lower Upper
112 100
64 61.7 55.5 83.3
63 32.9 31.3 46.9









#7
Phenol-d6
Concen: 88.186 ng
RT: 7.125 min Scan# 661
Delta R.T. -0.000 min
Lab File: BM033709.D
Acq: 12 Jan 2022 21:21




Instrument :
BNA_M
ClientSampleId :
VACLOT-BACKFILL-TILCON-10


Tgt Ion: 99 Resp: 1340845
Ion Ratio Lower Upper
99 100
42 19.4 20.3 30.5#
71 39.5 34.5 51.7



#9
Benzaldehyde
Concen: Below Cal
RT: 7.125 min Scan# 661
Delta R.T. 0.029 min
Lab File: BM033709.D
Acq: 12 Jan 2022 21:21

Tgt Ion: 77 Resp: 3223
Ion Ratio Lower Upper
77 100
105 13.1 67.1 107.1#
106 0.0 64.3 104.3#









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.783 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM033709.D
Acq: 12 Jan 2022 21:21




Instrument :
BNA_M
ClientSampleId :
VACLOT-BACKFILL-TILCON-10


Tgt Ion:136 Resp: 668258
Ion Ratio Lower Upper
136 100
137 11.0 8.6 13.0
54 8.0 8.3 12.5#
68 5.7 5.0 7.6



#23
Nitrobenzene-d5
Concen: 61.418 ng
RT: 9.130 min Scan# 1002
Delta R.T. -0.006 min
Lab File: BM033709.D
Acq: 12 Jan 2022 21:21

Tgt Ion: 82 Resp: 888709
Ion Ratio Lower Upper
82 100
128 43.7 28.1 42.1#
54 53.9 46.1 69.1









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.607 min Scan# 1936
Delta R.T. -0.000 min
Lab File: BM033709.D
Acq: 12 Jan 2022 21:21




Instrument :
BNA_M
ClientSampleId :
VACLOT-BACKFILL-TILCON-10


Tgt Ion:164 Resp: 332566
Ion Ratio Lower Upper
164 100
162 102.0 79.7 119.5
160 44.0 34.2 51.4



#42
2,4,6-Tribromophenol
Concen: 103.165 ng
RT: 16.083 min Scan# 2184
Delta R.T. -0.006 min
Lab File: BM033709.D
Acq: 12 Jan 2022 21:21

Tgt Ion:330 Resp: 358350
Ion Ratio Lower Upper
330 100
332 96.4 76.5 114.7
141 39.7 35.9 53.9









#45
2-Fluorobiphenyl
Concen: 65.946 ng
RT: 13.236 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM033709.D
Acq: 12 Jan 2022 21:21




Instrument :
BNA_M
ClientSampleId :
VACLOT-BACKFILL-TILCON-10


Tgt Ion:172 Resp: 1674423
Ion Ratio Lower Upper
172 100
171 35.5 27.2 40.8
170 23.0 18.7 28.1



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.342 min Scan# 2398
Delta R.T. -0.000 min
Lab File: BM033709.D
Acq: 12 Jan 2022 21:21

Tgt Ion:188 Resp: 582312
Ion Ratio Lower Upper
188 100
94 8.5 7.5 11.3
80 9.6 8.9 13.3









#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.500 min Scan# 3109
Delta R.T. -0.006 min
Lab File: BM033709.D
Acq: 12 Jan 2022 21:21




Instrument :
BNA_M
ClientSampleId :
VACLOT-BACKFILL-TILCON-10


Tgt Ion:240 Resp: 672289
Ion Ratio Lower Upper
240 100
120 10.2 7.8 11.6
236 26.7 20.3 30.5



#79
Terphenyl-d14
Concen: 52.791 ng
RT: 19.953 min Scan# 2842
Delta R.T. -0.006 min
Lab File: BM033709.D
Acq: 12 Jan 2022 21:21

Tgt Ion:244 Resp: 1971710
Ion Ratio Lower Upper
244 100
212 7.8 5.9 8.9
122 10.0 8.4 12.6





#86
Perylene-d12
Concen: 20.000 ng
RT: 23.924 min Scan# 31
Delta R.T. -0.006 min
Lab File: BM033709.D
Acq: 12 Jan 2022 21:21

Instrument :
BNA_M
ClientSampleId :
VACLOT-BACKFILL-TILCON-10

Tgt Ion:264 Resp: 732649
Ion Ratio Lower Upper
264 100
260 24.0 17.7 26.5
265 22.2 17.7 26.5

