


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011325\
Data File : BM049334.D
Acq On : 13 Jan 2025 12:26
Operator : RC/JU
Sample : DFTPP060
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP060

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM011325.MA.M
Title : SVOA CALIBRATION
Last Update : Mon Jan 13 16:56:06 2025



AutoFind: Scans 2447, 2448, 2449; Background Corrected with Scan 2441

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.6	161949	PASS
68	69	0.00	2	1.4	2492	PASS
69	198	0.00	100	32.4	183552	PASS
70	69	0.00	2	0.5	848	PASS
127	198	10	80	40.4	229013	PASS
197	198	0.00	2	0.3	1474	PASS
198	198	100	100	100.0	566400	PASS
199	198	5	9	6.7	38117	PASS
275	198	10	60	24.8	140584	PASS
365	198	1	100	3.3	18758	PASS
441	198	0.01	100	11.3	64189	PASS
442	198	50	100	71.5	404779	PASS
443	442	15	24	19.1	77208	PASS

m/z	Abundance
38.90	493.0
40.00	3485.0
40.90	594.0
41.90	180.0
43.10	549.0
44.00	5778.0
45.00	180.0
49.00	402.0
51.20	398.0
55.10	226.0
56.00	168.0
57.10	204.0
59.90	187.0
68.00	168.0
70.00	157.0
73.10	558.0
75.00	239.0
77.00	326.0
78.00	180.0
79.00	162.0
80.90	159.0
83.80	442.0
85.80	211.0
90.90	212.0
95.90	235.0
133.10	283.0
147.00	333.0
190.90	258.0
192.70	307.0
207.00	2466.0
208.00	465.0
209.10	285.0
221.10	171.0
267.00	189.0
280.90	843.0
281.90	201.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

m/z	Abundance
35.90	215.0
37.00	598.0
38.10	1750.0
39.10	12337.0
40.00	3811.0
41.00	906.0
42.10	319.0
43.00	856.0
44.00	6674.0
45.10	533.0
46.90	378.0
47.90	195.0
49.00	1164.0
50.00	39584.0
51.00	160192.0
52.00	8436.0
53.00	514.0
55.10	1056.0
56.00	4894.0
57.00	11592.0
57.90	545.0
59.90	269.0
61.00	1986.0
62.10	2290.0
63.10	6488.0
64.00	887.0
65.00	3327.0
66.10	342.0
67.00	593.0
68.00	2380.0
69.00	178176.0
70.00	1216.0
71.10	207.0
73.00	1657.0
74.00	16952.0
75.00	29232.0
76.10	9478.0
77.10	208640.0
78.00	15046.0
79.00	11588.0
80.00	9420.0
81.00	14239.0
82.00	3454.0
83.00	3735.0
83.90	889.0
85.00	2192.0
86.00	3795.0
87.00	1777.0
87.90	811.0
89.10	277.0
91.00	3270.0
92.00	3240.0
93.00	20488.0
94.00	1559.0
95.00	463.0
96.00	1194.0
97.00	562.0
98.00	16680.0
99.00	14863.0
100.00	1453.0
101.00	8189.0
102.00	545.0
102.90	2344.0
104.00	4394.0
105.00	4045.0
106.00	1612.0
107.00	59472.0
108.00	9620.0
109.00	1420.0
110.00	109304.0
111.00	16984.0
112.00	2140.0
112.90	732.0
113.90	176.0
115.00	245.0
116.00	3042.0
117.00	48824.0
118.00	3788.0
119.10	560.0
119.90	779.0
120.90	319.0
122.00	3553.0
123.00	5711.0
124.00	2268.0
125.00	2610.0
126.00	513.0
127.00	210624.0
128.00	16776.0
129.00	78416.0
130.00	6115.0
130.90	1094.0
132.00	654.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

133.00	635.0
134.00	2025.0
135.00	5746.0
136.00	2273.0
137.00	4314.0
137.90	842.0
139.00	368.0
139.90	851.0
141.00	9461.0
142.00	3279.0
143.00	2561.0
144.00	715.0
145.00	588.0
145.90	1653.0
147.00	5729.0
148.00	11946.0
149.00	2817.0
150.00	1020.0
151.00	1129.0
152.00	686.0
153.00	3756.0
154.00	2705.0
155.00	6043.0
156.00	8929.0
157.00	1628.0
158.00	1916.0
159.00	1743.0
160.00	3440.0
161.00	4826.0
161.90	1552.0
163.00	534.0
164.00	624.0
165.00	3926.0
166.00	3586.0
167.00	25320.0
168.00	17712.0
169.00	2292.0
169.90	751.0
170.90	1025.0
172.00	2142.0
173.00	2666.0
174.00	4652.0
175.00	9768.0
176.00	2067.0
176.90	3687.0
177.90	1281.0
179.00	16400.0
180.00	12501.0
181.00	5418.0
182.00	929.0
182.90	558.0
184.00	1332.0
185.00	7851.0
186.00	62832.0
187.00	19448.0
188.00	1931.0
189.00	3401.0
190.00	756.0
191.00	2030.0
192.00	6363.0
193.00	6158.0
193.90	1355.0
194.90	706.0
196.00	11777.0
196.90	1707.0
197.90	481600.0
198.90	33472.0
200.00	2697.0
201.50	1962.0
203.00	3331.0
204.00	16960.0
205.00	29592.0
206.00	124480.0
207.00	18448.0
208.00	4351.0
209.00	1507.0
210.20	2222.0
210.90	5035.0
211.70	994.0
213.00	347.0
213.90	281.0
214.90	1188.0
216.00	2451.0
217.00	34328.0
218.00	4741.0
219.00	469.0
220.10	587.0
221.00	17128.0
221.80	4739.0
223.00	8218.0
224.00	68648.0
225.00	18464.0
226.00	2073.0
227.00	27928.0
228.00	4401.0
229.00	6169.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

230.00	950.0
231.00	2585.0
231.90	332.0
232.80	448.0
233.90	1780.0
235.00	2032.0
235.90	1271.0
237.00	2553.0
238.00	354.0
238.90	966.0
240.00	749.0
241.00	1755.0
242.00	3371.0
243.10	3679.0
244.00	51040.0
245.00	6859.0
246.00	9951.0
247.00	2204.0
247.90	479.0
249.00	1770.0
249.90	294.0
250.90	417.0
251.90	476.0
252.90	984.0
253.90	1174.0
255.00	258496.0
256.00	37848.0
257.00	3363.0
258.00	14903.0
259.00	2429.0
260.00	595.0
260.90	316.0
264.00	323.0
265.00	5493.0
266.00	1027.0
267.00	235.0
270.00	299.0
271.00	1089.0
272.00	744.0
273.00	8140.0
274.00	20352.0
275.00	102456.0
276.00	13855.0
277.00	8818.0
278.00	2076.0
279.10	335.0
281.00	899.0
282.00	383.0
283.00	1229.0
284.00	825.0
285.00	1381.0
285.90	339.0
288.90	416.0
289.90	380.0
290.80	293.0
292.00	552.0
293.00	2296.0
294.10	558.0
295.00	629.0
296.00	31432.0
297.00	4461.0
298.00	229.0
299.10	166.0
301.00	431.0
302.10	563.0
303.00	3645.0
304.00	971.0
308.00	340.0
308.90	218.0
310.00	355.0
311.00	151.0
312.90	189.0
313.90	1269.0
315.00	3768.0
316.00	1987.0
317.00	369.0
320.10	150.0
321.00	815.0
321.90	436.0
323.00	9427.0
324.00	1768.0
324.90	298.0
327.00	1699.0
327.90	883.0
329.00	180.0
331.90	596.0
333.00	1084.0
334.00	6240.0
335.00	1497.0
336.10	177.0
341.00	1339.0
342.00	219.0
346.00	2359.0
347.00	459.0
349.90	168.0
350.90	320.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

352.00	2926.0
353.00	1809.0
354.00	2542.0
355.10	754.0
358.90	179.0
365.00	13266.0
365.90	1810.0
367.00	181.0
370.00	291.0
371.00	660.0
372.00	4602.0
373.10	1100.0
383.00	1425.0
383.90	423.0
390.00	487.0
391.10	372.0
392.00	293.0
400.90	422.0
402.00	2110.0
403.00	2583.0
404.00	977.0
404.90	167.0
415.10	179.0
421.00	2101.0
422.10	2045.0
423.00	12499.0
424.10	3570.0
425.00	424.0
437.90	165.0
441.10	38560.0
442.10	232384.0
443.10	47104.0
444.00	4962.0
445.00	331.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

m/z	Abundance
36.00	210.0
37.00	868.0
38.00	2003.0
39.00	13418.0
40.00	3893.0
41.00	809.0
42.10	259.0
43.00	598.0
44.00	6053.0
44.90	586.0
46.80	271.0
48.00	201.0
49.00	1567.0
50.10	46952.0
51.10	190016.0
52.10	10038.0
53.20	534.0
55.00	1382.0
56.00	5148.0
57.00	13608.0
58.00	704.0
58.90	263.0
60.00	179.0
61.10	2307.0
62.00	2911.0
63.00	7880.0
64.00	950.0
65.00	4182.0
66.00	354.0
67.00	528.0
68.00	3616.0
69.00	216448.0
70.00	1169.0
71.00	211.0
73.10	2174.0
74.00	20032.0
75.00	35080.0
76.10	12268.0
77.10	255424.0
78.10	17992.0
79.00	14003.0
80.00	10868.0
81.00	16322.0
82.00	4012.0
83.00	4101.0
83.90	906.0
85.00	2641.0
86.00	4822.0
87.00	2059.0
88.00	824.0
89.00	397.0
91.00	3817.0
92.00	4065.0
93.00	25336.0
94.00	1609.0
95.00	458.0
96.00	1218.0
97.00	662.0
98.00	18360.0
99.00	17408.0
100.00	1738.0
101.00	8988.0
102.00	603.0
103.00	2839.0
104.00	5836.0
105.00	5247.0
106.10	1952.0
107.00	75712.0
108.00	11717.0
109.00	1711.0
110.00	143232.0
111.00	21320.0
112.00	2545.0
113.00	761.0
114.20	201.0
114.90	264.0
116.00	3703.0
117.00	60824.0
118.00	4349.0
118.90	503.0
120.00	693.0
120.90	360.0
122.00	3974.0
123.00	7589.0
124.00	3328.0
125.00	3364.0
126.10	946.0
127.00	268288.0
128.00	19056.0
129.00	99064.0
130.00	8190.0
131.00	1556.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

132.00	690.0
133.00	555.0
134.00	2742.0
135.00	7888.0
136.00	3184.0
137.00	4632.0
138.00	936.0
138.90	471.0
140.00	1130.0
141.00	12402.0
142.00	4102.0
143.00	3431.0
144.00	769.0
145.00	882.0
146.10	2644.0
147.00	6860.0
148.00	16392.0
149.00	3789.0
150.00	1126.0
151.20	1765.0
151.70	1057.0
153.00	4325.0
154.00	2782.0
155.00	8080.0
156.00	11696.0
157.00	1992.0
158.00	1978.0
159.00	2063.0
160.00	3951.0
161.00	6330.0
162.10	1538.0
162.90	568.0
164.00	665.0
165.00	5239.0
166.00	4306.0
167.00	31944.0
168.00	21544.0
169.00	3030.0
170.10	1092.0
171.00	1078.0
172.00	2407.0
173.00	3129.0
174.00	6061.0
175.00	12290.0
176.10	3566.0
177.00	5030.0
178.00	1622.0
179.00	24624.0
180.00	16840.0
181.00	8075.0
182.00	1457.0
182.80	490.0
184.00	1762.0
185.00	11465.0
186.00	87288.0
187.00	26656.0
188.00	2755.0
189.00	5584.0
189.90	1085.0
191.00	3109.0
192.00	8182.0
193.00	9095.0
194.00	2159.0
195.00	1219.0
196.00	16400.0
196.90	2714.0
198.00	678656.0
199.00	44352.0
200.00	4096.0
201.40	2487.0
203.00	4815.0
204.00	26496.0
205.00	42712.0
206.00	182336.0
207.00	26440.0
208.00	6115.0
209.00	2311.0
210.10	2529.0
211.00	7167.0
212.90	606.0
213.90	216.0
215.00	1791.0
216.00	4037.0
217.00	49968.0
218.00	6272.0
218.90	488.0
220.20	1015.0
221.00	27880.0
221.80	6968.0
223.00	11081.0
224.00	103760.0
225.00	26648.0
226.00	2624.0
227.00	40312.0
228.00	6132.0
229.00	9549.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

230.00	1340.0
231.00	3592.0
232.00	613.0
233.00	700.0
234.00	2798.0
235.00	2875.0
236.00	2304.0
237.00	3432.0
237.90	630.0
239.00	1633.0
240.00	1184.0
241.00	2377.0
242.00	5647.0
243.00	5864.0
244.00	83128.0
245.00	11429.0
246.00	15876.0
246.90	3451.0
248.00	977.0
249.00	2647.0
250.00	506.0
250.90	661.0
252.00	521.0
253.00	1525.0
254.00	2008.0
255.00	407680.0
256.00	58568.0
257.00	5110.0
258.00	23240.0
259.00	3431.0
260.00	618.0
260.90	838.0
261.80	158.0
263.00	189.0
264.00	526.0
265.00	9069.0
265.90	1163.0
266.90	386.0
270.00	422.0
271.00	1632.0
272.00	1178.0
273.00	12637.0
274.00	32304.0
275.00	173248.0
276.00	22768.0
277.00	13775.0
278.00	2903.0
279.00	606.0
281.00	833.0
281.90	552.0
283.00	1998.0
284.00	1284.0
285.00	2754.0
285.90	408.0
288.00	174.0
289.00	533.0
290.00	471.0
290.90	538.0
292.00	674.0
293.00	3522.0
294.00	1036.0
295.00	1113.0
296.00	51864.0
297.00	7819.0
298.00	615.0
301.00	742.0
302.00	654.0
303.00	6304.0
304.00	1895.0
304.90	230.0
308.00	800.0
309.00	574.0
310.00	753.0
312.00	345.0
313.00	450.0
314.00	2027.0
315.00	5543.0
316.00	3316.0
317.00	405.0
320.00	302.0
321.00	1698.0
321.90	766.0
323.00	15616.0
324.00	2971.0
325.10	394.0
326.00	364.0
326.90	3231.0
328.00	1537.0
328.90	227.0
331.90	1229.0
333.10	1483.0
334.00	10492.0
335.00	2200.0
336.00	320.0
339.00	312.0
340.00	186.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

341.00	1908.0
342.00	482.0
346.00	3598.0
347.00	830.0
349.90	178.0
351.10	234.0
352.00	4031.0
353.10	3361.0
354.10	4574.0
355.00	981.0
356.00	199.0
358.90	452.0
365.00	21920.0
366.00	3284.0
366.90	347.0
370.00	328.0
371.00	1097.0
372.10	7961.0
373.00	2275.0
374.10	204.0
377.00	185.0
383.00	2155.0
383.90	615.0
385.10	207.0
390.00	1187.0
391.00	751.0
392.10	662.0
401.00	493.0
402.00	3410.0
403.00	4317.0
404.00	1945.0
404.90	205.0
410.10	173.0
415.00	304.0
421.00	3830.0
422.00	3647.0
423.10	24864.0
424.00	5697.0
425.10	622.0
438.00	255.0
439.90	166.0
441.10	75648.0
442.10	472384.0
443.10	92824.0
444.10	8138.0
445.00	419.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

m/z	Abundance
35.90	292.0
37.00	488.0
38.00	1194.0
39.00	9956.0
40.00	4444.0
41.00	801.0
41.90	255.0
43.00	466.0
44.00	6536.0
44.90	407.0
46.90	277.0
49.00	1519.0
50.00	33680.0
51.10	136832.0
52.00	6567.0
53.20	480.0
55.10	956.0
56.00	4044.0
57.00	9504.0
58.10	350.0
58.90	288.0
60.10	156.0
61.00	1984.0
62.00	2148.0
63.00	5642.0
64.00	971.0
65.10	2918.0
65.90	318.0
67.00	211.0
68.00	1985.0
69.00	156032.0
70.00	631.0
71.00	153.0
73.10	1802.0
74.00	14616.0
75.00	25520.0
76.10	8449.0
77.10	185088.0
78.00	12725.0
79.00	10358.0
80.00	8779.0
81.00	12360.0
82.00	3088.0
83.00	3589.0
84.00	737.0
85.00	2076.0
86.00	3547.0
87.00	1478.0
88.10	650.0
88.90	331.0
91.00	2857.0
92.00	2956.0
93.00	18760.0
94.00	1343.0
95.10	489.0
96.00	1237.0
97.10	502.0
98.00	14395.0
99.00	14020.0
100.00	1213.0
101.00	6770.0
102.00	339.0
103.00	2169.0
104.00	4095.0
105.00	4323.0
106.10	1225.0
107.00	55416.0
108.00	7863.0
109.00	1365.0
110.00	105600.0
111.00	15560.0
112.00	2218.0
112.90	608.0
114.90	412.0
116.00	2810.0
117.00	44936.0
118.00	3634.0
119.00	446.0
119.90	704.0
121.00	208.0
122.00	3208.0
123.00	5781.0
124.00	2552.0
125.00	2368.0
126.10	803.0
127.00	208128.0
128.00	14017.0
129.00	77472.0
130.00	6174.0
131.00	1513.0
132.00	524.0
133.10	482.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

133.90	1839.0
135.00	6133.0
136.00	2438.0
137.00	4254.0
137.90	651.0
139.10	497.0
140.00	948.0
141.00	9604.0
141.90	3441.0
143.00	2397.0
143.90	556.0
144.90	570.0
146.00	1645.0
147.00	6276.0
148.00	13175.0
149.00	3482.0
150.00	1000.0
151.00	1219.0
151.90	746.0
153.00	3463.0
154.00	2503.0
155.00	6155.0
156.00	8990.0
157.10	1840.0
158.00	2003.0
159.00	1533.0
159.90	3212.0
161.00	5326.0
162.00	1439.0
162.90	596.0
164.00	518.0
165.00	3807.0
166.00	3285.0
167.00	22656.0
168.00	16432.0
169.00	2387.0
169.90	810.0
170.90	855.0
171.90	2011.0
173.00	3042.0
174.00	5366.0
175.00	9612.0
176.00	2543.0
177.00	3918.0
177.90	1469.0
179.00	19712.0
180.00	12932.0
181.00	5755.0
182.00	758.0
182.80	485.0
183.80	1235.0
185.00	9387.0
186.00	72384.0
187.00	19536.0
188.00	2064.0
189.00	4275.0
190.00	630.0
191.00	2570.0
192.00	6870.0
193.00	7039.0
193.90	1319.0
195.00	744.0
196.00	13071.0
197.90	538944.0
198.90	36528.0
200.00	2690.0
201.50	2584.0
203.00	3607.0
204.00	20904.0
205.00	34488.0
206.00	143616.0
207.00	20096.0
208.00	5249.0
209.00	1775.0
209.90	1925.0
211.00	5624.0
211.80	1023.0
213.00	491.0
214.00	164.0
215.00	1396.0
216.00	2808.0
217.00	39880.0
218.00	5178.0
219.00	542.0
220.20	492.0
221.00	21576.0
221.90	5532.0
223.00	8501.0
224.00	86656.0
225.00	20608.0
226.10	2009.0
227.00	34176.0
228.00	4339.0
229.00	7869.0
230.10	1100.0
231.00	3017.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

232.00	530.0
232.80	707.0
234.00	2328.0
235.00	2369.0
236.00	2007.0
237.00	3434.0
237.90	564.0
239.00	1388.0
239.90	1310.0
241.00	1850.0
242.00	4801.0
243.00	4685.0
244.00	67288.0
245.00	9011.0
246.00	12562.0
247.00	2570.0
248.00	527.0
249.00	2140.0
249.80	374.0
250.90	538.0
252.00	867.0
252.90	1227.0
254.00	1847.0
255.00	345024.0
256.00	48888.0
257.00	3999.0
258.00	18920.0
259.00	3169.0
259.90	465.0
260.90	716.0
261.90	170.0
262.90	288.0
263.80	426.0
265.00	8111.0
265.90	1071.0
266.90	304.0
270.00	276.0
270.90	1231.0
272.00	832.0
273.00	10956.0
274.00	26144.0
275.00	146048.0
276.00	20432.0
277.00	12484.0
278.00	2545.0
279.00	400.0
281.10	999.0
281.90	406.0
283.10	1532.0
283.90	911.0
285.00	2416.0
285.90	443.0
288.90	425.0
290.00	425.0
291.00	492.0
292.00	734.0
293.00	3151.0
293.90	988.0
295.00	938.0
296.00	45936.0
297.00	6196.0
298.00	449.0
301.00	541.0
302.00	521.0
303.00	5434.0
304.00	1257.0
305.00	200.0
308.00	610.0
309.00	502.0
309.90	655.0
313.00	472.0
314.00	2123.0
315.00	5255.0
316.00	2639.0
316.90	435.0
319.90	200.0
321.00	1292.0
322.00	717.0
323.10	13954.0
324.00	2486.0
325.00	198.0
325.90	301.0
326.90	2882.0
328.00	1655.0
329.10	203.0
332.00	1416.0
333.00	1434.0
334.00	9382.0
335.00	2061.0
336.00	375.0
339.00	402.0
340.00	235.0
341.00	1696.0
342.10	413.0
346.00	3962.0
347.00	512.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060

350.90	233.0
352.00	4314.0
353.00	3197.0
354.10	4182.0
355.10	942.0
359.00	402.0
365.00	21088.0
366.00	3659.0
367.00	219.0
370.00	385.0
370.90	1151.0
372.10	7778.0
373.00	1775.0
377.00	204.0
383.00	2143.0
384.00	635.0
385.00	254.0
390.00	1061.0
391.00	770.0
392.00	504.0
401.00	354.0
402.00	3335.0
403.00	4657.0
404.00	1739.0
405.00	297.0
421.00	3973.0
422.10	3660.0
423.10	25552.0
424.10	6345.0
425.10	660.0
439.00	192.0
441.10	78360.0
442.10	509568.0
443.10	91696.0
444.10	8386.0
444.90	525.0

Instrument :
BNA_M
ClientSampleId :
DFTPP060