

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM011416\
 Data File : BM004012.D
 Acq On : 14 Jan 2016 08:12
 Operator : SJ/UM
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02056

Quant Time: Jan 14 13:18:40 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM010116.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 13 06:45:02 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2	1,4-Dioxane	0.519	0.469	9.6	73	0.00
3 S	1,4-Dioxane-d8	0.447	0.409	8.5	78	0.00
4	Benzaldehyde	0.974	1.244	-27.7#	96	0.00
5 S	Phenol-d5	1.661	1.881	-13.2	91	0.00
6	Phenol	1.752	2.037	-16.3	94	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.939	1.038	-10.5	88	0.00
8	Bis(2-Chloroethyl)ether	1.318	1.486	-12.7	90	0.00
9 S	2-Chlorophenol-d4	1.365	1.556	-14.0	92	0.00
10	2-Chlorophenol	1.427	1.631	-14.3	91	0.00
11	2-Methylphenol	1.339	1.564	-16.8	94	0.00
12	2,2'-oxybis(1-Chloropropane	1.449	1.675	-15.6	91	0.00
13 S	4-Methylphenol-d8	1.346	1.557	-15.7	92	0.00
14	Acetophenone	2.096	2.522	-20.3#	93	0.00
15 P	N-Nitroso-di-n-propylamine	1.052	1.251	-18.9	93	0.00
16	4-Methylphenol	1.452	1.733	-19.4	95	0.00
17	Hexachloroethane	0.551	0.615	-11.6	89	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
19 S	Nitrobenzene-d5	0.151	0.163	-7.9	95	0.00
20	Nitrobenzene	0.355	0.388	-9.3	96	0.00
21	Isophorone	0.659	0.729	-10.6	96	0.00
22 S	2-Nitrophenol-d4	0.165	0.176	-6.7	94	0.00
23 C	2-Nitrophenol	0.183	0.200	-9.3	96	0.00
24	2,4-Dimethylphenol	0.367	0.404	-10.1	95	0.00
25	Bis(2-Chloroethoxy)methane	0.408	0.435	-6.6	93	0.00
26 S	2,4-Dichlorophenol-d3	0.293	0.322	-9.9	96	0.00
27 C	2,4-Dichlorophenol	0.303	0.337	-11.2	97	0.00
28	Naphthalene	1.042	1.150	-10.4	97	0.00
29 S	4-Chloroaniline-d4	0.387	0.355	8.3	76	0.00
30	4-Chloroaniline	0.392	0.381	2.8	80	0.00
31 C	Hexachlorobutadiene	0.186	0.202	-8.6	95	0.00
32	Caprolactam	0.097	0.106	-9.3	95	0.00
33 C	4-Chloro-3-methylphenol	0.338	0.386	-14.2	98	0.00
34	2-Methylnaphthalene	0.737	0.838	-13.7	99	0.00
35	1-Methylnaphthalene	0.699	0.799	-14.3	99	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
37	1,2,4,5-Tetrachlorobenzene	0.640	0.689	-7.7	98	0.00
38	Hexachlorocyclopentadiene	0.335	0.164	51.0#	51	0.00
39 C	2,4,6-Trichlorophenol	0.412	0.443	-7.5	98	0.00
40	2,4,5-Trichlorophenol	0.447	0.491	-9.8	99	0.00
41	1,1'-Biphenyl	1.685	1.824	-8.2	98	0.00
42	2-Chloronaphthalene	1.263	1.398	-10.7	100	0.00
43	2-Nitroaniline	0.352	0.396	-12.5	101	0.00
44 S	Dimethylphthalate-d6	1.570	1.658	-5.6	95	0.00

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM011416\
 Data File : BM004012.D
 Acq On : 14 Jan 2016 08:12
 Operator : SJ/UM
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02056

Quant Time: Jan 14 13:18:40 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM010116.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 13 06:45:02 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.601	1.730	-8.1	97	0.00
46	2,6-Dinitrotoluene	0.315	0.361	-14.6	102	0.00
47 S	Acenaphthylene-d8	1.913	2.121	-10.9	100	0.00
48	Acenaphthylene	2.096	2.306	-10.0	99	0.00
49	3-Nitroaniline	0.342	0.366	-7.0	93	0.00
50 C	Acenaphthene	1.386	1.537	-10.9	100	0.00
51	2,4-Dinitrophenol	0.153	0.125	18.3	89	0.00
52 S	4-Nitrophenol-d4	0.282	0.290	-2.8	95	0.00
53	4-Nitrophenol	0.223	0.234	-4.9	97	0.00
54	Dibenzofuran	1.940	2.156	-11.1	100	0.00
55	2,4-Dinitrotoluene	0.458	0.538	-17.5	104	0.00
56	2,3,4,6-Tetrachlorophenol	0.358	0.395	-10.3	99	0.00
57	Diethylphthalate	1.623	1.789	-10.2	98	0.00
58 S	Fluorene-d10	1.332	1.469	-10.3	99	0.00
59	Fluorene	1.539	1.749	-13.6	101	0.00
60	4-Chlorophenyl-phenylether	0.742	0.822	-10.8	98	0.00
61	4-Nitroaniline	0.358	0.432	-20.7#	107	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.113	0.102	9.7	92	0.00
64	4,6-Dinitro-2-methylphenol	0.120	0.111	7.5	93	0.00
65	N-Nitrosodiphenylamine	0.628	0.682	-8.6	100	0.00
66	4-Bromophenyl-phenylether	0.207	0.221	-6.8	99	0.00
67	Hexachlorobenzene	0.227	0.250	-10.1	102	0.00
68	Atrazine	0.218	0.235	-7.8	99	0.00
69 C	Pentachlorophenol	0.110	0.112	-1.8	103	0.00
70	Phenanthrene	1.125	1.261	-12.1	103	0.00
71 S	Anthracene-d10	0.925	1.012	-9.4	102	0.00
72	Anthracene	1.154	1.284	-11.3	102	0.00
73	1,2,3,4-Tetrachlorobenzene	0.321	0.293	8.7	86	0.00
74	Pentachlorobenzene	0.285	0.289	-1.4	95	0.00
75	Carbazole	1.015	1.137	-12.0	101	0.00
76	Di-n-butylphthalate	1.239	1.396	-12.7	107	0.00
77 C	Fluoranthene	1.169	1.358	-16.2	105	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
79 S	Pyrene-d10	1.036	1.194	-15.3	101	0.00
80	Pyrene	1.358	1.623	-19.5	105	0.00
81	Butylbenzylphthalate	0.548	0.716	-30.7#	121	0.00
82	3,3'-Dichlorobenzidine	0.347	0.436	-25.6#	114	0.00
83	Benzo(a)anthracene	1.191	1.338	-12.3	103	0.00
84	Bis(2-ethylhexyl)phthalate	0.725	1.048	-44.6#	134	0.00
85	Chrysene	1.132	1.251	-10.5	100	0.00
86 I	Perylene-d12	1.000	1.000	0.0	98	0.00
87	Di-n-octyl phthalate	1.333	1.952	-46.4#	135	0.00

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM011416\
Data File : BM004012.D
Acq On : 14 Jan 2016 08:12
Operator : SJ/UM
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02056

Quant Time: Jan 14 13:18:40 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM010116.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jan 13 06:45:02 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.216	1.394	-14.6	106	0.00
89	Benzo(k)fluoranthene	1.157	1.254	-8.4	99	0.00
90 S	Benzo(a)pyrene-d12	0.999	1.090	-9.1	100	0.00
91 C	Benzo(a)pyrene	1.153	1.294	-12.2	102	0.00
92	Indeno(1,2,3-cd)pyrene	1.276	1.460	-14.4	102	0.00
93	Dibenzo(a,h)anthracene	1.072	1.219	-13.7	101	0.00
94	Benzo(g,h,i)perylene	1.071	1.231	-14.9	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0