

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM011421\
 Data File : BM028395.D
 Acq On : 14 Jan 2021 21:45
 Operator : CG/JU
 Sample : M1065-04
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BFXN0

Quant Time: Jan 14 23:02:10 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM010821MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 14 22:49:26 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.12	152	37377	20.00	ng/ul	0.00
18) Naphthalene-d8	10.95	136	156239	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.75	164	86225	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.47	188	160188	20.00	ng/ul	0.00
78) Chrysene-d12	21.59	240	128779	20.00	ng/ul	0.00
86) Perylene-d12	23.92	264	124434	20.00	ng/ul	-0.01

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.36	96	4020	4.43	ng/uL	0.00
5) Phenol-d5	7.26	99	15744	4.80	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.44	67	49988	25.12	ng/ul	0.00
9) 2-Chlorophenol-d4	7.64	132	54349	19.85	ng/ul	0.00
13) 4-Methylphenol-d8	8.83	113	32253	11.90	ng/ul	0.00
19) Nitrobenzene-d5	9.29	128	34000	26.42	ng/ul	0.00
22) 2-Nitrophenol-d4	10.02	143	35818	26.40	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.56	165	60588	23.41	ng/ul	0.00
29) 4-Chloroaniline-d4	11.07	131	4045	1.27	ng/ul	0.00
44) Dimethylphthalate-d6	14.16	166	186747	27.49	ng/ul	0.00
47) Acenaphthylene-d8	14.45	160	244925	28.52	ng/ul	0.00
52) 4-Nitrophenol-d4	14.93	143	4272	3.30	ng/ul	0.00
58) Fluorene-d10	15.74	176	159825	26.42	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.85	200	23452	22.18	ng/ul	0.00
71) Anthracene-d10	17.57	188	245151	30.08	ng/ul	0.00
79) Pyrene-d10	19.84	212	239740	31.48	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.77	264	219512	31.39	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM011421\
Data File : BM028395.D
Acq On : 14 Jan 2021 21:45
Operator : CG/JU
Sample : M1065-04
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BFXN0

Quant Time: Jan 14 23:02:10 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM010821MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jan 14 22:49:26 2021
Response via : Initial Calibration

