

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011422\
 Data File : BM033730.D
 Acq On : 14 Jan 2022 16:25
 Operator : CG/JU
 Sample : N1143-03
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 WEST-BOILER-ROOM-BOTTOM

Quant Time: Jan 17 00:25:54 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM010522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 13 07:49:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.966	152	148813	20.000	ng	0.00
21) Naphthalene-d8	10.778	136	623560	20.000	ng	-0.01
39) Acenaphthene-d10	14.601	164	404046	20.000	ng	0.00
64) Phenanthrene-d10	17.336	188	731494	20.000	ng	0.00
76) Chrysene-d12	21.500	240	601963	20.000	ng	0.00
86) Perylene-d12	23.918	264	650452	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.507	112	997960	105.765	ng	0.00
7) Phenol-d6	7.119	99	1255192	98.702	ng	0.00
23) Nitrobenzene-d5	9.131	82	860778	63.752	ng	0.00
42) 2,4,6-Tribromophenol	16.083	330	511146	121.121	ng	0.00
45) 2-Fluorobiphenyl	13.230	172	1908253	61.859	ng	-0.01
79) Terphenyl-d14	19.948	244	1962523	58.683	ng	-0.01
Target Compounds						
31) Naphthalene	10.831	128	164597	4.816	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011422\
Data File : BM033730.D
Acq On : 14 Jan 2022 16:25
Operator : CG/JU
Sample : N1143-03
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
WEST-BOILER-ROOM-BOTTOM

Quant Time: Jan 17 00:25:54 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM010522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 13 07:49:01 2022
Response via : Initial Calibration

