



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011422\
 Data File : BM033724.D
 Acq On : 14 Jan 2022 12:00
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM010522.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jan 13 07:49:01 2022






AutoFind: Scans 2472, 2473, 2474; Background Corrected with Scan 2465

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.0	134348	PASS
68	69	0.00	2	1.9	2587	PASS
69	198	0.00	100	37.9	137472	PASS
70	69	0.00	2	0.5	713	PASS
127	198	10	80	47.5	172267	PASS
197	198	0.00	2	0.7	2441	PASS
198	198	100	100	100.0	363029	PASS
199	198	5	9	6.6	23933	PASS
275	198	10	60	27.5	99888	PASS
365	198	1	100	3.7	13489	PASS
441	198	0.01	100	12.0	43395	PASS
442	442	50	100	100.0	274987	PASS
443	442	15	24	19.5	53544	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011422\
Data File : BM033724.D
Acq On : 14 Jan 2022 12:00
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Jan 17 02:49:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM010522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 13 07:49:01 2022
Response via : Initial Calibration



TIC: BM033724.D\data.ms

(70) Pentachlorophenol (C)

16.989min (-0.006) 0.00 ng




response 478850

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.60	63.89
264.00	61.40	63.44
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011422\
 Data File : BM033724.D
 Acq On : 14 Jan 2022 12:00
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Quant Time: Jan 17 02:49:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM010522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 13 07:49:01 2022
 Response via : Initial Calibration



TIC: BM033724.D\data.ms

(77) Benzidine

19.571min (+ 0.000) 191094.92 ng m

response 1921331

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.20	13.92
183.00	12.10	12.12
0.00	0.00	0.00

Instrument :
BNA_M
ClientSampleId :
DFTPP

DDT Breakdown

Date	Instrument Name	DFTPP Data File
1/14/2022	BNA_M	<u>BM033724.D</u>
Compound Name	Response	Retention Time
DDT	1086970	20.8
DDD	87797	20.365
DDE	2866	19.859
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
90663	1177633	7.70