

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM011516\
 Data File : BM004022.D
 Acq On : 14 Jan 2016 20:56
 Operator : SJ/UM
 Sample : H1099-05
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BC998

Quant Time: Jan 15 00:33:10 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM010116.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jan 15 00:28:21 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.66	152	40678	20.00	ng/ul	0.00
18) Naphthalene-d8	10.45	136	197718	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.32	164	115450	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.07	188	248985	20.00	ng/ul	0.00
78) Chrysene-d12	21.27	240	205840	20.00	ng/ul	0.00
86) Perylene-d12	23.51	264	180709	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.13	96	3427	3.77	ng/uL	0.00
5) Phenol-d5	6.85	99	83602	24.75	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.00	67	50407	26.40	ng/ul	0.00
9) 2-Chlorophenol-d4	7.20	132	72104	25.98	ng/ul	0.00
13) 4-Methylphenol-d8	8.37	113	71806	26.24	ng/ul	0.00
19) Nitrobenzene-d5	8.82	128	36707	24.65	ng/ul	0.00
22) 2-Nitrophenol-d4	9.54	143	40543	24.84	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.08	165	70490	24.30	ng/ul	0.00
29) 4-Chloroaniline-d4	10.59	131	99108	25.88	ng/ul	0.00
44) Dimethylphthalate-d6	13.73	166	236682	26.11	ng/ul	0.00
47) Acenaphthylene-d8	14.01	160	291994	26.44	ng/ul	0.00
52) 4-Nitrophenol-d4	14.54	143	31651	19.44	ng/ul	0.00
58) Fluorene-d10	15.32	176	210589	27.39	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.45	200	23270	16.57	ng/ul	0.00
71) Anthracene-d10	17.17	188	317687	27.59	ng/ul	0.00
79) Pyrene-d10	19.47	212	324485	30.44	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.37	264	259018	28.70	ng/ul	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
45) Dimethylphthalate	13.78	163	97523	10.55	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM011516\
Data File : BM004022.D
Acq On : 14 Jan 2016 20:56
Operator : SJ/UM
Sample : H1099-05
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampled :
BC998

Quant Time: Jan 15 00:33:10 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM010116.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jan 15 00:28:21 2016
Response via : Initial Calibration





#45

Dimethylphthalate

Concen: 10.55 ng/ul

RT: 13.78 min Scan# 1818

Delta R.T. 0.00 min

Lab File: BM004022.D


Acq: 14 Jan 2016 20:56

Instrument :

BNA_M

ClientSampleId :

BC998



Tgt Ion:163 Resp: 97523

Ion Ratio Lower Upper

163 100

194 3.9 3.2 4.8

164 10.0 7.8 11.8

