


Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_M\DATA\BM011518\
 Data File : BM013576.D
 Acq On : 15 Jan 2018 08:29
 Operator : SJ/JU
 Sample : DFTPP74
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP74

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM011018.M
 Title : SVOA CALIBRATION
 Last Update : Tue Jan 16 07:13:29 2018



AutoFind: Scans 2445, 2446, 2447; Background Corrected with Scan 2439

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.6	97258	PASS
68	69	0.00	2	1.9	1870	PASS
69	198	0.00	100	40.7	99818	PASS
70	69	0.00	2	0.5	469	PASS
127	198	10	80	45.7	112162	PASS
197	198	0.00	2	0.4	993	PASS
198	198	100	100	100.0	245333	PASS
199	198	5	9	7.0	17139	PASS
275	198	10	60	31.0	76090	PASS
365	198	1	100	4.6	11259	PASS
441	443	0.01	100	74.6	31992	PASS
442	198	50	100	87.0	213461	PASS
443	442	15	24	20.1	42890	PASS

m/z	Abundance
37.00	462.0
38.10	1954.0
39.10	7768.0
39.90	1304.0
40.90	483.0
44.10	1561.0
45.00	426.0
49.10	813.0
50.10	33424.0
51.10	119448.0
52.00	6657.0
55.10	884.0
56.00	3459.0
57.00	7018.0
58.20	302.0
61.00	1449.0
62.00	1376.0
63.10	5235.0
64.00	811.0
65.00	2369.0
66.90	494.0
68.00	2005.0
69.00	119312.0
70.10	587.0
73.20	1520.0
74.10	13120.0
75.00	19144.0
76.10	8000.0
77.10	128384.0
78.10	7325.0
79.10	9418.0
80.00	7273.0
81.00	10090.0
82.00	2340.0
83.00	2323.0
85.10	1959.0
86.00	2905.0
87.10	1432.0
87.80	401.0
91.00	2116.0
92.00	3521.0
93.00	17616.0
94.00	1211.0
94.90	405.0
96.10	540.0
98.00	14923.0
99.00	11149.0
99.90	862.0
101.00	5474.0
103.10	1712.0
104.10	3963.0
105.00	3414.0
105.90	1431.0
107.00	42200.0
108.00	7747.0
109.00	1422.0
110.00	69352.0
111.00	10845.0
111.90	1398.0
113.00	416.0
116.10	2700.0
117.00	47960.0
118.00	3307.0
119.00	393.0
120.00	1062.0
122.00	3283.0
123.00	4348.0
123.90	2014.0
124.90	1791.0
127.00	127952.0
128.00	9687.0
129.00	59896.0
130.00	4833.0
130.90	870.0
132.10	521.0
132.90	594.0
134.00	1539.0
135.00	4753.0
136.00	2407.0
137.10	3797.0
137.80	452.0
138.80	350.0
140.10	497.0
141.00	8356.0
142.00	2607.0
143.00	1703.0
144.00	564.0
145.00	395.0
146.00	1390.0
147.00	3760.0
148.10	10041.0

Instrument :
BNA_M
ClientSampleId :
DFTPP74

149.00	2301.0
150.80	397.0
151.20	380.0
151.90	640.0
152.90	2744.0
154.20	1715.0
155.00	4364.0
156.10	6390.0
156.90	551.0
157.80	1720.0
158.80	1734.0
159.90	3108.0
161.00	3208.0
161.90	1088.0
162.90	313.0
165.00	4162.0
166.10	3481.0
167.00	18240.0
168.00	8867.0
168.90	1816.0
170.20	432.0
170.90	812.0
172.00	1312.0
172.90	1858.0
174.00	2850.0
175.10	6119.0
176.20	1371.0
176.90	2644.0
178.00	978.0
179.00	11001.0
180.00	8533.0
181.10	3369.0
181.90	386.0
184.00	1243.0
184.90	5918.0
186.00	37704.0
187.00	12121.0
188.20	594.0
189.00	2404.0
189.90	704.0
190.90	1866.0
192.00	3972.0
193.10	3971.0
194.10	1739.0
195.20	812.0
196.00	7157.0
196.70	1399.0
198.00	270144.0
198.90	16952.0
199.80	832.0
200.20	830.0
201.20	1074.0
202.10	417.0
203.00	1825.0
204.00	8755.0
205.00	16384.0
206.00	77880.0
207.00	12314.0
208.00	3064.0
209.00	1306.0
210.10	955.0
211.10	2837.0
211.80	1150.0
212.80	334.0
215.10	888.0
216.10	1520.0
217.00	22424.0
218.00	2621.0
220.30	1015.0
221.00	11983.0
221.80	2893.0
223.10	5878.0
224.00	42744.0
225.00	11367.0
226.10	1008.0
227.00	22632.0
228.00	3457.0
228.90	3532.0
230.00	589.0
230.90	2570.0
232.10	351.0
232.90	360.0
234.00	1398.0
234.90	1323.0
235.90	946.0
237.00	1663.0
238.10	537.0
238.90	1000.0
240.00	712.0
240.90	1514.0
242.00	2546.0
243.00	1966.0
244.00	28792.0
245.00	6942.0
246.00	9191.0

Instrument :
BNA_M
ClientSampleId :
DFTPP74

246.90	1205.0
248.90	623.0
251.90	432.0
252.90	734.0
253.90	641.0
254.10	649.0
255.00	165056.0
256.00	24784.0
257.00	2834.0
258.00	10549.0
259.00	1537.0
264.10	366.0
265.00	4092.0
265.90	882.0
270.00	468.0
271.00	696.0
272.00	484.0
273.00	4603.0
274.10	13298.0
275.00	77624.0
276.10	9518.0
277.00	7653.0
277.90	1156.0
281.10	397.0
281.90	628.0
283.00	701.0
284.00	885.0
284.90	1080.0
288.80	416.0
289.70	393.0
291.00	547.0
292.00	494.0
292.90	1535.0
294.10	476.0
295.10	389.0
296.00	21776.0
297.20	2121.0
301.00	462.0
302.10	662.0
303.00	2060.0
304.10	806.0
308.00	449.0
310.10	553.0
314.10	987.0
315.00	2904.0
315.90	1567.0
321.00	976.0
322.00	329.0
323.10	6427.0
323.90	1370.0
326.90	998.0
328.00	721.0
328.90	377.0
332.10	1303.0
333.00	598.0
334.10	3988.0
335.10	766.0
335.90	528.0
340.90	899.0
346.00	1478.0
352.00	1812.0
353.00	1093.0
354.00	2269.0
354.90	576.0
365.00	9948.0
365.90	1656.0
371.00	447.0
371.90	2583.0
373.00	738.0
383.10	1087.0
390.20	501.0
401.10	614.0
402.00	1454.0
403.10	1892.0
403.90	419.0
404.10	410.0
421.00	1228.0
421.90	1461.0
423.10	10343.0
424.00	3421.0
441.00	26960.0
442.10	179968.0
443.10	35400.0
444.10	2340.0

Instrument :
BNA_M
ClientSampleId :
DFTPP74

m/z	Abundance
37.10	538.0
38.00	1361.0
39.10	9436.0
40.00	1453.0
40.80	303.0
44.00	1610.0
49.00	943.0
50.10	35072.0
51.10	114544.0
52.00	5650.0
55.00	600.0
56.00	3684.0
57.00	7094.0
58.20	362.0
60.90	1248.0
62.00	1836.0
63.00	4189.0
64.00	697.0
65.10	1896.0
68.00	2505.0
69.00	117704.0
70.00	423.0
73.10	1071.0
74.00	14317.0
75.00	22664.0
76.10	7640.0
77.10	131904.0
78.10	9095.0
79.00	9428.0
79.90	6814.0
81.10	11262.0
81.90	2267.0
83.00	2190.0
83.90	427.0
84.90	1911.0
86.00	2841.0
87.00	867.0
88.00	602.0
89.00	389.0
91.10	1860.0
92.00	4141.0
93.00	18808.0
93.90	937.0
94.70	437.0
95.90	572.0
97.00	701.0
98.00	16358.0
99.00	9861.0
99.80	801.0
100.10	675.0
101.00	5856.0
102.80	1382.0
103.90	3647.0
105.00	5132.0
105.90	887.0
107.00	44032.0
108.00	5860.0
109.10	966.0
110.00	69632.0
111.00	10765.0
112.00	1602.0
113.00	933.0
116.00	1871.0
117.00	49800.0
118.00	2527.0
119.00	369.0
119.90	520.0
121.90	4011.0
123.00	5581.0
124.00	1806.0
125.10	2311.0
126.10	822.0
127.00	131136.0
128.00	10451.0
129.00	61384.0
129.90	4313.0
130.90	1389.0
132.00	508.0
132.90	860.0
134.10	1891.0
135.00	4403.0
136.00	1982.0
137.00	3426.0
138.70	432.0
140.00	923.0
141.00	8206.0
142.10	3014.0
142.90	1782.0
144.10	654.0
145.00	616.0
145.90	1176.0

Instrument :
BNA_M
ClientSampleId :
DFTPP74

146.90	4261.0
148.00	11458.0
148.90	2102.0
149.80	369.0
151.30	747.0
152.00	620.0
153.00	2508.0
153.80	1129.0
155.00	4360.0
156.00	6775.0
157.00	989.0
158.00	1196.0
159.10	896.0
160.00	3429.0
161.10	3393.0
161.80	1316.0
165.00	3435.0
166.00	3056.0
167.00	14342.0
168.00	7991.0
169.00	1473.0
170.00	346.0
170.80	779.0
172.10	828.0
172.90	1951.0
174.10	3373.0
175.10	5787.0
176.00	1616.0
177.00	3579.0
178.10	856.0
179.00	11883.0
180.00	8181.0
181.10	4606.0
181.90	853.0
183.90	1365.0
185.00	5920.0
186.00	48184.0
187.00	12501.0
188.00	1353.0
189.00	3462.0
191.00	1800.0
192.00	3932.0
193.10	4729.0
193.90	1064.0
195.10	301.0
196.00	8318.0
196.80	1581.0
198.00	286848.0
199.00	20272.0
200.00	1343.0
200.80	329.0
201.30	1236.0
202.90	3237.0
204.10	9509.0
205.00	19024.0
206.00	85824.0
207.00	13529.0
207.90	3520.0
208.90	704.0
210.00	1609.0
210.90	3064.0
211.90	411.0
215.10	1858.0
216.00	3499.0
217.00	24728.0
217.90	2907.0
219.10	587.0
221.00	10999.0
221.70	2828.0
223.00	5287.0
224.00	47744.0
225.00	10835.0
225.90	1348.0
227.00	23104.0
228.10	4201.0
229.00	4120.0
230.00	379.0
231.10	2359.0
231.80	457.0
234.00	1366.0
235.10	2335.0
236.00	1110.0
237.10	2409.0
238.00	478.0
238.90	903.0
239.90	907.0
241.10	1307.0
242.10	2162.0
243.10	2753.0
244.10	36120.0
245.00	4879.0
246.00	7787.0
247.10	2261.0
248.80	951.0
249.80	362.0

Instrument :
BNA_M
ClientSampleId :
DFTPP74

250.00	358.0
252.00	754.0
253.00	1286.0
254.10	1638.0
255.00	200704.0
256.00	29408.0
257.00	2618.0
258.00	13213.0
259.10	2461.0
260.00	436.0
261.00	450.0
265.00	5809.0
265.90	921.0
270.00	397.0
271.00	1134.0
272.00	325.0
273.00	5680.0
274.00	15445.0
275.00	88440.0
275.90	10061.0
276.90	7455.0
278.10	1598.0
278.90	564.0
281.10	693.0
282.00	492.0
282.90	415.0
284.00	551.0
285.20	1195.0
285.90	412.0
289.00	598.0
289.80	541.0
291.80	306.0
293.00	2186.0
294.00	558.0
295.00	645.0
296.00	30392.0
297.00	3942.0
301.00	576.0
302.00	337.0
303.00	2683.0
304.00	742.0
307.90	427.0
310.20	427.0
314.10	1099.0
314.90	2641.0
315.90	1753.0
316.90	303.0
321.00	800.0
323.10	9418.0
324.00	2148.0
326.80	823.0
328.10	944.0
332.10	969.0
333.20	975.0
334.00	5310.0
334.90	1002.0
340.80	605.0
346.00	1653.0
352.10	3063.0
352.90	1521.0
354.00	2884.0
355.10	690.0
365.00	14545.0
366.00	1838.0
371.00	561.0
372.10	3906.0
373.00	896.0
383.10	1140.0
385.10	489.0
390.10	818.0
391.10	806.0
401.10	347.0
402.00	1456.0
403.00	1807.0
403.90	774.0
421.10	1909.0
422.10	2277.0
423.10	14844.0
424.00	3145.0
425.10	415.0
441.10	37216.0
442.10	264000.0
443.10	54344.0
444.10	4761.0

Instrument :
BNA_M
ClientSampleId :
DFTPP74

m/z	Abundance
38.00	843.0
39.10	3555.0
40.00	888.0
44.00	2090.0
48.90	1033.0
50.00	18656.0
51.10	57784.0
52.00	2984.0
55.00	533.0
56.10	1846.0
57.00	3561.0
60.90	979.0
62.00	301.0
63.00	3176.0
65.10	1242.0
68.10	1101.0
69.00	62440.0
69.90	398.0
73.10	1645.0
74.00	6425.0
75.10	9806.0
76.00	2872.0
77.10	72096.0
78.10	5434.0
79.00	6271.0
80.00	4295.0
80.90	6557.0
82.00	1249.0
83.00	1040.0
85.00	1027.0
85.90	2241.0
87.10	736.0
91.00	1622.0
92.00	1987.0
93.00	9372.0
93.80	682.0
96.10	635.0
97.10	301.0
98.00	8606.0
99.00	5745.0
100.20	477.0
101.10	3529.0
102.90	2054.0
104.00	2890.0
104.90	2093.0
106.10	818.0
107.00	23720.0
108.00	2793.0
109.10	485.0
110.00	37440.0
111.00	7200.0
111.80	973.0
113.10	310.0
114.70	506.0
115.90	1341.0
117.00	27400.0
118.10	2153.0
122.00	2307.0
123.10	2584.0
124.10	1165.0
125.10	954.0
126.10	546.0
127.00	77400.0
128.00	7051.0
129.00	32640.0
130.00	2453.0
131.00	611.0
134.00	1219.0
135.00	2740.0
136.00	1197.0
136.80	1115.0
137.90	408.0
140.20	337.0
141.00	5223.0
141.90	1412.0
142.90	1371.0
146.00	1195.0
147.00	3266.0
148.10	6403.0
149.10	1224.0
150.10	611.0
151.00	602.0
152.10	462.0
152.90	1409.0
154.10	1213.0
155.00	2569.0
156.00	2998.0
157.10	437.0
158.00	740.0
158.80	474.0
159.90	1200.0

Instrument :
BNA_M
ClientSampleId :
DFTPP74

161.20	1881.0
161.90	500.0
164.20	398.0
164.90	1982.0
166.00	1230.0
167.10	8357.0
168.00	5803.0
168.80	682.0
170.10	308.0
171.10	431.0
171.90	936.0
173.00	1393.0
174.00	1861.0
175.10	3778.0
175.80	697.0
176.50	438.0
177.10	2265.0
178.00	692.0
179.00	6642.0
180.00	5479.0
181.00	2102.0
182.70	408.0
183.80	707.0
185.00	4598.0
186.00	27528.0
187.00	7031.0
188.00	1583.0
188.90	2048.0
190.80	1975.0
191.90	2868.0
193.00	1920.0
194.10	978.0
196.00	4226.0
197.00	551.0
198.00	179008.0
199.00	14195.0
200.00	964.0
200.50	376.0
201.20	492.0
201.70	345.0
202.90	1814.0
204.00	7087.0
205.00	13969.0
206.00	50168.0
207.00	9447.0
208.00	2236.0
209.10	945.0
210.00	1047.0
210.90	2361.0
214.90	1045.0
216.10	1526.0
217.00	15835.0
218.00	1661.0
220.00	350.0
221.00	6542.0
221.70	2277.0
223.10	2465.0
224.00	30240.0
225.00	9436.0
226.10	762.0
227.00	16448.0
227.90	1632.0
229.10	2925.0
230.00	703.0
230.90	1430.0
231.80	310.0
233.00	347.0
233.90	737.0
234.90	1032.0
236.20	467.0
237.00	1304.0
238.10	360.0
239.00	595.0
239.90	975.0
241.00	1280.0
242.10	1807.0
243.00	1485.0
244.10	22304.0
245.00	2350.0
246.00	6306.0
246.80	531.0
247.10	542.0
249.00	886.0
252.90	841.0
254.10	382.0
255.00	131008.0
256.10	16536.0
257.00	1948.0
258.00	7663.0
259.00	1213.0
260.00	348.0
265.00	3512.0
265.90	381.0
267.10	405.0
271.00	740.0

Instrument :

BNA_M

ClientSampleId :

DFTPP74

271.90	815.0
272.90	3025.0
274.00	10431.0
275.00	62208.0
276.00	7676.0
277.00	6126.0
278.10	1095.0
279.10	356.0
283.10	775.0
285.10	805.0
285.90	644.0
291.90	498.0
293.10	1413.0
296.00	18296.0
297.10	2556.0
302.10	429.0
303.00	2005.0
304.20	391.0
313.90	1272.0
314.90	2133.0
316.00	1358.0
320.90	473.0
321.80	306.0
323.10	5820.0
323.90	1032.0
326.80	870.0
328.20	457.0
332.90	423.0
334.10	3588.0
335.10	1355.0
341.10	560.0
345.90	555.0
352.10	2000.0
353.00	972.0
354.00	1132.0
355.00	738.0
365.00	9285.0
366.00	1268.0
370.10	323.0
371.30	722.0
372.20	3433.0
373.00	918.0
373.20	947.0
383.20	890.0
383.90	321.0
400.90	325.0
402.00	954.0
403.20	1740.0
403.90	536.0
421.20	1046.0
422.10	2077.0
423.10	8679.0
424.00	2125.0
441.10	31800.0
442.10	196416.0
443.00	38928.0
444.10	4193.0

Instrument :
BNA_M
ClientSampleId :
DFTPP74

Scan 2439 (17.433 min): BM013576.D
DFTPP74

m/z	Abundance
39.90	398.0
43.90	1114.0
73.00	577.0
115.10	367.0
146.80	310.0
191.10	382.0
207.00	1747.0
208.10	823.0
281.10	548.0

Instrument :
BNA_M
ClientSampleId :
DFTPP74