

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011623\
 Data File : BM038450.D
 Acq On : 16 Jan 2023 13:10
 Operator : CG/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC2.5

Quant Time: Jan 17 02:14:04 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM011623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 02:09:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.981	152	42841	20.000	ng	0.00
21) Naphthalene-d8	10.792	136	166536	20.000	ng	0.00
39) Acenaphthene-d10	14.615	164	84356	20.000	ng	0.00
64) Phenanthrene-d10	17.356	188	193018	20.000	ng	0.00
76) Chrysene-d12	21.527	240	214046	20.000	ng	0.00
86) Perylene-d12	23.950	264	285679	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						Qvalue
19) n-Nitroso-di-n-propyla...	8.792	70	5519	1.789	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011623\
Data File : BM038450.D
Acq On : 16 Jan 2023 13:10
Operator : CG/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC2.5

Quant Time: Jan 17 02:14:04 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM011623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 17 02:09:55 2023
Response via : Initial Calibration

