

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011624\
 Data File : BM043849.D
 Acq On : 15 Jan 2024 20:24
 Operator : MA/JU
 Sample : P1091-16
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0B38

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/16/2024
 Supervised By :mohammad ahmed 01/18/2024

Quant Time: Jan 16 00:42:12 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM011024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 10 15:40:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.950	152	6423	0.400	ng/u1	0.00
4) Naphthalene-d8	10.757	136	20000	0.400	ng/u1	-0.02
9) Acenaphthene-d10	14.592	164	12182	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.341	188	25326m	0.400	ng/u1	-0.01
17) Chrysene-d12	21.542	240	15882m	0.400	ng/u1	0.00
23) Perylene-d12	23.959	264	22748	0.400	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.355	96	33658	4.334	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.352	152	9458	0.327	ng/u1	-0.02
18) Fluoranthene-d10	19.371	212	20752	0.404	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011624\
Data File : BM043849.D
Acq On : 15 Jan 2024 20:24
Operator : MA/JU
Sample : P1091-16
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0B38

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/16/2024
Supervised By :mohammad ahmed 01/18/2024

Quant Time: Jan 16 00:42:12 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM011024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 10 15:40:11 2024
Response via : Initial Calibration

