

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011624\
 Data File : BM043856.D
 Acq On : 16 Jan 2024 10:32
 Operator : MA/JU
 Sample : PB158427BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK427

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 01/17/2024
 Supervised By :mohammad ahmed 01/18/2024

Quant Time: Jan 16 11:04:08 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM011024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 10 15:40:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.946	152	8905	0.400	ng/u1	-0.01
4) Naphthalene-d8	10.752	136	27580	0.400	ng/u1	-0.02
9) Acenaphthene-d10	14.592	164	15439	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.341	188	32134m	0.400	ng/u1	-0.01
17) Chrysene-d12	21.553	240	19592m	0.400	ng/u1	0.01
23) Perylene-d12	23.965	264	18603m	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.355	96	55906	5.192	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.368	152	10354	0.260	ng/u1	0.00
18) Fluoranthene-d10	19.375	212	18734	0.296	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011624\
Data File : BM043856.D
Acq On : 16 Jan 2024 10:32
Operator : MA/JU
Sample : PB158427BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK427

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 01/17/2024
Supervised By :mohammad ahmed 01/18/2024

Quant Time: Jan 16 11:04:08 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM011024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 10 15:40:11 2024
Response via : Initial Calibration

