

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011725\
 Data File : BM049356.D
 Acq On : 17 Jan 2025 15:02
 Operator : RC/JU
 Sample : SSTD3.241
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.2025

Quant Time: Jan 17 15:39:34 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM011725.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 17 15:07:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.527	152	6287	0.400	ng/ul	0.00
4) Naphthalene-d8	10.287	136	18461	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.164	164	10330	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.917	188	19940	0.400	ng/ul	0.00
17) Chrysene-d12	21.122	240	13685	0.400	ng/ul	0.00
23) Perylene-d12	23.273	264	12845	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.126	96	21172	2.826	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.160	88	23224	2.738	ng/ul#	82
14) Pentachlorophenol	16.575	266	15682	4.381	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011725\
Data File : BM049356.D
Acq On : 17 Jan 2025 15:02
Operator : RC/JU
Sample : SST3.241
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD3.2025

Quant Time: Jan 17 15:39:34 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM011725.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 17 15:07:34 2025
Response via : Initial Calibration

