


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM011816\
Data File : BM004068.D
Acq On : 18 Jan 2016 12:45
Operator : SJ/UM
Sample : DFTPP40
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM010116.M
Title : SVOA CALIBRATION
Last Update : Sat Jan 16 03:00:58 2016



AutoFind: Scans 2452, 2453, 2454; Background Corrected with Scan 2447

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.1	16180	PASS
68	69	0.00	2	2.0	406	PASS
69	198	0.00	100	35.0	20120	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	46.7	26856	PASS
197	198	0.00	2	0.2	126	PASS
198	198	100	100	100.0	57520	PASS
199	198	5	9	6.8	3936	PASS
275	198	10	60	24.9	14307	PASS
365	198	1	100	2.8	1605	PASS
441	443	0.01	100	79.9	6755	PASS
442	198	50	100	76.2	43834	PASS
443	442	15	24	19.3	8456	PASS

Scan 2447 (17.480 min): BM004068.D
DFTPP40

m/z	Abundance
43.00	643.0
44.00	415.0
45.00	304.0

Instrument :
BNA_M
ClientSampleId :

m/z	Abundance
39.10	1246.0
41.10	337.0
43.00	578.0
44.00	367.0
50.10	3578.0
51.10	13847.0
52.10	802.0
55.00	334.0
56.00	485.0
57.00	1262.0
63.10	650.0
65.00	374.0
68.00	409.0
69.00	16968.0
73.00	331.0
74.00	1637.0
75.00	2651.0
76.10	918.0
77.10	19624.0
78.10	1456.0
79.00	1289.0
80.00	1057.0
81.00	1563.0
82.00	363.0
83.00	416.0
85.00	470.0
86.00	399.0
91.00	403.0
92.00	392.0
93.10	2073.0
98.00	1657.0
99.00	1238.0
101.00	834.0
104.00	466.0
105.00	587.0
107.10	6168.0
108.00	1081.0
110.00	11016.0
111.00	1748.0
116.00	337.0
117.00	4495.0
118.00	390.0
121.90	386.0
123.00	731.0
123.90	338.0
127.10	21000.0
128.00	1620.0
129.00	8258.0
130.00	704.0
135.00	767.0
137.00	360.0
141.00	1060.0
142.00	348.0
147.00	646.0
148.00	1201.0
149.00	316.0
153.00	333.0
155.00	617.0
156.10	911.0
160.00	333.0
161.00	530.0
165.00	462.0
166.00	347.0
167.10	2212.0
168.00	1093.0
174.00	486.0
175.10	853.0
177.00	380.0
179.00	1546.0
180.00	1185.0
181.00	490.0
185.00	730.0
186.10	5889.0
187.00	1858.0
189.00	363.0
192.00	461.0
193.00	572.0
196.00	1044.0
198.00	41592.0
199.00	2907.0
203.00	317.0
204.10	1448.0
205.10	2725.0
206.10	11046.0
207.10	1672.0
208.00	380.0
211.10	437.0
217.00	2796.0
218.00	339.0
221.10	1589.0
221.80	429.0

Instrument :
BNA_M
ClientSampleId :

223.00	601.0
224.10	5860.0
225.10	1467.0
227.00	2488.0
228.00	394.0
229.10	465.0
243.00	366.0
244.10	4252.0
245.10	617.0
246.00	885.0
255.10	22552.0
256.00	3541.0
258.10	1290.0
265.10	519.0
273.00	672.0
274.00	1762.0
275.10	9683.0
276.10	1393.0
277.00	930.0
296.10	2587.0
297.10	365.0
323.10	755.0
334.00	539.0
365.10	1009.0
372.10	371.0
423.10	1297.0
424.10	357.0
441.10	3504.0
442.10	23896.0
443.10	4675.0
444.20	529.0

Instrument :
BNA_M
ClientSampleId :

m/z	Abundance
39.10	1543.0
41.00	310.0
43.00	595.0
44.00	383.0
45.00	307.0
50.10	4759.0
51.10	19136.0
52.10	1054.0
55.00	328.0
56.00	668.0
57.00	1659.0
62.00	335.0
63.00	927.0
65.10	505.0
68.10	432.0
69.00	23448.0
73.00	317.0
74.00	2443.0
75.00	3768.0
76.10	1309.0
77.10	28176.0
78.10	1914.0
79.00	1795.0
80.00	1411.0
81.00	2042.0
82.00	535.0
83.00	500.0
85.00	635.0
86.00	586.0
91.00	496.0
92.00	496.0
93.00	3128.0
98.00	2495.0
99.00	1909.0
101.00	1226.0
103.00	406.0
104.00	771.0
105.00	727.0
107.00	9209.0
108.10	1371.0
109.00	305.0
110.00	16305.0
111.00	2388.0
112.00	315.0
116.00	509.0
117.00	6979.0
118.00	549.0
122.00	574.0
123.00	1147.0
124.00	450.0
125.00	403.0
127.00	31880.0
128.00	2476.0
129.00	12245.0
130.00	1161.0
134.00	370.0
135.00	1040.0
136.00	478.0
137.00	605.0
141.00	1519.0
142.00	509.0
142.90	366.0
146.00	318.0
147.00	947.0
148.00	1825.0
149.10	448.0
153.00	536.0
153.90	414.0
155.00	922.0
156.00	1367.0
157.90	343.0
160.00	455.0
161.00	756.0
165.00	676.0
166.00	503.0
167.00	3365.0
168.00	1556.0
173.00	447.0
174.00	728.0
175.10	1336.0
176.00	388.0
177.00	512.0
179.00	2374.0
180.00	1715.0
181.00	761.0
185.00	1202.0
186.10	9292.0
187.10	2629.0
189.00	651.0
192.00	789.0
193.10	935.0

Instrument :
BNA_M
ClientSampleId :

196.00	1949.0
196.80	329.0
198.00	67096.0
199.00	4573.0
200.00	348.0
203.00	484.0
204.00	2475.0
205.10	4168.0
206.10	17800.0
207.10	2529.0
208.00	565.0
210.20	310.0
211.10	564.0
216.10	399.0
217.00	4503.0
218.00	595.0
221.10	2580.0
221.80	624.0
223.00	924.0
224.10	9598.0
225.00	2368.0
227.10	3887.0
228.00	565.0
229.00	924.0
231.00	377.0
237.00	348.0
242.00	491.0
243.00	465.0
244.10	7197.0
245.10	984.0
246.00	1543.0
247.10	303.0
254.10	306.0
255.10	37720.0
256.10	5598.0
257.00	408.0
258.00	2209.0
259.00	324.0
265.00	843.0
273.00	1027.0
274.10	2820.0
275.10	16536.0
276.10	2203.0
277.00	1372.0
293.00	351.0
296.10	4459.0
297.00	554.0
303.10	422.0
315.00	477.0
323.10	1364.0
324.10	321.0
334.10	829.0
346.10	307.0
352.10	390.0
353.10	302.0
354.10	524.0
365.10	1877.0
372.10	631.0
403.10	466.0
421.00	341.0
422.10	353.0
423.10	2510.0
424.10	647.0
441.10	7612.0
442.10	49120.0
443.10	9653.0
444.10	944.0

Instrument :
BNA_M
ClientSampleId :

m/z	Abundance
39.00	1363.0
41.10	353.0
43.00	617.0
44.00	384.0
50.10	4046.0
51.10	15558.0
52.10	869.0
55.00	377.0
56.00	554.0
57.00	1313.0
63.10	731.0
65.10	423.0
68.10	377.0
69.00	19944.0
73.00	343.0
74.00	1950.0
75.00	3270.0
76.10	1177.0
77.10	23688.0
78.10	1557.0
79.00	1611.0
80.00	1274.0
81.00	1849.0
82.00	490.0
83.10	447.0
85.00	566.0
86.00	524.0
91.00	420.0
92.00	427.0
93.00	2577.0
98.00	2174.0
99.00	1728.0
101.00	1013.0
103.00	358.0
104.00	610.0
105.00	699.0
107.00	7538.0
108.00	1288.0
110.00	14287.0
111.00	2173.0
116.00	450.0
117.00	5949.0
118.00	413.0
122.00	566.0
123.00	1005.0
124.00	422.0
125.00	407.0
127.00	27688.0
128.10	2195.0
129.00	10627.0
130.00	906.0
134.00	333.0
135.00	957.0
136.00	336.0
137.00	519.0
141.00	1356.0
142.00	490.0
142.90	347.0
147.00	835.0
148.00	1686.0
149.00	444.0
153.00	462.0
154.00	330.0
155.00	804.0
156.00	1250.0
157.90	303.0
160.00	418.0
161.00	640.0
165.00	588.0
166.00	469.0
167.00	3042.0
168.00	1506.0
173.00	389.0
174.00	574.0
175.00	1244.0
176.00	367.0
177.00	613.0
179.00	2132.0
180.10	1523.0
181.00	676.0
185.10	1102.0
186.00	8505.0
187.00	2606.0
189.00	539.0
192.00	741.0
193.10	831.0
196.00	1809.0
197.10	378.0
198.00	63872.0
199.00	4328.0
203.00	427.0

Instrument :
BNA_M
ClientSampleId :

204.10	2316.0
205.10	3748.0
206.10	17224.0
207.10	2241.0
208.00	539.0
211.10	623.0
216.00	361.0
217.00	4442.0
218.00	503.0
221.10	2526.0
223.00	896.0
224.10	9648.0
225.10	2364.0
227.00	3738.0
228.00	574.0
229.00	887.0
231.00	341.0
237.00	317.0
242.00	453.0
243.10	543.0
244.10	7129.0
245.10	964.0
246.00	1492.0
249.00	321.0
254.20	324.0
255.10	37880.0
256.10	5444.0
257.00	428.0
258.00	2120.0
258.90	320.0
265.00	855.0
273.00	1051.0
274.10	2996.0
275.10	16704.0
276.10	2201.0
277.00	1427.0
296.10	4705.0
297.00	786.0
303.10	635.0
315.10	549.0
323.10	1520.0
334.10	1063.0
346.00	364.0
352.10	490.0
353.00	334.0
354.10	495.0
365.10	1930.0
372.10	904.0
402.10	337.0
403.00	485.0
421.00	467.0
422.10	426.0
423.10	2911.0
424.10	716.0
441.10	9149.0
442.10	58488.0
443.10	11040.0
444.10	1051.0

Instrument :
BNA_M
ClientSampleId :