

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011822\
 Data File : BM033761.D
 Acq On : 18 Jan 2022 13:25
 Operator : CG/JU
 Sample : PB142061BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB142061BL

Quant Time: Jan 19 04:18:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM011722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 17 16:30:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.954	152	163147	20.000	ng	0.00
21) Naphthalene-d8	10.766	136	672975	20.000	ng	0.00
39) Acenaphthene-d10	14.589	164	458223	20.000	ng	0.00
64) Phenanthrene-d10	17.324	188	986715	20.000	ng	0.00
76) Chrysene-d12	21.483	240	905142	20.000	ng	-0.01
86) Perylene-d12	23.894	264	861739	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.501	112	1314548	135.390	ng	0.00
7) Phenol-d6	7.113	99	1686669	123.307	ng	0.00
23) Nitrobenzene-d5	9.119	82	1111913	81.310	ng	0.00
42) 2,4,6-Tribromophenol	16.071	330	711919	114.596	ng	0.00
45) 2-Fluorobiphenyl	13.224	172	2600154	79.830	ng	0.00
79) Terphenyl-d14	19.936	244	4213363	77.193	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011822\
Data File : BM033761.D
Acq On : 18 Jan 2022 13:25
Operator : CG/JU
Sample : PB142061BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB142061BL

Quant Time: Jan 19 04:18:42 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM011722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 17 16:30:56 2022
Response via : Initial Calibration

