

Data Path : Z:\HPCHEM1\BNA_M\Data\BM012015\
 Data File : BM000272.D
 Acq On : 20 Jan 2015 20:24
 Operator : TP/IZ
 Sample : SSTD02039
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 LabSampleId :
 SSTD02039

Quant Time: Jan 21 01:33:48 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-BM122914.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 21 00:20:03 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.82	152	121427	20.00	ng/ul	0.00
18) Naphthalene-d8	10.61	136	472011	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.46	164	299175	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.20	188	630434	20.00	ng/ul	0.00
75) Chrysene-d12	21.39	240	722361	20.00	ng/ul	0.00
83) Perylene-d12	23.67	264	694786	20.00	ng/ul	-0.01

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.31	96	20628	9.85	ng/uL	0.00
5) Phenol-d5	6.98	99	187660	18.37	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.15	67	131161	20.31	ng/ul	0.00
9) 2-Chlorophenol-d4	7.35	132	139675	18.65	ng/ul	0.00
13) 4-Methylphenol-d8	8.52	113	144047	17.73	ng/ul	0.00
19) Nitrobenzene-d5	8.97	128	72586	22.24	ng/ul	0.00
22) 2-Nitrophenol-d4	9.69	143	69101	20.74	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.23	165	138153	18.77	ng/ul	0.00
29) 4-Chloroaniline-d4	10.75	131	178692	21.39	ng/ul	0.00
43) Dimethylphthalate-d6	13.86	166	427806	19.39	ng/ul	0.00
46) Acenaphthylene-d8	14.15	160	541208	20.47	ng/ul	0.00
51) 4-Nitrophenol-d4	14.65	143	60281	16.79	ng/ul	0.00
57) Fluorene-d10	15.45	176	379484	20.03	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.57	200	55667	18.63	ng/ul	0.00
70) Anthracene-d10	17.30	188	587071	20.14	ng/ul	0.00
76) Pyrene-d10	19.59	212	651734	19.56	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.53	264	644024	20.42	ng/ul	-0.01

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.35	88	27492	9.47	ng/uL	95
4) Benzaldehyde	6.96	77	152666	21.45	ng/ul	92
6) Phenol	7.00	94	198818	18.45	ng/ul	95
8) Bis(2-Chloroethyl)ether	7.25	93	162741	19.46	ng/ul	95
10) 2-Chlorophenol	7.39	128	146958	18.75	ng/ul	93
11) 2-Methylphenol	8.26	108	147049	17.73	ng/ul	96
12) 2,2'-oxybis(1-Chloropropan	8.35	45	304026	22.12	ng/ul	98
14) Acetophenone	8.64	105	262301	18.37	ng/ul#	93
15) N-Nitroso-di-n-propylamine	8.62	70	137010	18.45	ng/ul#	86
16) 4-Methylphenol	8.58	108	152986	17.08	ng/ul	98
17) Hexachloroethane	8.90	117	75391	21.05	ng/ul	93
20) Nitrobenzene	9.02	77	230911	23.73	ng/ul	97
21) Isophorone	9.53	82	365012	19.46	ng/ul	98
23) 2-Nitrophenol	9.73	139	77952	21.19	ng/ul	90
24) 2,4-Dimethylphenol	9.78	107	192608	19.29	ng/ul	94
25) Bis(2-Chloroethoxy)methane	10.02	93	207269	19.85	ng/ul	99
27) 2,4-Dichlorophenol	10.26	162	140948	19.18	ng/ul	99
28) Naphthalene	10.66	128	472663	19.64	ng/ul	99
30) 4-Chloroaniline	10.77	127	185455	21.58	ng/ul	99
31) Hexachlorobutadiene	10.95	225	109815	21.08	ng/ul	94
33) 4-Chloro-3-methylphenol	11.89	107	160577	17.81	ng/ul	99
34) 2-Methylnaphthalene	12.27	142	336541	18.86	ng/ul	97
36) 1,2,4,5-Tetrachlorobenzene	12.65	216	184455	22.24	ng/ul#	96


Data Path : Z:\HPCHEM1\BNA_M\Data\BM012015\
 Data File : BM000272.D
 Acq On : 20 Jan 2015 20:24
 Operator : TP/IZ
 Sample : SSTD02039
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 LabSampleId :
 SSTD02039

Quant Time: Jan 21 01:33:48 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-BM122914.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 21 00:20:03 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) Hexachlorocyclopentadiene	12.62	237	115098	22.31	ng/ul	99
38) 2,4,6-Trichlorophenol	12.89	196	107895	20.39	ng/ul	95
39) 2,4,5-Trichlorophenol	12.95	196	116053	20.20	ng/ul	98
40) 1,1'-Biphenyl	13.29	154	448336	20.79	ng/ul	99
41) 2-Chloronaphthalene	13.33	162	356931	21.60	ng/ul	97
42) 2-Nitroaniline	13.53	65	119538	21.22	ng/ul	95
44) Dimethylphthalate	13.91	163	431927	19.65	ng/ul	98
45) 2,6-Dinitrotoluene	14.03	165	79633	20.26	ng/ul#	82
47) Acenaphthylene	14.18	152	568663	20.40	ng/ul	99
48) 3-Nitroaniline	14.36	138	83829	20.09	ng/ul	99
49) Acenaphthene	14.52	153	363601	20.47	ng/ul	100
50) 2,4-Dinitrophenol	14.57	184	32885	17.15	ng/ul#	77
52) 4-Nitrophenol	14.66	109	96358	20.23	ng/ul	91
53) Dibenzofuran	14.86	168	522840	20.32	ng/ul	93
54) 2,4-Dinitrotoluene	14.82	165	124144	21.16	ng/ul	98
55) 2,3,4,6-Tetrachlorophenol	15.08	232	96447	19.46	ng/ul	98
56) Diethylphthalate	15.27	149	449685	19.47	ng/ul	98
58) Fluorene	15.50	166	435586	20.40	ng/ul	98
59) 4-Chlorophenyl-phenylether	15.50	204	214959	20.42	ng/ul	98
60) 4-Nitroaniline	15.52	138	91200	19.51	ng/ul	98
63) 4,6-Dinitro-2-methylphenol	15.58	198	56824	18.46	ng/ul#	88
64) N-Nitrosodiphenylamine	15.71	169	359870	19.68	ng/ul	96
65) 4-Bromophenyl-phenylether	16.39	248	131538	20.47	ng/ul#	89
66) Hexachlorobenzene	16.51	284	153414	20.71	ng/ul	96
67) Atrazine	16.66	200	126467	17.79	ng/ul	95
68) Pentachlorophenol	16.85	266	71387	16.00	ng/ul	92
69) Phenanthrene	17.24	178	691344	20.26	ng/ul	99
71) Anthracene	17.33	178	715875	20.40	ng/ul	99
72) Carbazole	17.60	167	627356	19.79	ng/ul	99
73) Di-n-butylphthalate	18.16	149	715028	18.96	ng/ul#	98
74) Fluoranthene	19.26	202	814644	20.42	ng/ul	99
77) Pyrene	19.62	202	875269	20.10	ng/ul	98
78) Butylbenzylphthalate	20.52	149	309605	17.50	ng/ul	91
79) 3,3'-Dichlorobenzidine	21.30	252	228839	17.97	ng/ul	97
80) Benzo(a)anthracene	21.37	228	842847	20.21	ng/ul	99
81) Bis(2-ethylhexyl)phthalate	21.30	149	469067	18.22	ng/ul	99
82) Chrysene	21.42	228	784835	20.42	ng/ul	97
84) Di-n-octyl phthalate	22.19	149	820839	18.20	ng/ul	100
85) Benzo(b)fluoranthene	22.99	252	832104	18.77	ng/ul	99
86) Benzo(k)fluoranthene	22.99	252	464672	12.28	ng/ul	99
88) Benzo(a)pyrene	23.57	252	789176	20.26	ng/ul	98
89) Indeno(1,2,3-cd)pyrene	25.99	276	901520	22.19	ng/ul	97
90) Dibenzo(a,h)anthracene	26.01	278	758537	22.17	ng/ul	97
91) Benzo(g,h,i)perylene	26.71	276	762683	22.85	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

1,4-Dioxane

Concen: 9.47 ng/uL

RT: 3.35 min Scan# 112





Delta R.T. 0.00 min

Lab File: BM000272.D

Acq: 20 Jan 2015 20:24

Instrument :
LabSampleId :
SSTD02039

Tgt Ion:	88	Resp:	27492
Ion	Ratio	Lower	Upper
88	100		
43	42.9	34.9	52.3
58	77.6	67.3	100.9



#4

Benzaldehyde

Concen: 21.45 ng/uL




RT: 6.96 min Scan# 727


Delta R.T. 0.00 min

Lab File: BM000272.D

Acq: 20 Jan 2015 20:24

Tgt Ion:	77	Resp:	152666
Ion	Ratio	Lower	Upper
77	100		
105	84.2	73.5	110.3
106	77.3	67.9	101.9





#6

Phenol

Concen: 18.45 ng/ul

RT: 7.00 min Scan# 734

Delta R.T. -0.01 min

Lab File: BM000272.D

Acq: 20 Jan 2015 20:24

Instrument :
LabSampled :
SSTD02039


Tgt Ion: 94 Resp: 198818

Ion Ratio Lower Upper

94 100

65 39.4 27.4 41.0




66 54.9 42.5 63.7



Abundance Ion 94.00 (93.70 to 94.70): BM000272.D (-701) (-)

Ion 65.00 (64.70 to 65.70): BM000272.D (-701) (-)

Ion 66.00 (65.70 to 66.70): BM000272.D (-701) (-)



#8

Bis(2-Chloroethyl)ether

Concen: 19.46 ng/ul

RT: 7.25 min Scan# 775

Delta R.T. 0.00 min

Lab File: BM000272.D

Acq: 20 Jan 2015 20:24


Tgt Ion: 93 Resp: 162741

Ion Ratio Lower Upper

93 100

63 92.0 68.9 103.3



95 31.9 24.6 37.0




Abundance Ion 93.00 (92.70 to 93.70): BM000272.D (-741) (-)

Ion 63.00 (62.70 to 63.70): BM000272.D (-741) (-)

Ion 95.00 (94.70 to 95.70): BM000272.D (-741) (-)





#10

2-Chlorophenol

Concen: 18.75 ng/ul

RT: 7.39 min Scan# 799

Delta R.T. 0.01 min

Lab File: BM000272.D

Acq: 20 Jan 2015 20:24

Instrument :
LabSampled :
SSTD02039


Tgt Ion:128 Resp: 146958

Ion Ratio Lower Upper




128 100

64 64.1 47.2 70.8

130 35.7 25.2 37.8



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 64.00 (63.70 to 64.70): BM0
Ion 130.00 (129.70 to 130.70): E



#11

2-Methylphenol

Concen: 17.73 ng/ul

RT: 8.26 min Scan# 947

Delta R.T. 0.00 min

Lab File: BM000272.D


Acq: 20 Jan 2015 20:24

Tgt Ion:108 Resp: 147049



Ion Ratio Lower Upper


108 100

107 85.0 70.7 106.1



Abundance Ion 108.00 (107.70 to 108.70): E
Ion 107.00 (106.70 to 107.70): E









#12
 2,2'-oxybis(1-Chloropropane)
 Concen: 22.12 ng/ul
 RT: 8.35 min Scan# 963
 Delta R.T. 0.00 min
 Lab File: BM000272.D
 Acq: 20 Jan 2015 20:24




Instrument :
 LabSampleId :
 SSTD02039




Tgt Ion	Ratio	Lower	Upper
45	100		
77	12.4	9.9	14.9
79	9.1	8.3	12.5



#14
 Acetophenone
 Concen: 18.37 ng/ul
 RT: 8.64 min Scan# 1012
 Delta R.T. 0.00 min
 Lab File: BM000272.D
 Acq: 20 Jan 2015 20:24

Tgt Ion	Ratio	Lower	Upper
105	100		
77	86.8	66.3	99.5
51	40.7	26.5	39.7









#15
N-Nitroso-di-n-propylamine
Concen: 18.45 ng/ul
RT: 8.62 min Scan# 1009
Delta R.T. 0.00 min
Lab File: BM000272.D
Acq: 20 Jan 2015 20:24


Instrument :
LabSampleId :
SSTD02039


Tgt Ion: 70 Resp: 137010
Ion Ratio Lower Upper
70 100
42 72.1 46.4 69.6#
101 8.9 7.8 11.8
130 15.6 14.6 21.8



#16
4-Methylphenol
Concen: 17.08 ng/ul
RT: 8.58 min Scan# 1002
Delta R.T. 0.00 min
Lab File: BM000272.D
Acq: 20 Jan 2015 20:24

Tgt Ion: 108 Resp: 152986
Ion Ratio Lower Upper
108 100
107 107.1 84.0 126.0





#17

Hexachloroethane

Concen: 21.05 ng/ul

RT: 8.90 min Scan# 1056


Delta R.T. 0.00 min

Lab File: BM000272.D




Acq: 20 Jan 2015 20:24

Instrument :
LabSampled :
SSTD02039

Tgt Ion:	117	Resp:	75391
Ion	Ratio	Lower	Upper
117	100		
201	100.1	85.2	127.8
199	61.2	55.1	82.7



Abundance Ion 117.00 (116.70 to 117.70): E
Ion 201.00 (200.70 to 201.70): E
Ion 199.00 (198.70 to 199.70): E



#20

Nitrobenzene

Concen: 23.73 ng/ul


RT: 9.02 min Scan# 1076

Delta R.T. -0.01 min



Lab File: BM000272.D


Acq: 20 Jan 2015 20:24

Tgt Ion:	77	Resp:	230911
Ion	Ratio	Lower	Upper
77	100		
123	34.9	29.2	43.8
65	17.0	12.4	18.6



Abundance Ion 77.00 (76.70 to 77.70): BM
Ion 123.00 (122.70 to 123.70): E
Ion 65.00 (64.70 to 65.70): BM





#21

Isophorone

Concen: 19.46 ng/ul

RT: 9.53 min Scan# 1164





Delta R.T. -0.01 min

Lab File: BM000272.D

Acq: 20 Jan 2015 20:24

Instrument :
LabSampled :
SSTD02039

Tgt Ion:	82	Resp:	365012
Ion	Ratio	Lower	Upper
82	100		
95	9.1	6.9	10.3
138	14.2	12.1	18.1



#23

2-Nitrophenol

Concen: 21.19 ng/ul




RT: 9.73 min Scan# 1197


Delta R.T. 0.00 min

Lab File: BM000272.D

Acq: 20 Jan 2015 20:24

Tgt Ion:	139	Resp:	77952
Ion	Ratio	Lower	Upper
139	100		
65	78.6	56.3	84.5
109	53.8	37.1	55.7





#24

2,4-Dimethylphenol

Concen: 19.29 ng/ul

RT: 9.78 min Scan# 1206

Delta R.T. -0.01 min

Lab File: BM000272.D

Acq: 20 Jan 2015 20:24

Instrument :
LabSampled :
SSTD02039


Tgt Ion: 107 Resp: 192608

Ion Ratio Lower Upper




107 100

121 39.3 33.0 49.6

122 68.1 59.3 88.9



Abundance Ion 107.00 (106.70 to 107.70): E
Ion 121.00 (120.70 to 121.70): E
Ion 122.00 (121.70 to 122.70): E



#25

Bis(2-Chloroethoxy)methane

Concen: 19.85 ng/ul

RT: 10.02 min Scan# 1247

Delta R.T. 0.00 min

Lab File: BM000272.D

Acq: 20 Jan 2015 20:24


Tgt Ion: 93 Resp: 207269

Ion Ratio Lower Upper



93 100


95 32.0 26.1 39.1

123 9.9 7.4 11.2



Abundance Ion 93.00 (92.70 to 93.70): BM0
Ion 95.00 (94.70 to 95.70): BM0
Ion 123.00 (122.70 to 123.70): E





#27

2,4-Dichlorophenol

Concen: 19.18 ng/ul

RT: 10.26 min Scan# 1287

Delta R.T. 0.00 min

Lab File: BM000272.D

Acq: 20 Jan 2015 20:24

Instrument :
LabSampled :
SSTD02039


Tgt Ion:162 Resp: 140948

Ion Ratio Lower Upper




162 100

164 62.9 50.2 75.4

98 40.2 31.4 47.2



Abundance Ion 162.00 (161.70 to 162.70): E
Ion 164.00 (163.70 to 164.70): E
Ion 98.00 (97.70 to 98.70): BM



#28

Naphthalene

Concen: 19.64 ng/ul

RT: 10.66 min Scan# 1356

Delta R.T. 0.00 min

Lab File: BM000272.D

Acq: 20 Jan 2015 20:24


Tgt Ion:128 Resp: 472663

Ion Ratio Lower Upper



128 100


129 11.0 8.6 13.0

127 13.2 10.3 15.5



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E









#30
4-Chloroaniline
Concen: 21.58 ng/ul
RT: 10.77 min Scan# 1374
Delta R.T. 0.00 min
Lab File: BM000272.D
Acq: 20 Jan 2015 20:24




Instrument :
LabSampleId :
SSTD02039


Tgt Ion:127 Resp: 185455
Ion Ratio Lower Upper
127 100
129 32.4 26.5 39.7



#31
Hexachlorobutadiene
Concen: 21.08 ng/ul
RT: 10.95 min Scan# 1404
Delta R.T. 0.00 min
Lab File: BM000272.D
Acq: 20 Jan 2015 20:24

Tgt Ion:225 Resp: 109815
Ion Ratio Lower Upper
225 100
223 64.6 54.4 81.6
227 70.3 51.8 77.8









#33
4-Chloro-3-methylphenol
Concen: 17.81 ng/ul
RT: 11.89 min Scan# 1565
Delta R.T. 0.00 min
Lab File: BM000272.D
Acq: 20 Jan 2015 20:24

Instrument :
LabSampleId :
SSTD02039

Tgt Ion:107 Resp: 160577
Ion Ratio Lower Upper
107 100
144 20.4 16.6 24.8
142 68.9 54.6 82.0




Abundance Ion 107.00 (106.70 to 107.70): E
Ion 144.00 (143.70 to 144.70): E
Ion 142.00 (141.70 to 142.70): E






#34
2-Methylnaphthalene
Concen: 18.86 ng/ul
RT: 12.27 min Scan# 1630
Delta R.T. 0.00 min
Lab File: BM000272.D
Acq: 20 Jan 2015 20:24

Tgt Ion:142 Resp: 336541
Ion Ratio Lower Upper
142 100
141 83.7 69.3 103.9



Abundance Ion 142.00 (141.70 to 142.70): E
Ion 141.00 (140.70 to 141.70): E









#36
 1,2,4,5-Tetrachlorobenzene
 Concen: 22.24 ng/ul
 RT: 12.65 min Scan# 1693
 Delta R.T. 0.00 min
 Lab File: BM000272.D
 Acq: 20 Jan 2015 20:24

Instrument :
 LabSampleId :
 SSTD02039

Tgt Ion:216 Resp: 184455
 Ion Ratio Lower Upper
 216 100
 214 81.6 62.6 94.0
 179 23.7 17.4 26.2
 108 20.4 13.4 20.0#




Abundance Ion 216.00 (215.70 to 216.70): E
 Ion 214.00 (213.70 to 214.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 108.00 (107.70 to 108.70): E






#37
 Hexachlorocyclopentadiene
 Concen: 22.31 ng/ul
 RT: 12.62 min Scan# 1689
 Delta R.T. -0.01 min
 Lab File: BM000272.D
 Acq: 20 Jan 2015 20:24

Tgt Ion:237 Resp: 115098
 Ion Ratio Lower Upper
 237 100
 235 61.8 50.0 75.0
 272 11.7 8.9 13.3



Abundance Ion 237.00 (236.70 to 237.70): E
 Ion 235.00 (234.70 to 235.70): E
 Ion 272.00 (271.70 to 272.70): E









#38
 2,4,6-Trichlorophenol
 Concen: 20.39 ng/ul
 RT: 12.89 min Scan# 1734
 Delta R.T. 0.00 min
 Lab File: BM000272.D
 Acq: 20 Jan 2015 20:24




Instrument :
 LabSampled :
 SSTD02039


Tgt Ion:196 Resp: 107895
 Ion Ratio Lower Upper
 196 100
 198 90.6 76.6 114.8
 200 28.1 24.5 36.7



#39
 2,4,5-Trichlorophenol
 Concen: 20.20 ng/ul
 RT: 12.95 min Scan# 1745
 Delta R.T. 0.00 min
 Lab File: BM000272.D
 Acq: 20 Jan 2015 20:24

Tgt Ion:196 Resp: 116053
 Ion Ratio Lower Upper
 196 100
 198 95.3 75.4 113.2
 200 29.2 25.0 37.4









#40
 1,1'-Biphenyl
 Concen: 20.79 ng/ul
 RT: 13.29 min Scan# 1802
 Delta R.T. -0.01 min
 Lab File: BM000272.D
 Acq: 20 Jan 2015 20:24




Instrument :
 LabSampleId :
 SSTD02039


Tgt Ion	Ratio	Lower	Upper
154	100		
153	38.3	30.5	45.7
76	12.2	8.9	13.3



#41
 2-Chloronaphthalene
 Concen: 21.60 ng/ul
 RT: 13.33 min Scan# 1809
 Delta R.T. -0.01 min
 Lab File: BM000272.D
 Acq: 20 Jan 2015 20:24

Tgt Ion	Ratio	Lower	Upper
162	100		
164	33.5	25.8	38.6
127	43.0	35.9	53.9





#42

2-Nitroaniline

Concen: 21.22 ng/ul

RT: 13.53 min Scan# 1844

Delta R.T. -0.01 min

Lab File: BM000272.D

Acq: 20 Jan 2015 20:24

Instrument :
LabSampleId :
SSTD02039


Tgt Ion: 65 Resp: 119538

Ion Ratio Lower Upper

65 100

92 62.6 50.4 75.6




138 75.9 66.9 100.3



Abundance Ion 65.00 (64.70 to 65.70): BM000272.D (-1811) (-)

Ion 92.00 (91.70 to 92.70): BM000272.D (-1811) (-)

Ion 138.00 (137.70 to 138.70): BM000272.D (-1811) (-)



#44

Dimethylphthalate

Concen: 19.65 ng/ul

RT: 13.91 min Scan# 1908

Delta R.T. -0.01 min

Lab File: BM000272.D

Acq: 20 Jan 2015 20:24


Tgt Ion: 163 Resp: 431927

Ion Ratio Lower Upper

163 100

194 3.8 3.0 4.4



164 9.6 8.5 12.7




Abundance Ion 163.00 (162.70 to 163.70): BM000272.D (-1875) (-)

Ion 194.00 (193.70 to 194.70): BM000272.D (-1875) (-)

Ion 164.00 (163.70 to 164.70): BM000272.D (-1875) (-)









#45
2,6-Dinitrotoluene
Concen: 20.26 ng/ul
RT: 14.03 min Scan# 1929
Delta R.T. 0.00 min
Lab File: BM000272.D
Acq: 20 Jan 2015 20:24




Instrument :
LabSampleId :
SSTD02039


Tgt Ion:165 Resp: 79633
Ion Ratio Lower Upper
165 100
89 86.5 54.4 81.6#
121 25.2 18.6 28.0



#47
Acenaphthylene
Concen: 20.40 ng/ul
RT: 14.18 min Scan# 1954
Delta R.T. 0.00 min
Lab File: BM000272.D
Acq: 20 Jan 2015 20:24

Tgt Ion:152 Resp: 568663
Ion Ratio Lower Upper
152 100
151 18.7 15.4 23.2
153 13.0 10.2 15.2









#48
3-Nitroaniline
Concen: 20.09 ng/ul
RT: 14.36 min Scan# 1985
Delta R.T. 0.00 min
Lab File: BM000272.D
Acq: 20 Jan 2015 20:24




Instrument :
LabSampled :
SSTD02039


Tgt Ion:138 Resp: 83829
Ion Ratio Lower Upper
138 100
108 12.3 11.2 16.8
92 137.6 110.6 165.8



#49
Acenaphthene
Concen: 20.47 ng/ul
RT: 14.52 min Scan# 1212
Delta R.T. 0.00 min
Lab File: BM000272.D
Acq: 20 Jan 2015 20:24

Tgt Ion:153 Resp: 363601
Ion Ratio Lower Upper
153 100
152 47.2 37.5 56.3
154 90.2 72.2 108.4









#50
2,4-Dinitrophenol
Concen: 17.15 ng/ul
RT: 14.57 min Scan# 2020
Delta R.T. 0.00 min
Lab File: BM000272.D
Acq: 20 Jan 2015 20:24




Instrument :
LabSampleId :
SSTD02039


Tgt Ion:184 Resp: 32885
Ion Ratio Lower Upper
184 100
63 98.2 56.1 84.1#
154 97.8 68.2 102.4



#52
4-Nitrophenol
Concen: 20.23 ng/ul
RT: 14.66 min Scan# 2036
Delta R.T. 0.00 min
Lab File: BM000272.D
Acq: 20 Jan 2015 20:24

Tgt Ion:109 Resp: 96358
Ion Ratio Lower Upper
109 100
139 69.7 62.5 93.7
65 102.4 88.7 133.1





#53

Dibenzofuran

Concen: 20.32 ng/ul

RT: 14.86 min Scan# 2069

Delta R.T. 0.00 min

Lab File: BM000272.D

Acq: 20 Jan 2015 20:24


Instrument :
LabSampled :
SSTD02039

Tgt Ion:168 Resp: 522840

Ion Ratio Lower Upper

168 100

139 46.2 33.6 50.4



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

400000



300000

200000

100000

0

Time--> 14.80 14.85 14.90



#54

2,4-Dinitrotoluene

Concen: 21.16 ng/ul

RT: 14.82 min Scan# 2063

Delta R.T. 0.00 min

Lab File: BM000272.D

Acq: 20 Jan 2015 20:24


Tgt Ion:165 Resp: 124144

Ion Ratio Lower Upper

165 100

63 62.3 48.8 73.2

182 2.5 2.1 3.1



Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BM

Ion 182.00 (181.70 to 182.70): E

120000

100000

80000


60000


40000

20000

0

Time--> 14.80 14.85





#55

2,3,4,6-Tetrachlorophenol

Concen: 19.46 ng/ul

RT: 15.08 min Scan# 2107

Delta R.T. 0.00 min

Lab File: BM000272.D

Acq: 20 Jan 2015 20:24

Instrument :
LabSampled :
SSTD02039

Tgt Ion:232 Resp: 96447


Ion Ratio Lower Upper

232 100

131 55.7 43.4 65.0

130 3.9 2.7 4.1

166 34.9 28.2 42.4






Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E



#56

Diethylphthalate

Concen: 19.47 ng/ul

RT: 15.27 min Scan# 2140

Delta R.T. -0.01 min

Lab File: BM000272.D

Acq: 20 Jan 2015 20:24


Tgt Ion:149 Resp: 449685

Ion Ratio Lower Upper

149 100

177 20.9 17.4 26.2



150 13.9 10.5 15.7




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E





#58

Fluorene

Concen: 20.40 ng/ul

RT: 15.50 min Scan# 2179

Delta R.T. 0.00 min

Lab File: BM000272.D

Acq: 20 Jan 2015 20:24

Instrument :
LabSampleId :
SSTD02039


Tgt Ion:166 Resp: 435586

Ion Ratio Lower Upper




166 100

165 96.8 76.0 114.0

167 12.7 10.2 15.4



Abundance Ion 166.00 (165.70 to 166.70): E
Ion 165.00 (164.70 to 165.70): E
Ion 167.00 (166.70 to 167.70): E



#59

4-Chlorophenyl-phenylether

Concen: 20.42 ng/ul

RT: 15.50 min Scan# 2178

Delta R.T. -0.01 min

Lab File: BM000272.D

Acq: 20 Jan 2015 20:24


Tgt Ion:204 Resp: 214959

Ion Ratio Lower Upper



204 100


206 33.7 25.8 38.8

141 77.8 61.3 91.9



Abundance Ion 204.00 (203.70 to 204.70): E
Ion 206.00 (205.70 to 206.70): E
Ion 141.00 (140.70 to 141.70): E









#60
4-Nitroaniline
Concen: 19.51 ng/ul
RT: 15.52 min Scan# 2182
Delta R.T. -0.01 min
Lab File: BM000272.D
Acq: 20 Jan 2015 20:24




Instrument :
LabSampled :
SSTD02039


Tgt Ion:138 Resp: 91200
Ion Ratio Lower Upper
138 100
92 63.8 49.3 73.9
108 140.6 110.3 165.5



#63
4,6-Dinitro-2-methylphenol
Concen: 18.46 ng/ul
RT: 15.58 min Scan# 2192
Delta R.T. 0.00 min
Lab File: BM000272.D
Acq: 20 Jan 2015 20:24

Tgt Ion:198 Resp: 56824
Ion Ratio Lower Upper
198 100
182 4.7 2.6 4.0#
77 38.6 25.2 37.8#






#64




N-Nitrosodiphenylamine
 Concen: 19.68 ng/ul
 RT: 15.71 min Scan# 2214
 Delta R.T. -0.01 min
 Lab File: BM000272.D
 Acq: 20 Jan 2015 20:24

Instrument :
 LabSampleId :
 SSTD02039

Tgt Ion:169 Resp: 359870
 Ion Ratio Lower Upper
 169 100
 168 60.7 51.8 77.6
 167 36.2 28.1 42.1




Abundance Ion 169.00 (168.70 to 169.70): E
 Ion 168.00 (167.70 to 168.70): E
 Ion 167.00 (166.70 to 167.70): E





#65


4-Bromophenyl-phenylether
 Concen: 20.47 ng/ul
 RT: 16.39 min Scan# 2330
 Delta R.T. 0.00 min
 Lab File: BM000272.D
 Acq: 20 Jan 2015 20:24

Tgt Ion:248 Resp: 131538
 Ion Ratio Lower Upper
 248 100
 250 96.6 78.4 117.6
 141 104.7 67.0 100.4#



Abundance Ion 248.00 (247.70 to 248.70): E
 Ion 250.00 (249.70 to 250.70): E
 Ion 141.00 (140.70 to 141.70): E





#66

Hexachlorobenzene

Concen: 20.71 ng/ul

RT: 16.51 min Scan# 2350

Delta R.T. 0.00 min

Lab File: BM000272.D

Acq: 20 Jan 2015 20:24

Instrument :
LabSampled :
SSTD02039


Tgt Ion:284 Resp: 153414

Ion Ratio Lower Upper




284 100

142 40.8 32.4 48.6

249 29.0 26.7 40.1



Abundance Ion 284.00 (283.70 to 284.70): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E



#67

Atrazine

Concen: 17.79 ng/ul

RT: 16.66 min Scan# 2376

Delta R.T. 0.00 min

Lab File: BM000272.D

Acq: 20 Jan 2015 20:24


Tgt Ion:200 Resp: 126467

Ion Ratio Lower Upper



200 100


173 25.0 19.5 29.3

215 41.6 36.6 54.8



Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E









#68
 Pentachlorophenol
 Concen: 16.00 ng/ul
 RT: 16.85 min Scan# 2408
 Delta R.T. 0.00 min
 Lab File: BM000272.D
 Acq: 20 Jan 2015 20:24




Instrument :
 LabSampled :
 SSTD02039


Tgt Ion:266 Resp: 71387
 Ion Ratio Lower Upper
 266 100
 264 62.4 49.8 74.6
 268 72.3 48.2 72.4



#69
 Phenanthrene
 Concen: 20.26 ng/ul
 RT: 17.24 min Scan# 2474
 Delta R.T. -0.01 min
 Lab File: BM000272.D
 Acq: 20 Jan 2015 20:24

Tgt Ion:178 Resp: 691344
 Ion Ratio Lower Upper
 178 100
 179 15.2 11.6 17.4
 176 18.6 14.6 22.0









#71
 Anthracene
 Concen: 20.40 ng/ul
 RT: 17.33 min Scan# 2490
 Delta R.T. 0.00 min
 Lab File: BM000272.D
 Acq: 20 Jan 2015 20:24




Instrument :
 LabSampleId :
 SSTD02039


Tgt Ion:178 Resp: 715875
 Ion Ratio Lower Upper
 178 100
 179 15.1 12.6 18.8
 176 18.3 14.6 21.8



#72
 Carbazole
 Concen: 19.79 ng/ul
 RT: 17.60 min Scan# 2536
 Delta R.T. 0.00 min
 Lab File: BM000272.D
 Acq: 20 Jan 2015 20:24

Tgt Ion:167 Resp: 627356
 Ion Ratio Lower Upper
 167 100
 166 20.0 16.2 24.4
 139 14.2 10.7 16.1









#73
Di-n-butylphthalate
Concen: 18.96 ng/ul
RT: 18.16 min Scan# 2631
Delta R.T. -0.01 min
Lab File: BM000272.D
Acq: 20 Jan 2015 20:24




Instrument :
LabSampled :
SSTD02039


Tgt Ion:149 Resp: 715028
Ion Ratio Lower Upper
149 100
150 8.8 7.1 10.7
104 7.4 4.8 7.2#



#74
Fluoranthene
Concen: 20.42 ng/ul
RT: 19.26 min Scan# 2817
Delta R.T. -0.01 min
Lab File: BM000272.D
Acq: 20 Jan 2015 20:24

Tgt Ion:202 Resp: 814644
Ion Ratio Lower Upper
202 100
101 8.4 6.4 9.6
100 6.2 4.6 7.0









#77
 Pyrene
 Concen: 20.10 ng/ul
 RT: 19.62 min Scan# 2879
 Delta R.T. -0.01 min
 Lab File: BM000272.D
 Acq: 20 Jan 2015 20:24




Instrument :
 LabSampleId :
 SSTD02039


Tgt Ion	Ratio	Lower	Upper
202	100		
101	9.2	6.9	10.3
100	7.7	5.5	8.3



#78
 Butylbenzylphthalate
 Concen: 17.50 ng/ul
 RT: 20.52 min Scan# 3032
 Delta R.T. -0.01 min
 Lab File: BM000272.D
 Acq: 20 Jan 2015 20:24

Tgt Ion	Ratio	Lower	Upper
149	100		
91	80.4	56.6	84.8
206	21.0	16.8	25.2





#79

3,3'-Dichlorobenzidine

Concen: 17.97 ng/ul

RT: 21.30 min Scan# 3165


Delta R.T. -0.01 min

Lab File: BM000272.D




Acq: 20 Jan 2015 20:24

Instrument :
LabSampleId :
SSTD02039

Tgt Ion:	252	Resp:	228839
Ion	Ratio	Lower	Upper
252	100		
254	65.9	51.2	76.8
126	11.8	7.9	11.9



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 254.00 (253.70 to 254.70): E
Ion 126.00 (125.70 to 126.70): E



#80

Benzo(a)anthracene

Concen: 20.21 ng/ul


RT: 21.37 min Scan# 3177

Delta R.T. -0.01 min



Lab File: BM000272.D


Acq: 20 Jan 2015 20:24

Tgt Ion:	228	Resp:	842847
Ion	Ratio	Lower	Upper
228	100		
229	20.0	15.8	23.8
226	26.3	20.7	31.1



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 229.00 (228.70 to 229.70): E
Ion 226.00 (225.70 to 226.70): E









#81
Bis(2-ethylhexyl)phthalate
Concen: 18.22 ng/ul
RT: 21.30 min Scan# 3164
Delta R.T. -0.01 min
Lab File: BM000272.D
Acq: 20 Jan 2015 20:24

Instrument :
LabSampleId :
SSTD02039

Tgt Ion	Ratio	Lower	Upper
149	100		
167	28.3	22.1	33.1
279	4.0	3.0	4.6




Abundance Ion 149.00 (148.70 to 149.70): E
Ion 167.00 (166.70 to 167.70): E
Ion 279.00 (278.70 to 279.70): E






#82
Chrysene
Concen: 20.42 ng/ul
RT: 21.42 min Scan# 3185
Delta R.T. -0.01 min
Lab File: BM000272.D
Acq: 20 Jan 2015 20:24

Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.2	23.0	34.4
229	20.7	15.6	23.4



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E









#84
Di-n-octyl phthalate
Concen: 18.20 ng/ul
RT: 22.19 min Scan# 3315
Delta R.T. -0.01 min
Lab File: BM000272.D
Acq: 20 Jan 2015 20:24




Instrument :
LabSampleId :
SSTD02039


Tgt Ion:149 Resp: 820839



#85
Benzo(b)fluoranthene
Concen: 18.77 ng/ul
RT: 22.99 min Scan# 3451
Delta R.T. -0.01 min
Lab File: BM000272.D
Acq: 20 Jan 2015 20:24

Tgt Ion:252 Resp: 832104
Ion Ratio Lower Upper
252 100
253 21.3 17.1 25.7
125 7.0 4.8 7.2









#86
Benzo(k)fluoranthene
Concen: 12.28 ng/ul
RT: 22.99 min Scan# 3451
Delta R.T. -0.05 min
Lab File: BM000272.D
Acq: 20 Jan 2015 20:24




Instrument :
LabSampled :
SSTD02039


Tgt Ion:252 Resp: 464672
Ion Ratio Lower Upper
252 100
253 21.3 17.2 25.8
125 7.0 4.9 7.3



#88
Benzo(a)pyrene
Concen: 20.26 ng/ul
RT: 23.57 min Scan# 3551
Delta R.T. -0.01 min
Lab File: BM000272.D
Acq: 20 Jan 2015 20:24

Tgt Ion:252 Resp: 789176
Ion Ratio Lower Upper
252 100
253 22.3 17.0 25.6
125 7.3 5.8 8.6





#89

Indeno(1,2,3-cd)pyrene

Concen: 22.19 ng/ul

RT: 25.99 min Scan# 3962


Delta R.T. -0.02 min

Lab File: BM000272.D




Acq: 20 Jan 2015 20:24

Instrument :
LabSampleId :
SSTD02039

Tgt Ion:	276	Resp:	901520
Ion	Ratio	Lower	Upper
276	100		
138	19.2	14.2	21.2
277	25.0	18.8	28.2



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E



#90

Dibenzo(a,h)anthracene

Concen: 22.17 ng/ul


RT: 26.01 min Scan# 3965

Delta R.T. -0.01 min



Lab File: BM000272.D


Acq: 20 Jan 2015 20:24

Tgt Ion:	278	Resp:	758537
Ion	Ratio	Lower	Upper
278	100		
139	10.7	10.2	15.2
279	22.9	19.2	28.8



Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E





#91

Benzo(g,h,i)perylene

Concen: 22.85 ng/ul

RT: 26.71 min Scan# 4084


Delta R.T. -0.01 min

Lab File: BM000272.D

Acq: 20 Jan 2015 20:24

Instrument :
LabSampleId :
SSTD02039

Tgt Ion:	276	Resp:	762683
Ion	Ratio	Lower	Upper
276	100		
138	14.6	12.1	18.1
277	23.2	19.0	28.4



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E

