

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM012015\
 Data File : BM000304.D
 Acq On : 21 Jan 2015 17:02
 Operator : TP/IZ
 Sample : SSTD02041
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 LabSampleId :
 SSTD02041

Quant Time: Jan 22 02:27:49 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-BM122914.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 22 02:18:27 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.82	152	108190	20.00	ng/ul	0.00
18) Naphthalene-d8	10.61	136	420972	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.46	164	273002	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.20	188	541254	20.00	ng/ul	0.00
75) Chrysene-d12	21.39	240	639157	20.00	ng/ul	0.00
83) Perylene-d12	23.67	264	624419	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.31	96	18776	10.06	ng/uL	0.00
5) Phenol-d5	6.98	99	167784	18.44	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.15	67	121063	21.04	ng/ul	0.00
9) 2-Chlorophenol-d4	7.35	132	128294	19.22	ng/ul	0.00
13) 4-Methylphenol-d8	8.52	113	127267	17.58	ng/ul	0.00
19) Nitrobenzene-d5	8.97	128	63512	21.82	ng/ul	0.00
22) 2-Nitrophenol-d4	9.69	143	62464	21.02	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.23	165	124796	19.01	ng/ul	0.00
29) 4-Chloroaniline-d4	10.75	131	154038	20.68	ng/ul	0.00
43) Dimethylphthalate-d6	13.86	166	376629	18.70	ng/ul	0.00
46) Acenaphthylene-d8	14.15	160	474673	19.68	ng/ul	0.00
51) 4-Nitrophenol-d4	14.64	143	53090	16.20	ng/ul	0.00
57) Fluorene-d10	15.44	176	329289	19.04	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.56	200	44566	17.37	ng/ul	0.00
70) Anthracene-d10	17.30	188	515015	20.58	ng/ul	0.00
76) Pyrene-d10	19.59	212	566372	19.21	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.53	264	575132	20.29	ng/ul	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.35	88	26285	10.16	ng/uL	97
4) Benzaldehyde	6.96	77	144471	22.79	ng/ul	99
6) Phenol	7.01	94	181648	18.92	ng/ul	94
8) Bis(2-Chloroethyl)ether	7.25	93	148444	19.92	ng/ul	96
10) 2-Chlorophenol	7.38	128	128581	18.41	ng/ul	92
11) 2-Methylphenol	8.26	108	131655	17.81	ng/ul	96
12) 2,2'-oxybis(1-Chloropropan	8.35	45	264787	21.62	ng/ul	97
14) Acetophenone	8.63	105	241571	18.99	ng/ul	92
15) N-Nitroso-di-n-propylamine	8.62	70	122527	18.52	ng/ul	88
16) 4-Methylphenol	8.58	108	139826	17.52	ng/ul	98
17) Hexachloroethane	8.89	117	70751	22.18	ng/ul	90
20) Nitrobenzene	9.02	77	210749	24.28	ng/ul	96
21) Isophorone	9.53	82	329653	19.70	ng/ul	97
23) 2-Nitrophenol	9.72	139	66723	20.33	ng/ul#	84
24) 2,4-Dimethylphenol	9.78	107	173557	19.49	ng/ul	98
25) Bis(2-Chloroethoxy)methane	10.02	93	180415	19.37	ng/ul#	97
27) 2,4-Dichlorophenol	10.25	162	123223	18.81	ng/ul	93
28) Naphthalene	10.66	128	423872	19.75	ng/ul	98
30) 4-Chloroaniline	10.77	127	158923	20.73	ng/ul	97
31) Hexachlorobutadiene	10.95	225	102675	22.10	ng/ul	95
33) 4-Chloro-3-methylphenol	11.89	107	142780	17.76	ng/ul	99
34) 2-Methylnaphthalene	12.27	142	300156	18.86	ng/ul	98
36) 1,2,4,5-Tetrachlorobenzene	12.65	216	163029	21.54	ng/ul#	94


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM012015\
 Data File : BM000304.D
 Acq On : 21 Jan 2015 17:02
 Operator : TP/IZ
 Sample : SST02041
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 LabSampleId :
 SST02041

Quant Time: Jan 22 02:27:49 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-BM122914.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 22 02:18:27 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) Hexachlorocyclopentadiene	12.62	237	101753	21.61	ng/ul	97
38) 2,4,6-Trichlorophenol	12.88	196	92595	19.18	ng/ul	99
39) 2,4,5-Trichlorophenol	12.95	196	99030	18.88	ng/ul	95
40) 1,1'-Biphenyl	13.29	154	394106	20.03	ng/ul	97
41) 2-Chloronaphthalene	13.33	162	305460	20.25	ng/ul	99
42) 2-Nitroaniline	13.53	65	97901	19.05	ng/ul	90
44) Dimethylphthalate	13.91	163	374119	18.65	ng/ul	99
45) 2,6-Dinitrotoluene	14.03	165	69580	19.40	ng/ul#	81
47) Acenaphthylene	14.17	152	499413	19.64	ng/ul	99
48) 3-Nitroaniline	14.36	138	69936	18.37	ng/ul	97
49) Acenaphthene	14.52	153	320622	19.78	ng/ul	99
50) 2,4-Dinitrophenol	14.56	184	25919	14.81	ng/ul#	87
52) 4-Nitrophenol	14.66	109	83699	19.26	ng/ul	90
53) Dibenzofuran	14.85	168	470334	20.04	ng/ul	89
54) 2,4-Dinitrotoluene	14.82	165	108118	20.20	ng/ul	94
55) 2,3,4,6-Tetrachlorophenol	15.08	232	85271	18.85	ng/ul	94
56) Diethylphthalate	15.27	149	392515	18.62	ng/ul	100
58) Fluorene	15.50	166	380070	19.51	ng/ul	100
59) 4-Chlorophenyl-phenylether	15.50	204	192563	20.04	ng/ul	99
60) 4-Nitroaniline	15.52	138	75389	17.68	ng/ul	86
63) 4,6-Dinitro-2-methylphenol	15.58	198	48218	18.25	ng/ul#	92
64) N-Nitrosodiphenylamine	15.71	169	313322	19.96	ng/ul	98
65) 4-Bromophenyl-phenylether	16.39	248	115613	20.96	ng/ul	89
66) Hexachlorobenzene	16.50	284	135152	21.25	ng/ul	98
67) Atrazine	16.66	200	106709	17.49	ng/ul	99
68) Pentachlorophenol	16.85	266	63660	16.62	ng/ul	97
69) Phenanthrene	17.24	178	614101	20.96	ng/ul	98
71) Anthracene	17.33	178	621576	20.63	ng/ul	99
72) Carbazole	17.60	167	535936	19.69	ng/ul	97
73) Di-n-butylphthalate	18.16	149	617310	19.07	ng/ul#	97
74) Fluoranthene	19.26	202	705716	20.61	ng/ul	99
77) Pyrene	19.62	202	762053	19.78	ng/ul	99
78) Butylbenzylphthalate	20.52	149	249572	15.94	ng/ul	89
79) 3,3'-Dichlorobenzidine	21.30	252	192287	17.06	ng/ul	98
80) Benzo(a)anthracene	21.37	228	739792	20.05	ng/ul	98
81) Bis(2-ethylhexyl)phthalate	21.30	149	385408	16.92	ng/ul	97
82) Chrysene	21.43	228	710137	20.88	ng/ul	99
84) Di-n-octyl phthalate	22.19	149	686281	16.93	ng/ul	100
85) Benzo(b)fluoranthene	22.99	252	736589	18.49	ng/ul	99
86) Benzo(k)fluoranthene	23.03	252	736723	21.66	ng/ul	98
88) Benzo(a)pyrene	23.57	252	726016	20.74	ng/ul	99
89) Indeno(1,2,3-cd)pyrene	25.99	276	838084	22.95	ng/ul	96
90) Dibenzo(a,h)anthracene	26.00	278	701821	22.82	ng/ul	98
91) Benzo(g,h,i)perylene	26.70	276	705783	23.52	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

1,4-Dioxane

Concen: 10.16 ng/uL

RT: 3.35 min Scan# 112

Delta R.T. 0.00 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02

Instrument :
LabSampled :
SSTD02041





Tgt Ion: 88 Resp: 26285

Ion Ratio Lower Upper

88 100

43 39.2 34.9 52.3

58 82.8 67.3 100.9



#4

Benzaldehyde

Concen: 22.79 ng/uL

RT: 6.96 min Scan# 727

Delta R.T. 0.00 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02




Tgt Ion: 77 Resp: 144471


Ion Ratio Lower Upper

77 100

105 91.4 73.5 110.3

106 84.5 67.9 101.9





#6

Phenol

Concen: 18.92 ng/ul

RT: 7.01 min Scan# 735

Delta R.T. 0.00 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02

Instrument :
LabSampleId :
SSTD02041





Tgt Ion: 94 Resp: 181648

Ion Ratio Lower Upper

94 100

65 38.0 27.4 41.0

66 56.9 42.5 63.7



#8

Bis(2-Chloroethyl)ether

Concen: 19.92 ng/ul

RT: 7.25 min Scan# 775

Delta R.T. 0.00 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02




Tgt Ion: 93 Resp: 148444


Ion Ratio Lower Upper

93 100

63 90.5 68.9 103.3

95 31.6 24.6 37.0









#10
2-Chlorophenol
Concen: 18.41 ng/ul
RT: 7.38 min Scan# 798
Delta R.T. 0.00 min
Lab File: BM000304.D
Acq: 21 Jan 2015 17:02

Instrument :
LabSampled :
SSTD02041

Tgt Ion:128 Resp: 128581
Ion Ratio Lower Upper
128 100
64 65.5 47.2 70.8
130 35.3 25.2 37.8




Abundance Ion 128.00 (127.70 to 128.70): E
Ion 64.00 (63.70 to 64.70): BM0
Ion 130.00 (129.70 to 130.70): E






#11
2-Methylphenol
Concen: 17.81 ng/ul
RT: 8.26 min Scan# 947
Delta R.T. 0.00 min
Lab File: BM000304.D
Acq: 21 Jan 2015 17:02

Tgt Ion:108 Resp: 131655
Ion Ratio Lower Upper
108 100
107 84.5 70.7 106.1



Abundance Ion 108.00 (107.70 to 108.70): E
Ion 107.00 (106.70 to 107.70): E





#12

2,2'-oxybis(1-Chloropropane)

Concen: 21.62 ng/ul

RT: 8.35 min Scan# 963





Delta R.T. 0.00 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02

Instrument :
LabSampled :
SSTD02041

Tgt Ion:	45	Resp:	264787
Ion	Ratio	Lower	Upper
45	100		
77	11.8	9.9	14.9
79	8.6	8.3	12.5



#14

Acetophenone

Concen: 18.99 ng/ul




RT: 8.63 min Scan# 1011


Delta R.T. -0.01 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02

Tgt Ion:	105	Resp:	241571
Ion	Ratio	Lower	Upper
105	100		
77	88.8	66.3	99.5
51	39.3	26.5	39.7









#15
N-Nitroso-di-n-propylamine
Concen: 18.52 ng/ul
RT: 8.62 min Scan# 1009
Delta R.T. 0.00 min
Lab File: BM000304.D
Acq: 21 Jan 2015 17:02




Instrument :
LabSampleId :
SSTD02041


Tgt Ion: 70 Resp: 122527
Ion Ratio Lower Upper
70 100
42 69.1 46.4 69.6
101 8.7 7.8 11.8
130 14.6 14.6 21.8



#16
4-Methylphenol
Concen: 17.52 ng/ul
RT: 8.58 min Scan# 1002
Delta R.T. 0.00 min
Lab File: BM000304.D
Acq: 21 Jan 2015 17:02

Tgt Ion: 108 Resp: 139826
Ion Ratio Lower Upper
108 100
107 103.4 84.0 126.0









#17
 Hexachloroethane
 Concen: 22.18 ng/ul
 RT: 8.89 min Scan# 1055
 Delta R.T. -0.01 min
 Lab File: BM000304.D
 Acq: 21 Jan 2015 17:02

Instrument :
 LabSampleId :
 SSTD02041

Tgt Ion: 117 Resp: 70751
 Ion Ratio Lower Upper
 117 100
 201 96.0 85.2 127.8
 199 61.1 55.1 82.7




Abundance Ion 117.00 (116.70 to 117.70): E
 Ion 201.00 (200.70 to 201.70): E
 Ion 199.00 (198.70 to 199.70): E






#20
 Nitrobenzene
 Concen: 24.28 ng/ul
 RT: 9.02 min Scan# 1076
 Delta R.T. 0.00 min
 Lab File: BM000304.D
 Acq: 21 Jan 2015 17:02

Tgt Ion: 77 Resp: 210749
 Ion Ratio Lower Upper
 77 100
 123 34.1 29.2 43.8
 65 16.9 12.4 18.6



Abundance Ion 77.00 (76.70 to 77.70): BM
 Ion 123.00 (122.70 to 123.70): E
 Ion 65.00 (64.70 to 65.70): BM





#21

Isophorone

Concen: 19.70 ng/ul

RT: 9.53 min Scan# 1164





Delta R.T. 0.00 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02

Instrument :
LabSampled :
SSTD02041

Tgt Ion:	82	Resp:	329653
Ion	Ratio	Lower	Upper
82	100		
95	10.0	6.9	10.3
138	14.2	12.1	18.1



#23

2-Nitrophenol

Concen: 20.33 ng/ul




RT: 9.72 min Scan# 1196


Delta R.T. 0.00 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02

Tgt Ion:	139	Resp:	66723
Ion	Ratio	Lower	Upper
139	100		
65	83.4	56.3	84.5
109	57.2	37.1	55.7#









#24
 2,4-Dimethylphenol
 Concen: 19.49 ng/ul
 RT: 9.78 min Scan# 1206
 Delta R.T. 0.00 min
 Lab File: BM000304.D
 Acq: 21 Jan 2015 17:02

Instrument :
 LabSampled :
 SSTD02041

Tgt Ion	Ratio	Lower	Upper
107	100		
121	41.0	33.0	49.6
122	72.3	59.3	88.9




Abundance Ion 107.00 (106.70 to 107.70): E
 Ion 121.00 (120.70 to 121.70): E
 Ion 122.00 (121.70 to 122.70): E






#25
 Bis(2-Chloroethoxy)methane
 Concen: 19.37 ng/ul
 RT: 10.02 min Scan# 1247
 Delta R.T. 0.00 min
 Lab File: BM000304.D
 Acq: 21 Jan 2015 17:02

Tgt Ion	Ratio	Lower	Upper
93	100		
95	31.7	26.1	39.1
123	11.7	7.4	11.2#



Abundance Ion 93.00 (92.70 to 93.70): BM0
 Ion 95.00 (94.70 to 95.70): BM0
 Ion 123.00 (122.70 to 123.70): E









#27
 2,4-Dichlorophenol
 Concen: 18.81 ng/ul
 RT: 10.25 min Scan# 1286
 Delta R.T. -0.01 min
 Lab File: BM000304.D
 Acq: 21 Jan 2015 17:02

Instrument :
 LabSampled :
 SSTD02041

Tgt Ion	Ratio	Lower	Upper
162	100		
164	69.9	50.2	75.4
98	42.0	31.4	47.2




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 98.00 (97.70 to 98.70): E






#28
 Naphthalene
 Concen: 19.75 ng/ul
 RT: 10.66 min Scan# 1355
 Delta R.T. -0.01 min
 Lab File: BM000304.D
 Acq: 21 Jan 2015 17:02

Tgt Ion	Ratio	Lower	Upper
128	100		
129	10.0	8.6	13.0
127	13.6	10.3	15.5



Abundance Ion 128.00 (127.70 to 128.70): E
 Ion 129.00 (128.70 to 129.70): E
 Ion 127.00 (126.70 to 127.70): E









#30
4-Chloroaniline
Concen: 20.73 ng/ul
RT: 10.77 min Scan# 1374
Delta R.T. 0.00 min
Lab File: BM000304.D
Acq: 21 Jan 2015 17:02

Instrument :
LabSampleId :
SSTD02041

Tgt Ion:127 Resp: 158923
Ion Ratio Lower Upper
127 100
129 31.5 26.5 39.7




Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E






#31
Hexachlorobutadiene
Concen: 22.10 ng/ul
RT: 10.95 min Scan# 1404
Delta R.T. 0.00 min
Lab File: BM000304.D
Acq: 21 Jan 2015 17:02

Tgt Ion:225 Resp: 102675
Ion Ratio Lower Upper
225 100
223 61.9 54.4 81.6
227 66.9 51.8 77.8



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E









#33
 4-Chloro-3-methylphenol
 Concen: 17.76 ng/ul
 RT: 11.89 min Scan# 1565
 Delta R.T. 0.00 min
 Lab File: BM000304.D
 Acq: 21 Jan 2015 17:02




Instrument :
 LabSampleId :
 SSTD02041


Tgt Ion:107 Resp: 142780
 Ion Ratio Lower Upper
 107 100
 144 21.9 16.6 24.8
 142 67.6 54.6 82.0



#34
 2-Methylnaphthalene
 Concen: 18.86 ng/ul
 RT: 12.27 min Scan# 1629
 Delta R.T. -0.01 min
 Lab File: BM000304.D
 Acq: 21 Jan 2015 17:02

Tgt Ion:142 Resp: 300156
 Ion Ratio Lower Upper
 142 100
 141 88.0 69.3 103.9









#36
 1,2,4,5-Tetrachlorobenzene
 Concen: 21.54 ng/ul
 RT: 12.65 min Scan# 1693
 Delta R.T. 0.01 min
 Lab File: BM000304.D
 Acq: 21 Jan 2015 17:02

Instrument :
 LabSampleId :
 SSTD02041

Tgt Ion:	216	Resp:	163029
Ion	Ratio	Lower	Upper
216	100		
214	72.9	62.6	94.0
179	22.8	17.4	26.2
108	20.6	13.4	20.0#




Abundance Ion 216.00 (215.70 to 216.70): E
 Ion 214.00 (213.70 to 214.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 108.00 (107.70 to 108.70): E






#37
 Hexachlorocyclopentadiene
 Concen: 21.61 ng/ul
 RT: 12.62 min Scan# 1689
 Delta R.T. 0.00 min
 Lab File: BM000304.D
 Acq: 21 Jan 2015 17:02

Tgt Ion:	237	Resp:	101753
Ion	Ratio	Lower	Upper
237	100		
235	60.2	50.0	75.0
272	11.3	8.9	13.3



Abundance Ion 237.00 (236.70 to 237.70): E
 Ion 235.00 (234.70 to 235.70): E
 Ion 272.00 (271.70 to 272.70): E









#38
 2,4,6-Trichlorophenol
 Concen: 19.18 ng/ul
 RT: 12.88 min Scan# 1733
 Delta R.T. 0.00 min
 Lab File: BM000304.D
 Acq: 21 Jan 2015 17:02




Instrument :
 LabSampleId :
 SSTD02041


Tgt Ion:196 Resp: 92595
 Ion Ratio Lower Upper
 196 100
 198 97.1 76.6 114.8
 200 30.3 24.5 36.7



#39
 2,4,5-Trichlorophenol
 Concen: 18.88 ng/ul
 RT: 12.95 min Scan# 1745
 Delta R.T. 0.00 min
 Lab File: BM000304.D
 Acq: 21 Jan 2015 17:02

Tgt Ion:196 Resp: 99030
 Ion Ratio Lower Upper
 196 100
 198 99.0 75.4 113.2
 200 35.0 25.0 37.4









#40
 1,1'-Biphenyl
 Concen: 20.03 ng/ul
 RT: 13.29 min Scan# 1802
 Delta R.T. 0.00 min
 Lab File: BM000304.D
 Acq: 21 Jan 2015 17:02




Instrument :
 LabSampleId :
 SSTD02041


Tgt Ion	Ratio	Lower	Upper
154	100		
153	40.0	30.5	45.7
76	11.1	8.9	13.3



#41
 2-Chloronaphthalene
 Concen: 20.25 ng/ul
 RT: 13.33 min Scan# 1809
 Delta R.T. 0.00 min
 Lab File: BM000304.D
 Acq: 21 Jan 2015 17:02

Tgt Ion	Ratio	Lower	Upper
162	100		
164	31.8	25.8	38.6
127	44.1	35.9	53.9









#42
2-Nitroaniline
Concen: 19.05 ng/ul
RT: 13.53 min Scan# 1844
Delta R.T. 0.00 min
Lab File: BM000304.D
Acq: 21 Jan 2015 17:02




Instrument :
LabSampled :
SSTD02041


Tgt Ion: 65 Resp: 97901
Ion Ratio Lower Upper
65 100
92 55.4 50.4 75.6
138 74.0 66.9 100.3



#44
Dimethylphthalate
Concen: 18.65 ng/ul
RT: 13.91 min Scan# 1908
Delta R.T. 0.00 min
Lab File: BM000304.D
Acq: 21 Jan 2015 17:02

Tgt Ion: 163 Resp: 374119
Ion Ratio Lower Upper
163 100
194 3.5 3.0 4.4
164 10.1 8.5 12.7









#45
2,6-Dinitrotoluene
Concen: 19.40 ng/ul
RT: 14.03 min Scan# 1928
Delta R.T. -0.01 min
Lab File: BM000304.D
Acq: 21 Jan 2015 17:02




Instrument :
LabSampled :
SSTD02041


Tgt Ion:165 Resp: 69580
Ion Ratio Lower Upper
165 100
89 84.8 54.4 81.6#
121 30.1 18.6 28.0#



#47
Acenaphthylene
Concen: 19.64 ng/ul
RT: 14.17 min Scan# 1953
Delta R.T. 0.00 min
Lab File: BM000304.D
Acq: 21 Jan 2015 17:02

Tgt Ion:152 Resp: 499413
Ion Ratio Lower Upper
152 100
151 18.5 15.4 23.2
153 12.9 10.2 15.2









#48
3-Nitroaniline
Concen: 18.37 ng/ul
RT: 14.36 min Scan# 1984
Delta R.T. 0.00 min
Lab File: BM000304.D
Acq: 21 Jan 2015 17:02




Instrument :
LabSampled :
SSTD02041


Tgt Ion:138 Resp: 69936
Ion Ratio Lower Upper
138 100
108 16.8 11.2 16.8
92 141.8 110.6 165.8



#49
Acenaphthene
Concen: 19.78 ng/ul
RT: 14.52 min Scan# 2011
Delta R.T. -0.01 min
Lab File: BM000304.D
Acq: 21 Jan 2015 17:02

Tgt Ion:153 Resp: 320622
Ion Ratio Lower Upper
153 100
152 48.2 37.5 56.3
154 89.9 72.2 108.4









#50
 2,4-Dinitrophenol
 Concen: 14.81 ng/ul
 RT: 14.56 min Scan# 2019
 Delta R.T. -0.01 min
 Lab File: BM000304.D
 Acq: 21 Jan 2015 17:02




Instrument :
 LabSampled :
 SSTD02041


Tgt Ion:184 Resp: 25919
 Ion Ratio Lower Upper
 184 100
 63 92.2 56.1 84.1#
 154 84.3 68.2 102.4



#52
 4-Nitrophenol
 Concen: 19.26 ng/ul
 RT: 14.66 min Scan# 2036
 Delta R.T. 0.01 min
 Lab File: BM000304.D
 Acq: 21 Jan 2015 17:02

Tgt Ion:109 Resp: 83699
 Ion Ratio Lower Upper
 109 100
 139 67.6 62.5 93.7
 65 101.9 88.7 133.1





#53

Dibenzenofuran

Concen: 20.04 ng/ul

RT: 14.85 min Scan# 2068

Delta R.T. -0.01 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02





Instrument :
LabSampleId :
SSTD02041

Tgt Ion:168 Resp: 470334

Ion Ratio Lower Upper

168 100

139 48.9 33.6 50.4



#54

2,4-Dinitrotoluene

Concen: 20.20 ng/ul

RT: 14.82 min Scan# 2062

Delta R.T. 0.00 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02




Tgt Ion:165 Resp: 108118


Ion Ratio Lower Upper

165 100

63 65.8 48.8 73.2

182 3.0 2.1 3.1









#55
 2,3,4,6-Tetrachlorophenol
 Concen: 18.85 ng/ul
 RT: 15.08 min Scan# 2107
 Delta R.T. 0.00 min
 Lab File: BM000304.D
 Acq: 21 Jan 2015 17:02

Instrument :
 LabSampleId :
 SSTD02041

Tgt Ion:232 Resp: 85271
 Ion Ratio Lower Upper
 232 100
 131 57.4 43.4 65.0
 130 4.0 2.7 4.1
 166 30.4 28.2 42.4




Abundance Ion 232.00 (231.70 to 232.70): E
 Ion 131.00 (130.70 to 131.70): E
 Ion 130.00 (129.70 to 130.70): E
 Ion 166.00 (165.70 to 166.70): E






#56
 Diethylphthalate
 Concen: 18.62 ng/ul
 RT: 15.27 min Scan# 2140
 Delta R.T. 0.00 min
 Lab File: BM000304.D
 Acq: 21 Jan 2015 17:02

Tgt Ion:149 Resp: 392515
 Ion Ratio Lower Upper
 149 100
 177 21.8 17.4 26.2
 150 13.6 10.5 15.7



Abundance Ion 149.00 (148.70 to 149.70): E
 Ion 177.00 (176.70 to 177.70): E
 Ion 150.00 (149.70 to 150.70): E





#58

Fluorene

Concen: 19.51 ng/ul

RT: 15.50 min Scan# 2179

Delta R.T. 0.00 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02

Instrument :
LabSampled :
SSTD02041


Tgt Ion:166 Resp: 380070

Ion Ratio Lower Upper




166 100

165 94.7 76.0 114.0

167 12.7 10.2 15.4



Abundance Ion 166.00 (165.70 to 166.70): E
Ion 165.00 (164.70 to 165.70): E
Ion 167.00 (166.70 to 167.70): E



#59

4-Chlorophenyl-phenylether

Concen: 20.04 ng/ul

RT: 15.50 min Scan# 2178

Delta R.T. 0.00 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02


Tgt Ion:204 Resp: 192563

Ion Ratio Lower Upper



204 100


206 33.6 25.8 38.8

141 75.8 61.3 91.9



Abundance Ion 204.00 (203.70 to 204.70): E
Ion 206.00 (205.70 to 206.70): E
Ion 141.00 (140.70 to 141.70): E









#60
4-Nitroaniline
Concen: 17.68 ng/ul
RT: 15.52 min Scan# 2182
Delta R.T. 0.00 min
Lab File: BM000304.D
Acq: 21 Jan 2015 17:02




Instrument :
LabSampleId :
SSTD02041


Tgt Ion:138 Resp: 75389
Ion Ratio Lower Upper
138 100
92 65.7 49.3 73.9
108 158.9 110.3 165.5



#63
4,6-Dinitro-2-methylphenol
Concen: 18.25 ng/ul
RT: 15.58 min Scan# 2192
Delta R.T. 0.00 min
Lab File: BM000304.D
Acq: 21 Jan 2015 17:02

Tgt Ion:198 Resp: 48218
Ion Ratio Lower Upper
198 100
182 5.4 2.6 4.0#
77 36.1 25.2 37.8





#64

N-Nitrosodiphenylamine

Concen: 19.96 ng/ul

RT: 15.71 min Scan# 2214

Delta R.T. 0.00 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02

Instrument :
LabSampleId :
SSTD02041


Tgt Ion:169 Resp: 313322

Ion Ratio Lower Upper




169 100

168 62.9 51.8 77.6

167 34.5 28.1 42.1



Abundance Ion 169.00 (168.70 to 169.70): E
Ion 168.00 (167.70 to 168.70): E
Ion 167.00 (166.70 to 167.70): E



#65

4-Bromophenyl-phenylether

Concen: 20.96 ng/ul

RT: 16.39 min Scan# 2330

Delta R.T. 0.00 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02


Tgt Ion:248 Resp: 115613

Ion Ratio Lower Upper



248 100


250 93.4 78.4 117.6

141 99.5 67.0 100.4



Abundance Ion 248.00 (247.70 to 248.70): E
Ion 250.00 (249.70 to 250.70): E
Ion 141.00 (140.70 to 141.70): E





#66

Hexachlorobenzene

Concen: 21.25 ng/ul

RT: 16.50 min Scan# 2349


Delta R.T. 0.00 min

Lab File: BM000304.D




Acq: 21 Jan 2015 17:02

Instrument :
LabSampled :
SSTD02041

Tgt Ion:	284	Resp:	135152
Ion Ratio	Lower	Upper	
284	100		
142	42.8	32.4	48.6
249	33.9	26.7	40.1



Abundance Ion 284.00 (283.70 to 284.70): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E



#67

Atrazine

Concen: 17.49 ng/ul


RT: 16.66 min Scan# 2375

Delta R.T. 0.00 min



Lab File: BM000304.D


Acq: 21 Jan 2015 17:02

Tgt Ion:	200	Resp:	106709
Ion Ratio	Lower	Upper	
200	100		
173	23.2	19.5	29.3
215	45.2	36.6	54.8



Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E





#68

Pentachlorophenol

Concen: 16.62 ng/ul

RT: 16.85 min Scan# 2408

Delta R.T. 0.00 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02

Instrument :
LabSampled :
SSTD02041


Tgt Ion:266 Resp: 63660

Ion Ratio Lower Upper




266 100

264 61.3 49.8 74.6

268 56.8 48.2 72.4



Abundance Ion 265.70 (265.40 to 266.40): E
Ion 264.00 (263.70 to 264.70): E
Ion 268.00 (267.70 to 268.70): E



#69

Phenanthrene

Concen: 20.96 ng/ul

RT: 17.24 min Scan# 2474

Delta R.T. 0.00 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02


Tgt Ion:178 Resp: 614101

Ion Ratio Lower Upper



178 100


179 15.3 11.6 17.4

176 19.6 14.6 22.0



Abundance Ion 178.00 (177.70 to 178.70): E
Ion 179.00 (178.70 to 179.70): E
Ion 176.00 (175.70 to 176.70): E









#71
 Anthracene
 Concen: 20.63 ng/ul
 RT: 17.33 min Scan# 2490
 Delta R.T. 0.00 min
 Lab File: BM000304.D
 Acq: 21 Jan 2015 17:02




Instrument :
 LabSampleId :
 SSTD02041


Tgt Ion:178 Resp: 621576
 Ion Ratio Lower Upper
 178 100
 179 15.5 12.6 18.8
 176 18.8 14.6 21.8



#72
 Carbazole
 Concen: 19.69 ng/ul
 RT: 17.60 min Scan# 2536
 Delta R.T. 0.00 min
 Lab File: BM000304.D
 Acq: 21 Jan 2015 17:02

Tgt Ion:167 Resp: 535936
 Ion Ratio Lower Upper
 167 100
 166 21.6 16.2 24.4
 139 14.3 10.7 16.1





#73

Di-n-butylphthalate

Concen: 19.07 ng/ul

RT: 18.16 min Scan# 2631





Delta R.T. 0.00 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02

Instrument :
LabSampleId :
SSTD02041

Tgt Ion:	149	Resp:	617310
Ion	Ratio	Lower	Upper
149	100		
150	8.1	7.1	10.7
104	7.3	4.8	7.2#



#74

Fluoranthene

Concen: 20.61 ng/ul




RT: 19.26 min Scan# 2817


Delta R.T. 0.00 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02

Tgt Ion:	202	Resp:	705716
Ion	Ratio	Lower	Upper
202	100		
101	7.6	6.4	9.6
100	6.1	4.6	7.0





#77

Pyrene

Concen: 19.78 ng/ul

RT: 19.62 min Scan# 2879





Delta R.T. 0.00 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02

Instrument :
LabSampled :
SSTD02041

Tgt Ion:	202	Resp:	762053
Ion	Ratio	Lower	Upper
202	100		
101	8.7	6.9	10.3
100	7.7	5.5	8.3



#78

Butylbenzylphthalate

Concen: 15.94 ng/ul




RT: 20.52 min Scan# 3032


Delta R.T. 0.00 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02

Tgt Ion:	149	Resp:	249572
Ion	Ratio	Lower	Upper
149	100		
91	82.0	56.6	84.8
206	21.0	16.8	25.2





#79

3,3'-Dichlorobenzidine

Concen: 17.06 ng/ul

RT: 21.30 min Scan# 3165


Delta R.T. 0.00 min

Lab File: BM000304.D




Acq: 21 Jan 2015 17:02

Instrument :
LabSampled :
SSTD02041

Tgt Ion:	252	Resp:	192287
Ion	Ratio	Lower	Upper
252	100		
254	62.4	51.2	76.8
126	11.7	7.9	11.9



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 254.00 (253.70 to 254.70): E
Ion 126.00 (125.70 to 126.70): E



#80

Benzo(a)anthracene

Concen: 20.05 ng/ul


RT: 21.37 min Scan# 3177

Delta R.T. 0.01 min



Lab File: BM000304.D


Acq: 21 Jan 2015 17:02

Tgt Ion:	228	Resp:	739792
Ion	Ratio	Lower	Upper
228	100		
229	19.2	15.8	23.8
226	27.2	20.7	31.1



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 229.00 (228.70 to 229.70): E
Ion 226.00 (225.70 to 226.70): E





#81

Bis(2-ethylhexyl)phthalate

Concen: 16.92 ng/ul

RT: 21.30 min Scan# 3164


Delta R.T. 0.00 min

Lab File: BM000304.D




Acq: 21 Jan 2015 17:02

Instrument :
LabSampled :
SSTD02041

Tgt Ion:	149	Resp:	385408
Ion	Ratio	Lower	Upper
149	100		
167	25.6	22.1	33.1
279	3.8	3.0	4.6



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 167.00 (166.70 to 167.70): E
Ion 279.00 (278.70 to 279.70): E



#82

Chrysene

Concen: 20.88 ng/ul


RT: 21.43 min Scan# 3186

Delta R.T. 0.01 min



Lab File: BM000304.D


Acq: 21 Jan 2015 17:02

Tgt Ion:	228	Resp:	710137
Ion	Ratio	Lower	Upper
228	100		
226	29.0	23.0	34.4
229	18.6	15.6	23.4



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E





#84

Di-n-octyl phthalate

Concen: 16.93 ng/ul

RT: 22.19 min Scan# 3315


Delta R.T. 0.00 min

Lab File: BM000304.D




Acq: 21 Jan 2015 17:02

Instrument :
LabSampleId :
SSTD02041

Tgt Ion:149 Resp: 686281



Abundance Ion 149.00 (148.70 to 149.70): E



#85

Benzo(b)fluoranthene

Concen: 18.49 ng/ul

RT: 22.99 min Scan# 3451


Delta R.T. 0.00 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02

Tgt Ion:252 Resp: 736589



Ion	Ratio	Lower	Upper
252	100		
253	21.4	17.1	25.7
125	6.7	4.8	7.2




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#86

Benzo(k)fluoranthene

Concen: 21.66 ng/ul

RT: 23.03 min Scan# 3458





Delta R.T. 0.00 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02

Instrument :
LabSampleId :
SSTD02041

Tgt Ion:	252	Resp:	736723
Ion	Ratio	Lower	Upper
252	100		
253	22.5	17.2	25.8
125	6.8	4.9	7.3



#88

Benzo(a)pyrene

Concen: 20.74 ng/ul




RT: 23.57 min Scan# 3551


Delta R.T. 0.00 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02

Tgt Ion:	252	Resp:	726016
Ion	Ratio	Lower	Upper
252	100		
253	21.4	17.0	25.6
125	6.4	5.8	8.6






#89




Indeno(1,2,3-cd)pyrene
Concen: 22.95 ng/ul
RT: 25.99 min Scan# 3962
Delta R.T. 0.00 min
Lab File: BM000304.D
Acq: 21 Jan 2015 17:02

Instrument :
LabSampleId :
SSTD02041

Tgt Ion	Ratio	Lower	Upper
276	100		
138	16.5	14.2	21.2
277	25.8	18.8	28.2




Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E





#90


Dibenzo(a,h)anthracene
Concen: 22.82 ng/ul
RT: 26.00 min Scan# 3964
Delta R.T. 0.00 min
Lab File: BM000304.D
Acq: 21 Jan 2015 17:02

Tgt Ion	Ratio	Lower	Upper
278	100		
139	10.6	10.2	15.2
279	24.0	19.2	28.8



Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E





#91

Benzo(g,h,i)perylene

Concen: 23.52 ng/ul

RT: 26.70 min Scan# 4083

Delta R.T. -0.01 min

Lab File: BM000304.D

Acq: 21 Jan 2015 17:02

Instrument :
LabSampleId :
SSTD02041

Tgt Ion: 276 Resp: 705783

Ion Ratio Lower Upper

276 100

138 14.7 12.1 18.1

277 24.4 19.0 28.4

