

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM012015\
 Data File : BM000285.D
 Acq On : 21 Jan 2015 04:42
 Operator : TP/IZ
 Sample : G1102-17
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ClientSampleId :
 C0AH5

Quant Time: Jan 21 06:14:22 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-BM122914.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 21 02:17:47 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.82	152	159029	20.00	ng/ul	0.00
18) Naphthalene-d8	10.61	136	615885	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.46	164	387240	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.20	188	745915	20.00	ng/ul	0.00
75) Chrysene-d12	21.39	240	825241	20.00	ng/ul	0.00
83) Perylene-d12	23.67	264	799549	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.31	96	18444	6.72	ng/uL	0.00
5) Phenol-d5	6.98	99	383722	28.68	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.15	67	268736	31.78	ng/ul	0.00
9) 2-Chlorophenol-d4	7.35	132	283088	28.86	ng/ul	0.00
13) 4-Methylphenol-d8	8.52	113	294678	27.69	ng/ul	0.00
19) Nitrobenzene-d5	8.97	128	153735	36.10	ng/ul	0.00
22) 2-Nitrophenol-d4	9.69	143	147302	33.88	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.23	165	273093	28.43	ng/ul	0.00
29) 4-Chloroaniline-d4	10.75	131	228170	20.93	ng/ul	0.00
43) Dimethylphthalate-d6	13.86	166	876777	30.70	ng/ul	0.00
46) Acenaphthylene-d8	14.15	160	1093157	31.95	ng/ul	0.00
51) 4-Nitrophenol-d4	14.65	143	105093	22.61	ng/ul	0.00
57) Fluorene-d10	15.45	176	733438	29.90	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.56	200	83802	23.71	ng/ul	0.00
70) Anthracene-d10	17.30	188	1095943	31.77	ng/ul	0.00
76) Pyrene-d10	19.59	212	1186807	31.18	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.53	264	1124875	31.00	ng/ul	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
6) Phenol	7.01	94	33043	2.34	ng/ul#	85
44) Dimethylphthalate	13.91	163	105957	3.72	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM012015\
Data File : BM000285.D
Acq On : 21 Jan 2015 04:42
Operator : TP/IZ
Sample : G1102-17
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ClientSampleId :
C0AH5

Quant Time: Jan 21 06:14:22 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-BM122914.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jan 21 02:17:47 2015
Response via : Initial Calibration





#6

Phenol

Concen: 2.34 ng/ul

RT: 7.01 min Scan# 735

Delta R.T. 0.01 min





Lab File: BM000285.D

Acq: 21 Jan 2015 04:42

Instrument :
ClientSampled :
C0AH5

Tgt Ion: 94 Resp: 33043

Ion	Ratio	Lower	Upper
94	100		
65	39.7	27.4	41.0
66	65.8	42.5	63.7#



#44

Dimethylphthalate

Concen: 3.72 ng/ul

RT: 13.91 min Scan# 1908

Delta R.T. -0.00 min

Lab File: BM000285.D

Acq: 21 Jan 2015 04:42

Tgt Ion: 163 Resp: 105957

Ion	Ratio	Lower	Upper
163	100		
194	3.9	3.0	4.4
164	10.2	8.5	12.7

