


Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_M\Data\BM012018\
 Data File : BM013670.D
 Acq On : 20 Jan 2018 08:07
 Operator : SJ/JU
 Sample : DFTPP78
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP78

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM011018.M
 Title : SVOA CALIBRATION
 Last Update : Mon Jan 22 02:14:14 2018



AutoFind: Scans 2460, 2461, 2462; Background Corrected with Scan 2455

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.3	67901	PASS
68	69	0.00	2	1.7	1172	PASS
69	198	0.00	100	43.2	69272	PASS
70	69	0.00	2	0.6	447	PASS
127	198	10	80	47.6	76312	PASS
197	198	0.00	2	0.5	769	PASS
198	198	100	100	100.0	160416	PASS
199	198	5	9	7.1	11444	PASS
275	198	10	60	28.8	46136	PASS
365	198	1	100	4.0	6366	PASS
441	443	0.01	100	78.0	19318	PASS
442	198	50	100	75.7	121464	PASS
443	442	15	24	20.4	24765	PASS

m/z	Abundance
38.00	744.0
39.00	5072.0
40.00	419.0
43.10	324.0
44.00	1037.0
49.00	342.0
50.10	13058.0
51.10	51344.0
52.00	2701.0
55.00	631.0
56.00	1654.0
57.00	2892.0
57.90	418.0
61.00	725.0
62.10	1109.0
63.10	1835.0
65.00	660.0
68.10	919.0
69.00	53384.0
69.90	634.0
73.00	826.0
74.10	6216.0
75.10	7851.0
76.00	2919.0
77.10	56328.0
78.10	3568.0
79.00	4568.0
80.10	4105.0
81.00	4450.0
82.10	1078.0
83.00	1174.0
85.10	807.0
86.00	1229.0
87.00	610.0
88.10	392.0
91.00	1223.0
91.90	1102.0
93.00	6947.0
93.90	658.0
96.00	590.0
97.20	442.0
98.00	6659.0
99.00	4828.0
100.00	552.0
100.90	2421.0
102.10	331.0
102.90	712.0
103.90	1622.0
105.00	1513.0
106.00	405.0
107.00	17312.0
107.90	1904.0
109.10	404.0
110.00	26248.0
110.90	3666.0
112.00	396.0
112.90	448.0
116.10	1449.0
117.00	19784.0
117.90	1594.0
119.90	316.0
122.00	1392.0
123.10	1769.0
123.90	1526.0
125.10	773.0
126.00	329.0
127.00	52528.0
128.00	4660.0
129.00	22216.0
130.00	2054.0
131.10	431.0
134.00	759.0
134.80	1110.0
136.00	1107.0
137.10	944.0
137.80	397.0
139.00	386.0
140.10	517.0
141.00	2838.0
142.10	1167.0
143.00	695.0
143.90	382.0
147.10	2282.0
148.00	4249.0
149.10	973.0
151.10	656.0
151.80	302.0
153.10	920.0
153.90	534.0
155.00	2711.0
155.90	1822.0

Instrument :
BNA_M
ClientSampleId :
DFTPP78

156.80	414.0
157.90	350.0
159.00	511.0
160.10	978.0
160.80	914.0
162.10	564.0
164.10	407.0
164.90	1246.0
165.90	1069.0
167.00	5389.0
168.00	4056.0
169.00	424.0
170.00	310.0
170.90	438.0
171.90	1079.0
172.90	724.0
174.00	1518.0
175.00	2971.0
175.80	1197.0
176.90	1748.0
178.00	428.0
178.90	4593.0
180.00	3655.0
181.00	2113.0
185.00	1922.0
186.00	16648.0
187.10	5243.0
187.90	634.0
189.10	1227.0
191.10	1212.0
191.80	1054.0
192.20	696.0
193.10	1550.0
194.20	430.0
196.00	3238.0
196.70	901.0
196.90	906.0
198.00	101152.0
198.90	8404.0
200.10	873.0
201.40	1280.0
202.80	814.0
203.10	721.0
204.00	3416.0
205.00	6747.0
206.00	29872.0
207.00	7362.0
208.00	1483.0
209.10	558.0
209.90	795.0
211.00	1130.0
212.00	554.0
215.90	729.0
217.00	8871.0
217.90	1006.0
220.90	4713.0
221.90	1238.0
223.10	1919.0
224.00	16424.0
225.10	3710.0
226.90	7004.0
228.00	1348.0
229.00	1431.0
230.10	550.0
231.00	1222.0
234.00	705.0
234.90	459.0
237.00	1261.0
240.70	368.0
241.90	940.0
243.00	869.0
244.00	12475.0
245.10	2466.0
246.00	4022.0
246.80	614.0
249.00	865.0
254.20	418.0
255.00	61816.0
256.00	10804.0
257.00	607.0
258.00	3826.0
259.10	787.0
265.00	2039.0
266.10	338.0
271.90	330.0
273.00	2316.0
274.00	5724.0
275.10	27096.0
276.00	4650.0
277.00	2890.0
278.00	788.0
281.20	371.0
283.20	513.0
285.00	361.0
293.10	518.0

Instrument :
BNA_M
ClientSampleId :
DFTPP78

294.00	340.0
296.00	8154.0
297.00	1820.0
302.80	577.0
308.00	324.0
314.90	523.0
316.00	547.0
323.00	2404.0
324.10	396.0
326.90	526.0
334.10	1526.0
346.00	712.0
351.80	825.0
354.00	697.0
355.10	467.0
364.90	2870.0
365.90	308.0
372.00	1335.0
383.00	501.0
401.90	308.0
403.00	765.0
421.00	577.0
421.80	467.0
422.20	313.0
423.10	3492.0
424.10	1254.0
425.20	349.0
441.00	8850.0
442.10	52264.0
443.10	10231.0
444.10	881.0

Instrument :
BNA_M
ClientSampleId :
DFTPP78

m/z	Abundance
37.10	409.0
38.00	1582.0
39.00	6530.0
40.00	598.0
41.10	424.0
43.90	668.0
44.90	369.0
48.90	692.0
50.10	25888.0
51.10	83120.0
52.00	4918.0
55.10	450.0
55.90	2524.0
57.10	5949.0
60.90	1067.0
61.90	1085.0
63.00	3612.0
63.90	329.0
65.00	1769.0
66.70	386.0
68.00	1393.0
69.00	84376.0
69.90	709.0
73.00	982.0
74.10	11653.0
75.00	13953.0
76.00	4745.0
77.00	89464.0
78.00	6630.0
79.00	8344.0
80.00	5949.0
81.00	7301.0
82.10	1722.0
83.00	1224.0
84.00	444.0
85.00	1121.0
85.90	2744.0
86.90	1178.0
87.90	312.0
90.90	1244.0
92.10	2065.0
93.00	10561.0
94.10	1128.0
96.00	777.0
96.80	490.0
98.00	11813.0
99.00	7753.0
99.80	1134.0
101.00	3563.0
102.00	325.0
103.10	1208.0
104.00	2705.0
105.00	2922.0
106.00	863.0
107.00	33688.0
108.00	5170.0
108.90	992.0
110.00	49232.0
111.00	8444.0
111.90	1364.0
113.10	598.0
115.90	1467.0
117.00	33400.0
118.00	1755.0
118.90	474.0
120.00	657.0
122.00	2110.0
123.00	3213.0
123.90	1722.0
125.10	1364.0
127.00	96424.0
128.10	6723.0
129.00	38856.0
130.00	3646.0
130.90	1183.0
132.10	406.0
133.00	419.0
133.90	952.0
135.00	3166.0
136.00	1739.0
136.90	1805.0
138.00	676.0
138.90	307.0
140.00	655.0
141.00	6897.0
141.90	1793.0
143.00	1600.0
144.00	536.0
145.00	377.0
145.90	1275.0
147.00	3148.0

Instrument :
BNA_M
ClientSampleId :
DFTPP78

147.90	6129.0
149.00	1294.0
149.90	503.0
151.20	1005.0
153.00	1594.0
154.00	1660.0
155.00	3104.0
156.00	5441.0
157.20	1030.0
157.80	1341.0
158.90	771.0
159.90	2141.0
161.00	2185.0
162.00	649.0
165.00	3514.0
166.00	1235.0
167.00	11849.0
168.00	6498.0
168.90	1742.0
169.80	388.0
170.80	496.0
171.90	1196.0
172.90	770.0
173.90	1961.0
175.10	3752.0
175.90	1546.0
177.00	2414.0
178.10	1171.0
179.00	9646.0
180.00	6282.0
180.90	2592.0
182.20	340.0
184.00	940.0
185.10	3573.0
186.00	27872.0
187.00	7487.0
188.10	1554.0
188.80	1846.0
190.00	385.0
191.00	1033.0
192.00	2399.0
193.00	3416.0
194.10	409.0
196.00	4937.0
197.00	1402.0
198.00	192896.0
198.90	13498.0
200.00	1322.0
201.50	1182.0
202.90	1782.0
204.00	8067.0
205.00	13333.0
206.10	54688.0
207.00	8999.0
208.00	3345.0
209.10	501.0
209.90	1012.0
211.20	1783.0
213.00	346.0
214.80	566.0
216.00	1075.0
217.00	14876.0
218.00	2477.0
219.80	417.0
221.00	8015.0
222.00	1874.0
223.00	3186.0
224.00	32688.0
225.00	7938.0
226.10	765.0
227.00	14808.0
228.10	2067.0
229.00	3435.0
230.00	579.0
231.00	1456.0
234.00	985.0
235.00	1197.0
235.90	1001.0
236.90	1306.0
239.00	562.0
239.90	621.0
240.90	567.0
242.10	1623.0
243.00	1362.0
244.10	25736.0
245.10	2585.0
246.00	6084.0
246.90	1146.0
248.90	1570.0
250.90	778.0
252.90	391.0
254.20	1351.0
255.00	125720.0
256.00	19264.0
257.00	1720.0

Instrument :
BNA_M
ClientSampleId :
DFTPP78

258.00	10783.0
259.10	1513.0
264.90	3312.0
266.00	503.0
271.20	501.0
272.00	492.0
273.00	4195.0
274.00	9046.0
275.00	55448.0
276.00	8810.0
277.00	5135.0
278.00	973.0
281.10	312.0
282.80	573.0
283.20	475.0
284.10	509.0
284.90	719.0
285.80	355.0
289.20	304.0
291.90	371.0
293.00	1226.0
293.90	636.0
295.10	506.0
296.00	18176.0
296.90	1943.0
302.10	326.0
302.90	1528.0
303.70	425.0
309.10	364.0
313.90	666.0
315.00	1943.0
315.90	606.0
320.90	499.0
322.20	328.0
323.00	5691.0
323.80	728.0
324.20	619.0
327.00	999.0
328.00	631.0
331.80	472.0
332.90	505.0
334.00	3552.0
334.90	792.0
341.10	791.0
346.10	1283.0
352.00	1028.0
353.00	1670.0
354.10	1904.0
355.20	759.0
365.00	7457.0
366.00	1532.0
372.10	1812.0
373.00	331.0
383.00	844.0
390.10	546.0
402.00	579.0
403.00	1314.0
421.10	1054.0
421.90	874.0
423.20	6266.0
424.00	2649.0
425.10	482.0
441.10	20424.0
442.10	137280.0
443.10	26280.0
444.00	2446.0

Instrument :
BNA_M
ClientSampleId :
DFTPP78

m/z	Abundance
37.10	525.0
38.10	1465.0
39.10	5038.0
39.90	499.0
44.10	1118.0
49.20	336.0
50.00	19688.0
51.00	69240.0
52.00	3131.0
52.90	347.0
56.00	1763.0
57.00	3635.0
60.80	1103.0
62.00	729.0
63.00	2297.0
63.90	473.0
64.90	1630.0
68.00	1206.0
69.00	70056.0
73.00	1290.0
74.00	7519.0
75.00	12275.0
76.10	4898.0
77.10	75160.0
78.10	5207.0
79.00	7766.0
80.00	4997.0
81.00	6205.0
81.90	1447.0
83.00	1175.0
84.00	464.0
85.00	1054.0
86.00	2290.0
86.90	457.0
91.00	1324.0
92.00	1285.0
93.00	10072.0
94.10	890.0
95.00	354.0
96.00	1119.0
97.00	520.0
98.00	9469.0
99.00	6731.0
100.00	645.0
101.00	3285.0
103.00	1070.0
103.90	2405.0
105.00	1931.0
106.10	762.0
107.00	26808.0
108.00	3386.0
110.00	41064.0
111.00	7800.0
111.90	954.0
113.00	351.0
114.90	372.0
116.00	1229.0
117.00	29888.0
118.00	2318.0
119.10	443.0
119.90	317.0
122.00	1886.0
123.00	3309.0
124.00	1331.0
124.90	708.0
127.00	79984.0
128.00	7124.0
129.00	34240.0
130.00	2260.0
131.00	397.0
133.90	1688.0
135.00	2316.0
136.00	1201.0
136.90	1335.0
137.80	392.0
138.30	378.0
140.00	513.0
140.90	3763.0
142.10	1669.0
142.90	1290.0
143.90	643.0
146.00	1017.0
147.00	3828.0
148.00	6173.0
149.00	1247.0
151.00	927.0
152.10	307.0
153.00	1609.0
153.90	632.0
154.90	2937.0
156.10	4092.0

Instrument :
BNA_M
ClientSampleId :
DFTPP78

157.00	962.0
157.80	792.0
158.90	685.0
159.90	1333.0
161.00	2783.0
162.00	896.0
165.00	1981.0
166.00	1423.0
167.00	10150.0
168.00	6464.0
168.90	406.0
170.00	527.0
171.00	425.0
172.00	1032.0
172.90	832.0
174.00	2009.0
175.00	5275.0
176.10	1263.0
176.70	1572.0
178.00	1235.0
179.00	7927.0
180.00	6622.0
181.00	2067.0
182.00	545.0
183.90	750.0
185.00	3798.0
186.00	30400.0
187.00	8488.0
188.10	499.0
188.90	1759.0
190.00	538.0
191.10	1331.0
192.00	2913.0
193.10	3466.0
193.90	434.0
196.00	4540.0
196.80	758.0
198.00	187200.0
198.90	12430.0
200.10	805.0
200.70	401.0
201.20	350.0
201.70	886.0
203.10	1941.0
204.00	8548.0
205.10	13346.0
206.00	58704.0
207.00	8238.0
208.10	3029.0
209.10	1277.0
210.00	1272.0
211.00	1978.0
212.10	633.0
213.00	448.0
214.90	1080.0
216.00	885.0
217.00	18216.0
218.00	2017.0
218.90	424.0
221.00	6119.0
221.90	2342.0
223.00	3449.0
224.00	32040.0
225.10	7045.0
225.90	850.0
227.00	15078.0
228.00	2717.0
229.00	2068.0
229.90	426.0
231.00	1296.0
232.20	407.0
233.00	412.0
233.90	1238.0
234.90	686.0
235.90	853.0
236.90	1428.0
239.10	541.0
239.90	517.0
241.00	712.0
241.90	1801.0
243.00	1248.0
244.00	24856.0
245.10	2896.0
246.00	5437.0
247.00	1716.0
248.00	364.0
249.00	923.0
249.80	421.0
252.70	417.0
254.20	1070.0
255.00	127888.0
256.00	17592.0
257.00	2004.0
258.00	9508.0
258.80	1045.0

Instrument :
BNA_M
ClientSampleId :
DFTPP78

265.00	3514.0
265.90	447.0
271.10	574.0
271.80	322.0
273.10	3771.0
274.10	9432.0
275.00	55864.0
276.00	5951.0
277.00	5183.0
278.10	573.0
280.90	446.0
283.00	730.0
284.90	808.0
289.90	459.0
293.10	1151.0
294.10	302.0
296.00	20840.0
297.00	3570.0
303.10	2480.0
303.90	618.0
308.00	315.0
314.00	549.0
315.00	1993.0
316.10	1515.0
320.80	774.0
322.10	515.0
323.00	3829.0
324.00	1044.0
327.00	1193.0
331.70	447.0
332.80	458.0
334.00	4499.0
334.80	1050.0
341.10	957.0
346.10	1139.0
352.00	1964.0
353.00	1256.0
354.10	1944.0
355.00	775.0
365.00	8773.0
365.90	1030.0
371.00	528.0
372.10	2274.0
373.20	877.0
383.20	851.0
389.90	524.0
390.20	511.0
391.10	555.0
400.90	314.0
402.00	498.0
403.10	1554.0
404.20	1176.0
421.00	884.0
422.10	1666.0
423.10	10422.0
424.00	3392.0
425.10	333.0
441.10	28680.0
442.10	174848.0
443.10	37784.0
444.00	2545.0

Instrument :
BNA_M
ClientSampleId :
DFTPP78

Scan 2455 (17.427 min): BM013670.D
DFTPP78

m/z	Abundance
43.90	781.0
73.00	402.0
91.00	482.0
103.10	329.0
191.10	344.0
192.80	471.0
207.10	2105.0
207.90	411.0
208.90	427.0
281.10	351.0
282.10	319.0

Instrument :
BNA_M
ClientSampleId :
DFTPP78