

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM012117\
 Data File : BM008787.D
 Acq On : 22 Jan 2017 01:01
 Operator : UM/SJ
 Sample : I1323-22
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DA9R7

Quant Time: Jan 23 05:02:53 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM010617.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 23 04:58:16 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.96	152	94630	20.00	ng/ul	0.00
18) Naphthalene-d8	10.77	136	417714	20.00	ng/ul	0.02
35) Acenaphthene-d10	14.59	164	528233	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.34	188	739057	20.00	ng/ul	0.02
75) Chrysene-d12	21.49	240	824518	20.00	ng/ul	0.00
83) Perylene-d12	23.85	264	748229	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.41	96	2318	1.13	ng/uL	0.01
5) Phenol-d5	7.09	99	25182	3.24	ng/ul	-0.01
7) Bis-(2-Chloroethyl)ether-d	7.29	67	194458	36.86	ng/ul	0.00
9) 2-Chlorophenol-d4	7.50	132	134871	24.91	ng/ul	0.02
13) 4-Methylphenol-d8	8.67	113	99556	16.51	ng/ul	0.03
19) Nitrobenzene-d5	9.13	128	100706	39.07	ng/ul	0.00
22) 2-Nitrophenol-d4	9.85	143	120417	42.99	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.39	165	220728	34.22	ng/ul	0.02
29) 4-Chloroaniline-d4	10.94	131	4531	0.63	ng/ul	0.05
43) Dimethylphthalate-d6	14.09	166	833646	23.92	ng/ul	0.10
46) Acenaphthylene-d8	14.28	160	977340	23.78	ng/ul	0.00
51) 4-Nitrophenol-d4	14.79	143	999	0.18	ng/ul	0.03
57) Fluorene-d10	15.57	176	732078	22.78	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.76	200	25311505	6644.32	ng/ul	0.07
70) Anthracene-d10	17.44	188	1208822	37.97	ng/ul	0.02
76) Pyrene-d10	19.72	212	1365777	37.59	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.70	264	1163624	38.13	ng/ul	0.00

Target Compounds

						Qvalue
4) Benzaldehyde	7.10	77	28277	4.99	ng/ul#	1
11) 2-Methylphenol	8.42	108	45717	7.45	ng/ul#	1
14) Acetophenone	8.79	105	63670	6.01	ng/ul#	60
16) 4-Methylphenol	8.74	108	51157	7.72	ng/ul	90
17) Hexachloroethane	9.06	117	51386	17.48	ng/ul#	4
20) Nitrobenzene	9.14	77	25467	2.78	ng/ul#	1
21) Isophorone	9.70	82	979930	60.54	ng/ul#	90
24) 2,4-Dimethylphenol	9.94	107	101450	11.95	ng/ul#	81
25) Bis(2-Chloroethoxy)methane	10.19	93	19144	2.24	ng/ul#	1
27) 2,4-Dichlorophenol	10.42	162	27921	4.37	ng/ul#	78
30) 4-Chloroaniline	10.85	127	7121879	972.36	ng/ul#	18
34) 2-Methylnaphthalene	12.42	142	4904890	331.44	ng/ul	99
38) 2,4,6-Trichlorophenol	13.11	196	10541546	1102.61	ng/ul	99
39) 2,4,5-Trichlorophenol	13.11	196	10541546	1023.16	ng/ul	99
40) 1,1'-Biphenyl	13.42	154	710344	21.44	ng/ul	99
41) 2-Chloronaphthalene	13.60	162	4465237	171.74	ng/ul#	43
47) Acenaphthylene	14.31	152	123778	3.07	ng/ul#	67
49) Acenaphthene	14.65	153	1325797	47.56	ng/ul	96
52) 4-Nitrophenol	14.79	109	31842	3.36	ng/ul#	62
53) Dibenzofuran	14.99	168	1983312	46.76	ng/ul	88
55) 2,3,4,6-Tetrachlorophenol	15.21	232	1040867	106.87	ng/ul#	90
58) Fluorene	15.63	166	1047642	30.13	ng/ul	98
60) 4-Nitroaniline	15.48	138	7961	1.18	ng/ul	98

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM012117\
Data File : BM008787.D
Acq On : 22 Jan 2017 01:01
Operator : UM/SJ
Sample : I1323-22
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DA9R7

Quant Time: Jan 23 05:02:53 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM010617.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jan 23 04:58:16 2017
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
64) N-Nitrosodiphenylamine	15.76	169	4590407	247.40	ng/ul#	33
68) Pentachlorophenol	17.02	266	13758893	2552.07	ng/ul	97
69) Phenanthrene	17.39	178	2939074	82.62	ng/ul	97
71) Anthracene	17.47	178	235837	6.45	ng/ul	95
72) Carbazole	17.75	167	5544540	170.39	ng/ul	98
74) Fluoranthene	19.38	202	195542	4.27	ng/ul	98
77) Pyrene	19.74	202	77187	1.74	ng/ul#	95
81) Bis(2-ethylhexyl)phthalate	21.41	149	26701	1.17	ng/ul#	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM012117\
Data File : BM008787.D
Acq On : 22 Jan 2017 01:01
Operator : UM/SJ
Sample : I1323-22
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DA9R7

Quant Time: Jan 23 05:02:53 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM010617.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jan 23 04:58:16 2017
Response via : Initial Calibration





#4

Benzaldehyde

Concen: 4.99 ng/ul

RT: 7.10 min Scan# 767

Delta R.T. 0.00 min

Lab File: BM008787.D

Acq: 22 Jan 2017 01:01

Instrument :

BNA_M

ClientSampleId :

DA9R7





Tgt Ion: 77 Resp: 28277

Ion Ratio Lower Upper

77 100

105 372.9 76.5 114.7#

106 36.9 71.3 106.9#



#11

2-Methylphenol

Concen: 7.45 ng/ul

RT: 8.42 min Scan# 991

Delta R.T. 0.04 min

Lab File: BM008787.D




Acq: 22 Jan 2017 01:01


Tgt Ion: 108 Resp: 45717

Ion Ratio Lower Upper

108 100

107 259.9 74.6 111.8#





#14

Acetophenone

Concen: 6.01 ng/ul

RT: 8.79 min Scan# 1054

Delta R.T. 0.01 min

Lab File: BM008787.D

Acq: 22 Jan 2017 01:01

Instrument :

BNA_M

ClientSampleId :

DA9R7


Tgt Ion:105 Resp: 63670

Ion Ratio Lower Upper

105 100

77 107.5 64.2 96.4#




51 59.6 19.8 29.6#



Abundance Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BM0

Ion 51.00 (50.70 to 51.70): BM0



#16

4-Methylphenol

Concen: 7.72 ng/ul

RT: 8.74 min Scan# 1046

Delta R.T. 0.03 min

Lab File: BM008787.D


Acq: 22 Jan 2017 01:01

Tgt Ion:108 Resp: 51157

Ion Ratio Lower Upper



108 100


107 102.3 90.3 135.5



Abundance Ion 108.00 (107.70 to 108.70): E

Ion 107.00 (106.70 to 107.70): E





#17

Hexachloroethane

Concen: 17.48 ng/ul

RT: 9.06 min Scan# 1101

Delta R.T. 0.03 min

Lab File: BM008787.D

Acq: 22 Jan 2017 01:01

Instrument :

BNA_M

ClientSampleId :

DA9R7


Tgt Ion:117 Resp: 51386

Ion Ratio Lower Upper

117 100

201 0.0 85.4 128.2#




199 0.0 54.1 81.1#



Abundance Ion 117.00 (116.70 to 117.70): E

Ion 201.00 (200.70 to 201.70): E

Ion 199.00 (198.70 to 199.70): E



#20

Nitrobenzene

Concen: 2.78 ng/ul

RT: 9.14 min Scan# 1114

Delta R.T. -0.02 min

Lab File: BM008787.D

Acq: 22 Jan 2017 01:01


Tgt Ion: 77 Resp: 25467

Ion Ratio Lower Upper

77 100

123 0.0 32.2 48.2#



65 102.1 11.0 16.6#




Abundance Ion 77.00 (76.70 to 77.70): BM0

Ion 123.00 (122.70 to 123.70): E

Ion 65.00 (64.70 to 65.70): BM0





#21

Isophorone

Concen: 60.54 ng/ul

RT: 9.70 min Scan# 1209

Delta R.T. 0.02 min

Lab File: BM008787.D

Acq: 22 Jan 2017 01:01

Instrument :

BNA_M

ClientSampleId :

DA9R7


Tgt Ion: 82 Resp: 979930

Ion Ratio Lower Upper

82 100

95 10.4 6.1 9.1#




138 12.0 13.4 20.2#



Abundance Ion 82.00 (81.70 to 82.70): BM008787.D (-1172) (-)

Ion 95.00 (94.70 to 95.70): BM008787.D (-1172) (-)

Ion 138.00 (137.70 to 138.70): BM008787.D (-1172) (-)



#24

2,4-Dimethylphenol

Concen: 11.95 ng/ul

RT: 9.94 min Scan# 1250

Delta R.T. 0.02 min

Lab File: BM008787.D

Acq: 22 Jan 2017 01:01


Tgt Ion: 107 Resp: 101450

Ion Ratio Lower Upper

107 100

121 38.2 38.0 57.0



122 58.6 62.2 93.2#




Abundance Ion 107.00 (106.70 to 107.70): BM008787.D (-1212) (-)

Ion 121.00 (120.70 to 121.70): BM008787.D (-1212) (-)

Ion 122.00 (121.70 to 122.70): BM008787.D (-1212) (-)









#25
 Bis(2-Chloroethoxy)methane
 Concen: 2.24 ng/ul
 RT: 10.19 min Scan# 1292
 Delta R.T. 0.02 min
 Lab File: BM008787.D
 Acq: 22 Jan 2017 01:01




Instrument :
 BNA_M
 ClientSampleId :
 DA9R7


Tgt Ion: 93 Resp: 19144
 Ion Ratio Lower Upper
 93 100
 95 448.5 24.2 36.2#
 123 17.9 8.2 12.2#



#27
 2,4-Dichlorophenol
 Concen: 4.37 ng/ul
 RT: 10.42 min Scan# 1332
 Delta R.T. 0.02 min
 Lab File: BM008787.D
 Acq: 22 Jan 2017 01:01

Tgt Ion: 162 Resp: 27921
 Ion Ratio Lower Upper
 162 100
 164 73.8 50.8 76.2
 98 58.7 29.8 44.6#





#30

4-Chloroaniline

Concen: 972.36 ng/ul

RT: 10.85 min Scan# 1405

Delta R.T. -0.06 min

Lab File: BM008787.D

Acq: 22 Jan 2017 01:01

Instrument :

BNA_M

ClientSampleId :


DA9R7

Tgt Ion:127 Resp: 7121879

Ion Ratio Lower Upper




127 100

129 80.5 27.1 40.7#



Abundance Ion 127.00 (126.70 to 127.70): E

Ion 129.00 (128.70 to 129.70): E



#34

2-Methylnaphthalene

Concen: 331.44 ng/ul

RT: 12.42 min Scan# 1672

Delta R.T. 0.01 min

Lab File: BM008787.D


Acq: 22 Jan 2017 01:01

Tgt Ion:142 Resp: 4904890

Ion Ratio Lower Upper



142 100


141 90.5 73.0 109.4



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E









#38
 2,4,6-Trichlorophenol
 Concen: 1102.61 ng/ul
 RT: 13.11 min Scan# 1789
 Delta R.T. 0.10 min
 Lab File: BM008787.D
 Acq: 22 Jan 2017 01:01

Instrument :
 BNA_M
 ClientSampleId :
 DA9R7

Tgt Ion:196 Resp:10541546
 Ion Ratio Lower Upper
 196 100
 198 97.0 78.8 118.2
 200 32.4 25.7 38.5




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E






#39
 2,4,5-Trichlorophenol
 Concen: 1023.16 ng/ul
 RT: 13.11 min Scan# 1789
 Delta R.T. 0.04 min
 Lab File: BM008787.D
 Acq: 22 Jan 2017 01:01

Tgt Ion:196 Resp:10541546
 Ion Ratio Lower Upper
 196 100
 198 97.0 77.8 116.6
 200 32.4 23.6 35.4



Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E









#40
 1,1'-Biphenyl
 Concen: 21.44 ng/ul
 RT: 13.42 min Scan# 1842
 Delta R.T. 0.01 min
 Lab File: BM008787.D
 Acq: 22 Jan 2017 01:01




Instrument :
 BNA_M
 ClientSampleId :
 DA9R7


Tgt Ion:154 Resp: 710344
 Ion Ratio Lower Upper
 154 100
 153 40.5 32.2 48.4
 76 13.8 9.8 14.6



#41
 2-Chloronaphthalene
 Concen: 171.74 ng/ul
 RT: 13.60 min Scan# 1873
 Delta R.T. 0.14 min
 Lab File: BM008787.D
 Acq: 22 Jan 2017 01:01

Tgt Ion:162 Resp: 4465237
 Ion Ratio Lower Upper
 162 100
 164 63.9 26.3 39.5#
 127 3.9 31.7 47.5#





#47

Acenaphthylene

Concen: 3.07 ng/ul

RT: 14.31 min Scan# 1993

Delta R.T. 0.01 min

Lab File: BM008787.D

Acq: 22 Jan 2017 01:01

Instrument :

BNA_M

ClientSampleId :

DA9R7





Tgt Ion:152 Resp: 123778

Ion Ratio Lower Upper

152 100

151 29.0 16.0 24.0#

153 34.8 10.4 15.6#



#49

Acenaphthene

Concen: 47.56 ng/ul

RT: 14.65 min Scan# 2051

Delta R.T. 0.01 min

Lab File: BM008787.D

Acq: 22 Jan 2017 01:01




Tgt Ion:153 Resp: 1325797


Ion Ratio Lower Upper

153 100

152 50.1 36.4 54.6

154 88.5 68.7 103.1





#52

4-Nitrophenol

Concen: 3.36 ng/ul

RT: 14.79 min Scan# 2075

Delta R.T. 0.02 min

Lab File: BM008787.D

Acq: 22 Jan 2017 01:01

Instrument :

BNA_M

ClientSampleId :

DA9R7


Tgt Ion:109 Resp: 31842

Ion Ratio Lower Upper

109 100

139 48.0 85.4 128.2#




65 146.3 97.7 146.5



Abundance Ion 109.00 (108.70 to 109.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 65.00 (64.70 to 65.70): BM



#53

Dibenzofuran

Concen: 46.76 ng/ul

RT: 14.99 min Scan# 2108

Delta R.T. 0.01 min

Lab File: BM008787.D


Acq: 22 Jan 2017 01:01

Tgt Ion:168 Resp: 1983312

Ion Ratio Lower Upper



168 100


139 44.8 30.3 45.5



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E









#55
 2,3,4,6-Tetrachlorophenol
 Concen: 106.87 ng/ul
 RT: 15.21 min Scan# 2146
 Delta R.T. 0.01 min
 Lab File: BM008787.D
 Acq: 22 Jan 2017 01:01




Instrument :
 BNA_M
 ClientSampleId :
 DA9R7


Tgt Ion:232 Resp: 1040867
 Ion Ratio Lower Upper
 232 100
 131 47.3 31.8 47.8
 130 3.9 1.8 2.6#
 166 30.4 21.6 32.4



#58
 Fluorene
 Concen: 30.13 ng/ul
 RT: 15.63 min Scan# 2218
 Delta R.T. 0.01 min
 Lab File: BM008787.D
 Acq: 22 Jan 2017 01:01

Tgt Ion:166 Resp: 1047642
 Ion Ratio Lower Upper
 166 100
 165 98.0 80.2 120.2
 167 13.7 10.9 16.3





#60

4-Nitroaniline

Concen: 1.18 ng/ul

RT: 15.48 min Scan# 2192

Delta R.T. -0.16 min

Lab File: BM008787.D

Acq: 22 Jan 2017 01:01

Instrument :

BNA_M

ClientSampleId :

DA9R7


Tgt Ion:138 Resp: 7961

Ion Ratio Lower Upper

138 100

92 57.6 44.9 67.3




108 94.2 73.8 110.6



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BM0

Ion 108.00 (107.70 to 108.70): E



#64

N-Nitrosodiphenylamine

Concen: 247.40 ng/ul

RT: 15.76 min Scan# 2239

Delta R.T. -0.08 min

Lab File: BM008787.D

Acq: 22 Jan 2017 01:01


Tgt Ion:169 Resp: 4590407

Ion Ratio Lower Upper

169 100

168 32.9 53.1 79.7#



167 102.5 28.6 43.0#




Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E









#68
Pentachlorophenol
Concen: 2552.07 ng/ul
RT: 17.02 min Scan# 2453
Delta R.T. 0.05 min
Lab File: BM008787.D
Acq: 22 Jan 2017 01:01

Instrument :
BNA_M
ClientSampleId :
DA9R7

Tgt Ion:266 Resp:13758893
Ion Ratio Lower Upper
266 100
264 63.5 50.6 75.8
268 65.1 49.1 73.7




Abundance Ion 265.70 (265.40 to 266.40): E
Ion 264.00 (263.70 to 264.70): E
Ion 268.00 (267.70 to 268.70): E






#69
Phenanthrene
Concen: 82.62 ng/ul
RT: 17.39 min Scan# 2516
Delta R.T. 0.02 min
Lab File: BM008787.D
Acq: 22 Jan 2017 01:01

Tgt Ion:178 Resp: 2939074
Ion Ratio Lower Upper
178 100
179 15.5 12.2 18.2
176 20.1 14.7 22.1



Abundance Ion 178.00 (177.70 to 178.70): E
Ion 179.00 (178.70 to 179.70): E
Ion 176.00 (175.70 to 176.70): E









#71
 Anthracene
 Concen: 6.45 ng/ul
 RT: 17.47 min Scan# 2531
 Delta R.T. 0.02 min
 Lab File: BM008787.D
 Acq: 22 Jan 2017 01:01




Instrument :
 BNA_M
 ClientSampleId :
 DA9R7


Tgt Ion:178 Resp: 235837
 Ion Ratio Lower Upper
 178 100
 179 16.4 11.8 17.8
 176 21.2 14.6 21.8



#72
 Carbazole
 Concen: 170.39 ng/ul
 RT: 17.75 min Scan# 2578
 Delta R.T. 0.02 min
 Lab File: BM008787.D
 Acq: 22 Jan 2017 01:01

Tgt Ion:167 Resp: 5544540
 Ion Ratio Lower Upper
 167 100
 166 21.3 16.8 25.2
 139 14.2 10.4 15.6









#74
 Fluoranthene
 Concen: 4.27 ng/ul
 RT: 19.38 min Scan# 2855
 Delta R.T. 0.01 min
 Lab File: BM008787.D
 Acq: 22 Jan 2017 01:01




Instrument :
 BNA_M
 ClientSampleId :
 DA9R7


Tgt Ion	Ratio	Lower	Upper
202	100		
101	9.4	7.1	10.7
100	8.0	5.4	8.0



#77
 Pyrene
 Concen: 1.74 ng/ul
 RT: 19.74 min Scan# 2917
 Delta R.T. 0.01 min
 Lab File: BM008787.D
 Acq: 22 Jan 2017 01:01

Tgt Ion	Ratio	Lower	Upper
202	100		
101	10.9	8.6	12.8
100	12.3	6.5	9.7#





#81
 Bis(2-ethylhexyl)phthalate
 Concen: 1.17 ng/ul
 RT: 21.41 min Scan# 3201
 Delta R.T. 0.03 min
 Lab File: BM008787.D
 Acq: 22 Jan 2017 01:01

Instrument :
 BNA_M
 ClientSampleId :
 DA9R7

Tgt Ion:149 Resp: 26701
 Ion Ratio Lower Upper
 149 100
 167 28.7 22.5 33.7
 279 2.1 4.3 6.5#

