


Data Path : Z:\HPCHEM1\BNA\_M\DATA\BM012117\  
 Data File : BM008769.D  
 Acq On : 21 Jan 2017 12:03  
 Operator : UM/SJ  
 Sample : DFTPP55  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DFTPP55

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA\_M\METHODS\SOM02.2-TEST.M  
 Title : SVOA CALIBRATION  
 Last Update : Wed Nov 30 00:32:18 2016



AutoFind: Scans 2576, 2577, 2578; Background Corrected with Scan 2571

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 67.0      | 58442   | PASS             |
| 68          | 69           | 0.00         | 2            | 2.0       | 968     | PASS             |
| 69          | 198          | 0.00         | 100          | 56.9      | 49573   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.2       | 111     | PASS             |
| 127         | 198          | 10           | 80           | 45.5      | 39704   | PASS             |
| 197         | 198          | 0.00         | 2            | 1.0       | 863     | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 87181   | PASS             |
| 199         | 198          | 5            | 9            | 7.3       | 6385    | PASS             |
| 275         | 198          | 10           | 60           | 28.2      | 24596   | PASS             |
| 365         | 198          | 1            | 100          | 6.8       | 5892    | PASS             |
| 441         | 443          | 0.01         | 100          | 77.3      | 9455    | PASS             |
| 442         | 198          | 50           | 100          | 67.5      | 58813   | PASS             |
| 443         | 442          | 15           | 24           | 20.8      | 12234   | PASS             |

Scan 2571 (17.709 min): BM008769.D  
DFTPP55

| m/z    | Abundance |
|--------|-----------|
| 73.10  | 1178.0    |
| 191.10 | 604.0     |
| 193.10 | 515.0     |
| 207.00 | 2361.0    |
| 207.90 | 590.0     |
| 281.10 | 602.0     |

**Instrument :**  
BNA\_M  
**ClientSampleId :**  
DFTPP55

| m/z    | Abundance |
|--------|-----------|
| 37.10  | 429.0     |
| 38.00  | 682.0     |
| 39.00  | 3429.0    |
| 40.10  | 522.0     |
| 43.90  | 346.0     |
| 49.00  | 402.0     |
| 50.10  | 14625.0   |
| 51.00  | 51568.0   |
| 52.00  | 2883.0    |
| 55.00  | 716.0     |
| 55.90  | 1371.0    |
| 57.00  | 3810.0    |
| 61.10  | 801.0     |
| 61.80  | 938.0     |
| 63.00  | 2409.0    |
| 64.00  | 309.0     |
| 65.00  | 438.0     |
| 68.00  | 748.0     |
| 69.00  | 41736.0   |
| 70.10  | 335.0     |
| 73.00  | 1397.0    |
| 74.10  | 5509.0    |
| 75.00  | 6104.0    |
| 76.00  | 2568.0    |
| 77.00  | 49992.0   |
| 78.00  | 3067.0    |
| 79.00  | 4859.0    |
| 79.90  | 3075.0    |
| 81.00  | 4081.0    |
| 82.10  | 850.0     |
| 83.00  | 611.0     |
| 84.90  | 897.0     |
| 85.90  | 1060.0    |
| 86.90  | 846.0     |
| 90.70  | 460.0     |
| 91.10  | 719.0     |
| 92.10  | 879.0     |
| 92.90  | 7927.0    |
| 95.80  | 381.0     |
| 98.00  | 6146.0    |
| 99.00  | 4750.0    |
| 100.00 | 300.0     |
| 100.80 | 844.0     |
| 101.20 | 919.0     |
| 102.80 | 665.0     |
| 104.00 | 984.0     |
| 104.90 | 1150.0    |
| 106.00 | 645.0     |
| 107.00 | 14740.0   |
| 108.00 | 2498.0    |
| 108.70 | 360.0     |
| 110.00 | 18856.0   |
| 110.90 | 3217.0    |
| 112.00 | 337.0     |
| 112.80 | 395.0     |
| 115.80 | 922.0     |
| 117.00 | 20128.0   |
| 117.90 | 1016.0    |
| 122.10 | 483.0     |
| 123.10 | 1317.0    |
| 124.00 | 1068.0    |
| 125.00 | 843.0     |
| 127.00 | 31112.0   |
| 128.00 | 2256.0    |
| 129.00 | 16424.0   |
| 130.00 | 1431.0    |
| 134.00 | 902.0     |
| 135.00 | 1675.0    |
| 135.90 | 1023.0    |
| 136.80 | 845.0     |
| 138.10 | 685.0     |
| 139.80 | 331.0     |
| 140.90 | 2203.0    |
| 141.90 | 1289.0    |
| 143.00 | 491.0     |
| 145.90 | 470.0     |
| 147.20 | 1876.0    |
| 148.00 | 4201.0    |
| 148.80 | 712.0     |
| 151.30 | 345.0     |
| 152.90 | 1087.0    |
| 153.80 | 696.0     |
| 154.90 | 1941.0    |
| 156.00 | 1647.0    |
| 156.90 | 428.0     |
| 157.80 | 569.0     |
| 159.10 | 584.0     |
| 160.00 | 761.0     |
| 161.00 | 932.0     |
| 164.90 | 1254.0    |
| 165.90 | 597.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP55

|        |         |
|--------|---------|
| 167.00 | 6915.0  |
| 168.00 | 3268.0  |
| 169.00 | 546.0   |
| 171.90 | 630.0   |
| 172.90 | 675.0   |
| 174.00 | 551.0   |
| 175.00 | 2020.0  |
| 177.00 | 1268.0  |
| 178.80 | 3820.0  |
| 180.00 | 2510.0  |
| 180.90 | 1100.0  |
| 183.80 | 366.0   |
| 185.10 | 1994.0  |
| 186.10 | 11256.0 |
| 186.90 | 4852.0  |
| 187.90 | 715.0   |
| 188.90 | 791.0   |
| 189.80 | 380.0   |
| 191.00 | 1228.0  |
| 191.80 | 1675.0  |
| 192.90 | 1745.0  |
| 193.90 | 1065.0  |
| 195.90 | 1570.0  |
| 197.10 | 793.0   |
| 197.90 | 64000.0 |
| 198.90 | 4364.0  |
| 203.00 | 698.0   |
| 204.00 | 3719.0  |
| 205.00 | 5428.0  |
| 206.00 | 19280.0 |
| 207.00 | 4557.0  |
| 208.10 | 665.0   |
| 208.90 | 618.0   |
| 210.00 | 648.0   |
| 211.00 | 1339.0  |
| 212.10 | 315.0   |
| 215.90 | 861.0   |
| 216.90 | 8332.0  |
| 217.90 | 1391.0  |
| 221.20 | 1191.0  |
| 222.00 | 861.0   |
| 223.10 | 1168.0  |
| 224.00 | 13115.0 |
| 224.90 | 3024.0  |
| 226.10 | 415.0   |
| 227.00 | 6810.0  |
| 228.10 | 1309.0  |
| 228.70 | 1040.0  |
| 229.10 | 928.0   |
| 231.00 | 653.0   |
| 233.90 | 535.0   |
| 234.90 | 446.0   |
| 236.00 | 389.0   |
| 237.00 | 1030.0  |
| 240.90 | 440.0   |
| 241.80 | 998.0   |
| 243.00 | 743.0   |
| 244.00 | 8175.0  |
| 245.00 | 1574.0  |
| 246.00 | 3400.0  |
| 246.70 | 957.0   |
| 248.80 | 467.0   |
| 251.20 | 483.0   |
| 252.90 | 420.0   |
| 254.00 | 791.0   |
| 255.00 | 45752.0 |
| 255.90 | 6122.0  |
| 257.00 | 396.0   |
| 257.90 | 3664.0  |
| 258.90 | 491.0   |
| 265.00 | 2345.0  |
| 271.90 | 353.0   |
| 273.00 | 989.0   |
| 273.90 | 5292.0  |
| 275.00 | 14022.0 |
| 276.10 | 3049.0  |
| 277.00 | 2862.0  |
| 278.00 | 534.0   |
| 280.80 | 1184.0  |
| 283.10 | 451.0   |
| 292.90 | 346.0   |
| 295.00 | 307.0   |
| 296.00 | 6734.0  |
| 296.90 | 829.0   |
| 310.00 | 326.0   |
| 314.10 | 319.0   |
| 314.80 | 846.0   |
| 322.90 | 1185.0  |
| 323.90 | 531.0   |
| 326.90 | 1004.0  |
| 334.00 | 989.0   |
| 334.80 | 567.0   |
| 345.90 | 659.0   |
| 352.00 | 612.0   |
| 353.90 | 428.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP55

|        |         |
|--------|---------|
| 355.10 | 343.0   |
| 364.90 | 2696.0  |
| 365.90 | 913.0   |
| 372.10 | 882.0   |
| 372.90 | 418.0   |
| 383.00 | 400.0   |
| 402.80 | 648.0   |
| 420.90 | 378.0   |
| 422.80 | 1232.0  |
| 424.00 | 714.0   |
| 441.00 | 4128.0  |
| 442.00 | 31696.0 |
| 443.00 | 7062.0  |
| 444.00 | 1412.0  |

**Instrument :**  
BNA\_M  
**ClientSampleId :**  
DFTPP55

| m/z    | Abundance |
|--------|-----------|
| 37.10  | 341.0     |
| 38.00  | 1409.0    |
| 39.00  | 5414.0    |
| 44.00  | 630.0     |
| 49.10  | 1053.0    |
| 50.10  | 18712.0   |
| 51.00  | 67152.0   |
| 52.00  | 4735.0    |
| 55.00  | 775.0     |
| 56.00  | 2136.0    |
| 57.00  | 4473.0    |
| 60.90  | 741.0     |
| 62.00  | 996.0     |
| 62.90  | 2350.0    |
| 64.00  | 460.0     |
| 64.90  | 1435.0    |
| 67.90  | 1248.0    |
| 69.00  | 58160.0   |
| 73.00  | 1427.0    |
| 74.00  | 10033.0   |
| 75.00  | 11603.0   |
| 76.00  | 4675.0    |
| 77.00  | 67192.0   |
| 78.10  | 6060.0    |
| 79.00  | 4939.0    |
| 79.90  | 3931.0    |
| 81.00  | 5071.0    |
| 81.80  | 1490.0    |
| 82.90  | 1253.0    |
| 84.90  | 608.0     |
| 86.00  | 1736.0    |
| 87.00  | 1179.0    |
| 91.00  | 1929.0    |
| 91.90  | 1522.0    |
| 93.00  | 12197.0   |
| 94.00  | 463.0     |
| 95.70  | 360.0     |
| 96.20  | 417.0     |
| 98.00  | 8762.0    |
| 99.00  | 5529.0    |
| 99.90  | 809.0     |
| 100.90 | 2309.0    |
| 102.80 | 1224.0    |
| 104.00 | 1872.0    |
| 105.10 | 1941.0    |
| 106.00 | 465.0     |
| 107.00 | 24312.0   |
| 108.00 | 2331.0    |
| 108.90 | 1050.0    |
| 110.00 | 29048.0   |
| 111.00 | 5323.0    |
| 111.90 | 1341.0    |
| 113.00 | 664.0     |
| 115.90 | 1537.0    |
| 117.00 | 34936.0   |
| 117.90 | 2496.0    |
| 120.10 | 347.0     |
| 122.00 | 1153.0    |
| 123.00 | 1943.0    |
| 124.00 | 1382.0    |
| 125.00 | 963.0     |
| 126.20 | 396.0     |
| 127.00 | 45128.0   |
| 127.90 | 4996.0    |
| 129.00 | 26384.0   |
| 129.90 | 2545.0    |
| 130.90 | 727.0     |
| 133.10 | 661.0     |
| 133.90 | 1201.0    |
| 135.00 | 3042.0    |
| 136.00 | 1435.0    |
| 137.00 | 1095.0    |
| 141.00 | 3683.0    |
| 141.80 | 1038.0    |
| 143.00 | 464.0     |
| 146.10 | 686.0     |
| 147.00 | 2901.0    |
| 147.90 | 6352.0    |
| 148.90 | 1403.0    |
| 150.70 | 462.0     |
| 151.70 | 759.0     |
| 152.90 | 1044.0    |
| 154.00 | 1019.0    |
| 155.00 | 3098.0    |
| 156.00 | 2791.0    |
| 156.80 | 709.0     |
| 158.00 | 1036.0    |
| 158.70 | 425.0     |
| 159.10 | 502.0     |
| 160.00 | 1132.0    |
| 160.90 | 2002.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP55

|        |          |
|--------|----------|
| 162.10 | 578.0    |
| 164.90 | 2072.0   |
| 165.90 | 946.0    |
| 166.90 | 9467.0   |
| 167.90 | 4602.0   |
| 169.00 | 1056.0   |
| 171.90 | 891.0    |
| 173.10 | 1127.0   |
| 174.00 | 1645.0   |
| 175.00 | 3189.0   |
| 175.70 | 722.0    |
| 176.10 | 957.0    |
| 176.90 | 1724.0   |
| 177.70 | 658.0    |
| 178.90 | 6381.0   |
| 180.00 | 3471.0   |
| 180.90 | 2364.0   |
| 184.10 | 949.0    |
| 184.90 | 3584.0   |
| 186.00 | 21240.0  |
| 186.90 | 6063.0   |
| 187.90 | 1024.0   |
| 188.80 | 1702.0   |
| 190.80 | 765.0    |
| 192.00 | 2380.0   |
| 193.00 | 2698.0   |
| 193.90 | 624.0    |
| 194.90 | 710.0    |
| 195.90 | 3209.0   |
| 197.00 | 861.0    |
| 197.90 | 102376.0 |
| 199.00 | 7056.0   |
| 200.00 | 888.0    |
| 201.50 | 415.0    |
| 202.90 | 1878.0   |
| 203.90 | 6295.0   |
| 205.00 | 8870.0   |
| 206.00 | 35936.0  |
| 207.00 | 7079.0   |
| 208.00 | 2320.0   |
| 209.00 | 914.0    |
| 209.70 | 623.0    |
| 211.00 | 1120.0   |
| 215.10 | 397.0    |
| 216.10 | 1072.0   |
| 216.90 | 10359.0  |
| 217.90 | 1248.0   |
| 220.80 | 2911.0   |
| 222.00 | 1478.0   |
| 223.00 | 1828.0   |
| 224.00 | 18048.0  |
| 225.10 | 7177.0   |
| 226.10 | 644.0    |
| 227.00 | 13324.0  |
| 227.80 | 1729.0   |
| 229.00 | 2543.0   |
| 230.80 | 798.0    |
| 234.10 | 585.0    |
| 234.90 | 1263.0   |
| 235.70 | 402.0    |
| 236.20 | 552.0    |
| 237.10 | 1426.0   |
| 237.80 | 336.0    |
| 239.80 | 374.0    |
| 240.80 | 701.0    |
| 242.00 | 1027.0   |
| 243.10 | 1212.0   |
| 244.00 | 16968.0  |
| 245.00 | 2035.0   |
| 245.90 | 4413.0   |
| 246.90 | 787.0    |
| 247.90 | 390.0    |
| 249.00 | 634.0    |
| 252.10 | 541.0    |
| 252.80 | 375.0    |
| 255.00 | 76088.0  |
| 256.00 | 12642.0  |
| 256.90 | 1076.0   |
| 257.90 | 7238.0   |
| 258.90 | 1856.0   |
| 265.00 | 2807.0   |
| 265.90 | 444.0    |
| 271.90 | 410.0    |
| 273.00 | 2737.0   |
| 273.90 | 6095.0   |
| 275.00 | 31224.0  |
| 276.00 | 4054.0   |
| 277.00 | 3886.0   |
| 277.90 | 1205.0   |
| 281.10 | 442.0    |
| 283.10 | 377.0    |
| 284.00 | 427.0    |
| 284.90 | 509.0    |
| 293.00 | 714.0    |
| 295.00 | 860.0    |

Instrument :

BNA\_M

ClientSampleId :

DFTPP55

|        |         |
|--------|---------|
| 295.90 | 12658.0 |
| 297.10 | 1462.0  |
| 301.10 | 373.0   |
| 301.90 | 525.0   |
| 303.00 | 1985.0  |
| 303.90 | 602.0   |
| 313.90 | 653.0   |
| 314.80 | 926.0   |
| 321.00 | 381.0   |
| 322.90 | 3852.0  |
| 324.00 | 412.0   |
| 327.00 | 993.0   |
| 334.00 | 2261.0  |
| 334.90 | 682.0   |
| 341.00 | 1125.0  |
| 345.90 | 1437.0  |
| 352.00 | 1470.0  |
| 353.10 | 1020.0  |
| 354.00 | 969.0   |
| 354.90 | 402.0   |
| 365.00 | 6247.0  |
| 365.90 | 960.0   |
| 371.80 | 1292.0  |
| 373.00 | 627.0   |
| 401.90 | 674.0   |
| 402.90 | 1302.0  |
| 404.10 | 787.0   |
| 421.00 | 753.0   |
| 421.80 | 922.0   |
| 423.10 | 2650.0  |
| 423.90 | 1127.0  |
| 441.00 | 10186.0 |
| 442.00 | 67904.0 |
| 443.00 | 15117.0 |
| 444.00 | 1010.0  |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP55



| m/z    | Abundance |
|--------|-----------|
| 38.00  | 1401.0    |
| 39.00  | 3934.0    |
| 40.10  | 659.0     |
| 43.00  | 320.0     |
| 48.70  | 349.0     |
| 49.20  | 625.0     |
| 50.00  | 16800.0   |
| 51.00  | 56608.0   |
| 52.00  | 3364.0    |
| 54.90  | 431.0     |
| 55.10  | 437.0     |
| 55.90  | 1411.0    |
| 57.00  | 2841.0    |
| 60.90  | 509.0     |
| 62.10  | 1154.0    |
| 62.90  | 1590.0    |
| 65.10  | 793.0     |
| 67.80  | 910.0     |
| 69.00  | 48824.0   |
| 73.00  | 1572.0    |
| 74.00  | 6391.0    |
| 75.00  | 8456.0    |
| 76.00  | 3392.0    |
| 77.00  | 54024.0   |
| 78.00  | 3562.0    |
| 79.00  | 4667.0    |
| 80.10  | 4273.0    |
| 80.90  | 3694.0    |
| 81.90  | 869.0     |
| 82.90  | 1219.0    |
| 84.90  | 949.0     |
| 86.00  | 1812.0    |
| 87.00  | 788.0     |
| 87.90  | 409.0     |
| 88.90  | 399.0     |
| 91.00  | 1183.0    |
| 91.80  | 1258.0    |
| 93.00  | 9184.0    |
| 96.10  | 696.0     |
| 97.10  | 320.0     |
| 98.00  | 7198.0    |
| 98.90  | 6263.0    |
| 100.00 | 303.0     |
| 101.00 | 1979.0    |
| 103.10 | 804.0     |
| 103.80 | 1368.0    |
| 105.00 | 1770.0    |
| 105.80 | 506.0     |
| 107.00 | 22120.0   |
| 108.00 | 3550.0    |
| 108.90 | 380.0     |
| 109.90 | 22784.0   |
| 111.00 | 4426.0    |
| 112.00 | 392.0     |
| 116.10 | 923.0     |
| 117.00 | 30296.0   |
| 118.00 | 1723.0    |
| 121.90 | 2915.0    |
| 122.90 | 2130.0    |
| 123.80 | 958.0     |
| 125.00 | 761.0     |
| 127.00 | 42872.0   |
| 128.00 | 3024.0    |
| 128.90 | 23432.0   |
| 130.00 | 2096.0    |
| 130.80 | 480.0     |
| 134.00 | 844.0     |
| 134.90 | 1350.0    |
| 136.30 | 501.0     |
| 137.00 | 479.0     |
| 138.00 | 527.0     |
| 140.90 | 3984.0    |
| 141.80 | 1467.0    |
| 143.00 | 406.0     |
| 145.80 | 425.0     |
| 147.00 | 2204.0    |
| 148.00 | 5553.0    |
| 149.10 | 1404.0    |
| 151.00 | 525.0     |
| 153.00 | 1271.0    |
| 154.00 | 801.0     |
| 155.00 | 1524.0    |
| 156.10 | 2334.0    |
| 157.10 | 338.0     |
| 158.00 | 564.0     |
| 159.00 | 520.0     |
| 159.90 | 2031.0    |
| 161.10 | 1964.0    |
| 161.90 | 658.0     |
| 165.00 | 1221.0    |
| 165.80 | 795.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP55

|        |         |
|--------|---------|
| 166.10 | 699.0   |
| 167.00 | 10111.0 |
| 168.10 | 5445.0  |
| 169.00 | 778.0   |
| 172.00 | 878.0   |
| 172.90 | 816.0   |
| 174.00 | 1643.0  |
| 175.10 | 2763.0  |
| 177.00 | 1588.0  |
| 178.00 | 377.0   |
| 178.90 | 5950.0  |
| 180.00 | 3954.0  |
| 180.90 | 1581.0  |
| 182.00 | 308.0   |
| 185.00 | 3079.0  |
| 186.00 | 20128.0 |
| 187.00 | 4875.0  |
| 188.10 | 329.0   |
| 188.90 | 1552.0  |
| 191.00 | 1079.0  |
| 192.00 | 2795.0  |
| 192.90 | 2622.0  |
| 194.10 | 319.0   |
| 196.00 | 2962.0  |
| 196.90 | 935.0   |
| 197.90 | 95168.0 |
| 198.90 | 7735.0  |
| 199.80 | 472.0   |
| 201.10 | 498.0   |
| 202.80 | 1662.0  |
| 204.00 | 6087.0  |
| 205.00 | 9766.0  |
| 206.00 | 34920.0 |
| 207.00 | 5432.0  |
| 207.80 | 1922.0  |
| 208.90 | 1502.0  |
| 210.10 | 864.0   |
| 211.00 | 1531.0  |
| 212.10 | 469.0   |
| 214.70 | 584.0   |
| 216.00 | 1116.0  |
| 217.00 | 11061.0 |
| 218.00 | 1616.0  |
| 221.00 | 3904.0  |
| 221.60 | 1481.0  |
| 223.00 | 3836.0  |
| 224.00 | 20800.0 |
| 225.00 | 4790.0  |
| 227.00 | 11368.0 |
| 228.00 | 2250.0  |
| 228.90 | 2092.0  |
| 230.00 | 380.0   |
| 230.90 | 1186.0  |
| 233.90 | 1607.0  |
| 235.00 | 807.0   |
| 235.90 | 1190.0  |
| 236.90 | 1149.0  |
| 239.90 | 582.0   |
| 240.90 | 1264.0  |
| 242.00 | 823.0   |
| 242.90 | 1550.0  |
| 244.00 | 15955.0 |
| 245.10 | 2567.0  |
| 246.00 | 3836.0  |
| 247.00 | 885.0   |
| 249.00 | 1022.0  |
| 253.10 | 697.0   |
| 254.20 | 1000.0  |
| 255.00 | 79880.0 |
| 256.00 | 12727.0 |
| 257.10 | 906.0   |
| 257.90 | 5648.0  |
| 259.10 | 1113.0  |
| 263.80 | 407.0   |
| 264.90 | 3181.0  |
| 265.90 | 575.0   |
| 270.90 | 426.0   |
| 273.00 | 2365.0  |
| 274.00 | 7171.0  |
| 275.00 | 28544.0 |
| 275.90 | 3547.0  |
| 276.90 | 4165.0  |
| 277.90 | 849.0   |
| 280.90 | 938.0   |
| 283.10 | 546.0   |
| 285.00 | 808.0   |
| 292.90 | 936.0   |
| 294.80 | 468.0   |
| 296.00 | 14024.0 |
| 297.00 | 2006.0  |
| 297.90 | 337.0   |
| 303.00 | 1414.0  |
| 303.90 | 360.0   |
| 305.20 | 387.0   |
| 310.00 | 460.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP55

|        |         |
|--------|---------|
| 313.90 | 653.0   |
| 315.00 | 1882.0  |
| 316.00 | 621.0   |
| 321.00 | 640.0   |
| 323.00 | 2912.0  |
| 323.90 | 323.0   |
| 325.00 | 311.0   |
| 326.90 | 844.0   |
| 333.00 | 761.0   |
| 334.00 | 2917.0  |
| 334.80 | 709.0   |
| 340.80 | 656.0   |
| 342.10 | 330.0   |
| 346.00 | 1660.0  |
| 352.00 | 958.0   |
| 353.10 | 1403.0  |
| 354.00 | 1139.0  |
| 354.80 | 356.0   |
| 365.00 | 8734.0  |
| 366.10 | 887.0   |
| 370.90 | 346.0   |
| 372.00 | 1741.0  |
| 373.00 | 512.0   |
| 382.80 | 582.0   |
| 401.90 | 508.0   |
| 403.00 | 1108.0  |
| 404.10 | 423.0   |
| 421.10 | 820.0   |
| 422.00 | 841.0   |
| 423.00 | 3400.0  |
| 423.90 | 627.0   |
| 441.00 | 14052.0 |
| 442.00 | 76840.0 |
| 443.00 | 14525.0 |
| 444.00 | 1807.0  |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP55

| m/z    | Abundance |
|--------|-----------|
| 38.00  | 1401.0    |
| 39.00  | 3934.0    |
| 40.10  | 659.0     |
| 43.00  | 320.0     |
| 48.70  | 349.0     |
| 49.20  | 625.0     |
| 50.00  | 16800.0   |
| 51.00  | 56608.0   |
| 52.00  | 3364.0    |
| 54.90  | 431.0     |
| 55.10  | 437.0     |
| 55.90  | 1411.0    |
| 57.00  | 2841.0    |
| 60.90  | 509.0     |
| 62.10  | 1154.0    |
| 62.90  | 1590.0    |
| 65.10  | 793.0     |
| 67.80  | 910.0     |
| 69.00  | 48824.0   |
| 73.00  | 1572.0    |
| 74.00  | 6391.0    |
| 75.00  | 8456.0    |
| 76.00  | 3392.0    |
| 77.00  | 54024.0   |
| 78.00  | 3562.0    |
| 79.00  | 4667.0    |
| 80.10  | 4273.0    |
| 80.90  | 3694.0    |
| 81.90  | 869.0     |
| 82.90  | 1219.0    |
| 84.90  | 949.0     |
| 86.00  | 1812.0    |
| 87.00  | 788.0     |
| 87.90  | 409.0     |
| 88.90  | 399.0     |
| 91.00  | 1183.0    |
| 91.80  | 1258.0    |
| 93.00  | 9184.0    |
| 96.10  | 696.0     |
| 97.10  | 320.0     |
| 98.00  | 7198.0    |
| 98.90  | 6263.0    |
| 100.00 | 303.0     |
| 101.00 | 1979.0    |
| 103.10 | 804.0     |
| 103.80 | 1368.0    |
| 105.00 | 1770.0    |
| 105.80 | 506.0     |
| 107.00 | 22120.0   |
| 108.00 | 3550.0    |
| 108.90 | 380.0     |
| 109.90 | 22784.0   |
| 111.00 | 4426.0    |
| 112.00 | 392.0     |
| 116.10 | 923.0     |
| 117.00 | 30296.0   |
| 118.00 | 1723.0    |
| 121.90 | 2915.0    |
| 122.90 | 2130.0    |
| 123.80 | 958.0     |
| 125.00 | 761.0     |
| 127.00 | 42872.0   |
| 128.00 | 3024.0    |
| 128.90 | 23432.0   |
| 130.00 | 2096.0    |
| 130.80 | 480.0     |
| 134.00 | 844.0     |
| 134.90 | 1350.0    |
| 136.30 | 501.0     |
| 137.00 | 479.0     |
| 138.00 | 527.0     |
| 140.90 | 3984.0    |
| 141.80 | 1467.0    |
| 143.00 | 406.0     |
| 145.80 | 425.0     |
| 147.00 | 2204.0    |
| 148.00 | 5553.0    |
| 149.10 | 1404.0    |
| 151.00 | 525.0     |
| 153.00 | 1271.0    |
| 154.00 | 801.0     |
| 155.00 | 1524.0    |
| 156.10 | 2334.0    |
| 157.10 | 338.0     |
| 158.00 | 564.0     |
| 159.00 | 520.0     |
| 159.90 | 2031.0    |
| 161.10 | 1964.0    |
| 161.90 | 658.0     |
| 165.00 | 1221.0    |
| 165.80 | 795.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP55

|        |         |
|--------|---------|
| 166.10 | 699.0   |
| 167.00 | 10111.0 |
| 168.10 | 5445.0  |
| 169.00 | 778.0   |
| 172.00 | 878.0   |
| 172.90 | 816.0   |
| 174.00 | 1643.0  |
| 175.10 | 2763.0  |
| 177.00 | 1588.0  |
| 178.00 | 377.0   |
| 178.90 | 5950.0  |
| 180.00 | 3954.0  |
| 180.90 | 1581.0  |
| 182.00 | 308.0   |
| 185.00 | 3079.0  |
| 186.00 | 20128.0 |
| 187.00 | 4875.0  |
| 188.10 | 329.0   |
| 188.90 | 1552.0  |
| 191.00 | 1079.0  |
| 192.00 | 2795.0  |
| 192.90 | 2622.0  |
| 194.10 | 319.0   |
| 196.00 | 2962.0  |
| 196.90 | 935.0   |
| 197.90 | 95168.0 |
| 198.90 | 7735.0  |
| 199.80 | 472.0   |
| 201.10 | 498.0   |
| 202.80 | 1662.0  |
| 204.00 | 6087.0  |
| 205.00 | 9766.0  |
| 206.00 | 34920.0 |
| 207.00 | 5432.0  |
| 207.80 | 1922.0  |
| 208.90 | 1502.0  |
| 210.10 | 864.0   |
| 211.00 | 1531.0  |
| 212.10 | 469.0   |
| 214.70 | 584.0   |
| 216.00 | 1116.0  |
| 217.00 | 11061.0 |
| 218.00 | 1616.0  |
| 221.00 | 3904.0  |
| 221.60 | 1481.0  |
| 223.00 | 3836.0  |
| 224.00 | 20800.0 |
| 225.00 | 4790.0  |
| 227.00 | 11368.0 |
| 228.00 | 2250.0  |
| 228.90 | 2092.0  |
| 230.00 | 380.0   |
| 230.90 | 1186.0  |
| 233.90 | 1607.0  |
| 235.00 | 807.0   |
| 235.90 | 1190.0  |
| 236.90 | 1149.0  |
| 239.90 | 582.0   |
| 240.90 | 1264.0  |
| 242.00 | 823.0   |
| 242.90 | 1550.0  |
| 244.00 | 15955.0 |
| 245.10 | 2567.0  |
| 246.00 | 3836.0  |
| 247.00 | 885.0   |
| 249.00 | 1022.0  |
| 253.10 | 697.0   |
| 254.20 | 1000.0  |
| 255.00 | 79880.0 |
| 256.00 | 12727.0 |
| 257.10 | 906.0   |
| 257.90 | 5648.0  |
| 259.10 | 1113.0  |
| 263.80 | 407.0   |
| 264.90 | 3181.0  |
| 265.90 | 575.0   |
| 270.90 | 426.0   |
| 273.00 | 2365.0  |
| 274.00 | 7171.0  |
| 275.00 | 28544.0 |
| 275.90 | 3547.0  |
| 276.90 | 4165.0  |
| 277.90 | 849.0   |
| 280.90 | 938.0   |
| 283.10 | 546.0   |
| 285.00 | 808.0   |
| 292.90 | 936.0   |
| 294.80 | 468.0   |
| 296.00 | 14024.0 |
| 297.00 | 2006.0  |
| 297.90 | 337.0   |
| 303.00 | 1414.0  |
| 303.90 | 360.0   |
| 305.20 | 387.0   |
| 310.00 | 460.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP55

|        |         |
|--------|---------|
| 313.90 | 653.0   |
| 315.00 | 1882.0  |
| 316.00 | 621.0   |
| 321.00 | 640.0   |
| 323.00 | 2912.0  |
| 323.90 | 323.0   |
| 325.00 | 311.0   |
| 326.90 | 844.0   |
| 333.00 | 761.0   |
| 334.00 | 2917.0  |
| 334.80 | 709.0   |
| 340.80 | 656.0   |
| 342.10 | 330.0   |
| 346.00 | 1660.0  |
| 352.00 | 958.0   |
| 353.10 | 1403.0  |
| 354.00 | 1139.0  |
| 354.80 | 356.0   |
| 365.00 | 8734.0  |
| 366.10 | 887.0   |
| 370.90 | 346.0   |
| 372.00 | 1741.0  |
| 373.00 | 512.0   |
| 382.80 | 582.0   |
| 401.90 | 508.0   |
| 403.00 | 1108.0  |
| 404.10 | 423.0   |
| 421.10 | 820.0   |
| 422.00 | 841.0   |
| 423.00 | 3400.0  |
| 423.90 | 627.0   |
| 441.00 | 14052.0 |
| 442.00 | 76840.0 |
| 443.00 | 14525.0 |
| 444.00 | 1807.0  |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP55