


Data Path : Z:\HPCHEM1\BNA_M\Data\BM012218\
 Data File : BM013684.D
 Acq On : 22 Jan 2018 09:01
 Operator : SJ/JU
 Sample : DFTPP79
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP79

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\Methods\SOM-EPA-BM011018.M
 Title : SVOA CALIBRATION
 Last Update : Mon Jan 29 04:03:39 2018



AutoFind: Scans 2460, 2461, 2462; Background Corrected with Scan 2455

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.1	105194	PASS
68	69	0.00	2	1.8	1966	PASS
69	198	0.00	100	42.2	107941	PASS
70	69	0.00	2	0.5	518	PASS
127	198	10	80	48.0	122930	PASS
197	198	0.00	2	0.6	1597	PASS
198	198	100	100	100.0	255957	PASS
199	198	5	9	6.6	17010	PASS
275	198	10	60	28.7	73373	PASS
365	198	1	100	4.1	10597	PASS
441	443	0.01	100	75.7	32346	PASS
442	198	50	100	84.7	216832	PASS
443	442	15	24	19.7	42704	PASS

Scan 2455 (17.427 min): BM013684.D
DFTPP79

m/z	Abundance
39.90	547.0
43.90	749.0
72.90	484.0
147.00	333.0
207.10	1544.0
208.10	621.0
281.00	641.0

Instrument :
BNA_M
ClientSampleId :
DFTPP79

m/z	Abundance
37.00	536.0
38.00	1485.0
39.00	9546.0
40.00	456.0
44.00	1161.0
48.90	676.0
50.00	32008.0
51.10	106120.0
52.10	5171.0
53.00	410.0
56.00	2466.0
57.10	6973.0
61.00	2280.0
62.00	1635.0
63.00	3631.0
64.00	957.0
65.00	1868.0
68.00	1982.0
69.00	103920.0
69.90	1007.0
73.10	1247.0
74.00	10641.0
75.00	18392.0
76.10	5788.0
77.10	103552.0
78.00	6670.0
79.00	8980.0
80.00	7173.0
81.00	9652.0
81.90	2253.0
83.00	1148.0
85.00	1561.0
86.00	3576.0
86.90	1234.0
87.90	520.0
88.90	560.0
91.00	3133.0
92.10	3342.0
93.00	17896.0
93.90	664.0
95.90	783.0
97.20	638.0
98.00	14950.0
99.00	8912.0
99.80	726.0
101.00	5203.0
102.90	1889.0
104.00	3135.0
105.10	2751.0
106.10	1188.0
107.00	35096.0
108.00	5969.0
109.00	796.0
110.00	59608.0
111.00	8723.0
111.90	1201.0
112.90	725.0
116.10	1970.0
117.00	41776.0
118.00	3638.0
120.00	549.0
121.00	521.0
122.00	2212.0
123.00	4037.0
124.00	1538.0
124.90	1419.0
126.10	440.0
127.00	109248.0
128.00	8263.0
129.00	48592.0
130.00	3832.0
131.10	1044.0
132.80	328.0
133.90	2225.0
135.00	4097.0
136.00	1422.0
136.90	2944.0
137.60	853.0
138.90	363.0
139.90	764.0
140.90	7465.0
142.00	1535.0
142.90	1598.0
144.20	402.0
145.10	513.0
145.90	1137.0
147.10	3295.0
148.00	8611.0
149.00	1811.0
151.00	484.0
152.90	1969.0

Instrument :
BNA_M
ClientSampleId :
DFTPP79

154.00	1635.0
155.00	3562.0
156.10	5757.0
156.90	642.0
157.30	634.0
158.20	1374.0
158.90	1054.0
160.10	1703.0
161.00	2954.0
162.20	985.0
163.10	382.0
165.00	2316.0
165.90	2186.0
167.10	12172.0
168.00	7175.0
169.10	1221.0
169.90	457.0
172.00	1127.0
173.10	1806.0
173.90	2063.0
175.00	3883.0
175.90	1316.0
177.00	2512.0
177.90	1251.0
179.00	10816.0
180.00	6301.0
181.00	3540.0
181.90	878.0
184.00	638.0
185.00	4284.0
186.00	32976.0
187.00	11144.0
188.00	806.0
188.90	2935.0
189.90	486.0
191.10	1710.0
192.10	2965.0
193.00	2672.0
194.10	588.0
194.80	668.0
196.00	6032.0
197.00	1799.0
197.90	213824.0
199.00	16134.0
200.00	1403.0
201.80	664.0
203.10	1382.0
204.00	9242.0
205.00	16054.0
206.10	60200.0
207.00	10311.0
208.00	3804.0
209.00	1557.0
210.20	1609.0
211.10	2881.0
212.00	355.0
215.00	1134.0
216.10	1647.0
216.90	17976.0
217.90	1734.0
219.00	329.0
220.00	518.0
221.00	8676.0
221.90	2786.0
223.00	2748.0
224.00	35496.0
225.00	8993.0
226.00	712.0
227.00	15789.0
227.90	2746.0
229.00	2702.0
230.10	524.0
231.00	1447.0
234.00	1410.0
235.00	702.0
236.00	703.0
236.90	1341.0
239.00	655.0
239.80	477.0
241.10	786.0
242.10	1340.0
243.00	1812.0
244.00	25640.0
245.10	3732.0
246.00	6272.0
247.00	2081.0
247.80	341.0
249.10	843.0
251.10	437.0
252.90	728.0
254.10	859.0
255.00	133568.0
256.00	20768.0
257.00	1974.0
258.00	8512.0

Instrument :
BNA_M
ClientSampleId :
DFTPP79

259.00	1583.0
265.00	3879.0
265.90	714.0
267.10	380.0
272.10	542.0
273.00	4798.0
274.00	9209.0
275.00	59752.0
276.00	7745.0
277.00	4729.0
277.90	1340.0
281.30	564.0
282.90	894.0
285.00	625.0
292.00	321.0
293.00	1418.0
293.80	407.0
294.20	313.0
296.00	19536.0
296.90	2834.0
301.90	384.0
303.00	1768.0
308.00	345.0
309.00	420.0
313.90	373.0
315.00	2142.0
316.00	957.0
321.10	1018.0
323.10	4973.0
324.10	526.0
327.10	1183.0
331.90	659.0
333.00	499.0
334.10	3250.0
335.10	772.0
335.80	351.0
341.00	368.0
345.90	1393.0
351.90	1123.0
353.10	1185.0
354.10	2144.0
355.20	717.0
365.00	6801.0
366.00	921.0
370.80	400.0
372.10	2348.0
373.10	1245.0
382.90	455.0
389.90	350.0
402.10	1329.0
403.00	1522.0
421.00	1499.0
422.00	1154.0
423.00	7509.0
424.10	2795.0
441.10	17736.0
442.10	133120.0
443.10	28896.0
444.00	2571.0

Instrument :
BNA_M
ClientSampleId :
DFTPP79

m/z	Abundance
37.00	598.0
38.10	1944.0
39.10	8665.0
40.10	714.0
40.90	399.0
44.00	1206.0
48.90	848.0
50.00	35464.0
51.10	123264.0
52.00	5244.0
55.00	654.0
56.00	2964.0
57.00	7435.0
57.80	374.0
61.00	1423.0
62.10	2018.0
63.10	5186.0
63.90	773.0
65.00	1501.0
68.00	2311.0
69.00	130232.0
70.20	496.0
73.00	1919.0
74.00	15057.0
75.00	20992.0
76.10	7155.0
77.10	139136.0
78.10	10543.0
79.00	10227.0
80.00	9660.0
81.00	10965.0
82.00	2726.0
82.90	2528.0
84.00	416.0
85.00	1774.0
86.00	3884.0
87.10	1590.0
87.90	562.0
91.10	3244.0
92.00	2275.0
93.00	22760.0
94.00	1167.0
95.00	364.0
95.90	1045.0
97.20	421.0
98.00	16984.0
99.00	11914.0
100.20	892.0
101.00	5417.0
102.00	302.0
103.00	2286.0
104.00	3209.0
105.00	3775.0
106.00	1561.0
107.00	46344.0
108.00	7475.0
109.00	1697.0
110.00	76168.0
111.00	9763.0
112.00	1997.0
113.00	713.0
115.10	739.0
116.00	2375.0
117.00	52552.0
117.90	4388.0
119.00	491.0
120.10	814.0
121.90	3294.0
123.00	5604.0
124.00	2294.0
125.00	2090.0
126.20	1010.0
127.00	149760.0
128.00	12857.0
129.00	63040.0
130.00	4637.0
130.90	1016.0
131.90	817.0
133.00	354.0
133.90	2935.0
135.00	4653.0
136.00	2090.0
137.00	3017.0
139.00	309.0
140.00	796.0
141.00	9239.0
142.00	2931.0
143.00	2087.0
143.80	317.0
145.00	439.0
145.90	1402.0

Instrument :
BNA_M
ClientSampleId :
DFTPP79

147.00	4484.0
148.00	11413.0
149.00	3078.0
149.90	757.0
151.10	767.0
151.70	508.0
152.20	335.0
153.00	3230.0
154.00	2482.0
155.00	5371.0
156.10	7160.0
157.00	1232.0
157.80	1738.0
159.10	748.0
160.00	2647.0
161.00	4129.0
162.00	1506.0
165.00	3473.0
166.00	2639.0
167.00	18968.0
168.00	10526.0
169.00	1525.0
169.90	883.0
170.70	834.0
172.00	2217.0
173.10	3095.0
174.00	3756.0
175.00	6280.0
175.90	2272.0
177.00	3680.0
177.90	888.0
178.90	12878.0
180.00	8586.0
181.00	3875.0
182.10	600.0
184.00	1357.0
185.10	6273.0
186.00	47792.0
187.10	12599.0
188.00	2152.0
189.00	3596.0
189.90	647.0
190.90	2210.0
192.00	4062.0
193.00	5356.0
194.00	1143.0
194.80	959.0
196.00	11569.0
197.00	1762.0
198.00	312704.0
198.90	19144.0
200.00	1477.0
200.90	527.0
201.60	951.0
203.00	3167.0
204.00	11811.0
205.00	23128.0
206.00	87352.0
207.00	13962.0
208.00	3582.0
209.00	2209.0
209.90	1303.0
211.10	3562.0
211.80	897.0
215.10	778.0
215.90	2017.0
217.00	27792.0
218.00	3103.0
219.80	566.0
221.10	11199.0
223.00	4215.0
224.00	49056.0
225.10	13239.0
226.00	2013.0
227.00	25256.0
228.10	4259.0
229.00	4779.0
230.10	555.0
231.00	2085.0
233.10	338.0
234.10	1476.0
235.00	2505.0
236.00	693.0
237.00	2258.0
237.90	894.0
239.00	1249.0
240.00	909.0
241.00	1011.0
242.00	2914.0
243.00	3030.0
244.10	37016.0
245.00	5366.0
246.00	8332.0
247.10	1514.0
248.10	490.0

Instrument :
BNA_M
ClientSampleId :
DFTPP79

249.00	1628.0
251.90	494.0
252.90	1228.0
254.10	1431.0
255.00	210816.0
256.00	29488.0
257.00	2969.0
258.00	13488.0
259.00	2792.0
259.90	576.0
261.00	319.0
263.80	524.0
265.00	5332.0
265.90	1009.0
271.00	901.0
271.90	500.0
273.00	5515.0
274.00	16311.0
275.00	85208.0
276.00	11583.0
276.90	9648.0
278.00	1459.0
278.90	513.0
281.10	491.0
281.80	309.0
282.20	339.0
283.10	1778.0
283.90	452.0
285.10	1378.0
289.00	546.0
290.30	403.0
292.20	427.0
293.00	2076.0
295.10	825.0
296.00	28728.0
297.00	4009.0
297.90	435.0
302.00	929.0
303.00	4360.0
304.00	1221.0
309.10	446.0
314.10	1263.0
315.00	3007.0
316.10	1784.0
316.90	443.0
321.10	1375.0
323.00	7288.0
324.00	1249.0
326.10	414.0
326.90	2064.0
327.90	878.0
332.90	1020.0
334.00	4945.0
335.00	1787.0
341.10	843.0
345.90	1233.0
351.90	1849.0
353.00	2556.0
354.10	2358.0
355.30	608.0
365.00	13384.0
366.00	1807.0
370.20	304.0
371.20	567.0
372.00	3390.0
373.00	1015.0
383.10	691.0
383.90	426.0
391.90	406.0
402.00	1916.0
403.20	1699.0
403.90	1352.0
405.00	341.0
421.00	2707.0
422.10	1903.0
423.00	12170.0
424.10	3835.0
425.00	1021.0
441.10	39792.0
442.00	248000.0
443.10	49792.0
444.00	3814.0

Instrument :
BNA_M
ClientSampleId :
DFTPP79

m/z	Abundance
37.10	498.0
37.90	729.0
39.10	6325.0
40.00	398.0
43.10	330.0
44.00	856.0
49.00	1014.0
50.10	26872.0
51.10	86200.0
52.00	4568.0
55.10	828.0
56.00	2634.0
57.00	4405.0
58.00	841.0
61.10	1215.0
62.00	1828.0
63.00	4046.0
64.10	416.0
65.10	1643.0
66.00	307.0
67.90	1606.0
69.00	89672.0
70.00	549.0
73.10	1587.0
74.10	10008.0
75.00	13974.0
76.10	4576.0
77.10	98840.0
78.00	5611.0
79.00	7409.0
80.10	6822.0
81.00	8852.0
82.00	1804.0
83.00	1800.0
85.00	2035.0
86.00	3528.0
87.00	1248.0
88.10	423.0
89.00	331.0
91.10	2261.0
92.10	2677.0
93.00	15086.0
94.00	1458.0
95.10	318.0
96.20	547.0
97.00	438.0
98.00	12502.0
99.00	7887.0
100.10	775.0
100.90	3845.0
103.10	1594.0
104.00	2409.0
104.90	3107.0
106.10	960.0
107.00	36424.0
108.00	4334.0
108.90	931.0
110.00	58440.0
111.00	9728.0
112.00	1178.0
113.00	374.0
116.00	1617.0
117.00	39968.0
118.00	2207.0
119.10	569.0
119.90	526.0
121.90	3207.0
123.10	3203.0
124.00	2767.0
125.00	1714.0
126.20	777.0
127.00	109784.0
128.10	8175.0
129.00	46664.0
130.10	4792.0
130.90	784.0
134.00	1356.0
135.00	3949.0
136.00	1335.0
137.10	2008.0
137.90	436.0
139.90	435.0
141.00	6786.0
142.00	1907.0
143.00	1606.0
143.90	305.0
144.70	479.0
145.90	1155.0
147.00	3622.0
148.00	10891.0
149.00	2370.0

Instrument :
BNA_M
ClientSampleId :
DFTPP79

149.90	579.0
150.90	661.0
151.90	454.0
153.00	1916.0
154.10	1932.0
155.00	3238.0
156.10	5524.0
157.20	1026.0
157.90	1164.0
158.90	815.0
159.90	2258.0
161.10	3080.0
161.90	865.0
162.90	330.0
165.00	2537.0
166.00	3053.0
167.00	13588.0
168.00	7595.0
169.20	1012.0
169.90	766.0
170.80	603.0
171.20	490.0
172.00	1472.0
173.00	1838.0
174.10	3409.0
175.00	4781.0
175.90	1187.0
176.90	2347.0
178.00	1720.0
179.00	10827.0
180.00	7010.0
181.00	2727.0
182.10	422.0
183.90	1578.0
185.00	4628.0
186.00	39832.0
187.00	10801.0
187.90	1455.0
189.00	3269.0
189.90	428.0
191.10	913.0
192.00	2845.0
193.00	4912.0
194.00	1312.0
195.00	370.0
195.90	6700.0
197.10	1230.0
198.00	241344.0
198.90	15752.0
200.10	1799.0
201.00	954.0
201.80	1061.0
202.30	542.0
203.00	2100.0
204.00	12698.0
205.00	16800.0
206.10	72112.0
207.00	11701.0
208.00	2929.0
208.90	1395.0
210.10	1189.0
211.00	3005.0
212.00	426.0
215.10	1831.0
216.00	1790.0
217.00	22640.0
218.00	2452.0
220.00	348.0
221.00	9597.0
221.90	3184.0
223.10	4048.0
224.00	43928.0
225.00	10603.0
226.10	1924.0
227.00	18920.0
228.00	3594.0
229.00	4128.0
229.90	482.0
231.00	1812.0
232.10	507.0
233.00	405.0
234.00	1593.0
234.90	1033.0
236.00	1310.0
236.90	2038.0
239.00	1382.0
240.10	598.0
241.00	963.0
242.00	2529.0
242.90	2989.0
244.10	33056.0
245.00	4144.0
246.00	6500.0
247.00	1388.0
249.00	1487.0

Instrument :
BNA_M
ClientSampleId :
DFTPP79

250.00	344.0
252.00	638.0
252.80	935.0
254.10	1241.0
255.00	179904.0
256.00	24376.0
257.10	3015.0
258.00	12510.0
258.90	1300.0
260.10	517.0
260.90	375.0
264.10	483.0
265.00	5930.0
265.90	877.0
271.90	548.0
273.00	4456.0
274.10	14142.0
275.00	75160.0
276.10	10727.0
277.00	6300.0
278.00	1187.0
279.00	354.0
280.90	589.0
282.90	917.0
284.00	417.0
284.90	1467.0
285.80	477.0
288.80	439.0
291.90	527.0
292.90	1335.0
294.00	599.0
295.00	581.0
296.00	29560.0
296.90	2942.0
297.80	321.0
301.10	637.0
302.10	351.0
303.10	2677.0
303.90	513.0
310.10	312.0
314.10	1237.0
315.00	3928.0
315.90	1234.0
321.00	984.0
321.90	621.0
323.00	6213.0
323.90	1188.0
327.00	2085.0
327.90	1196.0
329.00	428.0
332.10	764.0
333.20	667.0
334.10	4961.0
335.00	1110.0
339.00	345.0
341.00	709.0
346.10	1377.0
347.00	308.0
351.90	2939.0
353.00	1252.0
354.10	2892.0
354.90	434.0
365.00	11608.0
366.00	1653.0
371.00	520.0
372.10	4441.0
372.90	1080.0
383.00	978.0
401.10	572.0
402.00	1528.0
403.10	1945.0
404.20	747.0
421.10	1596.0
422.10	1726.0
423.10	11958.0
424.00	3501.0
425.10	536.0
441.10	39512.0
442.10	269376.0
443.10	49424.0
444.10	4544.0

Instrument :
BNA_M
ClientSampleId :
DFTPP79