

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012221\
 Data File : BM028510.D
 Acq On : 22 Jan 2021 09:12
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 22 10:47:51 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM012021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 22 10:45:04 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	130	0.00
2	1,4-Dioxane	40.000	33.339	16.7	119	0.00
3	Pyridine	40.000	36.304	9.2	126	0.00
4	n-Nitrosodimethylamine	40.000	33.419	16.5	118	0.00
5 S	2-Fluorophenol	80.000	70.410	12.0	124	0.00
6	Aniline	40.000	38.723	3.2	137	0.00
7 S	Phenol-d6	80.000	75.918	5.1	134	0.00
8	2-Chlorophenol	40.000	37.488	6.3	133	0.00
9	Benzaldehyde	40.000	35.784	10.5	131	0.00
10 C	Phenol	40.000	38.065	4.8	135	0.00
11	bis(2-Chloroethyl)ether	40.000	38.110	4.7	135	0.00
12	1,3-Dichlorobenzene	40.000	36.217	9.5	129	0.00
13 C	1,4-Dichlorobenzene	40.000	36.679	8.3	131	0.00
14	1,2-Dichlorobenzene	40.000	36.954	7.6	132	0.00
15	Benzyl Alcohol	40.000	39.551	1.1	140	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.544	8.6	131	0.00
17	2-Methylphenol	40.000	39.294	1.8	138	0.00
18	Hexachloroethane	40.000	37.230	6.9	131	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.230	-3.1	142	0.00
20	3+4-Methylphenols	40.000	40.697	-1.7	144	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	141	0.00
22	Acetophenone	40.000	35.638	10.9	142	0.00
23 S	Nitrobenzene-d5	80.000	69.802	12.7	138	0.00
24	Nitrobenzene	40.000	34.852	12.9	137	0.00
25	Isophorone	40.000	38.225	4.4	149	0.00
26 C	2-Nitrophenol	40.000	38.176	4.6	148	0.00
27	2,4-Dimethylphenol	40.000	36.590	8.5	142	0.00
28	bis(2-Chloroethoxy)methane	40.000	36.589	8.5	145	0.00
29 C	2,4-Dichlorophenol	40.000	37.125	7.2	146	0.00
30	1,2,4-Trichlorobenzene	40.000	34.720	13.2	140	0.00
31	Naphthalene	40.000	35.679	10.8	142	0.00
32	Benzoic acid	40.000	40.447	-1.1	160	0.00
33	4-Chloroaniline	40.000	37.369	6.6	147	0.00
34 C	Hexachlorobutadiene	40.000	34.416	14.0	138	0.00
35	Caprolactam	40.000	39.490	1.3	157	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.648	3.4	150	0.00
37	2-Methylnaphthalene	40.000	37.185	7.0	147	0.00
38	1-Methylnaphthalene	40.000	37.355	6.6	147	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	153	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	33.534	16.2	145	0.00
41 P	Hexachlorocyclopentadiene	40.000	33.482	16.3	143	0.00
42 S	2,4,6-Tribromophenol	80.000	75.868	5.2	159	0.00
43 C	2,4,6-Trichlorophenol	40.000	36.198	9.5	151	0.00
44	2,4,5-Trichlorophenol	40.000	35.994	10.0	152	0.00
45 S	2-Fluorobiphenyl	80.000	69.031	13.7	150	0.00
46	1,1'-Biphenyl	40.000	34.712	13.2	150	0.00
47	2-Chloronaphthalene	40.000	34.326	14.2	150	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012221\
 Data File : BM028510.D
 Acq On : 22 Jan 2021 09:12
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 22 10:47:51 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM012021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 22 10:45:04 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	36.002	10.0	151	0.00
49	Acenaphthylene	40.000	35.834	10.4	153	0.00
50	Dimethylphthalate	40.000	35.994	10.0	154	0.00
51	2,6-Dinitrotoluene	40.000	36.902	7.7	156	0.00
52 C	Acenaphthene	40.000	35.443	11.4	153	0.00
53	3-Nitroaniline	40.000	37.487	6.3	156	0.00
54 P	2,4-Dinitrophenol	40.000	36.766	8.1	158	0.00
55	Dibenzofuran	40.000	35.306	11.7	153	0.00
56 P	4-Nitrophenol	40.000	36.998	7.5	151	0.00
57	2,4-Dinitrotoluene	40.000	38.187	4.5	156	0.00
58	Fluorene	40.000	35.638	10.9	152	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.170	7.1	156	0.00
60	Diethylphthalate	40.000	36.161	9.6	154	0.00
61	4-Chlorophenyl-phenylether	40.000	35.014	12.5	151	0.00
62	4-Nitroaniline	40.000	37.188	7.0	152	0.00
63	Azobenzene	40.000	35.954	10.1	151	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	151	0.00
65	4,6-Dinitro-2-methylphenol	40.000	37.625	5.9	161	0.00
66 c	n-Nitrosodiphenylamine	40.000	36.465	8.8	155	0.00
67	4-Bromophenyl-phenylether	40.000	36.253	9.4	152	0.00
68	Hexachlorobenzene	40.000	36.096	9.8	155	0.00
69	Atrazine	40.000	37.006	7.5	152	0.00
70 C	Pentachlorophenol	40.000	38.002	5.0	155	0.00
71	Phenanthrene	40.000	35.338	11.7	151	0.00
72	Anthracene	40.000	35.943	10.1	152	0.00
73	Carbazole	40.000	35.143	12.1	148	0.00
74	Di-n-butylphthalate	40.000	36.276	9.3	149	0.00
75 C	Fluoranthene	40.000	33.389	16.5	141	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	118	0.00
77	Benzydine	40.000	47.470	-18.7	134	0.00
78	Pyrene	40.000	42.386	-6.0	139	0.00
79 S	Terphenyl-d14	80.000	81.938	-2.4	134	0.00
80	Butylbenzylphthalate	40.000	39.805	0.5	128	0.00
81	Benzo(a)anthracene	40.000	36.061	9.8	120	0.00
82	3,3'-Dichlorobenzidine	40.000	36.195	9.5	113	0.00
83	Chrysene	40.000	35.710	10.7	117	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	36.572	8.6	116	0.00
85 c	Di-n-octyl phthalate	40.000	32.695	18.3	105	0.00
86 I	Perylene-d12	20.000	20.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	35.569	11.1	98	0.00
88	Benzo(b)fluoranthene	40.000	38.891	2.8	104	0.00
89	Benzo(k)fluoranthene	40.000	37.463	6.3	103	0.00
90 C	Benzo(a)pyrene	40.000	37.746	5.6	103	0.00
91	Dibenzo(a,h)anthracene	40.000	35.515	11.2	97	0.00
92	Benzo(g,h,i)perylene	40.000	35.496	11.3	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012221\
Data File : BM028510.D
Acq On : 22 Jan 2021 09:12
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Jan 22 10:47:51 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM012021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 22 10:45:04 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0