

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012224\
 Data File : BM043882.D
 Acq On : 22 Jan 2024 10:39
 Operator : MA/JU
 Sample : PB158502BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK502

Quant Time: Jan 23 00:43:25 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM011924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 19 22:45:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.976	152	3053	0.400	ng/ul	# 0.05
4) Naphthalene-d8	10.897	136	9572	0.400	ng/ul	# 0.13
9) Acenaphthene-d10	14.627	164	5478m	0.400	ng/ul	0.04
13) Phenanthrene-d10	17.420	188	9432m	0.400	ng/ul	0.07
17) Chrysene-d12	21.560	240	6521	0.400	ng/ul	# 0.03
23) Perylene-d12	23.963	264	8107	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.327	96	4010	0.964	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.492	152	3585	0.295	ng/ul	0.13
18) Fluoranthene-d10	19.411	212	5979	0.278	ng/ul	0.04

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012224\
Data File : BM043882.D
Acq On : 22 Jan 2024 10:39
Operator : MA/JU
Sample : PB158502BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK502

Quant Time: Jan 23 00:43:25 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM011924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 19 22:45:24 2024
Response via : Initial Calibration

