

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012225\
 Data File : BM049373.D
 Acq On : 21 Jan 2025 20:20
 Operator : RC/JU
 Sample : Q1139-22
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DCZJ9

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/22/2025
 Supervised By :mohammad ahmed 01/24/2025

Quant Time: Jan 22 02:21:10 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM011325.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 13 16:56:06 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.522	152	177171	20.000	ng/ul	0.00
20) Naphthalene-d8	10.281	136	625046	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.163	164	364321	20.000	ng/ul	0.00
64) Phenanthrene-d10	16.910	188	697646	20.000	ng/ul	0.00
79) Chrysene-d12	21.139	240	596924	20.000	ng/ul	0.00
88) Perylene-d12	23.915	264	656492	20.000	ng/ul	-0.02

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.122	96	23986	5.601	ng/uL	0.00
4) Pyridine-d5	3.516	84	86343	6.334	ng/ul	0.00
7) Phenol-d5	6.716	99	97373	6.547	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.869	67	322520	32.567	ng/ul	0.00
11) 2-Chlorophenol-d4	7.063	132	312976	27.509	ng/ul	0.00
15) 4-Methylphenol-d8	8.234	113	180454	15.647	ng/ul	0.00
21) Nitrobenzene-d5	8.663	128	152272	31.947	ng/ul	0.00
24) 2-Nitrophenol-d4	9.381	143	153778	28.860	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.910	165	302958	27.864	ng/ul	0.00
31) 4-Chloroaniline-d4	10.433	131	13357m	0.934	ng/ul	0.00
46) Dimethylphthalate-d6	13.580	166	911313	33.608	ng/ul	0.00
49) Acenaphthylene-d8	13.851	160	1025182	33.118	ng/ul	0.00
54) 4-Nitrophenol-d4	14.404	143	12074m	2.689	ng/ul	0.01
60) Fluorene-d10	15.157	176	844598	36.056	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.286	200	100785	22.668	ng/ul	0.00
73) Anthracene-d10	17.010	188	1349809	39.139	ng/ul	0.00
81) Pyrene-d10	19.315	212	1536762	42.302	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.733	264	1357643	38.524	ng/ul	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012225\
Data File : BM049373.D
Acq On : 21 Jan 2025 20:20
Operator : RC/JU
Sample : Q1139-22
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DCZJ9

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/22/2025
Supervised By :mohammad ahmed 01/24/2025

Quant Time: Jan 22 02:21:10 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM011325.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jan 13 16:56:06 2025
Response via : Initial Calibration

