

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM012317\
 Data File : BM008840.D
 Acq On : 24 Jan 2017 15:13
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02045

Quant Time: Jan 24 15:54:17 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM012317.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 24 04:28:48 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	66	0.00
2	1,4-Dioxane	0.490	0.643	-31.2#	82	0.00
3 S	1,4-Dioxane-d8	0.405	0.501	-23.7#	84	0.00
4	Benzaldehyde	1.394	1.863	-33.6#	73	0.00
5 S	Phenol-d5	1.694	1.884	-11.2	72	0.00
6	Phenol	1.847	2.017	-9.2	72	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.347	1.540	-14.3	72	0.00
8	Bis(2-Chloroethyl)ether	1.384	1.515	-9.5	68	0.00
9 S	2-Chlorophenol-d4	1.104	1.294	-17.2	76	0.00
10	2-Chlorophenol	1.140	1.279	-12.2	71	0.00
11	2-Methylphenol	1.324	1.427	-7.8	71	0.00
12	2,2'-oxybis(1-Chloropropane	2.600	2.991	-15.0	71	0.00
13 S	4-Methylphenol-d8	1.326	1.377	-3.8	66	0.00
14	Acetophenone	2.431	2.710	-11.5	73	0.00
15 P	N-Nitroso-di-n-propylamine	1.584	1.803	-13.8	72	0.00
16	4-Methylphenol	1.427	1.598	-12.0	72	0.00
17	Hexachloroethane	0.703	0.811	-15.4	76	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	68	0.00
19 S	Nitrobenzene-d5	0.127	0.137	-7.9	70	0.00
20	Nitrobenzene	0.561	0.663	-18.2	74	0.00
21	Isophorone	0.905	1.002	-10.7	69	0.00
22 S	2-Nitrophenol-d4	0.155	0.159	-2.6	66	0.00
23 C	2-Nitrophenol	0.160	0.176	-10.0	70	0.00
24	2,4-Dimethylphenol	0.458	0.510	-11.4	69	0.00
25	Bis(2-Chloroethoxy)methane	0.451	0.500	-10.9	69	0.00
26 S	2,4-Dichlorophenol-d3	0.324	0.356	-9.9	68	0.00
27 C	2,4-Dichlorophenol	0.318	0.340	-6.9	66	0.00
28	Naphthalene	0.923	1.023	-10.8	71	0.00
29 S	4-Chloroaniline-d4	0.353	0.394	-11.6	68	0.00
30	4-Chloroaniline	0.361	0.412	-14.1	71	0.00
31 C	Hexachlorobutadiene	0.308	0.339	-10.1	68	0.00
32	Caprolactam	0.098	0.096	2.0	61	0.00
33 C	4-Chloro-3-methylphenol	0.413	0.460	-11.4	72	0.00
34	2-Methylnaphthalene	0.750	0.811	-8.1	68	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	65	0.00
36	1,2,4,5-Tetrachlorobenzene	0.614	0.662	-7.8	69	0.00
37	Hexachlorocyclopentadiene	0.475	0.499	-5.1	70	0.00
38 C	2,4,6-Trichlorophenol	0.371	0.392	-5.7	67	0.00
39	2,4,5-Trichlorophenol	0.398	0.412	-3.5	65	0.00
40	1,1'-Biphenyl	1.236	1.334	-7.9	68	0.00
41	2-Chloronaphthalene	0.972	1.058	-8.8	70	0.00
42	2-Nitroaniline	0.483	0.539	-11.6	69	0.00
43 S	Dimethylphthalate-d6	1.310	1.399	-6.8	66	0.00
44	Dimethylphthalate	1.310	1.433	-9.4	67	0.00

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM012317\
 Data File : BM008840.D
 Acq On : 24 Jan 2017 15:13
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02045

Quant Time: Jan 24 15:54:17 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM012317.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 24 04:28:48 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.251	0.267	-6.4	65	0.00
46 S	Acenaphthylene-d8	1.550	1.685	-8.7	68	0.00
47	Acenaphthylene	1.494	1.612	-7.9	69	0.00
48	3-Nitroaniline	0.237	0.265	-11.8	64	0.00
49 C	Acenaphthene	1.051	1.179	-12.2	70	0.00
50	2,4-Dinitrophenol	0.180	0.176	2.2	65	0.00
51 S	4-Nitrophenol-d4	0.219	0.227	-3.7	64	0.00
52	4-Nitrophenol	0.452	0.496	-9.7	68	0.00
53	Dibenzofuran	1.616	1.797	-11.2	68	0.00
54	2,4-Dinitrotoluene	0.385	0.408	-6.0	65	0.00
55	2,3,4,6-Tetrachlorophenol	0.400	0.430	-7.5	65	0.00
56	Diethylphthalate	1.425	1.578	-10.7	68	0.00
57 S	Fluorene-d10	1.245	1.376	-10.5	68	0.00
58	Fluorene	1.350	1.465	-8.5	66	0.00
59	4-Chlorophenyl-phenylether	0.782	0.871	-11.4	69	0.00
60	4-Nitroaniline	0.265	0.295	-11.3	67	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	65	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.119	0.107	10.1	62	0.00
63	4,6-Dinitro-2-methylphenol	0.123	0.116	5.7	63	0.00
64	N-Nitrosodiphenylamine	0.506	0.554	-9.5	70	0.00
65	4-Bromophenyl-phenylether	0.217	0.241	-11.1	69	0.00
66	Hexachlorobenzene	0.236	0.250	-5.9	67	0.00
67	Atrazine	0.232	0.242	-4.3	65	0.00
68 C	Pentachlorophenol	0.155	0.145	6.5	61	0.00
69	Phenanthrene	0.985	1.073	-8.9	68	0.00
70 S	Anthracene-d10	0.875	0.958	-9.5	68	0.00
71	Anthracene	1.008	1.101	-9.2	67	0.00
72	Carbazole	0.884	0.942	-6.6	67	0.00
73	Di-n-butylphthalate	1.063	1.155	-8.7	67	0.00
74 C	Fluoranthene	1.280	1.367	-6.8	67	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	68	0.00
76 S	Pyrene-d10	0.877	0.897	-2.3	66	0.00
77	Pyrene	1.073	1.117	-4.1	69	0.00
78	Butylbenzylphthalate	0.400	0.414	-3.5	66	0.00
79	3,3'-Dichlorobenzidine	0.374	0.399	-6.7	65	0.00
80	Benzo(a)anthracene	1.075	1.160	-7.9	70	0.00
81	Bis(2-ethylhexyl)phthalate	0.555	0.552	0.5	64	0.00
82	Chrysene	1.000	1.094	-9.4	70	0.00
83 I	Perylene-d12	1.000	1.000	0.0	74	0.00
84	Di-n-octyl phthalate	1.161	1.129	2.8	72	0.00
85	Benzo(b)fluoranthene	1.162	1.217	-4.7	71	0.00
86	Benzo(k)fluoranthene	1.104	1.151	-4.3	72	0.00
87 S	Benzo(a)pyrene-d12	0.847	0.914	-7.9	75	0.00

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM012317\
Data File : BM008840.D
Acq On : 24 Jan 2017 15:13
Operator : UM/SJ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02045

Quant Time: Jan 24 15:54:17 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM012317.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jan 24 04:28:48 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.066	1.172	-9.9	77	0.00
89	Indeno(1,2,3-cd)pyrene	1.070	1.199	-12.1	81	0.00
90	Dibenzo(a,h)anthracene	0.926	1.045	-12.9	82	0.00
91	Benzo(g,h,i)perylene	0.870	0.978	-12.4	80	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0