

Data Path : Z:\HPCHEM1\BNA M\DATA\BM012417\
 Data File : BM008856.D
 Acq On : 25 Jan 2017 01:53
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02047

Quant Time: Jan 25 04:13:06 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM012317.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 25 03:34:06 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
2	1,4-Dioxane	0.490	0.533	-8.8	86	0.00
3 S	1,4-Dioxane-d8	0.405	0.444	-9.6	94	0.00
4	Benzaldehyde	1.394	1.782	-27.8#	88	0.00
5 S	Phenol-d5	1.694	1.777	-4.9	85	0.00
6	Phenol	1.847	1.915	-3.7	86	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.347	1.510	-12.1	89	0.00
8	Bis(2-Chloroethyl)ether	1.384	1.440	-4.0	82	0.00
9 S	2-Chlorophenol-d4	1.104	1.159	-5.0	86	0.00
10	2-Chlorophenol	1.140	1.259	-10.4	88	0.00
11	2-Methylphenol	1.324	1.380	-4.2	86	0.00
12	2,2'-oxybis(1-Chloropropane	2.600	2.742	-5.5	82	0.00
13 S	4-Methylphenol-d8	1.326	1.347	-1.6	81	0.00
14	Acetophenone	2.431	2.544	-4.6	86	0.00
15 P	N-Nitroso-di-n-propylamine	1.584	1.638	-3.4	83	0.00
16	4-Methylphenol	1.427	1.506	-5.5	85	0.00
17	Hexachloroethane	0.703	0.748	-6.4	88	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	85	0.00
19 S	Nitrobenzene-d5	0.127	0.139	-9.4	88	0.00
20	Nitrobenzene	0.561	0.616	-9.8	86	0.00
21	Isophorone	0.905	0.954	-5.4	82	0.00
22 S	2-Nitrophenol-d4	0.155	0.157	-1.3	81	0.00
23 C	2-Nitrophenol	0.160	0.168	-5.0	83	0.00
24	2,4-Dimethylphenol	0.458	0.496	-8.3	83	0.00
25	Bis(2-Chloroethoxy)methane	0.451	0.491	-8.9	84	0.00
26 S	2,4-Dichlorophenol-d3	0.324	0.347	-7.1	82	0.00
27 C	2,4-Dichlorophenol	0.318	0.348	-9.4	84	0.00
28	Naphthalene	0.923	0.980	-6.2	84	0.00
29 S	4-Chloroaniline-d4	0.353	0.368	-4.2	79	0.00
30	4-Chloroaniline	0.361	0.408	-13.0	87	0.00
31 C	Hexachlorobutadiene	0.308	0.344	-11.7	86	0.00
32	Caprolactam	0.098	0.098	0.0	77	0.00
33 C	4-Chloro-3-methylphenol	0.413	0.432	-4.6	84	0.00
34	2-Methylnaphthalene	0.750	0.790	-5.3	82	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	76	0.00
36	1,2,4,5-Tetrachlorobenzene	0.614	0.656	-6.8	80	0.00
37	Hexachlorocyclopentadiene	0.475	0.491	-3.4	81	0.00
38 C	2,4,6-Trichlorophenol	0.371	0.402	-8.4	80	0.00
39	2,4,5-Trichlorophenol	0.398	0.425	-6.8	79	0.00
40	1,1'-Biphenyl	1.236	1.371	-10.9	82	0.00
41	2-Chloronaphthalene	0.972	1.062	-9.3	83	0.00
42	2-Nitroaniline	0.483	0.542	-12.2	81	0.00
43 S	Dimethylphthalate-d6	1.310	1.431	-9.2	80	0.00
44	Dimethylphthalate	1.310	1.374	-4.9	76	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM012417\
 Data File : BM008856.D
 Acq On : 25 Jan 2017 01:53
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02047

Quant Time: Jan 25 04:13:06 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM012317.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 25 03:34:06 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.251	0.260	-3.6	75	0.00
46 S	Acenaphthylene-d8	1.550	1.704	-9.9	81	0.00
47	Acenaphthylene	1.494	1.635	-9.4	82	0.00
48	3-Nitroaniline	0.237	0.262	-10.5	74	0.00
49 C	Acenaphthene	1.051	1.139	-8.4	80	0.00
50	2,4-Dinitrophenol	0.180	0.147	18.3	64	0.00
51 S	4-Nitrophenol-d4	0.219	0.219	0.0	72	0.00
52	4-Nitrophenol	0.452	0.458	-1.3	74	0.00
53	Dibenzofuran	1.616	1.735	-7.4	77	0.00
54	2,4-Dinitrotoluene	0.385	0.412	-7.0	76	0.00
55	2,3,4,6-Tetrachlorophenol	0.400	0.428	-7.0	76	0.00
56	Diethylphthalate	1.425	1.525	-7.0	78	0.00
57 S	Fluorene-d10	1.245	1.336	-7.3	77	0.00
58	Fluorene	1.350	1.446	-7.1	77	0.00
59	4-Chlorophenyl-phenylether	0.782	0.853	-9.1	79	0.00
60	4-Nitroaniline	0.265	0.291	-9.8	77	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	74	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.119	0.111	6.7	72	0.00
63	4,6-Dinitro-2-methylphenol	0.123	0.122	0.8	74	0.00
64	N-Nitrosodiphenylamine	0.506	0.551	-8.9	78	0.00
65	4-Bromophenyl-phenylether	0.217	0.243	-12.0	78	0.00
66	Hexachlorobenzene	0.236	0.255	-8.1	77	0.00
67	Atrazine	0.232	0.239	-3.0	72	0.00
68 C	Pentachlorophenol	0.155	0.158	-1.9	75	0.00
69	Phenanthrene	0.985	1.058	-7.4	76	0.00
70 S	Anthracene-d10	0.875	0.949	-8.5	76	0.00
71	Anthracene	1.008	1.098	-8.9	75	0.00
72	Carbazole	0.884	0.944	-6.8	76	0.00
73	Di-n-butylphthalate	1.063	1.115	-4.9	73	0.00
74 C	Fluoranthene	1.280	1.321	-3.2	73	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	71	0.00
76 S	Pyrene-d10	0.877	0.930	-6.0	72	0.00
77	Pyrene	1.073	1.171	-9.1	76	0.00
78	Butylbenzylphthalate	0.400	0.422	-5.5	71	0.00
79	3,3'-Dichlorobenzidine	0.374	0.407	-8.8	70	0.00
80	Benzo(a)anthracene	1.075	1.159	-7.8	73	0.00
81	Bis(2-ethylhexyl)phthalate	0.555	0.582	-4.9	71	0.00
82	Chrysene	1.000	1.061	-6.1	72	0.00
83 I	Perylene-d12	1.000	1.000	0.0	75	0.00
84	Di-n-octyl phthalate	1.161	1.134	2.3	73	0.00
85	Benzo(b)fluoranthene	1.162	1.289	-10.9	76	0.00
86	Benzo(k)fluoranthene	1.104	1.178	-6.7	75	0.00
87 S	Benzo(a)pyrene-d12	0.847	0.929	-9.7	78	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM012417\
Data File : BM008856.D
Acq On : 25 Jan 2017 01:53
Operator : UM/SJ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02047

Quant Time: Jan 25 04:13:06 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM012317.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jan 25 03:34:06 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.066	1.171	-9.8	78	0.00
89	Indeno(1,2,3-cd)pyrene	1.070	1.240	-15.9	85	0.00
90	Dibenzo(a,h)anthracene	0.926	1.044	-12.7	83	0.00
91	Benzo(a,h,i)perylene	0.870	1.026	-17.9	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0