


Data Path : Z:\HPCHEM1\BNA_M\Data\BM012418\
 Data File : BM013766.D
 Acq On : 24 Jan 2018 14:02
 Operator : SJ/JU
 Sample : DFTPP81
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP81

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM011018.M
 Title : SVOA CALIBRATION
 Last Update : Mon Jan 29 03:14:13 2018



AutoFind: Scans 2458, 2459, 2460; Background Corrected with Scan 2453

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.8	80109	PASS
68	69	0.00	2	1.7	1376	PASS
69	198	0.00	100	41.0	80595	PASS
70	69	0.00	2	0.0	16	PASS
127	198	10	80	44.9	88309	PASS
197	198	0.00	2	0.5	1043	PASS
198	198	100	100	100.0	196565	PASS
199	198	5	9	6.0	11877	PASS
275	198	10	60	29.4	57725	PASS
365	198	1	100	4.3	8381	PASS
441	443	0.01	100	76.9	24897	PASS
442	198	50	100	81.9	161082	PASS
443	442	15	24	20.1	32389	PASS

m/z	Abundance
39.00	652.0
40.10	502.0
41.00	1271.0
42.00	659.0
43.20	1327.0
44.00	1457.0
44.80	799.0
55.10	870.0
56.00	305.0
57.00	762.0
67.00	352.0
68.10	343.0
69.20	338.0
70.20	589.0
71.20	307.0
73.10	1882.0
74.80	417.0
81.10	437.0
82.00	514.0
82.90	581.0
91.10	594.0
95.00	315.0
98.00	416.0
115.10	400.0
129.10	306.0
135.00	548.0
146.80	310.0
164.90	352.0
192.90	431.0
193.80	539.0
207.10	1197.0
207.90	368.0
209.00	399.0
281.10	446.0
283.00	307.0

Instrument :
BNA_M
ClientSampleId :
DFTPP81

m/z	Abundance
37.00	620.0
37.90	1422.0
39.00	7172.0
40.10	855.0
41.00	871.0
42.00	346.0
43.00	1407.0
43.90	1647.0
45.10	392.0
49.10	990.0
50.10	20560.0
51.10	78248.0
52.10	3284.0
53.10	441.0
55.10	1519.0
56.10	1791.0
57.00	5330.0
58.10	475.0
60.00	478.0
60.90	945.0
62.00	1479.0
63.10	2720.0
63.90	400.0
65.00	1781.0
67.00	406.0
68.10	1850.0
69.00	78400.0
70.10	862.0
71.20	626.0
73.00	2858.0
74.00	8378.0
75.00	11657.0
76.10	3555.0
77.00	79800.0
78.10	4683.0
79.00	7983.0
80.00	4451.0
81.00	5912.0
82.00	1341.0
83.00	1956.0
84.10	619.0
85.10	1554.0
85.90	2075.0
87.00	1071.0
91.00	2159.0
92.00	2214.0
93.00	11754.0
94.00	447.0
95.00	558.0
95.90	427.0
97.00	655.0
98.00	10016.0
99.00	7150.0
99.80	575.0
101.00	3586.0
102.70	498.0
103.10	768.0
104.00	2930.0
105.10	2304.0
107.00	29840.0
108.00	4091.0
109.00	756.0
110.00	41448.0
111.00	5732.0
112.10	619.0
115.10	329.0
115.80	1705.0
116.20	1549.0
117.00	30176.0
118.00	2317.0
119.10	354.0
120.00	696.0
122.10	2145.0
122.90	3698.0
124.00	1169.0
125.00	1239.0
126.20	380.0
127.00	79920.0
128.00	6058.0
129.00	36504.0
130.00	3034.0
131.00	752.0
133.00	633.0
134.00	1277.0
135.00	3857.0
135.80	1037.0
136.20	849.0
136.90	1697.0
140.00	656.0
141.00	5165.0
141.90	1517.0

Instrument :
BNA_M
ClientSampleId :
DFTPP81

143.00	1055.0
143.90	644.0
146.10	766.0
147.10	3561.0
148.00	7189.0
149.10	1224.0
150.20	623.0
151.10	962.0
152.20	482.0
152.90	2047.0
153.90	873.0
155.00	2612.0
155.90	3424.0
156.90	511.0
158.00	904.0
159.00	897.0
160.00	2047.0
160.80	1966.0
162.10	489.0
164.90	1750.0
166.00	1240.0
167.00	10285.0
168.00	4679.0
168.90	760.0
169.20	739.0
169.90	423.0
171.00	408.0
171.90	1880.0
172.90	1104.0
174.10	2057.0
175.10	2900.0
175.90	1139.0
177.00	2863.0
178.10	582.0
178.90	6422.0
179.90	5281.0
181.00	3058.0
181.70	580.0
182.20	301.0
182.40	305.0
184.10	513.0
185.00	3135.0
186.00	23200.0
187.10	7832.0
188.20	410.0
189.00	1021.0
191.00	1056.0
191.90	1643.0
192.90	2498.0
194.00	968.0
194.90	690.0
196.00	5283.0
197.00	1077.0
197.90	165824.0
199.00	11683.0
200.10	783.0
201.30	575.0
202.10	317.0
202.90	1814.0
204.00	6296.0
205.00	10325.0
206.00	47136.0
207.00	8101.0
208.10	3037.0
209.00	1052.0
210.10	616.0
210.90	2519.0
215.10	960.0
216.00	1100.0
217.00	12964.0
218.00	1210.0
219.90	380.0
221.00	5662.0
222.00	1865.0
223.00	2442.0
224.00	27584.0
225.00	8484.0
226.10	1296.0
227.00	11601.0
228.00	2297.0
229.00	1819.0
231.00	1540.0
233.20	433.0
234.10	847.0
234.90	1704.0
236.10	355.0
236.90	1357.0
239.00	485.0
239.90	496.0
241.00	383.0
242.00	1879.0
242.90	1271.0
244.00	20264.0
245.00	3128.0
245.90	3488.0

Instrument :
BNA_M
ClientSampleId :
DFTPP81

247.00	978.0
249.00	673.0
251.00	358.0
253.00	975.0
254.00	654.0
255.00	98992.0
256.00	16520.0
256.90	1334.0
258.00	6970.0
258.90	895.0
265.00	2328.0
271.00	335.0
272.00	342.0
273.00	2928.0
274.00	7891.0
275.00	42504.0
276.00	6258.0
277.00	4652.0
277.90	990.0
281.10	544.0
282.00	387.0
283.10	440.0
284.00	455.0
284.90	955.0
289.00	390.0
290.10	309.0
293.00	821.0
294.20	399.0
295.10	741.0
296.00	16009.0
297.00	2250.0
297.90	479.0
301.00	641.0
301.90	470.0
303.10	1135.0
304.10	447.0
313.90	454.0
314.90	1351.0
315.80	866.0
320.90	316.0
323.00	3684.0
324.00	776.0
327.00	554.0
328.00	359.0
331.80	810.0
332.20	404.0
333.10	375.0
334.00	3133.0
335.20	330.0
345.70	441.0
346.00	520.0
351.90	1461.0
353.90	916.0
355.20	675.0
365.00	5260.0
366.00	835.0
372.20	1972.0
382.80	411.0
402.00	398.0
403.00	1719.0
403.80	395.0
421.00	805.0
422.00	1132.0
423.00	5068.0
424.00	1396.0
441.10	14668.0
442.10	99184.0
443.10	22280.0
444.10	1748.0

Instrument :
BNA_M
ClientSampleId :
DFTPP81

m/z	Abundance
38.00	2037.0
39.00	7001.0
40.00	997.0
41.00	1134.0
42.10	737.0
43.10	1282.0
43.90	1021.0
45.00	604.0
49.20	1030.0
50.10	25920.0
51.00	98568.0
52.10	4457.0
54.00	348.0
55.10	1295.0
56.00	3370.0
57.10	5945.0
58.00	312.0
59.00	326.0
60.10	456.0
60.90	959.0
62.00	1236.0
63.10	5389.0
65.20	1427.0
67.10	761.0
68.10	1731.0
69.00	97616.0
69.90	511.0
71.10	596.0
73.10	1710.0
74.00	11786.0
75.00	17128.0
76.10	6963.0
77.10	102000.0
78.00	7985.0
79.00	9196.0
80.10	5953.0
81.00	8905.0
81.90	2594.0
82.90	2424.0
83.80	655.0
85.00	1914.0
86.00	3163.0
87.00	2005.0
88.00	501.0
91.00	2219.0
92.00	2036.0
93.10	14703.0
94.00	951.0
94.90	459.0
96.10	699.0
97.10	659.0
98.00	14143.0
98.90	10095.0
99.90	767.0
101.00	3679.0
102.00	310.0
103.00	2197.0
104.00	2915.0
104.90	3464.0
106.10	1293.0
107.00	35496.0
108.00	6563.0
109.10	1770.0
110.00	54648.0
111.00	9899.0
112.00	1502.0
112.80	733.0
115.10	374.0
116.00	1748.0
117.00	36160.0
117.90	3317.0
121.90	2925.0
123.00	4046.0
124.00	2153.0
125.00	1374.0
125.90	788.0
127.00	108360.0
128.00	8219.0
129.00	48144.0
130.00	4634.0
131.00	1340.0
131.90	670.0
134.00	1709.0
134.90	4706.0
136.00	2003.0
136.90	2353.0
138.00	556.0
139.90	568.0
141.00	7440.0
141.90	2164.0
142.90	1459.0

Instrument :
BNA_M
ClientSampleId :
DFTPP81

144.00	477.0
144.90	435.0
145.70	580.0
146.10	761.0
147.00	3861.0
148.00	9534.0
149.00	2088.0
150.10	570.0
151.20	1029.0
152.00	566.0
153.00	2770.0
153.90	1191.0
155.10	3557.0
156.00	4795.0
157.00	855.0
157.90	1335.0
158.90	1092.0
160.00	1268.0
161.00	4198.0
161.90	877.0
162.20	1027.0
164.20	341.0
165.00	1843.0
166.00	2714.0
167.10	12816.0
168.00	7298.0
169.00	1013.0
169.90	1004.0
170.90	1076.0
172.10	1633.0
173.00	1799.0
174.10	3606.0
175.00	4875.0
176.00	1556.0
176.90	1853.0
178.10	833.0
179.00	11132.0
180.00	5651.0
181.00	4140.0
182.00	696.0
183.00	357.0
184.10	1400.0
185.00	5163.0
186.00	34176.0
187.00	11501.0
188.10	1363.0
188.90	3032.0
189.80	581.0
191.00	714.0
192.00	4624.0
193.00	3446.0
194.00	975.0
195.10	410.0
195.90	4652.0
196.10	4563.0
196.90	1354.0
198.00	232000.0
198.90	14068.0
200.10	1312.0
200.70	366.0
201.30	381.0
201.80	388.0
203.00	2119.0
204.10	8771.0
205.10	18432.0
206.00	69720.0
207.00	9559.0
208.10	3145.0
208.90	945.0
209.90	1127.0
210.90	2481.0
213.30	300.0
214.90	1190.0
216.00	2000.0
217.00	19792.0
218.00	2011.0
220.00	317.0
221.00	10833.0
222.00	1960.0
222.90	3523.0
224.00	38448.0
225.00	9459.0
225.90	1204.0
227.00	17600.0
228.10	2389.0
229.00	3017.0
230.10	508.0
231.00	1410.0
231.80	325.0
233.90	1742.0
234.80	1660.0
235.90	703.0
237.00	1433.0
239.00	753.0
240.00	321.0

Instrument :
BNA_M
ClientSampleId :
DFTPP81

240.90	911.0
242.10	2083.0
243.20	2168.0
244.00	27936.0
245.00	5012.0
246.00	7269.0
246.90	1530.0
247.80	308.0
249.00	1399.0
250.90	562.0
253.00	877.0
253.90	783.0
255.00	158848.0
256.00	23952.0
257.00	1715.0
258.00	10430.0
259.00	2058.0
264.00	463.0
265.00	4685.0
265.90	654.0
270.80	343.0
271.90	628.0
273.10	4792.0
274.00	11718.0
275.00	71424.0
276.10	8372.0
277.00	6532.0
278.00	1670.0
280.90	633.0
282.00	663.0
282.90	747.0
284.10	415.0
285.10	1070.0
288.90	371.0
293.10	1722.0
294.10	967.0
296.00	20032.0
296.90	1913.0
300.90	627.0
302.10	549.0
303.00	2969.0
303.80	672.0
309.80	410.0
313.80	413.0
314.20	357.0
315.00	3027.0
316.00	1812.0
320.90	664.0
323.10	6213.0
324.00	599.0
325.20	325.0
326.90	1348.0
328.00	522.0
332.00	660.0
332.90	504.0
334.00	5130.0
335.00	1682.0
341.10	1099.0
341.90	368.0
346.00	1647.0
346.90	320.0
352.10	1684.0
352.90	1471.0
354.10	2011.0
355.10	736.0
365.00	8925.0
366.00	1368.0
371.00	536.0
372.00	3122.0
373.20	1027.0
382.90	617.0
390.00	435.0
390.90	474.0
401.90	1587.0
403.10	1599.0
404.00	720.0
421.10	1638.0
422.00	1741.0
423.10	11392.0
424.00	3486.0
441.00	28872.0
442.10	178816.0
443.10	36736.0
443.90	2079.0
445.20	390.0

Instrument :
BNA_M
ClientSampleId :
DFTPP81

m/z	Abundance
38.00	798.0
39.00	5876.0
39.90	940.0
41.10	1247.0
42.00	676.0
43.10	1412.0
44.00	1505.0
45.00	557.0
49.00	933.0
50.10	18136.0
51.10	63512.0
52.10	2717.0
53.10	667.0
55.20	895.0
56.00	3046.0
57.00	4965.0
60.30	313.0
61.10	953.0
62.00	1157.0
63.10	1843.0
65.10	1503.0
67.10	633.0
67.90	1576.0
69.00	66784.0
70.10	443.0
71.10	530.0
73.00	2581.0
74.00	6979.0
75.00	10873.0
76.10	4247.0
77.00	72664.0
78.10	4328.0
79.10	7239.0
80.00	4462.0
81.00	5413.0
82.00	1428.0
83.00	1290.0
83.70	360.0
84.20	353.0
84.90	1366.0
86.10	1773.0
86.90	619.0
91.00	1578.0
92.10	1637.0
93.00	11215.0
93.90	601.0
94.90	356.0
95.90	671.0
97.10	596.0
97.90	9623.0
99.00	6281.0
100.00	505.0
100.90	3770.0
103.00	1752.0
104.00	2327.0
105.00	2658.0
105.80	523.0
106.20	677.0
107.00	27544.0
107.90	3225.0
109.00	1153.0
110.00	39944.0
111.00	7967.0
111.90	1293.0
115.70	508.0
116.20	1198.0
117.00	27744.0
118.00	2714.0
118.80	443.0
119.90	349.0
120.90	353.0
122.00	2079.0
123.00	3229.0
123.90	1232.0
125.00	990.0
126.10	362.0
127.00	76648.0
127.90	6091.0
129.00	32928.0
130.00	3799.0
130.80	924.0
131.90	663.0
132.50	346.0
134.10	903.0
135.00	2492.0
136.00	1312.0
136.90	1699.0
138.10	662.0
139.90	745.0
141.00	5379.0
141.90	1475.0

Instrument :

BNA_M

ClientSampleId :

DFTPP81

142.90	1292.0
144.10	496.0
145.20	485.0
145.90	1170.0
146.90	2276.0
147.90	6941.0
148.90	1341.0
149.90	489.0
150.90	524.0
152.00	677.0
152.90	1617.0
153.90	1313.0
155.00	3070.0
156.00	3457.0
157.10	457.0
158.00	1112.0
159.10	1282.0
160.10	1167.0
161.00	3057.0
161.80	885.0
165.10	1915.0
165.90	1826.0
167.00	10829.0
168.00	4439.0
168.80	603.0
169.90	486.0
171.00	306.0
172.00	1209.0
172.90	1497.0
174.00	2235.0
175.00	5690.0
176.00	989.0
177.00	1966.0
177.90	791.0
179.00	8405.0
180.00	5449.0
181.00	2346.0
181.90	406.0
183.80	1030.0
184.20	548.0
185.00	3867.0
186.00	28624.0
187.00	7909.0
188.00	931.0
189.10	1996.0
190.90	1322.0
191.90	2591.0
193.10	3284.0
194.00	643.0
196.10	6847.0
196.90	700.0
197.90	191872.0
199.00	9881.0
200.10	1148.0
201.60	1466.0
202.90	2093.0
204.00	9208.0
205.00	14230.0
206.00	56896.0
207.10	6661.0
208.00	2570.0
209.00	1068.0
210.90	1549.0
211.70	744.0
214.90	622.0
216.00	585.0
217.00	16992.0
218.00	2410.0
218.90	330.0
219.90	581.0
221.00	8043.0
221.70	1975.0
222.90	3451.0
224.00	33664.0
225.00	9418.0
226.10	823.0
227.00	15419.0
228.00	1715.0
229.10	2277.0
230.00	816.0
231.10	1369.0
234.00	1362.0
234.90	1211.0
235.90	1209.0
237.00	1614.0
237.80	363.0
238.90	835.0
241.10	896.0
242.00	2036.0
243.00	2244.0
244.00	24328.0
245.10	3035.0
245.90	4693.0
247.00	877.0
249.00	1038.0

Instrument :

BNA_M

ClientSampleId :

DFTPP81

249.80	498.0
250.90	364.0
252.00	380.0
253.00	457.0
253.90	1133.0
255.00	130032.0
256.00	18456.0
257.00	1309.0
258.00	10496.0
259.10	838.0
264.00	378.0
265.00	3528.0
266.00	536.0
272.00	387.0
273.00	4107.0
274.00	10910.0
275.00	59248.0
276.00	7786.0
277.00	6855.0
278.10	659.0
281.10	514.0
282.90	1441.0
284.10	602.0
285.10	997.0
293.20	1091.0
294.20	489.0
295.10	557.0
296.00	19536.0
297.00	2707.0
300.90	403.0
303.10	2682.0
304.00	931.0
309.00	337.0
314.00	1109.0
315.00	1638.0
316.10	1210.0
321.10	532.0
322.30	581.0
323.10	4314.0
324.10	912.0
325.00	327.0
327.00	1433.0
327.80	425.0
331.90	745.0
333.10	851.0
334.10	3521.0
334.90	1093.0
341.00	987.0
345.90	1820.0
352.10	1971.0
353.10	1285.0
354.00	1946.0
355.00	414.0
365.00	10958.0
365.90	1198.0
371.20	494.0
372.10	3370.0
372.90	343.0
373.20	446.0
382.90	685.0
390.00	371.0
390.90	308.0
391.90	381.0
402.10	2196.0
403.00	1358.0
403.90	355.0
405.10	313.0
421.00	1998.0
422.10	1869.0
423.10	11029.0
424.00	2327.0
425.10	460.0
441.10	31152.0
442.10	205248.0
443.10	38152.0
444.00	3282.0

Instrument :
BNA_M
ClientSampleId :
DFTPP81