

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM012420\
 Data File : BM024623.D
 Acq On : 24 Jan 2020 16:42
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 24 17:22:18 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM012320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 23 14:57:07 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	121	0.00
2	1,4-Dioxane	0.416	0.443	-6.5	127	0.00
3	Pyridine	1.206	1.223	-1.4	124	0.00
4	n-Nitrosodimethylamine	0.535	0.558	-4.3	123	0.00
5 S	2-Fluorophenol	1.052	1.084	-3.0	122	0.00
6	Aniline	1.764	1.786	-1.2	119	0.00
7 S	Phenol-d6	1.449	1.481	-2.2	120	0.00
8	2-Chlorophenol	1.197	1.225	-2.3	121	0.00
9	Benzaldehyde	0.856	0.814	4.9	116	0.00
10 C	Phenol	1.444	1.474	-2.1	120	0.00
11	bis(2-Chloroethyl)ether	1.156	1.189	-2.9	119	0.00
12	1,3-Dichlorobenzene	1.471	1.494	-1.6	120	0.00
13 C	1,4-Dichlorobenzene	1.492	1.511	-1.3	119	0.00
14	1,2-Dichlorobenzene	1.440	1.463	-1.6	120	0.00
15	Benzyl Alcohol	1.194	1.224	-2.5	121	0.00
16	2,2'-oxybis(1-Chloropropane	1.058	1.081	-2.2	118	0.00
17	2-Methylphenol	1.018	1.027	-0.9	118	0.00
18	Hexachloroethane	0.576	0.587	-1.9	120	0.00
19 P	n-Nitroso-di-n-propylamine	1.088	1.092	-0.4	119	0.00
20	3+4-Methylphenols	1.409	1.432	-1.6	118	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	117	0.00
22	Acetophenone	0.516	0.525	-1.7	117	0.00
23 S	Nitrobenzene-d5	0.408	0.419	-2.7	118	0.00
24	Nitrobenzene	0.391	0.403	-3.1	119	0.00
25	Isophorone	0.693	0.710	-2.5	117	0.00
26 C	2-Nitrophenol	0.181	0.185	-2.2	117	0.00
27	2,4-Dimethylphenol	0.246	0.252	-2.4	119	0.00
28	bis(2-Chloroethoxy)methane	0.423	0.432	-2.1	117	0.00
29 C	2,4-Dichlorophenol	0.332	0.342	-3.0	118	0.00
30	1,2,4-Trichlorobenzene	0.401	0.408	-1.7	119	0.00
31	Naphthalene	1.010	1.034	-2.4	118	0.00
32	Benzoic acid	0.225	0.239	-6.2	119	0.01
33	4-Chloroaniline	0.443	0.460	-3.8	119	0.00
34 C	Hexachlorobutadiene	0.277	0.284	-2.5	118	0.00
35	Caprolactam	0.107	0.110	-2.8	118	0.00
36 C	4-Chloro-3-methylphenol	0.353	0.362	-2.5	118	0.00
37	2-Methylnaphthalene	0.780	0.799	-2.4	118	0.00
38	1-Methylnaphthalene	0.742	0.755	-1.8	118	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
40	1,2,4,5-Tetrachlorobenzene	0.714	0.730	-2.2	119	0.00
41 P	Hexachlorocyclopentadiene	0.400	0.420	-5.0	120	0.00
42 S	2,4,6-Tribromophenol	0.325	0.328	-0.9	119	0.00
43 C	2,4,6-Trichlorophenol	0.454	0.461	-1.5	116	0.00
44	2,4,5-Trichlorophenol	0.464	0.476	-2.6	119	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM012420\
 Data File : BM024623.D
 Acq On : 24 Jan 2020 16:42
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 24 17:22:18 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM012320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 23 14:57:07 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.471	1.495	-1.6	118	0.00
46	1,1'-Biphenyl	1.514	1.536	-1.5	118	0.00
47	2-Chloronaphthalene	1.222	1.244	-1.8	118	0.00
48	2-Nitroaniline	0.373	0.384	-2.9	118	0.00
49	Acenaphthylene	1.829	1.863	-1.9	118	0.00
50	Dimethylphthalate	1.617	1.622	-0.3	116	0.00
51	2,6-Dinitrotoluene	0.344	0.349	-1.5	117	0.00
52 C	Acenaphthene	1.135	1.146	-1.0	117	0.00
53	3-Nitroaniline	0.357	0.360	-0.8	114	0.00
54 P	2,4-Dinitrophenol	0.206	0.205	0.5	115	0.00
55	Dibenzofuran	1.851	1.868	-0.9	118	0.00
56 P	4-Nitrophenol	0.279	0.282	-1.1	115	0.00
57	2,4-Dinitrotoluene	0.493	0.498	-1.0	116	0.00
58	Fluorene	1.556	1.562	-0.4	117	0.00
59	2,3,4,6-Tetrachlorophenol	0.470	0.475	-1.1	118	0.00
60	Diethylphthalate	1.642	1.646	-0.2	116	0.00
61	4-Chlorophenyl-phenylether	0.897	0.897	0.0	117	0.00
62	4-Nitroaniline	0.394	0.399	-1.3	115	0.00
63	Azobenzene	1.401	1.427	-1.9	118	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
65	4,6-Dinitro-2-methylphenol	0.118	0.120	-1.7	115	0.00
66 c	n-Nitrosodiphenylamine	0.574	0.594	-3.5	118	0.00
67	4-Bromophenyl-phenylether	0.249	0.252	-1.2	116	0.00
68	Hexachlorobenzene	0.285	0.291	-2.1	118	0.00
69	Atrazine	0.221	0.225	-1.8	114	0.00
70 C	Pentachlorophenol	0.174	0.178	-2.3	116	0.00
71	Phenanthrene	1.083	1.107	-2.2	117	0.00
72	Anthracene	1.060	1.077	-1.6	117	0.00
73	Carbazole	0.965	0.976	-1.1	115	0.00
74	Di-n-butylphthalate	1.233	1.243	-0.8	114	0.00
75 C	Fluoranthene	1.352	1.337	1.1	114	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
77	Benzidine	0.466	0.577	-23.8	121	0.00
78	Pyrene	1.318	1.511	-14.6	112	0.00
79 S	Terphenyl-d14	1.071	1.289	-20.4	110	0.00
80	Butylbenzylphthalate	0.553	0.593	-7.2	104	0.00
81	Benzo(a)anthracene	1.388	1.426	-2.7	101	0.00
82	3,3'-Dichlorobenzidine	0.499	0.498	0.2	94	0.00
83	Chrysene	1.325	1.361	-2.7	100	0.00
84	Bis(2-ethylhexyl)phthalate	0.816	0.821	-0.6	98	0.00
85 c	Di-n-octyl phthalate	1.349	1.232	8.7	87	0.00
86	Indeno(1,2,3-cd)pyrene	1.444	1.200	16.9	79	0.00
87 I	Perylene-d12	1.000	1.000	0.0	82	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM012420\
Data File : BM024623.D
Acq On : 24 Jan 2020 16:42
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Jan 24 17:22:18 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM012320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 23 14:57:07 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.297	1.420	-9.5	87	0.00
89	Benzo(k)fluoranthene	1.266	1.304	-3.0	84	0.00
90 C	Benzo(a)pyrene	1.179	1.218	-3.3	83	0.00
91	Dibenzo(a,h)anthracene	1.119	1.117	0.2	79	0.00
92	Benzo(q,h,i)perylene	1.016	1.039	-2.3	80	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0