

Data Path : Z:\HPCHEM1\BNA M\DATA\BM012517\
 Data File : BM008858.D
 Acq On : 25 Jan 2017 09:22
 Operator : UM/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.448

Quant Time: Jan 25 16:38:57 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM010917.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 25 16:35:23 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	69	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	64	0.00
3	Naphthalene	1.097	1.067	2.7	62	0.00
4 SURR	2-Methylnaphthalene-d10	0.663	0.666	-0.5	64	0.00
5	2-Methylnaphthalene	0.796	0.783	1.6	62	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	62	0.00
7	Acenaphthylene	1.763	1.719	2.5	60	0.00
8 C	Acenaphthene	1.260	1.259	0.1	61	0.00
9	Fluorene	1.577	1.546	2.0	60	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	60	0.00
11	Pentachlorophenol	0.184	0.170	7.6	54	0.00
12	Phenanthrene	1.221	1.176	3.7	58	0.00
13	Anthracene	1.191	1.178	1.1	60	0.00
14 SURR	Fluoranthene-d10	1.180	1.168	1.0	59	0.00
15 C	Fluoranthene	1.529	1.429	6.5	56	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	61	0.00
17	Pyrene	1.594	1.521	4.6	56	0.00
18	Benzo(a)anthracene	1.483	1.402	5.5	56	0.00
19	Chrysene	1.405	1.266	9.9	54	0.00
20 I	Perylene-d12	1.000	1.000	0.0	58	0.00
21	Benzo(b)fluoranthene	1.713	1.539	10.2	51	0.00
22	Benzo(k)fluoranthene	1.495	1.480	1.0	59	0.00
23 C	Benzo(a)pyrene	1.560	1.432	8.2	53	0.00
24	Indeno(1,2,3-cd)pyrene	1.821	1.685	7.5	54	0.00
25	Dibenzo(a,h)anthracene	1.449	1.419	2.1	55	0.00
26	Benzo(a,h,i)perylene	1.588	1.448	8.8	53	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0