

Data Path : Z:\HPCHEM1\BNA M\DATA\BM012517\
 Data File : BM008866.D
 Acq On : 25 Jan 2017 17:27
 Operator : SJ/MA
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.449

Quant Time: Jan 25 18:03:05 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM010917.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 25 16:35:23 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	61	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	58	0.00
3	Naphthalene	1.097	1.080	1.5	58	0.01
4 SURR	2-Methylnaphthalene-d10	0.663	0.654	1.4	57	0.00
5	2-Methylnaphthalene	0.796	0.779	2.1	56	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	57	0.00
7	Acenaphthylene	1.763	1.713	2.8	55	0.00
8 C	Acenaphthene	1.260	1.242	1.4	56	0.00
9	Fluorene	1.577	1.487	5.7	53	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	54	0.00
11	Pentachlorophenol	0.184	0.174	5.4	50	0.00
12	Phenanthrene	1.221	1.162	4.8	52	0.00
13	Anthracene	1.191	1.175	1.3	54	0.00
14 SURR	Fluoranthene-d10	1.180	1.187	-0.6	55	0.00
15 C	Fluoranthene	1.529	1.459	4.6	52	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	60	0.00
17	Pyrene	1.594	1.432	10.2	51	0.00
18	Benzo(a)anthracene	1.483	1.385	6.6	54	0.00
19	Chrysene	1.405	1.271	9.5	53	0.00
20 I	Perylene-d12	1.000	1.000	0.0	58	0.00
21	Benzo(b)fluoranthene	1.713	1.553	9.3	52	0.00
22	Benzo(k)fluoranthene	1.495	1.402	6.2	56	0.00
23 C	Benzo(a)pyrene	1.560	1.402	10.1	53	0.00
24	Indeno(1,2,3-cd)pyrene	1.821	1.687	7.4	54	0.00
25	Dibenzo(a,h)anthracene	1.449	1.425	1.7	56	0.00
26	Benzo(a,h,i)perylene	1.588	1.452	8.6	54	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0