

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012524\
 Data File : BM043905.D
 Acq On : 25 Jan 2024 16:06
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM012524

Quant Time: Jan 26 00:39:38 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM012524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 26 00:26:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	108	0.00
2	1,4-Dioxane	40.000	41.234	-3.1	108	0.00
3	Pyridine	40.000	41.895	-4.7	109	0.00
4	n-Nitrosodimethylamine	40.000	40.826	-2.1	105	0.00
5 S	2-Fluorophenol	80.000	86.311	-7.9	112	0.00
6	Aniline	40.000	41.408	-3.5	109	0.00
7 S	Phenol-d6	80.000	85.085	-6.4	110	0.00
8	2-Chlorophenol	40.000	43.368	-8.4	111	0.00
9	Benzaldehyde	40.000	42.754	-6.9	109	0.00
10 C	Phenol	40.000	42.508	-6.3	111	0.00
11	bis(2-Chloroethyl)ether	40.000	41.406	-3.5	109	0.00
12	1,3-Dichlorobenzene	40.000	41.082	-2.7	108	0.00
13 C	1,4-Dichlorobenzene	40.000	41.047	-2.6	108	0.00
14	1,2-Dichlorobenzene	40.000	41.121	-2.8	108	0.00
15	Benzyl Alcohol	40.000	42.841	-7.1	110	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.311	-3.3	108	0.00
17	2-Methylphenol	40.000	42.090	-5.2	109	0.00
18	Hexachloroethane	40.000	42.581	-6.5	110	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.492	-3.7	107	0.00
20	3+4-Methylphenols	40.000	43.039	-7.6	110	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	107	0.00
22	Acetophenone	40.000	41.126	-2.8	108	0.00
23 S	Nitrobenzene-d5	80.000	87.698	-9.6	111	0.00
24	Nitrobenzene	40.000	42.890	-7.2	110	0.00
25	Isophorone	40.000	41.478	-3.7	107	0.00
26 C	2-Nitrophenol	40.000	43.032	-7.6	131	0.00
27	2,4-Dimethylphenol	40.000	42.909	-7.3	109	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.617	-4.0	108	0.00
29 C	2,4-Dichlorophenol	40.000	44.260	-10.6	110	0.00
30	1,2,4-Trichlorobenzene	40.000	41.027	-2.6	106	0.00
31	Naphthalene	40.000	40.926	-2.3	107	0.00
32	Benzoic acid	40.000	46.427	-16.1	140	0.00
33	4-Chloroaniline	40.000	40.760	-1.9	106	0.00
34 C	Hexachlorobutadiene	40.000	41.341	-3.4	108	0.00
35	Caprolactam	40.000	45.033	-12.6	110	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.309	-8.3	107	0.00
37	2-Methylnaphthalene	40.000	41.026	-2.6	107	0.00
38	1-Methylnaphthalene	40.000	40.834	-2.1	107	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	105	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.816	-4.5	107	0.00
41 P	Hexachlorocyclopentadiene	40.000	45.852	-14.6	112	0.00
42 S	2,4,6-Tribromophenol	80.000	87.233	-9.0	110	0.00
43 C	2,4,6-Trichlorophenol	40.000	46.183	-15.5	110	0.00
44	2,4,5-Trichlorophenol	40.000	46.437	-16.1	110	0.00
45 S	2-Fluorobiphenyl	80.000	83.423	-4.3	105	0.00
46	1,1'-Biphenyl	40.000	41.321	-3.3	106	0.00
47	2-Chloronaphthalene	40.000	41.352	-3.4	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012524\
 Data File : BM043905.D
 Acq On : 25 Jan 2024 16:06
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM012524

Quant Time: Jan 26 00:39:38 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM012524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 26 00:26:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.477	-6.2	117	0.00
49	Acenaphthylene	40.000	41.342	-3.4	105	0.00
50	Dimethylphthalate	40.000	41.730	-4.3	106	0.00
51	2,6-Dinitrotoluene	40.000	42.173	-5.4	111	0.00
52 C	Acenaphthene	40.000	41.521	-3.8	105	0.00
53	3-Nitroaniline	40.000	41.293	-3.2	109	0.00
54 P	2,4-Dinitrophenol	40.000	45.970	-14.9	133	0.00
55	Dibenzofuran	40.000	41.050	-2.6	105	0.00
56 P	4-Nitrophenol	40.000	41.754	-4.4	112	0.00
57	2,4-Dinitrotoluene	40.000	41.301	-3.3	112	0.00
58	Fluorene	40.000	41.393	-3.5	104	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.972	-4.9	111	0.00
60	Diethylphthalate	40.000	41.432	-3.6	103	0.00
61	4-Chlorophenyl-phenylether	40.000	41.366	-3.4	105	0.00
62	4-Nitroaniline	40.000	41.549	-3.9	110	0.00
63	Azobenzene	40.000	41.169	-2.9	104	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	40.000	45.413	-13.5	131	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.755	-4.4	104	0.00
67	4-Bromophenyl-phenylether	40.000	41.106	-2.8	103	0.00
68	Hexachlorobenzene	40.000	40.833	-2.1	103	0.00
69	Atrazine	40.000	42.878	-7.2	102	0.00
70 C	Pentachlorophenol	40.000	41.527	-3.8	114	0.00
71	Phenanthrene	40.000	41.134	-2.8	103	0.00
72	Anthracene	40.000	41.368	-3.4	104	0.00
73	Carbazole	40.000	41.485	-3.7	103	0.00
74	Di-n-butylphthalate	40.000	42.922	-7.3	103	0.00
75 C	Fluoranthene	40.000	40.989	-2.5	103	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	102	0.00
77	Benzydine	40.000	44.413	-11.0	96	0.00
78	Pyrene	40.000	41.521	-3.8	103	0.00
79 S	Terphenyl-d14	80.000	85.636	-7.0	103	0.00
80	Butylbenzylphthalate	40.000	40.450	-1.1	107	0.00
81	Benzo(a)anthracene	40.000	41.276	-3.2	102	0.00
82	3,3'-Dichlorobenzidine	40.000	44.154	-10.4	103	0.00
83	Chrysene	40.000	41.604	-4.0	103	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.621	-1.6	103	0.00
85 c	Di-n-octyl phthalate	40.000	43.136	-7.8	105	0.00
86 I	Perylene-d12	20.000	20.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.996	-5.0	102	0.00
88	Benzo(b)fluoranthene	40.000	41.633	-4.1	101	0.00
89	Benzo(k)fluoranthene	40.000	41.873	-4.7	102	0.00
90 C	Benzo(a)pyrene	40.000	42.112	-5.3	102	0.00
91	Dibenzo(a,h)anthracene	40.000	42.155	-5.4	102	0.00
92	Benzo(g,h,i)perylene	40.000	42.577	-6.4	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012524\
Data File : BM043905.D
Acq On : 25 Jan 2024 16:06
Operator : MA/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM012524

Quant Time: Jan 26 00:39:38 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM012524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 26 00:26:01 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0