


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012621\
 Data File : BM028536.D
 Acq On : 26 Jan 2021 09:15
 Operator : CG/JU
 Sample : DFTPP22
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP022

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM011921.M
 Title : SVOA CALIBRATION
 Last Update : Mon Jan 25 12:52:04 2021



AutoFind: Scans 2480, 2481, 2482; Background Corrected with Scan 2475

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	47.2	21408	PASS
68	69	0.00	2	1.5	306	PASS
69	198	0.00	100	45.0	20406	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	53.2	24109	PASS
197	198	0.00	2	0.3	149	PASS
198	198	100	100	100.0	45341	PASS
199	198	5	9	7.0	3179	PASS
275	198	10	60	27.9	12652	PASS
365	198	1	100	3.4	1532	PASS
441	198	0.01	100	13.4	6090	PASS
442	198	50	100	86.9	39381	PASS
443	442	15	24	19.6	7717	PASS

m/z	Abundance
43.90	403.0
73.10	160.0
207.00	470.0

Instrument :
BNA_M
ClientSampleId :
DFTPP022

m/z	Abundance
38.00	243.0
39.10	1515.0
40.00	172.0
44.00	425.0
49.00	203.0
50.10	4686.0
51.10	17792.0
52.10	927.0
56.00	580.0
57.00	1086.0
61.00	229.0
61.90	236.0
63.00	594.0
65.00	300.0
69.00	15913.0
73.00	342.0
74.00	1842.0
75.00	2789.0
76.10	921.0
77.10	17336.0
78.10	1117.0
79.00	1236.0
80.00	949.0
81.00	1206.0
81.90	272.0
82.90	357.0
84.90	177.0
86.00	421.0
86.90	223.0
90.90	326.0
92.00	295.0
93.00	2203.0
98.00	1768.0
99.00	1298.0
100.00	151.0
101.00	732.0
102.90	256.0
104.00	486.0
104.90	488.0
105.90	168.0
107.00	5056.0
108.00	867.0
109.00	169.0
110.00	9218.0
111.00	1363.0
116.00	262.0
117.00	4597.0
118.00	348.0
122.00	334.0
123.00	580.0
124.00	257.0
125.00	190.0
127.00	17920.0
128.00	1367.0
129.00	6853.0
130.00	580.0
131.00	162.0
134.00	194.0
135.00	643.0
135.90	197.0
136.90	290.0
141.00	838.0
142.00	353.0
142.90	192.0
146.00	166.0
147.00	536.0
148.00	1015.0
148.90	247.0
152.90	255.0
153.90	182.0
155.00	524.0
156.00	736.0
157.80	175.0
160.00	236.0
161.00	390.0
164.90	335.0
165.90	243.0
167.00	1876.0
168.00	1132.0
169.00	167.0
171.90	167.0
173.00	170.0
174.00	377.0
175.00	620.0
175.90	206.0
176.90	456.0
179.00	1355.0
180.00	935.0
181.00	398.0
185.00	645.0
186.00	4258.0
187.00	1368.0

Instrument :
BNA_M
ClientSampleId :
DFTPP022

187.90	157.0
189.00	292.0
191.00	180.0
192.00	387.0
193.00	471.0
195.90	861.0
198.00	30208.0
199.00	2236.0
203.00	268.0
204.00	1144.0
205.00	2061.0
206.10	8154.0
207.00	1504.0
207.90	345.0
209.00	158.0
210.40	152.0
211.00	384.0
216.00	269.0
217.00	2191.0
218.00	325.0
221.00	1346.0
221.70	260.0
223.00	453.0
224.00	4114.0
225.00	1171.0
227.00	1834.0
227.90	242.0
228.90	369.0
235.00	153.0
237.00	235.0
241.90	202.0
243.00	240.0
244.00	3344.0
245.10	431.0
246.00	611.0
247.00	172.0
254.10	153.0
255.00	16792.0
256.00	2584.0
256.90	185.0
258.00	877.0
258.90	175.0
264.90	352.0
273.00	478.0
274.00	1295.0
275.00	7764.0
276.00	1042.0
277.00	590.0
292.90	179.0
296.00	2190.0
297.00	303.0
302.90	285.0
315.00	204.0
323.00	699.0
333.90	366.0
354.00	155.0
364.90	883.0
365.90	151.0
372.10	275.0
402.90	226.0
421.90	153.0
423.00	1164.0
424.00	360.0
441.00	3044.0
442.00	19616.0
443.00	3776.0
444.00	326.0

Instrument :
BNA_M
ClientSampleId :
DFTPP022

m/z	Abundance
36.00	150.0
38.00	341.0
39.00	1915.0
40.00	178.0
44.00	347.0
49.10	229.0
50.10	7198.0
51.10	25656.0
52.00	1388.0
56.00	783.0
57.00	1758.0
61.00	286.0
61.90	342.0
63.00	1016.0
65.00	467.0
68.10	532.0
69.00	24512.0
73.10	335.0
74.10	2548.0
75.00	3838.0
76.00	1396.0
77.10	26632.0
78.10	1769.0
79.00	2015.0
80.00	1505.0
81.00	1835.0
82.00	523.0
83.00	421.0
85.00	321.0
86.00	568.0
87.00	250.0
91.00	406.0
92.00	459.0
93.00	3575.0
94.00	192.0
98.00	2719.0
99.00	1905.0
101.00	1099.0
102.90	418.0
104.00	825.0
104.90	672.0
105.90	224.0
107.00	8389.0
108.00	1297.0
109.10	274.0
110.00	14107.0
111.00	2015.0
111.90	253.0
116.00	503.0
117.00	7309.0
117.90	516.0
121.90	561.0
123.00	863.0
124.00	414.0
125.00	328.0
127.00	28872.0
128.00	2235.0
129.00	11126.0
130.00	867.0
131.00	177.0
133.90	322.0
135.00	840.0
136.10	347.0
137.00	396.0
140.00	171.0
141.00	1472.0
142.00	433.0
143.00	394.0
146.00	309.0
147.00	761.0
148.00	1844.0
149.00	422.0
153.00	399.0
154.00	379.0
155.00	737.0
156.00	1037.0
157.00	212.0
158.00	300.0
159.00	189.0
159.90	448.0
161.00	648.0
161.90	178.0
165.00	482.0
166.00	468.0
167.00	2814.0
168.00	1575.0
168.90	221.0
172.00	271.0
173.00	365.0
174.00	543.0
175.00	1015.0
176.00	337.0

Instrument :
BNA_M
ClientSampleId :
DFTPP022

177.00	512.0
177.90	212.0
179.00	2195.0
180.00	1492.0
181.00	715.0
184.00	156.0
185.00	1055.0
186.00	7495.0
187.00	2186.0
188.00	352.0
188.90	488.0
190.80	241.0
192.00	662.0
193.00	916.0
196.00	1683.0
196.70	263.0
198.00	53992.0
199.00	3887.0
200.00	302.0
201.50	194.0
203.00	349.0
204.00	1937.0
205.00	3136.0
206.10	14269.0
207.00	2387.0
208.00	509.0
208.90	234.0
210.00	288.0
211.00	507.0
215.00	170.0
216.00	310.0
217.00	3866.0
218.00	545.0
221.00	2649.0
221.90	437.0
223.00	810.0
224.00	7803.0
225.00	2185.0
226.10	234.0
227.00	3247.0
228.00	534.0
229.00	677.0
231.00	276.0
234.00	196.0
234.90	254.0
237.00	382.0
241.00	235.0
242.00	458.0
243.00	444.0
244.00	6487.0
245.00	732.0
246.00	1179.0
247.00	267.0
248.90	276.0
254.20	287.0
255.00	31776.0
256.00	4574.0
257.00	342.0
258.00	1850.0
259.00	284.0
265.00	769.0
273.00	1052.0
274.00	2544.0
275.00	14605.0
276.00	1948.0
277.00	1106.0
277.90	222.0
282.90	153.0
285.00	220.0
292.90	252.0
296.00	4046.0
297.00	590.0
303.00	498.0
304.00	167.0
314.00	158.0
315.00	485.0
316.00	244.0
320.90	152.0
323.00	1234.0
324.00	235.0
327.00	283.0
333.00	168.0
334.00	803.0
335.00	195.0
341.00	179.0
345.90	295.0
352.00	427.0
353.00	230.0
354.10	333.0
365.00	1795.0
365.90	260.0
372.00	636.0
372.90	181.0
383.00	191.0
402.00	234.0
403.00	350.0

Instrument :
BNA_M
ClientSampleId :
DFTPP022

404.00	193.0
421.00	431.0
422.00	372.0
423.00	2509.0
424.00	787.0
441.00	6960.0
442.00	44280.0
443.00	8813.0
444.00	799.0

Instrument :
BNA_M
ClientSampleId :
DFTPP022

m/z	Abundance
38.00	287.0
39.10	1595.0
40.00	173.0
44.00	370.0
49.00	282.0
50.10	5754.0
51.10	20776.0
52.00	1136.0
56.00	587.0
57.00	1501.0
60.90	266.0
62.00	276.0
63.00	845.0
65.00	404.0
68.10	386.0
69.00	20792.0
73.00	329.0
74.00	2163.0
75.00	3225.0
76.00	1179.0
77.10	22456.0
78.10	1561.0
79.00	1623.0
80.00	1158.0
81.00	1674.0
82.00	393.0
83.00	352.0
85.00	216.0
86.00	487.0
87.00	224.0
90.90	310.0
92.00	416.0
93.00	2793.0
93.90	187.0
98.00	2277.0
99.00	1596.0
101.00	933.0
102.90	324.0
103.90	728.0
105.00	624.0
107.00	7039.0
108.00	1004.0
109.00	221.0
110.00	12184.0
111.00	1809.0
111.90	215.0
116.00	401.0
117.00	6654.0
118.00	416.0
122.00	577.0
123.00	771.0
124.00	318.0
125.00	347.0
126.20	163.0
127.00	25536.0
128.00	1837.0
129.00	9943.0
130.00	842.0
130.90	183.0
134.00	269.0
135.00	859.0
136.00	294.0
137.00	449.0
139.90	155.0
141.00	1417.0
142.00	391.0
143.00	259.0
146.00	221.0
147.00	693.0
148.00	1622.0
149.00	273.0
151.00	171.0
153.00	428.0
153.90	312.0
155.00	747.0
156.00	1145.0
157.00	171.0
157.90	269.0
158.90	178.0
160.00	336.0
161.00	601.0
162.00	187.0
165.00	503.0
166.10	333.0
167.00	2810.0
168.00	1449.0
168.90	234.0
171.90	249.0
173.00	347.0
174.00	640.0
175.00	980.0
175.90	289.0

Instrument :
BNA_M
ClientSampleId :
DFTPP022

177.00	553.0
178.00	176.0
179.00	2019.0
180.00	1244.0
181.00	623.0
183.90	158.0
185.00	913.0
186.00	7194.0
187.00	1900.0
188.10	211.0
189.00	474.0
190.90	315.0
191.90	646.0
193.00	727.0
194.00	183.0
196.00	1628.0
197.10	447.0
198.00	51824.0
199.00	3414.0
200.00	274.0
201.60	247.0
203.00	391.0
204.00	1959.0
205.00	3086.0
206.10	13719.0
207.00	2085.0
208.00	507.0
209.00	174.0
210.40	333.0
211.10	540.0
214.90	201.0
216.00	303.0
217.00	4022.0
218.00	446.0
221.00	2412.0
221.70	578.0
223.00	846.0
224.00	7598.0
225.00	2090.0
226.00	207.0
227.00	3091.0
228.00	487.0
229.00	627.0
231.00	340.0
234.00	213.0
234.90	296.0
235.90	182.0
237.00	300.0
241.00	178.0
241.90	392.0
243.00	525.0
244.00	5883.0
245.10	955.0
246.00	1180.0
246.90	269.0
248.90	247.0
254.20	322.0
255.00	32448.0
256.00	4811.0
256.90	334.0
258.00	1774.0
259.00	347.0
265.00	641.0
273.00	887.0
274.00	2588.0
275.00	15588.0
276.00	2075.0
277.00	1210.0
278.00	190.0
285.10	160.0
293.00	262.0
296.00	3993.0
296.90	540.0
303.00	496.0
314.00	178.0
314.90	524.0
315.90	192.0
320.90	150.0
323.00	1309.0
324.00	206.0
326.90	264.0
334.00	912.0
335.00	242.0
341.00	174.0
346.00	297.0
352.00	354.0
353.00	288.0
354.00	427.0
364.90	1917.0
366.00	262.0
372.00	644.0
373.00	257.0
383.00	252.0
402.00	377.0
403.00	449.0
404.00	181.0

Instrument :
BNA_M
ClientSampleId :
DFTPP022

420.90	354.0
422.00	438.0
423.00	2977.0
424.00	731.0
441.00	8266.0
442.00	54248.0
443.00	10563.0
444.00	999.0

Instrument :
BNA_M
ClientSampleId :
DFTPP022