

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012623\
 Data File : BM038595.D
 Acq On : 27 Jan 2023 21:13
 Operator : CG/JU
 Sample : 01226-13
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EX9G0

Quant Time: Jan 27 22:01:18 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM012623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 27 00:18:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.962	152	1813	0.400	ng/ul	0.00
4) Naphthalene-d8	10.768	136	5083	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.597	164	3513	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.337	188	8180	0.400	ng/ul	0.00
17) Chrysene-d12	21.512	240	7519	0.400	ng/ul	0.00
23) Perylene-d12	23.921	264	7645	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.338	96	7924	3.328	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.357	152	2369	0.300	ng/ul	0.00
18) Fluoranthene-d10	19.362	212	7832	0.365	ng/ul	0.00
Target Compounds						
16) Anthracene	17.468	178	688	0.031	ng/ul#	Qvalue 61

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012623\
Data File : BM038595.D
Acq On : 27 Jan 2023 21:13
Operator : CG/JU
Sample : 01226-13
Misc :
ALS Vial : 57 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EX9G0

Quant Time: Jan 27 22:01:18 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM012623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 27 00:18:15 2023
Response via : Initial Calibration

