

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM012716\
 Data File : BM004130.D
 Acq On : 27 Jan 2016 11:03
 Operator : SJ/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Jan 28 00:01:47 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM012216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 22 14:46:41 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
2	1,4-Dioxane	0.547	0.466	14.8	83	0.00
3	n-Nitrosodimethylamine	0.446	0.397	11.0	80	0.00
4 S	2-Fluorophenol	1.043	0.945	9.4	82	0.00
5 S	Phenol-d6	1.259	1.131	10.2	82	0.00
6	bis(2-Chloroethyl)ether	1.167	1.065	8.7	83	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	85	0.00
8 S	Nitrobenzene-d5	0.219	0.201	8.2	86	0.00
9	Nitrobenzene	0.250	0.227	9.2	87	0.00
10	Naphthalene	1.165	1.066	8.5	84	0.00
11	Hexachlorobutadiene	0.180	0.163	9.4	84	0.00
12	2-Methylnaphthalene	0.762	0.704	7.6	86	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	90	0.00
14 S	2,4,6-Tribromophenol	0.145	0.124	14.5	89	0.00
15 S	2-Fluorobiphenyl	1.532	1.355	11.6	87	0.00
16	Acenaphthylene	2.158	1.930	10.6	89	0.00
17	Acenaphthene	1.634	1.416	13.3	88	0.00
18	Fluorene	1.987	1.738	12.5	88	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
20	4-Bromophenyl-phenylether	0.163	0.147	9.8	90	0.00
21	Hexachlorobenzene	0.171	0.156	8.8	93	0.00
22	Pentachlorophenol	0.043	0.035	18.6	107	0.00
23	Phenanthrene	1.074	0.989	7.9	97	0.00
24	Anthracene	1.142	1.038	9.1	95	0.00
25	Fluoranthene	1.355	1.232	9.1	96	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
27	Benzydine	0.241	0.202	16.2	93	0.02
28	Pyrene	1.423	1.235	13.2	94	0.00
29 S	Terphenyl-d14	0.643	0.557	13.4	95	0.00
30	Benzo(a)anthracene	1.518	1.197	21.1	87	0.00
31	3,3'-Dichlorobenzidine	0.338	0.254	24.9	89	0.00
32	Chrysene	1.372	1.313	4.3	106	0.00
33	Bis(2-ethylhexyl)phthalate	0.595	0.434	27.1#	101	-0.02
34	Indeno(1,2,3-cd)pyrene	1.463	1.175	19.7	89	0.00
35 I	Perylene-d12	1.000	1.000	0.0	92	0.00
36	Benzo(b)fluoranthene	1.551	1.466	5.5	94	0.00
37	Benzo(k)fluoranthene	1.404	1.238	11.8	92	0.00
38 C	Benzo(a)pyrene	1.366	1.246	8.8	94	0.00
39	Dibenzo(a,h)anthracene	1.227	1.034	15.7	86	0.00
40	Benzo(g,h,i)perylene	1.339	1.194	10.8	92	0.00

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM012716\
Data File : BM004130.D
Acq On : 27 Jan 2016 11:03
Operator : SJ/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Jan 28 00:01:47 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM012216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 22 14:46:41 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				