

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012721\
 Data File : BM028546.D
 Acq On : 26 Jan 2021 18:21
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 27 03:25:33 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM012021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 27 03:11:35 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	104	0.00
2	1,4-Dioxane	40.000	33.626	15.9	96	0.00
3	Pyridine	40.000	35.256	11.9	98	0.00
4	n-Nitrosodimethylamine	40.000	30.873	22.8	88	0.00
5 S	2-Fluorophenol	80.000	70.862	11.4	100	0.00
6	Aniline	40.000	36.643	8.4	104	0.00
7 S	Phenol-d6	80.000	72.553	9.3	103	0.00
8	2-Chlorophenol	40.000	36.336	9.2	103	0.00
9	Benzaldehyde	40.000	34.526	13.7	101	0.00
10 C	Phenol	40.000	36.296	9.3	103	0.00
11	bis(2-Chloroethyl)ether	40.000	36.735	8.2	105	0.00
12	1,3-Dichlorobenzene	40.000	36.433	8.9	104	0.00
13 C	1,4-Dichlorobenzene	40.000	36.371	9.1	104	0.00
14	1,2-Dichlorobenzene	40.000	36.432	8.9	104	0.00
15	Benzyl Alcohol	40.000	36.045	9.9	102	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	34.118	14.7	98	0.00
17	2-Methylphenol	40.000	37.019	7.5	104	0.00
18	Hexachloroethane	40.000	36.501	8.7	103	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.856	7.9	102	0.00
20	3+4-Methylphenols	40.000	37.444	6.4	106	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	104	0.00
22	Acetophenone	40.000	35.332	11.7	104	0.00
23 S	Nitrobenzene-d5	80.000	69.123	13.6	101	0.00
24	Nitrobenzene	40.000	34.342	14.1	100	0.00
25	Isophorone	40.000	36.708	8.2	106	0.00
26 C	2-Nitrophenol	40.000	37.479	6.3	107	0.00
27	2,4-Dimethylphenol	40.000	36.649	8.4	105	0.00
28	bis(2-Chloroethoxy)methane	40.000	36.268	9.3	106	0.00
29 C	2,4-Dichlorophenol	40.000	37.125	7.2	107	0.00
30	1,2,4-Trichlorobenzene	40.000	35.280	11.8	104	0.00
31	Naphthalene	40.000	35.723	10.7	105	0.00
32	Benzoic acid	40.000	38.451	3.9	112	0.00
33	4-Chloroaniline	40.000	36.230	9.4	105	0.00
34 C	Hexachlorobutadiene	40.000	35.522	11.2	105	0.00
35	Caprolactam	40.000	36.614	8.5	107	0.00
36 C	4-Chloro-3-methylphenol	40.000	36.452	8.9	104	0.00
37	2-Methylnaphthalene	40.000	36.170	9.6	105	0.00
38	1-Methylnaphthalene	40.000	35.945	10.1	104	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	105	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	35.449	11.4	105	0.00
41 P	Hexachlorocyclopentadiene	40.000	33.362	16.6	98	0.00
42 S	2,4,6-Tribromophenol	80.000	74.605	6.7	107	0.00
43 C	2,4,6-Trichlorophenol	40.000	36.635	8.4	105	0.00
44	2,4,5-Trichlorophenol	40.000	36.671	8.3	106	0.00
45 S	2-Fluorobiphenyl	80.000	71.147	11.1	106	0.00
46	1,1'-Biphenyl	40.000	35.353	11.6	105	0.00
47	2-Chloronaphthalene	40.000	35.315	11.7	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012721\
 Data File : BM028546.D
 Acq On : 26 Jan 2021 18:21
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 27 03:25:33 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM012021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 27 03:11:35 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	35.007	12.5	100	0.00
49	Acenaphthylene	40.000	36.049	9.9	106	0.00
50	Dimethylphthalate	40.000	35.637	10.9	105	0.00
51	2,6-Dinitrotoluene	40.000	36.759	8.1	107	0.00
52 C	Acenaphthene	40.000	35.466	11.3	105	0.00
53	3-Nitroaniline	40.000	36.864	7.8	105	0.00
54 P	2,4-Dinitrophenol	40.000	36.077	9.8	106	0.00
55	Dibenzofuran	40.000	35.316	11.7	105	0.00
56 P	4-Nitrophenol	40.000	36.064	9.8	101	0.00
57	2,4-Dinitrotoluene	40.000	37.108	7.2	104	0.00
58	Fluorene	40.000	35.386	11.5	104	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	36.682	8.3	105	0.00
60	Diethylphthalate	40.000	36.044	9.9	105	0.00
61	4-Chlorophenyl-phenylether	40.000	35.202	12.0	104	0.00
62	4-Nitroaniline	40.000	35.914	10.2	101	0.00
63	Azobenzene	40.000	35.046	12.4	101	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	40.000	37.098	7.3	107	0.00
66 c	n-Nitrosodiphenylamine	40.000	36.756	8.1	105	0.00
67	4-Bromophenyl-phenylether	40.000	36.840	7.9	104	0.00
68	Hexachlorobenzene	40.000	36.215	9.5	105	0.00
69	Atrazine	40.000	37.343	6.6	103	0.00
70 C	Pentachlorophenol	40.000	38.324	4.2	105	0.00
71	Phenanthrene	40.000	35.716	10.7	103	0.00
72	Anthracene	40.000	35.981	10.0	102	0.00
73	Carbazole	40.000	35.403	11.5	100	0.00
74	Di-n-butylphthalate	40.000	38.161	4.6	106	0.00
75 C	Fluoranthene	40.000	34.933	12.7	99	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	95	0.00
77	Benzydine	40.000	46.104	-15.3	105	0.00
78	Pyrene	40.000	37.149	7.1	99	0.00
79 S	Terphenyl-d14	80.000	74.692	6.6	99	0.00
80	Butylbenzylphthalate	40.000	39.094	2.3	101	0.00
81	Benzo(a)anthracene	40.000	35.876	10.3	96	0.00
82	3,3'-Dichlorobenzidine	40.000	38.163	4.6	96	0.00
83	Chrysene	40.000	35.477	11.3	94	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.376	-0.9	104	0.00
85 c	Di-n-octyl phthalate	40.000	39.505	1.2	102	0.00
86 I	Perylene-d12	20.000	20.000	0.0	94	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.720	8.2	93	0.00
88	Benzo(b)fluoranthene	40.000	36.589	8.5	91	0.00
89	Benzo(k)fluoranthene	40.000	37.333	6.7	95	0.00
90 C	Benzo(a)pyrene	40.000	36.768	8.1	93	0.00
91	Dibenzo(a,h)anthracene	40.000	36.963	7.6	94	0.00
92	Benzo(g,h,i)perylene	40.000	36.659	8.4	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012721\
Data File : BM028546.D
Acq On : 26 Jan 2021 18:21
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Jan 27 03:25:33 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM012021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 27 03:11:35 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0