




Data File : BM013920.D
Acq On : 28 Jan 2018 13:52
Operator : SJ/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jan 29 02:13:45 2018
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM011018.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun Jan 28 02:49:11 2018
Response via : Initial Calibration



TIC: BM013920.D

(85) Benzo(b)fluoranthene

22.721min (-0.006) 19.59ng/ul m

response 732096

Ion	Exp%	Act%
252.00	100	100
253.00	22.40	20.29
125.00	4.50	6.01#
0.00	0.00	0.00