

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012825\
 Data File : BM049518.D
 Acq On : 30 Jan 2025 00:48
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020083

Quant Time: Jan 30 07:36:14 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM012825.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 28 14:45:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.510	152	199353	20.000	ng/u1	0.00
20) Naphthalene-d8	10.269	136	741329	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.151	164	485793	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.904	188	1028396	20.000	ng/u1	0.00
79) Chrysene-d12	21.133	240	1083363	20.000	ng/u1	0.00
88) Perylene-d12	23.909	264	1125082	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.116	96	42979	7.988	ng/uL	0.00
4) Pyridine-d5	3.505	84	318604	20.154	ng/u1	0.00
7) Phenol-d5	6.728	99	337735	20.877	ng/u1	0.02
9) Bis-(2-Chloroethyl)eth...	6.863	67	222670	20.129	ng/u1	0.00
11) 2-Chlorophenol-d4	7.057	132	266611	21.298	ng/u1	0.00
15) 4-Methylphenol-d8	8.240	113	247300	20.865	ng/u1	0.01
21) Nitrobenzene-d5	8.651	128	120988	21.333	ng/u1	0.00
24) 2-Nitrophenol-d4	9.369	143	135761	21.946	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.922	165	274840	21.772	ng/u1	0.02
31) 4-Chloroaniline-d4	10.422	131	324990	20.070	ng/u1	0.00
46) Dimethylphthalate-d6	13.574	166	740849	20.799	ng/u1	0.00
49) Acenaphthylene-d8	13.839	160	873087	20.862	ng/u1	0.00
54) 4-Nitrophenol-d4	14.427	143	99821	17.084	ng/u1	0.05
60) Fluorene-d10	15.151	176	664882	21.226	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.280	200	122278	18.869	ng/u1	0.00
73) Anthracene-d10	17.004	188	1063315	20.705	ng/u1	0.00
81) Pyrene-d10	19.304	212	1270803	20.066	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.721	264	1251110	20.849	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.146	88	45240	8.001	ng/uL	94
5) Pyridine	3.528	79	325691	20.089	ng/u1	97
6) Benzaldehyde	6.669	77	226043	25.442	ng/u1	99
8) Phenol	6.757	94	352106	20.737	ng/u1	97
10) Bis(2-Chloroethyl)ether	6.951	93	280056	20.226	ng/u1	96
12) 2-Chlorophenol	7.093	128	275721	21.329	ng/u1	98
13) 2-Methylphenol	7.975	108	249326	20.954	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.028	45	436384	20.460	ng/u1	99
16) Acetophenone	8.328	105	412187	20.913	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.316	70	219054	21.192	ng/u1	99
18) 4-Methylphenol	8.304	108	273502	21.297	ng/u1	98
19) Hexachloroethane	8.569	117	116236	20.464	ng/u1	97
22) Nitrobenzene	8.698	77	307858	20.323	ng/u1	98
23) Isophorone	9.222	82	566365	20.909	ng/u1	99
25) 2-Nitrophenol	9.398	139	149511	21.767	ng/u1	98
26) 2,4-Dimethylphenol	9.487	107	254314	20.119	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.704	93	348459	20.539	ng/u1	99
29) 2,4-Dichlorophenol	9.945	162	268727	21.176	ng/u1	98
30) Naphthalene	10.316	128	815755	20.668	ng/u1	100
32) 4-Chloroaniline	10.445	127	314594	20.304	ng/u1	98
33) Hexachlorobutadiene	10.610	225	192459	20.231	ng/u1	97
34) Caprolactam	11.216	113	83905	23.969	ng/u1	96
35) 4-Chloro-3-methylphenol	11.604	107	245194	22.363	ng/u1	98
36) 2-Methylnaphthalene	11.939	142	569909	21.171	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012825\
 Data File : BM049518.D
 Acq On : 30 Jan 2025 00:48
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 21 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD020083


Quant Time: Jan 30 07:36:14 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM012825.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 28 14:45:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.163	142	562721	21.205	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.316	216	357408	19.959	ng/ul	98
40) Hexachlorocyclopentadiene	12.298	237	205575	19.569	ng/ul	99
41) 2,4,6-Trichlorophenol	12.575	196	221668	20.990	ng/ul	95
42) 2,4,5-Trichlorophenol	12.669	196	236363	21.296	ng/ul	97
43) 1,1'-Biphenyl	12.975	154	741325	20.113	ng/ul	100
44) 2-Chloronaphthalene	13.010	162	612943	20.512	ng/ul	99
45) 2-Nitroaniline	13.233	65	157643	21.000	ng/ul	99
47) Dimethylphthalate	13.622	163	724586	20.478	ng/ul	99
48) 2,6-Dinitrotoluene	13.739	165	138398	21.748	ng/ul	95
50) Acenaphthylene	13.869	152	891172	20.727	ng/ul	99
51) 3-Nitroaniline	14.074	138	137027	21.677	ng/ul	95
52) Acenaphthene	14.216	153	629612	20.511	ng/ul	99
53) 2,4-Dinitrophenol	14.280	184	84146	21.385	ng/ul	93
55) 4-Nitrophenol	14.439	109	89943	21.038	ng/ul	97
56) Dibenzofuran	14.551	168	916490	20.939	ng/ul	99
57) 2,4-Dinitrotoluene	14.533	165	212364	22.316	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	14.798	232	196685	21.760	ng/ul	94
59) Diethylphthalate	14.998	149	702409	20.912	ng/ul	100
61) Fluorene	15.210	166	745075	21.066	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.210	204	401285	20.917	ng/ul	96
63) 4-Nitroaniline	15.239	138	126923	21.744	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.298	198	132902	18.711	ng/ul	98
67) N-Nitrosodiphenylamine	15.427	169	611203	19.937	ng/ul	97
68) 4-Bromophenyl-phenylether	16.104	248	236854	20.170	ng/ul	95
69) Hexachlorobenzene	16.210	284	274688	20.257	ng/ul	98
70) Atrazine	16.392	200	241194	19.988	ng/ul	98
71) Pentachlorophenol	16.568	266	175557	20.561	ng/ul	98
72) Phenanthrene	16.945	178	1181186	20.404	ng/ul	100
74) Anthracene	17.033	178	1205160	20.668	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	12.933	216	353711	19.030	ng/ul	99
76) Pentachlorobenzene	14.474	250	346554	19.499	ng/ul	98
77) Carbazole	17.315	167	1070846	20.548	ng/ul	99
78) Di-n-butylphthalate	17.892	149	1205461	20.252	ng/ul	99
80) Fluoranthene	18.974	202	1433185	19.813	ng/ul	99
82) Pyrene	19.339	202	1548449	19.895	ng/ul	99
83) Butylbenzylphthalate	20.262	149	546076	20.833	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.056	252	461218	22.422	ng/ul	100
85) Benzo(a)anthracene	21.115	228	1564537	20.558	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.068	149	808792	21.000	ng/ul	100
87) Chrysene	21.174	228	1469150	20.788	ng/ul	99
89) Di-n-octyl phthalate	22.127	149	1360108	19.763	ng/ul	100
90) Benzo(b)fluoranthene	23.039	252	1496652	20.523	ng/ul	99
91) Benzo(k)fluoranthene	23.097	252	1501750	20.379	ng/ul	99
93) Benzo(a)pyrene	23.780	252	1404273	20.527	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.991	276	1722529	20.314	ng/ul	98
95) Dibenzo(a,h)anthracene	27.038	278	1385752	20.168	ng/ul	99
96) Benzo(g,h,i)perylene	27.944	276	1312681	20.361	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jan 30 07:36:14 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM012825.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jan 28 14:45:18 2025
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.001 ng/uL
RT: 3.146 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48




Instrument :
BNA_M
ClientSampleId :
SSTD020083


Tgt Ion: 88 Resp: 45240
Ion Ratio Lower Upper
88 100
43 28.5 23.4 35.0
58 78.6 57.3 85.9



#5
Pyridine
Concen: 20.089 ng/uL
RT: 3.528 min Scan# 92
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion: 79 Resp: 325691
Ion Ratio Lower Upper
79 100
52 72.0 60.7 91.1
51 35.9 29.3 43.9









#6
Benzaldehyde
Concen: 25.442 ng/ul
RT: 6.669 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48




Instrument :
BNA_M
ClientSampleId :
SSTD020083


Tgt Ion: 77 Resp: 226043
Ion Ratio Lower Upper
77 100
105 93.8 74.2 111.4
106 92.3 72.7 109.1



#8
Phenol
Concen: 20.737 ng/ul
RT: 6.757 min Scan# 641
Delta R.T. 0.023 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion: 94 Resp: 352106
Ion Ratio Lower Upper
94 100
65 26.8 23.4 35.0
66 33.4 27.8 41.8









#10
Bis(2-Chloroethyl)ether
Concen: 20.226 ng/ul
RT: 6.951 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48




Instrument :
BNA_M
ClientSampleId :
SSTD020083


Tgt Ion: 93 Resp: 280056
Ion Ratio Lower Upper
93 100
63 79.1 66.3 99.5
95 32.8 25.0 37.6



#12
2-Chlorophenol
Concen: 21.329 ng/ul
RT: 7.093 min Scan# 698
Delta R.T. 0.006 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion:128 Resp: 275721
Ion Ratio Lower Upper
128 100
64 48.7 40.3 60.5
130 32.8 26.2 39.4









#13
2-Methylphenol
Concen: 20.954 ng/ul
RT: 7.975 min Scan# 845
Delta R.T. 0.012 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48




Instrument :
BNA_M
ClientSampleId :
SSTD020083


Tgt Ion:108 Resp: 249326
Ion Ratio Lower Upper
108 100
107 85.3 69.7 104.5



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.460 ng/ul
RT: 8.028 min Scan# 857
Delta R.T. -0.006 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion: 45 Resp: 436384
Ion Ratio Lower Upper
45 100
77 13.6 10.6 15.8
79 9.9 7.8 11.6









#16
Acetophenone
Concen: 20.913 ng/ul
RT: 8.328 min Scan# 907
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48




Instrument :
BNA_M
ClientSampleId :
SSTD020083


Tgt Ion:105 Resp: 412187
Ion Ratio Lower Upper
105 100
77 82.6 64.6 97.0
51 28.4 23.3 34.9



#17
N-Nitroso-di-n-propylamine
Concen: 21.192 ng/ul
RT: 8.316 min Scan# 906
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion: 70 Resp: 219054
Ion Ratio Lower Upper
70 100
42 57.9 47.0 70.6
101 9.1 7.2 10.8
130 19.1 14.6 21.8









#18
4-Methylphenol
Concen: 21.297 ng/ul
RT: 8.304 min Scan# 901
Delta R.T. 0.018 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48




Instrument :
BNA_M
ClientSampleId :
SSTD020083


Tgt Ion:108 Resp: 273502
Ion Ratio Lower Upper
108 100
107 110.2 90.0 135.0



#19
Hexachloroethane
Concen: 20.464 ng/ul
RT: 8.569 min Scan# 949
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion:117 Resp: 116236
Ion Ratio Lower Upper
117 100
201 119.5 94.9 142.3
199 76.2 57.3 85.9









#22
Nitrobenzene
Concen: 20.323 ng/ul
RT: 8.698 min Scan# 91
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48




Instrument :
BNA_M
ClientSampleId :
SSTD020083


Tgt Ion: 77 Resp: 307858
Ion Ratio Lower Upper
77 100
123 44.0 33.9 50.9
65 12.5 10.6 16.0



#23
Isophorone
Concen: 20.909 ng/ul
RT: 9.222 min Scan# 1060
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion: 82 Resp: 566365
Ion Ratio Lower Upper
82 100
95 6.2 5.1 7.7
138 17.1 13.3 19.9









#25
2-Nitrophenol
Concen: 21.767 ng/ul
RT: 9.398 min Scan# 1105
Delta R.T. -0.006 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48




Instrument :
BNA_M
ClientSampleId :
SSTD020083


Tgt Ion:139 Resp: 149511
Ion Ratio Lower Upper
139 100
65 49.4 40.6 60.8
109 26.7 20.7 31.1



#26
2,4-Dimethylphenol
Concen: 20.119 ng/ul
RT: 9.487 min Scan# 1105
Delta R.T. 0.012 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion:107 Resp: 254314
Ion Ratio Lower Upper
107 100
121 49.3 39.6 59.4
122 86.6 68.7 103.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.539 ng/ul
 RT: 9.704 min Scan# 1142
 Delta R.T. -0.000 min
 Lab File: BM049518.D
 Acq: 30 Jan 2025 00:48




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020083


Tgt Ion: 93 Resp: 348459
 Ion Ratio Lower Upper
 93 100
 95 32.3 25.0 37.6
 123 10.7 8.6 13.0



#29
 2,4-Dichlorophenol
 Concen: 21.176 ng/ul
 RT: 9.945 min Scan# 1183
 Delta R.T. 0.012 min
 Lab File: BM049518.D
 Acq: 30 Jan 2025 00:48

Tgt Ion: 162 Resp: 268727
 Ion Ratio Lower Upper
 162 100
 164 66.1 53.7 80.5
 98 32.7 27.2 40.8





#30

Naphthalene

Concen: 20.668 ng/ul

RT: 10.316 min Scan# 1246

Delta R.T. -0.006 min

Lab File: BM049518.D

Acq: 30 Jan 2025 00:48

Instrument :

BNA_M

ClientSampleId :

SSTD020083





Tgt Ion:128 Resp: 815755

Ion Ratio Lower Upper

128 100

129 11.3 8.8 13.2

127 13.2 10.6 15.8



#32

4-Chloroaniline

Concen: 20.304 ng/ul

RT: 10.445 min Scan# 1268

Delta R.T. -0.000 min

Lab File: BM049518.D




Acq: 30 Jan 2025 00:48


Tgt Ion:127 Resp: 314594

Ion Ratio Lower Upper

127 100

129 33.1 25.4 38.2





#33

Hexachlorobutadiene

Concen: 20.231 ng/ul

RT: 10.610 min Scan# 1296

Delta R.T. -0.000 min

Lab File: BM049518.D

Acq: 30 Jan 2025 00:48

Instrument :

BNA_M

ClientSampleId :

SSTD020083





Tgt Ion:225 Resp: 192459

Ion Ratio Lower Upper

225 100

223 65.3 50.1 75.1

227 64.5 50.2 75.4



#34

Caprolactam

Concen: 23.969 ng/ul

RT: 11.216 min Scan# 1399

Delta R.T. -0.000 min

Lab File: BM049518.D

Acq: 30 Jan 2025 00:48




Tgt Ion:113 Resp: 83905


Ion Ratio Lower Upper

113 100

55 187.7 146.0 219.0

56 140.0 107.8 161.8









#35
4-Chloro-3-methylphenol
Concen: 22.363 ng/ul
RT: 11.604 min Scan# 1465
Delta R.T. 0.023 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48




Instrument :
BNA_M
ClientSampleId :
SSTD020083


Tgt Ion:107 Resp: 245194
Ion Ratio Lower Upper
107 100
144 27.4 22.6 34.0
142 86.4 67.5 101.3



#36
2-Methylnaphthalene
Concen: 21.171 ng/ul
RT: 11.939 min Scan# 1522
Delta R.T. -0.006 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion:142 Resp: 569909
Ion Ratio Lower Upper
142 100
141 86.3 70.4 105.6





#37

1-Methylnaphthalene

Concen: 21.205 ng/ul

RT: 12.163 min Scan# 1560

Delta R.T. -0.000 min

Lab File: BM049518.D

Acq: 30 Jan 2025 00:48

Instrument :

BNA_M

ClientSampleId :

SSTD020083





Tgt Ion:142 Resp: 562721

Ion Ratio Lower Upper

142 100

141 91.2 72.8 109.2

116 3.3 3.0 4.6



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.959 ng/ul

RT: 12.316 min Scan# 1586

Delta R.T. -0.006 min

Lab File: BM049518.D

Acq: 30 Jan 2025 00:48

Tgt Ion:216 Resp: 357408




Ion Ratio Lower Upper


216 100

214 79.6 61.7 92.5

179 19.3 15.4 23.2

108 14.2 11.2 16.8









#40
Hexachlorocyclopentadiene
Concen: 19.569 ng/ul
RT: 12.298 min Scan# 1583
Delta R.T. -0.006 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48




Instrument :
BNA_M
ClientSampleId :
SSTD020083


Tgt Ion:237 Resp: 205575
Ion Ratio Lower Upper
237 100
235 61.9 50.1 75.1
272 13.9 10.6 15.8



#41
2,4,6-Trichlorophenol
Concen: 20.990 ng/ul
RT: 12.575 min Scan# 1630
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion:196 Resp: 221668
Ion Ratio Lower Upper
196 100
198 98.8 74.9 112.3
200 31.7 24.3 36.5









#42
2,4,5-Trichlorophenol
Concen: 21.296 ng/ul
RT: 12.669 min Scan# 1646
Delta R.T. 0.023 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48




Instrument :
BNA_M
ClientSampleId :
SSTD020083


Tgt Ion:196 Resp: 236363
Ion Ratio Lower Upper
196 100
198 95.4 79.4 119.0
200 30.0 25.1 37.7



#43
1,1'-Biphenyl
Concen: 20.113 ng/ul
RT: 12.975 min Scan# 1698
Delta R.T. -0.006 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion:154 Resp: 741325
Ion Ratio Lower Upper
154 100
153 39.5 31.7 47.5
76 12.1 9.8 14.8









#44
2-Chloronaphthalene
Concen: 20.512 ng/ul
RT: 13.010 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48




Instrument :
BNA_M
ClientSampleId :
SSTD020083


Tgt Ion:162 Resp: 612943
Ion Ratio Lower Upper
162 100
164 32.4 26.5 39.7
127 33.3 26.8 40.2



#45
2-Nitroaniline
Concen: 21.000 ng/ul
RT: 13.233 min Scan# 1742
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion: 65 Resp: 157643
Ion Ratio Lower Upper
65 100
92 66.1 52.6 79.0
138 101.4 79.6 119.4









#47
Dimethylphthalate
Concen: 20.478 ng/ul
RT: 13.622 min Scan# 1808
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48




Instrument :
BNA_M
ClientSampleId :
SSTD020083


Tgt Ion:163 Resp: 724586
Ion Ratio Lower Upper
163 100
194 4.2 3.3 4.9
164 10.3 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 21.748 ng/ul
RT: 13.739 min Scan# 1828
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion:165 Resp: 138398
Ion Ratio Lower Upper
165 100
89 51.2 44.5 66.7
121 19.0 16.3 24.5









#50
Acenaphthylene
Concen: 20.727 ng/ul
RT: 13.869 min Scan# 1850
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48




Instrument :
BNA_M
ClientSampleId :
SSTD020083


Tgt Ion:152 Resp: 891172
Ion Ratio Lower Upper
152 100
151 19.8 16.0 24.0
153 13.3 10.4 15.6



#51
3-Nitroaniline
Concen: 21.677 ng/ul
RT: 14.074 min Scan# 1885
Delta R.T. 0.006 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion:138 Resp: 137027
Ion Ratio Lower Upper
138 100
108 10.6 9.0 13.6
92 111.7 93.7 140.5









#52
Acenaphthene
Concen: 20.511 ng/ul
RT: 14.216 min Scan# 1909
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48




Instrument :
BNA_M
ClientSampleId :
SSTD020083


Tgt Ion:153 Resp: 629612
Ion Ratio Lower Upper
153 100
152 47.9 39.0 58.4
154 89.8 71.5 107.3



#53
2,4-Dinitrophenol
Concen: 21.385 ng/ul
RT: 14.280 min Scan# 1920
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion:184 Resp: 84146
Ion Ratio Lower Upper
184 100
63 58.3 51.0 76.6
154 56.9 49.8 74.6









#55
4-Nitrophenol
Concen: 21.038 ng/ul
RT: 14.439 min Scan# 1939
Delta R.T. 0.047 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48




Instrument :
BNA_M
ClientSampleId :
SSTD020083


Tgt Ion:109 Resp: 89943
Ion Ratio Lower Upper
109 100
139 149.8 115.4 173.2
65 149.1 118.6 178.0



#56
Dibenzofuran
Concen: 20.939 ng/ul
RT: 14.551 min Scan# 1966
Delta R.T. -0.006 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion:168 Resp: 916490
Ion Ratio Lower Upper
168 100
139 33.5 27.4 41.0





#57

2,4-Dinitrotoluene

Concen: 22.316 ng/ul

RT: 14.533 min Scan# 1963

Delta R.T. -0.000 min

Lab File: BM049518.D

Acq: 30 Jan 2025 00:48

Instrument :

BNA_M

ClientSampleId :

SSTD020083





Tgt Ion:165 Resp: 212364

Ion Ratio Lower Upper

165 100

63 41.2 35.5 53.3

182 2.8 1.9 2.9



#58

2,3,4,6-Tetrachlorophenol

Concen: 21.760 ng/ul

RT: 14.798 min Scan# 2008

Delta R.T. 0.012 min

Lab File: BM049518.D

Acq: 30 Jan 2025 00:48

Tgt Ion:232 Resp: 196685




Ion Ratio Lower Upper


232 100

131 36.4 32.8 49.2

130 1.7 1.5 2.3

166 27.5 24.3 36.5





#59

Diethylphthalate

Concen: 20.912 ng/ul

RT: 14.998 min Scan# 2042

Delta R.T. -0.000 min

Lab File: BM049518.D

Acq: 30 Jan 2025 00:48

Instrument :

BNA_M

ClientSampleId :

SSTD020083





Tgt Ion:149 Resp: 702409

Ion Ratio Lower Upper

149 100

177 22.7 18.3 27.5

150 12.2 9.9 14.9



#61

Fluorene

Concen: 21.066 ng/ul

RT: 15.210 min Scan# 2078

Delta R.T. -0.000 min

Lab File: BM049518.D

Acq: 30 Jan 2025 00:48




Tgt Ion:166 Resp: 745075


Ion Ratio Lower Upper

166 100

165 95.9 78.5 117.7

167 12.9 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 20.917 ng/ul
RT: 15.210 min Scan# 2078
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48




Instrument :
BNA_M
ClientSampleId :
SSTD020083


Tgt Ion:204 Resp: 401285
Ion Ratio Lower Upper
204 100
206 32.9 27.0 40.6
141 47.7 41.2 61.8



#63
4-Nitroaniline
Concen: 21.744 ng/ul
RT: 15.239 min Scan# 2083
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion:138 Resp: 126923
Ion Ratio Lower Upper
138 100
92 52.1 42.7 64.1
108 73.7 63.9 95.9









#66
4,6-Dinitro-2-methylphenol
Concen: 18.711 ng/ul
RT: 15.298 min Scan# 2092
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48




Instrument :
BNA_M
ClientSampleId :
SSTD020083


Tgt Ion:198 Resp: 132902
Ion Ratio Lower Upper
198 100
182 5.5 4.2 6.4
77 20.9 17.7 26.5



#67
N-Nitrosodiphenylamine
Concen: 19.937 ng/ul
RT: 15.427 min Scan# 2115
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion:169 Resp: 611203
Ion Ratio Lower Upper
169 100
168 66.6 55.1 82.7
167 36.1 29.9 44.9









#68
4-Bromophenyl-phenylether
Concen: 20.170 ng/ul
RT: 16.104 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48




Instrument :
BNA_M
ClientSampleId :
SSTD020083


Tgt Ion:248 Resp: 236854
Ion Ratio Lower Upper
248 100
250 98.9 75.9 113.9
141 59.6 43.8 65.8



#69
Hexachlorobenzene
Concen: 20.257 ng/ul
RT: 16.210 min Scan# 2248
Delta R.T. -0.006 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion:284 Resp: 274688
Ion Ratio Lower Upper
284 100
142 31.5 24.4 36.6
249 30.5 23.4 35.2





#70

Atrazine

Concen: 19.988 ng/ul

RT: 16.392 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM049518.D

Acq: 30 Jan 2025 00:48

Instrument :

BNA_M

ClientSampleId :

SSTD020083





Tgt Ion:200 Resp: 241194

Ion Ratio Lower Upper

200 100

173 24.9 19.8 29.8

215 50.2 38.7 58.1



#71

Pentachlorophenol

Concen: 20.561 ng/ul

RT: 16.568 min Scan# 2309

Delta R.T. 0.006 min

Lab File: BM049518.D

Acq: 30 Jan 2025 00:48




Tgt Ion:266 Resp: 175557


Ion Ratio Lower Upper

266 100

264 61.4 50.5 75.7

268 64.3 52.0 78.0





#72

Phenanthrene

Concen: 20.404 ng/ul

RT: 16.945 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM049518.D

Acq: 30 Jan 2025 00:48

Instrument :

BNA_M

ClientSampleId :

SSTD020083





Tgt Ion:178 Resp: 1181186

Ion Ratio Lower Upper

178 100

179 15.3 12.4 18.6

176 19.4 15.6 23.4



#74

Anthracene

Concen: 20.668 ng/ul

RT: 17.033 min Scan# 2388

Delta R.T. -0.006 min

Lab File: BM049518.D

Acq: 30 Jan 2025 00:48




Tgt Ion:178 Resp: 1205160


Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.4

176 18.5 15.0 22.4





#75

1,2,3,4-Tetrachlorobenzene

Concen: 19.030 ng/uL

RT: 12.933 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM049518.D

Acq: 30 Jan 2025 00:48

Instrument :

BNA_M

ClientSampleId :

SSTD020083





Tgt Ion:216 Resp: 353711

Ion Ratio Lower Upper

216 100

214 76.5 62.5 93.7

218 48.2 38.4 57.6



#76

Pentachlorobenzene

Concen: 19.499 ng/uL

RT: 14.474 min Scan# 1953

Delta R.T. -0.000 min

Lab File: BM049518.D

Acq: 30 Jan 2025 00:48




Tgt Ion:250 Resp: 346554


Ion Ratio Lower Upper

250 100

248 62.9 50.8 76.2

252 62.9 51.7 77.5









#77
 Carbazole
 Concen: 20.548 ng/ul
 RT: 17.315 min Scan# 2436
 Delta R.T. -0.000 min
 Lab File: BM049518.D
 Acq: 30 Jan 2025 00:48




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020083


Tgt Ion:167 Resp: 1070846
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.2 25.8
 139 12.0 9.3 13.9



#78
 Di-n-butylphthalate
 Concen: 20.252 ng/ul
 RT: 17.892 min Scan# 2534
 Delta R.T. -0.006 min
 Lab File: BM049518.D
 Acq: 30 Jan 2025 00:48

Tgt Ion:149 Resp: 1205461
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.2 10.8
 104 4.5 3.8 5.6









#80
Fluoranthene
Concen: 19.813 ng/ul
RT: 18.974 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48




Instrument :
BNA_M
ClientSampleId :
SSTD020083


Tgt Ion:202 Resp: 1433185
Ion Ratio Lower Upper
202 100
101 8.4 6.9 10.3
100 6.1 5.1 7.7



#82
Pyrene
Concen: 19.895 ng/ul
RT: 19.339 min Scan# 2780
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion:202 Resp: 1548449
Ion Ratio Lower Upper
202 100
101 10.0 8.3 12.5
100 8.0 6.6 9.8





#83

Butylbenzylphthalate

Concen: 20.833 ng/ul

RT: 20.262 min Scan# 2938

Delta R.T. -0.006 min

Lab File: BM049518.D

Acq: 30 Jan 2025 00:48

Instrument :

BNA_M

ClientSampleId :

SSTD020083





Tgt Ion:149 Resp: 546076

Ion Ratio Lower Upper

149 100

91 68.6 55.3 82.9

206 22.1 17.8 26.8



#84

3,3'-Dichlorobenzidine

Concen: 22.422 ng/ul

RT: 21.056 min Scan# 3072

Delta R.T. -0.000 min

Lab File: BM049518.D

Acq: 30 Jan 2025 00:48




Tgt Ion:252 Resp: 461218


Ion Ratio Lower Upper

252 100

254 63.8 51.1 76.7

126 8.7 7.4 11.0









#85
Benzo(a)anthracene
Concen: 20.558 ng/ul
RT: 21.115 min Scan# 3082
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48




Instrument :
BNA_M
ClientSampleId :
SSTD020083


Tgt Ion:228 Resp: 1564537
Ion Ratio Lower Upper
228 100
229 19.9 15.7 23.5
226 27.6 22.1 33.1



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.000 ng/ul
RT: 21.068 min Scan# 3074
Delta R.T. -0.006 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion:149 Resp: 808792
Ion Ratio Lower Upper
149 100
167 28.6 23.0 34.4
279 4.2 3.3 4.9





#87

Chrysene

Concen: 20.788 ng/ul

RT: 21.174 min Scan# 3092

Delta R.T. -0.000 min

Lab File: BM049518.D

Acq: 30 Jan 2025 00:48

Instrument :

BNA_M

ClientSampleId :

SSTD020083





Tgt Ion:228 Resp: 1469150

Ion Ratio Lower Upper

228 100

226 30.0 24.6 37.0

229 19.5 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 19.763 ng/ul




RT: 22.127 min Scan# 3254


Delta R.T. -0.006 min

Lab File: BM049518.D

Acq: 30 Jan 2025 00:48

Tgt Ion:149 Resp: 1360108









#90
Benzo(b)fluoranthene
Concen: 20.523 ng/ul
RT: 23.039 min Scan# 3410
Delta R.T. -0.006 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48




Instrument :
BNA_M
ClientSampleId :
SSTD020083


Tgt Ion:252 Resp: 1496652
Ion Ratio Lower Upper
252 100
253 21.6 17.6 26.4
125 7.4 6.3 9.5



#91
Benzo(k)fluoranthene
Concen: 20.379 ng/ul
RT: 23.097 min Scan# 3419
Delta R.T. -0.000 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion:252 Resp: 1501750
Ion Ratio Lower Upper
252 100
253 22.0 17.5 26.3
125 7.5 6.5 9.7





#93

Benzo(a)pyrene

Concen: 20.527 ng/ul

RT: 23.780 min Scan# 3

Delta R.T. -0.000 min

Lab File: BM049518.D

Acq: 30 Jan 2025 00:48

Instrument :

BNA_M

ClientSampleId :

SSTD020083





Tgt Ion:252 Resp: 1404273

Ion Ratio Lower Upper

252 100

253 21.8 17.3 25.9

125 8.4 7.0 10.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.314 ng/ul

RT: 26.991 min Scan# 4081

Delta R.T. -0.000 min

Lab File: BM049518.D

Acq: 30 Jan 2025 00:48




Tgt Ion:276 Resp: 1722529


Ion Ratio Lower Upper

276 100

138 15.8 13.8 20.6

277 24.0 19.0 28.6









#95
Dibenzo(a,h)anthracene
Concen: 20.168 ng/ul
RT: 27.038 min Scan# 4090
Delta R.T. -0.012 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Instrument :
BNA_M
ClientSampleId :
SSTD020083

Tgt Ion:278 Resp: 1385752
Ion Ratio Lower Upper
278 100
139 12.3 10.5 15.7
279 24.2 19.3 28.9



#96
Benzo(g,h,i)perylene
Concen: 20.361 ng/ul
RT: 27.944 min Scan# 4243
Delta R.T. -0.006 min
Lab File: BM049518.D
Acq: 30 Jan 2025 00:48

Tgt Ion:276 Resp: 1312681
Ion Ratio Lower Upper
276 100
138 17.2 14.1 21.1
277 23.5 19.1 28.7

