

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP013025\
 Data File : BP023793.D
 Acq On : 30 Jan 2025 10:10
 Operator : RC/JU
 Sample : PB166284BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Jan 30 11:11:14 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP012925.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 29 17:13:20 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.787	152	469568	20.000	ng/ul	0.00
20) Naphthalene-d8	10.563	136	1851006	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.410	164	1213266	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.204	188	2595293	20.000	ng/ul	0.00
79) Chrysene-d12	21.639	240	2805405	20.000	ng/ul	-0.04
88) Perylene-d12	25.021	264	3114365	20.000	ng/ul	-0.07
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.299	96	64543	5.711	ng/uL	0.00
4) Pyridine-d5	3.710	84	907047	27.814	ng/ul	0.00
7) Phenol-d5	6.975	99	1111039	26.736	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.122	67	571488	25.982	ng/ul	0.00
11) 2-Chlorophenol-d4	7.328	132	904580	27.720	ng/ul	0.00
15) 4-Methylphenol-d8	8.493	113	880784	25.946	ng/ul	0.00
21) Nitrobenzene-d5	8.928	128	398452	26.906	ng/ul	0.00
24) 2-Nitrophenol-d4	9.645	143	420557	26.319	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.192	165	793897	27.082	ng/ul	0.00
31) 4-Chloroaniline-d4	10.692	131	1092919	24.930	ng/ul	0.00
46) Dimethylphthalate-d6	13.816	166	2564845	28.341	ng/ul	0.00
49) Acenaphthylene-d8	14.098	160	2784726	27.283	ng/ul	0.00
54) 4-Nitrophenol-d4	14.627	143	446247	24.830	ng/ul	-0.01
60) Fluorene-d10	15.422	176	2146006	28.159	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.527	200	325255	20.315	ng/ul	-0.01
73) Anthracene-d10	17.304	188	3455994	27.121	ng/ul	-0.01
81) Pyrene-d10	19.668	212	4159993	27.675	ng/ul	-0.02
92) Benzo(a)pyrene-d12	24.786	264	4554983	28.032	ng/ul	-0.07

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP013025\
Data File : BP023793.D
Acq On : 30 Jan 2025 10:10
Operator : RC/JU
Sample : PB166284BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Quant Time: Jan 30 11:11:14 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP012925.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jan 29 17:13:20 2025
Response via : Initial Calibration

