

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012825\
 Data File : BM049484.D
 Acq On : 29 Jan 2025 01:15
 Operator : RC/JU
 Sample : PB166287BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS287

Quant Time: Jan 29 07:34:15 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM012825.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 28 14:45:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.510	152	197355	20.000	ng/u1	0.00
20) Naphthalene-d8	10.269	136	747094	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.151	164	499676	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.904	188	991503	20.000	ng/u1	0.00
79) Chrysene-d12	21.139	240	962735	20.000	ng/u1	0.00
88) Perylene-d12	23.915	264	932860	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.122	96	28071	5.270	ng/uL	0.00
4) Pyridine-d5	3.516	84	444942	28.431	ng/u1	0.00
7) Phenol-d5	6.710	99	499333	31.179	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.863	67	314446	28.714	ng/u1	0.00
11) 2-Chlorophenol-d4	7.057	132	394547	31.837	ng/u1	0.00
15) 4-Methylphenol-d8	8.228	113	369996	31.533	ng/u1	0.00
21) Nitrobenzene-d5	8.657	128	166125	29.066	ng/u1	0.00
24) 2-Nitrophenol-d4	9.369	143	188456	30.229	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.904	165	405805	31.898	ng/u1	0.00
31) 4-Chloroaniline-d4	10.416	131	428303	26.246	ng/u1	0.00
46) Dimethylphthalate-d6	13.575	166	1085693	29.634	ng/u1	0.00
49) Acenaphthylene-d8	13.839	160	1257425	29.211	ng/u1	0.00
54) 4-Nitrophenol-d4	14.380	143	173896	28.934	ng/u1	0.00
60) Fluorene-d10	15.151	176	960900	29.824	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.280	200	176357	28.226	ng/u1	0.00
73) Anthracene-d10	17.004	188	1475803	29.807	ng/u1	0.00
81) Pyrene-d10	19.309	212	1682277	29.891	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.727	264	1511205	30.373	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.152	88	57537	10.279	ng/uL	96
5) Pyridine	3.534	79	459919	28.656	ng/u1	98
6) Benzaldehyde	6.675	77	26227	2.982	ng/u1	96
8) Phenol	6.734	94	525511	31.263	ng/u1	99
10) Bis(2-Chloroethyl)ether	6.957	93	394401	28.772	ng/u1	97
12) 2-Chlorophenol	7.087	128	413886	32.341	ng/u1	99
13) 2-Methylphenol	7.963	108	368553	31.287	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.034	45	633098	29.983	ng/u1	100
16) Acetophenone	8.328	105	574187	29.427	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.322	70	320642	31.333	ng/u1	99
18) 4-Methylphenol	8.287	108	404678	31.830	ng/u1	99
19) Hexachloroethane	8.569	117	163746	29.121	ng/u1	97
22) Nitrobenzene	8.698	77	447677	29.325	ng/u1	99
23) Isophorone	9.222	82	833901	30.548	ng/u1	99
25) 2-Nitrophenol	9.398	139	207579	29.988	ng/u1	96
26) 2,4-Dimethylphenol	9.475	107	402291	31.581	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.704	93	511186	29.899	ng/u1	98
29) 2,4-Dichlorophenol	9.934	162	408596	31.949	ng/u1	98
30) Naphthalene	10.322	128	1152593	28.977	ng/u1	99
32) 4-Chloroaniline	10.439	127	230869	14.786	ng/u1	99
33) Hexachlorobutadiene	10.610	225	275442	28.731	ng/u1	98
34) Caprolactam	11.216	113	101234	28.696	ng/u1	88
35) 4-Chloro-3-methylphenol	11.581	107	361278	32.697	ng/u1	98
36) 2-Methylnaphthalene	11.945	142	820276	30.237	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012825\
 Data File : BM049484.D
 Acq On : 29 Jan 2025 01:15
 Operator : RC/JU
 Sample : PB166287BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS287

Quant Time: Jan 29 07:34:15 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM012825.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 28 14:45:18 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.163	142	818313	30.598	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.322	216	510734	27.729	ng/ul	99
40) Hexachlorocyclopentadiene	12.298	237	266882	24.699	ng/ul	100
41) 2,4,6-Trichlorophenol	12.569	196	330517	30.428	ng/ul	98
42) 2,4,5-Trichlorophenol	12.645	196	355601	31.149	ng/ul	98
43) 1,1'-Biphenyl	12.975	154	1080559	28.502	ng/ul	99
44) 2-Chloronaphthalene	13.010	162	866339	28.186	ng/ul	99
45) 2-Nitroaniline	13.233	65	243497	31.536	ng/ul	98
47) Dimethylphthalate	13.622	163	1098358	30.179	ng/ul	100
48) 2,6-Dinitrotoluene	13.739	165	208973	31.926	ng/ul	98
50) Acenaphthylene	13.869	152	1286487	29.090	ng/ul	99
51) 3-Nitroaniline	14.069	138	193841	29.813	ng/ul	97
52) Acenaphthene	14.216	153	917999	29.075	ng/ul	98
53) 2,4-Dinitrophenol	14.274	184	109816	27.133	ng/ul	97
55) 4-Nitrophenol	14.392	109	133719	30.409	ng/ul	97
56) Dibenzofuran	14.557	168	1319049	29.299	ng/ul	99
57) 2,4-Dinitrotoluene	14.533	165	313288	32.007	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	14.786	232	293773	31.598	ng/ul	99
59) Diethylphthalate	15.004	149	1064612	30.814	ng/ul	98
61) Fluorene	15.210	166	1099774	30.231	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.210	204	586686	29.732	ng/ul	97
63) 4-Nitroaniline	15.239	138	212819	35.446	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.298	198	195542	28.554	ng/ul	99
67) N-Nitrosodiphenylamine	15.427	169	893693	30.237	ng/ul	99
68) 4-Bromophenyl-phenylether	16.104	248	344664	30.443	ng/ul	95
69) Hexachlorobenzene	16.216	284	390514	29.871	ng/ul	99
70) Atrazine	16.392	200	322007	27.678	ng/ul	98
71) Pentachlorophenol	16.563	266	240243	29.184	ng/ul	99
72) Phenanthrene	16.945	178	1673211	29.979	ng/ul	100
74) Anthracene	17.039	178	1693122	30.117	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	12.939	216	499933	27.897	ng/ul	99
76) Pentachlorobenzene	14.474	250	476972	27.836	ng/ul	99
77) Carbazole	17.315	167	1481076	29.477	ng/ul	99
78) Di-n-butylphthalate	17.892	149	1804494	31.443	ng/ul	99
80) Fluoranthene	18.974	202	1952247	30.371	ng/ul	100
82) Pyrene	19.339	202	2066646	29.880	ng/ul	99
83) Butylbenzylphthalate	20.268	149	730591	31.365	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.056	252	535403	29.289	ng/ul	99
85) Benzo(a)anthracene	21.121	228	2021394	29.890	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.074	149	1108264	32.381	ng/ul	100
87) Chrysene	21.174	228	1863922	29.678	ng/ul	99
89) Di-n-octyl phthalate	22.133	149	1717430	30.097	ng/ul	100
90) Benzo(b)fluoranthene	23.044	252	1854538	30.670	ng/ul	100
91) Benzo(k)fluoranthene	23.103	252	1846940	30.228	ng/ul	99
93) Benzo(a)pyrene	23.786	252	1708286	30.116	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.997	276	2154102	30.639	ng/ul	99
95) Dibenzo(a,h)anthracene	27.050	278	1758218	30.862	ng/ul	100
96) Benzo(g,h,i)perylene	27.956	276	1624040	30.381	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012825\
 Data File : BM049484.D
 Acq On : 29 Jan 2025 01:15
 Operator : RC/JU
 Sample : PB166287BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS287

Quant Time: Jan 29 07:34:15 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM012825.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 28 14:45:18 2025
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 10.279 ng/uL
RT: 3.152 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15




Instrument :
BNA_M
ClientSampleId :
SLCS287


Tgt Ion: 88 Resp: 57537
Ion Ratio Lower Upper
88 100
43 30.7 23.4 35.0
58 74.8 57.3 85.9



#5
Pyridine
Concen: 28.656 ng/uL
RT: 3.534 min Scan# 93
Delta R.T. 0.006 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15

Tgt Ion: 79 Resp: 459919
Ion Ratio Lower Upper
79 100
52 74.1 60.7 91.1
51 35.3 29.3 43.9









#6
Benzaldehyde
Concen: 2.982 ng/ul
RT: 6.675 min Scan# 61
Delta R.T. 0.006 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15




Instrument :
BNA_M
ClientSampleId :
SLCS287


Tgt Ion: 77 Resp: 26227
Ion Ratio Lower Upper
77 100
105 90.8 74.2 111.4
106 85.9 72.7 109.1



#8
Phenol
Concen: 31.263 ng/ul
RT: 6.734 min Scan# 637
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15

Tgt Ion: 94 Resp: 525511
Ion Ratio Lower Upper
94 100
65 28.7 23.4 35.0
66 35.1 27.8 41.8









#10
Bis(2-Chloroethyl)ether
Concen: 28.772 ng/ul
RT: 6.957 min Scan# 61
Delta R.T. 0.006 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15




Instrument :
BNA_M
ClientSampleId :
SLCS287


Tgt Ion: 93 Resp: 394401
Ion Ratio Lower Upper
93 100
63 79.6 66.3 99.5
95 32.0 25.0 37.6



#12
2-Chlorophenol
Concen: 32.341 ng/ul
RT: 7.087 min Scan# 697
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15

Tgt Ion: 128 Resp: 413886
Ion Ratio Lower Upper
128 100
64 49.3 40.3 60.5
130 33.1 26.2 39.4









#13
2-Methylphenol
Concen: 31.287 ng/ul
RT: 7.963 min Scan# 846
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15




Instrument :
BNA_M
ClientSampleId :
SLCS287


Tgt Ion:108 Resp: 368553
Ion Ratio Lower Upper
108 100
107 87.6 69.7 104.5



#14
2,2'-oxybis(1-Chloropropane)
Concen: 29.983 ng/ul
RT: 8.034 min Scan# 858
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15

Tgt Ion: 45 Resp: 633098
Ion Ratio Lower Upper
45 100
77 13.2 10.6 15.8
79 9.5 7.8 11.6









#16
Acetophenone
Concen: 29.427 ng/ul
RT: 8.328 min Scan# 908
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15




Instrument :
BNA_M
ClientSampleId :
SLCS287


Tgt Ion:105 Resp: 574187
Ion Ratio Lower Upper
105 100
77 79.4 64.6 97.0
51 28.0 23.3 34.9



#17
N-Nitroso-di-n-propylamine
Concen: 31.333 ng/ul
RT: 8.322 min Scan# 907
Delta R.T. 0.006 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15

Tgt Ion: 70 Resp: 320642
Ion Ratio Lower Upper
70 100
42 58.0 47.0 70.6
101 8.7 7.2 10.8
130 18.8 14.6 21.8









#18
4-Methylphenol
Concen: 31.830 ng/ul
RT: 8.287 min Scan# 901
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15




Instrument :
BNA_M
ClientSampleId :
SLCS287


Tgt Ion:108 Resp: 404678
Ion Ratio Lower Upper
108 100
107 113.6 90.0 135.0



#19
Hexachloroethane
Concen: 29.121 ng/ul
RT: 8.569 min Scan# 949
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15

Tgt Ion:117 Resp: 163746
Ion Ratio Lower Upper
117 100
201 112.8 94.9 142.3
199 72.0 57.3 85.9









#22
Nitrobenzene
Concen: 29.325 ng/ul
RT: 8.698 min Scan# 971
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15




Instrument :
BNA_M
ClientSampleId :
SLCS287


Tgt Ion: 77 Resp: 447677
Ion Ratio Lower Upper
77 100
123 42.9 33.9 50.9
65 13.3 10.6 16.0



#23
Isophorone
Concen: 30.548 ng/ul
RT: 9.222 min Scan# 1060
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15

Tgt Ion: 82 Resp: 833901
Ion Ratio Lower Upper
82 100
95 6.2 5.1 7.7
138 16.9 13.3 19.9









#25
2-Nitrophenol
Concen: 29.988 ng/ul
RT: 9.398 min Scan# 1090
Delta R.T. -0.006 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15




Instrument :
BNA_M
ClientSampleId :
SLCS287


Tgt Ion:139 Resp: 207579
Ion Ratio Lower Upper
139 100
65 53.4 40.6 60.8
109 27.4 20.7 31.1



#26
2,4-Dimethylphenol
Concen: 31.581 ng/ul
RT: 9.475 min Scan# 1103
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15

Tgt Ion:107 Resp: 402291
Ion Ratio Lower Upper
107 100
121 49.3 39.6 59.4
122 87.0 68.7 103.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 29.899 ng/ul
 RT: 9.704 min Scan# 1143
 Delta R.T. -0.000 min
 Lab File: BM049484.D
 Acq: 29 Jan 2025 01:15




Instrument :
 BNA_M
 ClientSampleId :
 SLCS287


Tgt Ion: 93 Resp: 511186
 Ion Ratio Lower Upper
 93 100
 95 32.5 25.0 37.6
 123 10.4 8.6 13.0



#29
 2,4-Dichlorophenol
 Concen: 31.949 ng/ul
 RT: 9.934 min Scan# 1181
 Delta R.T. -0.000 min
 Lab File: BM049484.D
 Acq: 29 Jan 2025 01:15

Tgt Ion:162 Resp: 408596
 Ion Ratio Lower Upper
 162 100
 164 65.3 53.7 80.5
 98 33.7 27.2 40.8









#30
Naphthalene
Concen: 28.977 ng/ul
RT: 10.322 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15




Instrument :
BNA_M
ClientSampleId :
SLCS287


Tgt Ion:128 Resp: 1152593
Ion Ratio Lower Upper
128 100
129 10.6 8.8 13.2
127 12.9 10.6 15.8



#32
4-Chloroaniline
Concen: 14.786 ng/ul
RT: 10.439 min Scan# 1267
Delta R.T. -0.006 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15

Tgt Ion:127 Resp: 230869
Ion Ratio Lower Upper
127 100
129 32.5 25.4 38.2









#33
Hexachlorobutadiene
Concen: 28.731 ng/ul
RT: 10.610 min Scan# 1296
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15




Instrument :
BNA_M
ClientSampleId :
SLCS287


Tgt Ion:225 Resp: 275442
Ion Ratio Lower Upper
225 100
223 61.3 50.1 75.1
227 64.5 50.2 75.4



#34
Caprolactam
Concen: 28.696 ng/ul
RT: 11.216 min Scan# 1399
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15

Tgt Ion:113 Resp: 101234
Ion Ratio Lower Upper
113 100
55 196.7 146.0 219.0
56 151.3 107.8 161.8









#35
4-Chloro-3-methylphenol
Concen: 32.697 ng/ul
RT: 11.581 min Scan# 1461
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15




Instrument :
BNA_M
ClientSampleId :
SLCS287


Tgt Ion:107 Resp: 361278
Ion Ratio Lower Upper
107 100
144 27.7 22.6 34.0
142 86.0 67.5 101.3



#36
2-Methylnaphthalene
Concen: 30.237 ng/ul
RT: 11.945 min Scan# 1523
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15

Tgt Ion:142 Resp: 820276
Ion Ratio Lower Upper
142 100
141 86.8 70.4 105.6





#37

1-Methylnaphthalene

Concen: 30.598 ng/ul

RT: 12.163 min Scan# 1560

Delta R.T. -0.000 min

Lab File: BM049484.D

Acq: 29 Jan 2025 01:15

Instrument :

BNA_M

ClientSampleId :

SLCS287





Tgt Ion:142 Resp: 818313

Ion Ratio Lower Upper

142 100

141 90.8 72.8 109.2

116 3.5 3.0 4.6



#39

1,2,4,5-Tetrachlorobenzene

Concen: 27.729 ng/ul

RT: 12.322 min Scan# 1587

Delta R.T. -0.000 min

Lab File: BM049484.D

Acq: 29 Jan 2025 01:15

Tgt Ion:216 Resp: 510734




Ion Ratio Lower Upper


216 100

214 77.9 61.7 92.5

179 19.8 15.4 23.2

108 13.6 11.2 16.8









#40
Hexachlorocyclopentadiene
Concen: 24.699 ng/ul
RT: 12.298 min Scan# 1583
Delta R.T. -0.006 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15




Instrument :
BNA_M
ClientSampleId :
SLCS287


Tgt Ion:237 Resp: 266882
Ion Ratio Lower Upper
237 100
235 62.4 50.1 75.1
272 12.7 10.6 15.8



#41
2,4,6-Trichlorophenol
Concen: 30.428 ng/ul
RT: 12.569 min Scan# 1629
Delta R.T. -0.006 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15

Tgt Ion:196 Resp: 330517
Ion Ratio Lower Upper
196 100
198 95.8 74.9 112.3
200 30.6 24.3 36.5









#42
2,4,5-Trichlorophenol
Concen: 31.149 ng/ul
RT: 12.645 min Scan# 1642
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15




Instrument :
BNA_M
ClientSampleId :
SLCS287


Tgt Ion:196 Resp: 355601
Ion Ratio Lower Upper
196 100
198 96.3 79.4 119.0
200 31.5 25.1 37.7



#43
1,1'-Biphenyl
Concen: 28.502 ng/ul
RT: 12.975 min Scan# 1698
Delta R.T. -0.006 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15

Tgt Ion:154 Resp: 1080559
Ion Ratio Lower Upper
154 100
153 39.0 31.7 47.5
76 12.5 9.8 14.8









#44
2-Chloronaphthalene
Concen: 28.186 ng/ul
RT: 13.010 min Scan# 1705
Delta R.T. -0.006 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15




Instrument :
BNA_M
ClientSampleId :
SLCS287


Tgt Ion:162 Resp: 866339
Ion Ratio Lower Upper
162 100
164 33.8 26.5 39.7
127 33.9 26.8 40.2



#45
2-Nitroaniline
Concen: 31.536 ng/ul
RT: 13.233 min Scan# 1742
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15

Tgt Ion: 65 Resp: 243497
Ion Ratio Lower Upper
65 100
92 65.9 52.6 79.0
138 103.1 79.6 119.4









#47
Dimethylphthalate
Concen: 30.179 ng/ul
RT: 13.622 min Scan# 1808
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15




Instrument :
BNA_M
ClientSampleId :
SLCS287


Tgt Ion:163 Resp: 1098358
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.2 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 31.926 ng/ul
RT: 13.739 min Scan# 1828
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15

Tgt Ion:165 Resp: 208973
Ion Ratio Lower Upper
165 100
89 54.7 44.5 66.7
121 18.9 16.3 24.5









#50
Acenaphthylene
Concen: 29.090 ng/ul
RT: 13.869 min Scan# 1850
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15




Instrument :
BNA_M
ClientSampleId :
SLCS287


Tgt Ion:152 Resp: 1286487
Ion Ratio Lower Upper
152 100
151 19.5 16.0 24.0
153 12.9 10.4 15.6



#51
3-Nitroaniline
Concen: 29.813 ng/ul
RT: 14.069 min Scan# 1884
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15

Tgt Ion:138 Resp: 193841
Ion Ratio Lower Upper
138 100
108 9.5 9.0 13.6
92 113.8 93.7 140.5





#52

Acenaphthene

Concen: 29.075 ng/ul

RT: 14.216 min Scan# 1909

Delta R.T. -0.000 min

Lab File: BM049484.D

Acq: 29 Jan 2025 01:15

Instrument :

BNA_M

ClientSampleId :

SLCS287





Tgt Ion:153 Resp: 917999

Ion Ratio Lower Upper

153 100

152 49.2 39.0 58.4

154 91.3 71.5 107.3



#53

2,4-Dinitrophenol

Concen: 27.133 ng/ul

RT: 14.274 min Scan# 1919

Delta R.T. -0.006 min

Lab File: BM049484.D

Acq: 29 Jan 2025 01:15




Tgt Ion:184 Resp: 109816


Ion Ratio Lower Upper

184 100

63 65.1 51.0 76.6

154 58.6 49.8 74.6









#55
4-Nitrophenol
Concen: 30.409 ng/ul
RT: 14.392 min Scan# 1939
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15




Instrument :
BNA_M
ClientSampleId :
SLCS287


Tgt Ion:109 Resp: 133719
Ion Ratio Lower Upper
109 100
139 139.2 115.4 173.2
65 145.9 118.6 178.0



#56
Dibenzofuran
Concen: 29.299 ng/ul
RT: 14.557 min Scan# 1967
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15

Tgt Ion:168 Resp: 1319049
Ion Ratio Lower Upper
168 100
139 33.7 27.4 41.0









#57
2,4-Dinitrotoluene
Concen: 32.007 ng/ul
RT: 14.533 min Scan# 1963
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15




Instrument :
BNA_M
ClientSampleId :
SLCS287


Tgt Ion:165 Resp: 313288
Ion Ratio Lower Upper
165 100
63 43.6 35.5 53.3
182 2.5 1.9 2.9



#58
2,3,4,6-Tetrachlorophenol
Concen: 31.598 ng/ul
RT: 14.786 min Scan# 2006
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15

Tgt Ion:232 Resp: 293773
Ion Ratio Lower Upper
232 100
131 40.1 32.8 49.2
130 2.0 1.5 2.3
166 29.8 24.3 36.5





#59

Diethylphthalate

Concen: 30.814 ng/ul

RT: 15.004 min Scan# 2043

Delta R.T. 0.006 min

Lab File: BM049484.D

Acq: 29 Jan 2025 01:15

Instrument :

BNA_M

ClientSampleId :

SLCS287





Tgt Ion:149 Resp: 1064612

Ion Ratio Lower Upper

149 100

177 24.1 18.3 27.5

150 12.0 9.9 14.9



#61

Fluorene

Concen: 30.231 ng/ul

RT: 15.210 min Scan# 2078

Delta R.T. -0.000 min

Lab File: BM049484.D

Acq: 29 Jan 2025 01:15




Tgt Ion:166 Resp: 1099774


Ion Ratio Lower Upper

166 100

165 96.6 78.5 117.7

167 13.4 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 29.732 ng/ul
RT: 15.210 min Scan# 2078
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15




Instrument :
BNA_M
ClientSampleId :
SLCS287


Tgt Ion:204 Resp: 586686
Ion Ratio Lower Upper
204 100
206 32.4 27.0 40.6
141 49.0 41.2 61.8



#63
4-Nitroaniline
Concen: 35.446 ng/ul
RT: 15.239 min Scan# 2083
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15

Tgt Ion:138 Resp: 212819
Ion Ratio Lower Upper
138 100
92 50.6 42.7 64.1
108 76.1 63.9 95.9









#66
 4,6-Dinitro-2-methylphenol
 Concen: 28.554 ng/ul
 RT: 15.298 min Scan# 2093
 Delta R.T. -0.000 min
 Lab File: BM049484.D
 Acq: 29 Jan 2025 01:15




Instrument :
 BNA_M
 ClientSampleId :
 SLCS287


Tgt Ion:198 Resp: 195542
 Ion Ratio Lower Upper
 198 100
 182 4.9 4.2 6.4
 77 21.6 17.7 26.5



#67
 N-Nitrosodiphenylamine
 Concen: 30.237 ng/ul
 RT: 15.427 min Scan# 2115
 Delta R.T. -0.000 min
 Lab File: BM049484.D
 Acq: 29 Jan 2025 01:15

Tgt Ion:169 Resp: 893693
 Ion Ratio Lower Upper
 169 100
 168 68.3 55.1 82.7
 167 37.3 29.9 44.9









#68
4-Bromophenyl-phenylether
Concen: 30.443 ng/ul
RT: 16.104 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15




Instrument :
BNA_M
ClientSampleId :
SLCS287


Tgt Ion:248 Resp: 344664
Ion Ratio Lower Upper
248 100
250 96.2 75.9 113.9
141 63.4 43.8 65.8



#69
Hexachlorobenzene
Concen: 29.871 ng/ul
RT: 16.216 min Scan# 2249
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15

Tgt Ion:284 Resp: 390514
Ion Ratio Lower Upper
284 100
142 29.5 24.4 36.6
249 29.7 23.4 35.2





#70

Atrazine

Concen: 27.678 ng/ul

RT: 16.392 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM049484.D

Acq: 29 Jan 2025 01:15

Instrument :

BNA_M

ClientSampleId :

SLCS287





Tgt Ion:200 Resp: 322007

Ion Ratio Lower Upper

200 100

173 25.3 19.8 29.8

215 50.0 38.7 58.1



#71

Pentachlorophenol

Concen: 29.184 ng/ul

RT: 16.563 min Scan# 2308

Delta R.T. -0.000 min

Lab File: BM049484.D

Acq: 29 Jan 2025 01:15




Tgt Ion:266 Resp: 240243


Ion Ratio Lower Upper

266 100

264 62.3 50.5 75.7

268 64.9 52.0 78.0





#72

Phenanthrene

Concen: 29.979 ng/u1

RT: 16.945 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM049484.D

Acq: 29 Jan 2025 01:15

Instrument :

BNA_M

ClientSampleId :

SLCS287





Tgt Ion:178 Resp: 1673211

Ion Ratio Lower Upper

178 100

179 15.3 12.4 18.6

176 19.6 15.6 23.4



#74

Anthracene

Concen: 30.117 ng/u1

RT: 17.039 min Scan# 2389

Delta R.T. -0.000 min

Lab File: BM049484.D

Acq: 29 Jan 2025 01:15




Tgt Ion:178 Resp: 1693122


Ion Ratio Lower Upper

178 100

179 15.2 12.2 18.4

176 18.4 15.0 22.4





#75

1,2,3,4-Tetrachlorobenzene

Concen: 27.897 ng/uL

RT: 12.939 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM049484.D

Acq: 29 Jan 2025 01:15

Instrument :

BNA_M

ClientSampleId :

SLCS287





Tgt Ion:216 Resp: 499933

Ion Ratio Lower Upper

216 100

214 79.0 62.5 93.7

218 48.8 38.4 57.6



#76

Pentachlorobenzene

Concen: 27.836 ng/uL

RT: 14.474 min Scan# 1953

Delta R.T. -0.000 min

Lab File: BM049484.D

Acq: 29 Jan 2025 01:15




Tgt Ion:250 Resp: 476972


Ion Ratio Lower Upper

250 100

248 62.9 50.8 76.2

252 64.3 51.7 77.5









#77
 Carbazole
 Concen: 29.477 ng/ul
 RT: 17.315 min Scan# 2436
 Delta R.T. -0.000 min
 Lab File: BM049484.D
 Acq: 29 Jan 2025 01:15

Instrument :
 BNA_M
 ClientSampleId :
 SLCS287

Tgt Ion:167 Resp: 1481076




Ion	Ratio	Lower	Upper
167	100		
166	21.2	17.2	25.8
139	12.1	9.3	13.9




#78
 Di-n-butylphthalate
 Concen: 31.443 ng/ul
 RT: 17.892 min Scan# 2534
 Delta R.T. -0.006 min
 Lab File: BM049484.D
 Acq: 29 Jan 2025 01:15

Tgt Ion:149 Resp: 1804494

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.2	10.8
104	4.9	3.8	5.6









#80
 Fluoranthene
 Concen: 30.371 ng/ul
 RT: 18.974 min Scan# 2718
 Delta R.T. -0.000 min
 Lab File: BM049484.D
 Acq: 29 Jan 2025 01:15




Instrument :
 BNA_M
 ClientSampleId :
 SLCS287


Tgt Ion:202 Resp: 1952247
 Ion Ratio Lower Upper
 202 100
 101 8.8 6.9 10.3
 100 6.5 5.1 7.7



#82
 Pyrene
 Concen: 29.880 ng/ul
 RT: 19.339 min Scan# 2780
 Delta R.T. -0.000 min
 Lab File: BM049484.D
 Acq: 29 Jan 2025 01:15

Tgt Ion:202 Resp: 2066646
 Ion Ratio Lower Upper
 202 100
 101 10.5 8.3 12.5
 100 8.7 6.6 9.8





#83

Butylbenzylphthalate

Concen: 31.365 ng/ul

RT: 20.268 min Scan# 2938

Delta R.T. -0.000 min

Lab File: BM049484.D

Acq: 29 Jan 2025 01:15

Instrument :

BNA_M

ClientSampleId :

SLCS287





Tgt Ion:149 Resp: 730591

Ion Ratio Lower Upper

149 100

91 67.4 55.3 82.9

206 22.1 17.8 26.8



#84

3,3'-Dichlorobenzidine

Concen: 29.289 ng/ul

RT: 21.056 min Scan# 3072

Delta R.T. -0.000 min

Lab File: BM049484.D

Acq: 29 Jan 2025 01:15




Tgt Ion:252 Resp: 535403


Ion Ratio Lower Upper

252 100

254 64.5 51.1 76.7

126 9.0 7.4 11.0





#85

Benzo(a)anthracene

Concen: 29.890 ng/ul

RT: 21.121 min Scan# 3083

Delta R.T. 0.006 min

Lab File: BM049484.D

Acq: 29 Jan 2025 01:15

Instrument :

BNA_M

ClientSampleId :

SLCS287





Tgt Ion:228 Resp: 2021394

Ion Ratio Lower Upper

228 100

229 19.2 15.7 23.5

226 27.3 22.1 33.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 32.381 ng/ul

RT: 21.074 min Scan# 3075

Delta R.T. -0.000 min

Lab File: BM049484.D

Acq: 29 Jan 2025 01:15




Tgt Ion:149 Resp: 1108264


Ion Ratio Lower Upper

149 100

167 28.8 23.0 34.4

279 4.3 3.3 4.9





#87

Chrysene

Concen: 29.678 ng/ul

RT: 21.174 min Scan# 3092

Delta R.T. -0.000 min

Lab File: BM049484.D

Acq: 29 Jan 2025 01:15

Instrument :

BNA_M

ClientSampleId :

SLCS287





Tgt Ion:228 Resp: 1863922

Ion Ratio Lower Upper

228 100

226 30.1 24.6 37.0

229 19.5 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 30.097 ng/ul




RT: 22.133 min Scan# 3255


Delta R.T. -0.000 min

Lab File: BM049484.D

Acq: 29 Jan 2025 01:15

Tgt Ion:149 Resp: 1717430









#90
Benzo(b)fluoranthene
Concen: 30.670 ng/ul
RT: 23.044 min Scan# 3410
Delta R.T. -0.000 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15




Instrument :
BNA_M
ClientSampleId :
SLCS287


Tgt Ion:252 Resp: 1854538
Ion Ratio Lower Upper
252 100
253 21.9 17.6 26.4
125 8.2 6.3 9.5



#91
Benzo(k)fluoranthene
Concen: 30.228 ng/ul
RT: 23.103 min Scan# 3420
Delta R.T. 0.006 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15

Tgt Ion:252 Resp: 1846940
Ion Ratio Lower Upper
252 100
253 21.6 17.5 26.3
125 7.8 6.5 9.7









#93
Benzo(a)pyrene
Concen: 30.116 ng/u1
RT: 23.786 min Scan# 3
Delta R.T. 0.006 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15




Instrument :
BNA_M
ClientSampleId :
SLCS287


Tgt Ion:252 Resp: 1708286
Ion Ratio Lower Upper
252 100
253 22.2 17.3 25.9
125 9.0 7.0 10.6



#94
Indeno(1,2,3-cd)pyrene
Concen: 30.639 ng/u1
RT: 26.997 min Scan# 4082
Delta R.T. 0.006 min
Lab File: BM049484.D
Acq: 29 Jan 2025 01:15

Tgt Ion:276 Resp: 2154102
Ion Ratio Lower Upper
276 100
138 16.6 13.8 20.6
277 24.2 19.0 28.6









#95
 Dibenzo(a,h)anthracene
 Concen: 30.862 ng/ul
 RT: 27.050 min Scan# 4090
 Delta R.T. -0.000 min
 Lab File: BM049484.D
 Acq: 29 Jan 2025 01:15

Instrument :
 BNA_M
 ClientSampleId :
 SLCS287

Tgt Ion:278 Resp: 1758218
 Ion Ratio Lower Upper
 278 100
 139 13.1 10.5 15.7
 279 24.5 19.3 28.9



#96
 Benzo(g,h,i)perylene
 Concen: 30.381 ng/ul
 RT: 27.956 min Scan# 4245
 Delta R.T. 0.006 min
 Lab File: BM049484.D
 Acq: 29 Jan 2025 01:15

Tgt Ion:276 Resp: 1624040
 Ion Ratio Lower Upper
 276 100
 138 17.6 14.1 21.1
 277 23.9 19.1 28.7

