


Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA\_M\Data\BM012918\  
Data File : BM013954.D  
Acq On : 29 Jan 2018 12:08  
Operator : SJ/JU  
Sample : DFTPP86  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP86

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA\_M\METHODS\SOM-EPA-BM011018.M  
Title : SVOA CALIBRATION  
Last Update : Tue Jan 30 07:49:34 2018



AutoFind: Scans 2454, 2455, 2456; Background Corrected with Scan 2449

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 40.0      | 78872   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.1       | 873     | PASS             |
| 69          | 198          | 0.00         | 100          | 40.4      | 79736   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.4       | 299     | PASS             |
| 127         | 198          | 10           | 80           | 46.4      | 91554   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.6       | 1137    | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 197354  | PASS             |
| 199         | 198          | 5            | 9            | 7.0       | 13872   | PASS             |
| 275         | 198          | 10           | 60           | 31.6      | 62301   | PASS             |
| 365         | 198          | 1            | 100          | 4.9       | 9623    | PASS             |
| 441         | 443          | 0.01         | 100          | 75.5      | 25010   | PASS             |
| 442         | 198          | 50           | 100          | 86.6      | 170896  | PASS             |
| 443         | 442          | 15           | 24           | 19.4      | 33120   | PASS             |

| m/z    | Abundance |
|--------|-----------|
| 37.00  | 741.0     |
| 38.00  | 1818.0    |
| 39.00  | 7298.0    |
| 40.00  | 584.0     |
| 41.10  | 1239.0    |
| 42.00  | 324.0     |
| 43.00  | 2298.0    |
| 44.00  | 1622.0    |
| 44.90  | 1012.0    |
| 49.10  | 1084.0    |
| 50.00  | 22968.0   |
| 51.10  | 84072.0   |
| 52.00  | 5534.0    |
| 53.20  | 373.0     |
| 55.10  | 820.0     |
| 56.20  | 2468.0    |
| 57.00  | 7462.0    |
| 57.80  | 379.0     |
| 60.00  | 1170.0    |
| 61.10  | 864.0     |
| 62.10  | 1455.0    |
| 63.00  | 2901.0    |
| 64.00  | 803.0     |
| 65.00  | 1545.0    |
| 65.90  | 370.0     |
| 66.80  | 387.0     |
| 68.10  | 1496.0    |
| 69.00  | 82096.0   |
| 70.10  | 1037.0    |
| 71.10  | 470.0     |
| 73.10  | 1835.0    |
| 74.10  | 9398.0    |
| 75.00  | 14294.0   |
| 76.10  | 4736.0    |
| 77.10  | 86768.0   |
| 78.10  | 7124.0    |
| 79.00  | 7942.0    |
| 80.00  | 5752.0    |
| 81.00  | 7641.0    |
| 82.00  | 2018.0    |
| 83.00  | 1231.0    |
| 84.20  | 591.0     |
| 85.00  | 1402.0    |
| 86.10  | 3214.0    |
| 86.90  | 606.0     |
| 91.00  | 1597.0    |
| 92.00  | 1771.0    |
| 93.00  | 14429.0   |
| 94.10  | 1425.0    |
| 95.00  | 301.0     |
| 96.00  | 1288.0    |
| 96.90  | 678.0     |
| 97.30  | 688.0     |
| 98.00  | 10887.0   |
| 99.00  | 8019.0    |
| 99.90  | 459.0     |
| 101.00 | 3361.0    |
| 102.00 | 422.0     |
| 102.90 | 1768.0    |
| 104.00 | 2280.0    |
| 105.00 | 2664.0    |
| 106.10 | 1228.0    |
| 107.10 | 30056.0   |
| 107.90 | 5253.0    |
| 109.20 | 738.0     |
| 110.00 | 45944.0   |
| 111.00 | 7191.0    |
| 111.90 | 1383.0    |
| 112.90 | 308.0     |
| 115.20 | 331.0     |
| 115.90 | 2091.0    |
| 117.00 | 38272.0   |
| 117.90 | 2642.0    |
| 118.90 | 415.0     |
| 120.00 | 776.0     |
| 121.20 | 358.0     |
| 121.90 | 3148.0    |
| 123.00 | 3284.0    |
| 124.00 | 2040.0    |
| 125.00 | 1210.0    |
| 126.10 | 929.0     |
| 127.00 | 84472.0   |
| 128.00 | 7486.0    |
| 129.00 | 36856.0   |
| 130.00 | 4018.0    |
| 131.10 | 644.0     |
| 133.90 | 1349.0    |
| 135.00 | 4053.0    |
| 136.10 | 2057.0    |
| 137.00 | 1439.0    |
| 137.70 | 331.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP86

|        |          |
|--------|----------|
| 138.80 | 374.0    |
| 139.90 | 469.0    |
| 141.00 | 7286.0   |
| 142.10 | 1972.0   |
| 142.90 | 1770.0   |
| 143.90 | 574.0    |
| 145.00 | 422.0    |
| 145.90 | 806.0    |
| 147.10 | 2716.0   |
| 148.00 | 7073.0   |
| 149.00 | 1618.0   |
| 150.30 | 485.0    |
| 150.90 | 651.0    |
| 152.00 | 464.0    |
| 152.80 | 2575.0   |
| 154.10 | 932.0    |
| 154.90 | 2698.0   |
| 156.00 | 3841.0   |
| 157.00 | 715.0    |
| 157.70 | 675.0    |
| 158.10 | 1010.0   |
| 159.00 | 617.0    |
| 160.10 | 1199.0   |
| 161.00 | 2639.0   |
| 161.90 | 833.0    |
| 163.00 | 522.0    |
| 164.90 | 2536.0   |
| 166.00 | 1349.0   |
| 167.00 | 10871.0  |
| 168.00 | 5620.0   |
| 169.10 | 1050.0   |
| 169.80 | 603.0    |
| 171.00 | 648.0    |
| 172.00 | 672.0    |
| 172.90 | 1221.0   |
| 174.10 | 1814.0   |
| 174.90 | 4928.0   |
| 176.10 | 1993.0   |
| 177.00 | 1867.0   |
| 178.10 | 1655.0   |
| 179.00 | 9944.0   |
| 180.00 | 5088.0   |
| 181.00 | 3228.0   |
| 182.10 | 350.0    |
| 182.90 | 368.0    |
| 183.90 | 711.0    |
| 185.00 | 3763.0   |
| 186.00 | 30760.0  |
| 187.00 | 6529.0   |
| 188.00 | 1019.0   |
| 188.90 | 2500.0   |
| 190.20 | 492.0    |
| 190.90 | 877.0    |
| 191.90 | 2072.0   |
| 192.90 | 2651.0   |
| 193.90 | 588.0    |
| 195.90 | 3794.0   |
| 197.10 | 1067.0   |
| 197.90 | 172352.0 |
| 198.90 | 13313.0  |
| 200.10 | 1532.0   |
| 201.40 | 1184.0   |
| 202.00 | 491.0    |
| 203.00 | 2332.0   |
| 204.10 | 6845.0   |
| 205.10 | 12858.0  |
| 206.00 | 53632.0  |
| 207.10 | 7706.0   |
| 207.90 | 1876.0   |
| 208.90 | 600.0    |
| 210.00 | 1283.0   |
| 210.90 | 2747.0   |
| 211.70 | 473.0    |
| 213.00 | 408.0    |
| 215.00 | 1098.0   |
| 215.90 | 2508.0   |
| 217.00 | 16744.0  |
| 217.90 | 2243.0   |
| 219.00 | 480.0    |
| 221.00 | 6533.0   |
| 221.80 | 1801.0   |
| 223.00 | 3617.0   |
| 224.00 | 29480.0  |
| 225.10 | 6744.0   |
| 227.00 | 13627.0  |
| 227.90 | 1620.0   |
| 229.00 | 3319.0   |
| 230.00 | 465.0    |
| 231.00 | 1101.0   |
| 232.90 | 684.0    |
| 234.00 | 1158.0   |
| 234.90 | 1077.0   |
| 235.80 | 722.0    |
| 236.90 | 1294.0   |
| 238.80 | 523.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP86

|        |          |
|--------|----------|
| 240.00 | 778.0    |
| 241.10 | 943.0    |
| 242.10 | 1486.0   |
| 243.10 | 1669.0   |
| 244.00 | 22568.0  |
| 245.00 | 3098.0   |
| 245.90 | 5224.0   |
| 246.80 | 600.0    |
| 249.00 | 1472.0   |
| 251.90 | 322.0    |
| 253.00 | 759.0    |
| 254.10 | 1021.0   |
| 255.00 | 120344.0 |
| 256.00 | 18544.0  |
| 256.90 | 1631.0   |
| 258.00 | 7819.0   |
| 258.80 | 726.0    |
| 259.30 | 835.0    |
| 260.00 | 335.0    |
| 261.10 | 361.0    |
| 264.90 | 2876.0   |
| 265.90 | 416.0    |
| 271.00 | 581.0    |
| 273.10 | 4688.0   |
| 274.00 | 8228.0   |
| 275.00 | 51248.0  |
| 275.90 | 6927.0   |
| 276.90 | 4539.0   |
| 278.10 | 1180.0   |
| 279.00 | 512.0    |
| 281.00 | 377.0    |
| 282.10 | 480.0    |
| 283.00 | 562.0    |
| 284.20 | 412.0    |
| 284.90 | 707.0    |
| 293.10 | 1416.0   |
| 293.90 | 562.0    |
| 295.00 | 455.0    |
| 296.00 | 19984.0  |
| 297.10 | 3259.0   |
| 301.20 | 355.0    |
| 301.90 | 321.0    |
| 303.10 | 1901.0   |
| 304.00 | 896.0    |
| 309.10 | 428.0    |
| 313.90 | 327.0    |
| 315.00 | 1574.0   |
| 316.00 | 1529.0   |
| 321.00 | 701.0    |
| 323.10 | 5467.0   |
| 324.10 | 656.0    |
| 327.00 | 1451.0   |
| 329.00 | 425.0    |
| 331.90 | 638.0    |
| 332.90 | 398.0    |
| 333.90 | 3189.0   |
| 335.10 | 786.0    |
| 341.10 | 698.0    |
| 342.10 | 435.0    |
| 346.00 | 1273.0   |
| 352.10 | 1483.0   |
| 353.00 | 1001.0   |
| 354.00 | 1440.0   |
| 365.00 | 5939.0   |
| 366.10 | 1218.0   |
| 371.00 | 457.0    |
| 371.90 | 2063.0   |
| 372.90 | 712.0    |
| 382.90 | 655.0    |
| 391.10 | 568.0    |
| 392.10 | 522.0    |
| 402.10 | 995.0    |
| 403.10 | 1324.0   |
| 404.20 | 362.0    |
| 421.20 | 937.0    |
| 421.80 | 779.0    |
| 423.10 | 6170.0   |
| 424.10 | 1622.0   |
| 441.10 | 15816.0  |
| 442.10 | 111088.0 |
| 443.10 | 23104.0  |
| 444.00 | 2454.0   |
| 461.10 | 336.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP86

| m/z    | Abundance |
|--------|-----------|
| 38.00  | 1997.0    |
| 39.00  | 6695.0    |
| 39.90  | 529.0     |
| 41.00  | 1165.0    |
| 42.00  | 503.0     |
| 43.00  | 2108.0    |
| 44.10  | 1415.0    |
| 44.90  | 1394.0    |
| 49.10  | 789.0     |
| 50.00  | 29832.0   |
| 51.10  | 93184.0   |
| 52.00  | 4572.0    |
| 55.10  | 1470.0    |
| 56.00  | 2341.0    |
| 57.00  | 7566.0    |
| 57.90  | 512.0     |
| 60.00  | 1440.0    |
| 61.00  | 1573.0    |
| 62.00  | 1467.0    |
| 63.00  | 4221.0    |
| 63.90  | 622.0     |
| 65.00  | 1878.0    |
| 66.90  | 353.0     |
| 67.90  | 1474.0    |
| 69.00  | 94216.0   |
| 70.00  | 744.0     |
| 71.20  | 355.0     |
| 73.00  | 2844.0    |
| 74.00  | 12944.0   |
| 75.00  | 17040.0   |
| 76.10  | 5651.0    |
| 77.10  | 110152.0  |
| 78.10  | 6876.0    |
| 79.10  | 9901.0    |
| 79.90  | 6561.0    |
| 81.00  | 7514.0    |
| 82.10  | 1815.0    |
| 83.10  | 1579.0    |
| 84.00  | 425.0     |
| 85.00  | 1372.0    |
| 86.00  | 3140.0    |
| 87.10  | 1662.0    |
| 88.00  | 526.0     |
| 88.90  | 438.0     |
| 91.10  | 2503.0    |
| 92.10  | 2454.0    |
| 93.00  | 17664.0   |
| 94.00  | 1354.0    |
| 94.90  | 511.0     |
| 96.10  | 805.0     |
| 97.20  | 739.0     |
| 98.00  | 15806.0   |
| 99.00  | 9060.0    |
| 99.90  | 854.0     |
| 101.00 | 4846.0    |
| 103.00 | 2062.0    |
| 104.00 | 3832.0    |
| 105.00 | 3384.0    |
| 106.10 | 774.0     |
| 107.00 | 35520.0   |
| 107.90 | 5244.0    |
| 108.90 | 1013.0    |
| 110.00 | 59880.0   |
| 110.90 | 9897.0    |
| 112.00 | 1470.0    |
| 112.90 | 449.0     |
| 115.00 | 307.0     |
| 116.00 | 3574.0    |
| 117.00 | 42656.0   |
| 118.00 | 3645.0    |
| 118.80 | 387.0     |
| 120.00 | 509.0     |
| 122.00 | 1855.0    |
| 123.00 | 6026.0    |
| 124.00 | 2535.0    |
| 125.00 | 1662.0    |
| 125.70 | 476.0     |
| 126.20 | 472.0     |
| 127.00 | 113200.0  |
| 128.00 | 7853.0    |
| 129.00 | 49744.0   |
| 130.10 | 5251.0    |
| 130.90 | 687.0     |
| 132.00 | 558.0     |
| 132.90 | 354.0     |
| 133.90 | 1439.0    |
| 135.00 | 3328.0    |
| 135.90 | 2047.0    |
| 137.10 | 2176.0    |
| 138.30 | 412.0     |
| 140.10 | 715.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP86

|        |          |
|--------|----------|
| 140.90 | 6416.0   |
| 142.10 | 1913.0   |
| 143.00 | 1848.0   |
| 144.00 | 612.0    |
| 146.00 | 1549.0   |
| 147.00 | 3388.0   |
| 148.00 | 10425.0  |
| 149.00 | 2268.0   |
| 150.00 | 766.0    |
| 151.00 | 881.0    |
| 151.90 | 605.0    |
| 153.00 | 1828.0   |
| 154.00 | 1791.0   |
| 155.00 | 3709.0   |
| 156.00 | 7338.0   |
| 156.80 | 642.0    |
| 157.30 | 683.0    |
| 157.90 | 1668.0   |
| 158.90 | 1125.0   |
| 160.00 | 1905.0   |
| 161.00 | 3059.0   |
| 161.80 | 670.0    |
| 162.90 | 701.0    |
| 164.20 | 347.0    |
| 165.00 | 2352.0   |
| 166.00 | 2571.0   |
| 167.00 | 15126.0  |
| 168.10 | 6700.0   |
| 169.00 | 1410.0   |
| 169.90 | 325.0    |
| 170.90 | 622.0    |
| 171.80 | 1283.0   |
| 173.00 | 1298.0   |
| 173.90 | 2695.0   |
| 175.00 | 4466.0   |
| 176.00 | 1134.0   |
| 177.10 | 2325.0   |
| 178.10 | 1547.0   |
| 179.00 | 11490.0  |
| 179.90 | 7047.0   |
| 181.00 | 3710.0   |
| 183.90 | 947.0    |
| 185.10 | 4980.0   |
| 186.00 | 39960.0  |
| 187.00 | 11263.0  |
| 187.90 | 1549.0   |
| 189.00 | 2461.0   |
| 189.90 | 513.0    |
| 191.00 | 1339.0   |
| 192.10 | 3657.0   |
| 193.00 | 3851.0   |
| 193.80 | 619.0    |
| 194.90 | 357.0    |
| 196.00 | 8607.0   |
| 197.00 | 1567.0   |
| 197.90 | 241728.0 |
| 198.90 | 16608.0  |
| 200.10 | 1342.0   |
| 200.80 | 549.0    |
| 201.60 | 750.0    |
| 203.00 | 2852.0   |
| 204.00 | 10938.0  |
| 205.00 | 18512.0  |
| 206.00 | 65576.0  |
| 207.00 | 11479.0  |
| 208.00 | 4145.0   |
| 208.80 | 909.0    |
| 209.90 | 1427.0   |
| 211.00 | 2553.0   |
| 211.80 | 1023.0   |
| 213.20 | 375.0    |
| 214.90 | 637.0    |
| 216.00 | 1875.0   |
| 217.00 | 21064.0  |
| 217.90 | 3026.0   |
| 220.10 | 375.0    |
| 220.90 | 7922.0   |
| 222.00 | 2987.0   |
| 222.90 | 4517.0   |
| 224.00 | 45496.0  |
| 225.10 | 10170.0  |
| 225.90 | 463.0    |
| 227.00 | 20848.0  |
| 228.00 | 2676.0   |
| 229.10 | 3849.0   |
| 230.10 | 464.0    |
| 230.90 | 1539.0   |
| 234.00 | 1971.0   |
| 235.00 | 1439.0   |
| 236.00 | 783.0    |
| 237.00 | 1070.0   |
| 238.90 | 918.0    |
| 240.00 | 1223.0   |
| 241.10 | 1820.0   |
| 242.00 | 2008.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP86

|        |          |
|--------|----------|
| 242.90 | 2570.0   |
| 244.00 | 31136.0  |
| 245.10 | 3901.0   |
| 246.00 | 8294.0   |
| 246.90 | 1442.0   |
| 247.90 | 405.0    |
| 248.90 | 1132.0   |
| 250.80 | 402.0    |
| 252.10 | 317.0    |
| 252.90 | 1397.0   |
| 254.10 | 1608.0   |
| 255.00 | 169152.0 |
| 256.00 | 28672.0  |
| 257.00 | 1894.0   |
| 258.00 | 10877.0  |
| 258.80 | 2128.0   |
| 260.00 | 339.0    |
| 260.80 | 450.0    |
| 264.90 | 5147.0   |
| 266.00 | 873.0    |
| 269.90 | 325.0    |
| 271.00 | 387.0    |
| 271.90 | 456.0    |
| 273.00 | 4915.0   |
| 274.00 | 12848.0  |
| 275.00 | 73120.0  |
| 276.00 | 9861.0   |
| 277.00 | 7928.0   |
| 278.00 | 1029.0   |
| 280.80 | 417.0    |
| 283.00 | 809.0    |
| 283.90 | 768.0    |
| 285.10 | 1673.0   |
| 290.00 | 362.0    |
| 293.10 | 1628.0   |
| 293.80 | 654.0    |
| 296.00 | 24944.0  |
| 297.00 | 2714.0   |
| 301.10 | 328.0    |
| 302.00 | 577.0    |
| 303.00 | 2379.0   |
| 304.00 | 769.0    |
| 307.90 | 339.0    |
| 309.10 | 391.0    |
| 309.90 | 403.0    |
| 313.00 | 306.0    |
| 314.00 | 1385.0   |
| 315.10 | 3815.0   |
| 316.20 | 1613.0   |
| 321.00 | 660.0    |
| 323.00 | 7819.0   |
| 324.20 | 816.0    |
| 326.90 | 1866.0   |
| 328.00 | 585.0    |
| 331.90 | 652.0    |
| 333.20 | 723.0    |
| 334.00 | 4562.0   |
| 335.00 | 1521.0   |
| 340.80 | 776.0    |
| 341.20 | 751.0    |
| 342.00 | 332.0    |
| 346.00 | 1786.0   |
| 352.10 | 1713.0   |
| 352.90 | 859.0    |
| 354.00 | 2362.0   |
| 355.20 | 301.0    |
| 365.10 | 11823.0  |
| 366.10 | 2226.0   |
| 370.90 | 743.0    |
| 372.10 | 2733.0   |
| 373.10 | 1090.0   |
| 383.00 | 977.0    |
| 384.00 | 538.0    |
| 389.80 | 392.0    |
| 391.20 | 481.0    |
| 392.20 | 498.0    |
| 401.00 | 792.0    |
| 402.10 | 1449.0   |
| 403.10 | 1917.0   |
| 403.90 | 824.0    |
| 420.90 | 1349.0   |
| 422.00 | 2136.0   |
| 423.10 | 11252.0  |
| 424.20 | 2342.0   |
| 439.90 | 304.0    |
| 441.00 | 29584.0  |
| 442.10 | 198976.0 |
| 443.00 | 38456.0  |
| 444.00 | 3812.0   |
| 445.20 | 351.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP86

| m/z    | Abundance |
|--------|-----------|
| 38.10  | 822.0     |
| 39.10  | 4836.0    |
| 39.90  | 804.0     |
| 41.00  | 1631.0    |
| 42.30  | 526.0     |
| 43.10  | 2214.0    |
| 44.00  | 1688.0    |
| 45.00  | 1473.0    |
| 49.30  | 445.0     |
| 50.00  | 17792.0   |
| 51.10  | 60992.0   |
| 52.00  | 2808.0    |
| 53.10  | 530.0     |
| 54.90  | 1460.0    |
| 56.00  | 2620.0    |
| 57.00  | 4024.0    |
| 60.10  | 1086.0    |
| 61.00  | 1039.0    |
| 62.00  | 1581.0    |
| 63.00  | 2934.0    |
| 64.00  | 558.0     |
| 65.00  | 1521.0    |
| 65.90  | 402.0     |
| 68.00  | 1563.0    |
| 69.00  | 64496.0   |
| 70.10  | 531.0     |
| 73.10  | 1246.0    |
| 74.00  | 8071.0    |
| 75.00  | 12312.0   |
| 76.00  | 3054.0    |
| 77.10  | 74584.0   |
| 78.10  | 4446.0    |
| 79.00  | 6732.0    |
| 80.00  | 4593.0    |
| 81.00  | 5734.0    |
| 82.00  | 1507.0    |
| 82.90  | 1389.0    |
| 84.20  | 604.0     |
| 85.20  | 901.0     |
| 86.00  | 1665.0    |
| 87.10  | 1146.0    |
| 88.10  | 633.0     |
| 91.10  | 2421.0    |
| 92.10  | 1470.0    |
| 93.00  | 11564.0   |
| 93.90  | 1154.0    |
| 96.20  | 427.0     |
| 97.00  | 391.0     |
| 98.00  | 9409.0    |
| 99.00  | 6171.0    |
| 99.90  | 1109.0    |
| 101.10 | 2753.0    |
| 102.90 | 777.0     |
| 104.00 | 2337.0    |
| 105.00 | 2265.0    |
| 105.90 | 985.0     |
| 107.00 | 28272.0   |
| 108.10 | 3860.0    |
| 109.00 | 1576.0    |
| 110.00 | 39672.0   |
| 110.90 | 5775.0    |
| 112.10 | 1168.0    |
| 112.90 | 393.0     |
| 115.10 | 594.0     |
| 116.20 | 2265.0    |
| 117.00 | 30048.0   |
| 118.00 | 2017.0    |
| 119.30 | 356.0     |
| 120.10 | 451.0     |
| 121.90 | 1684.0    |
| 123.10 | 3709.0    |
| 123.80 | 1137.0    |
| 124.90 | 1207.0    |
| 127.00 | 76992.0   |
| 128.10 | 7475.0    |
| 129.00 | 35576.0   |
| 129.90 | 2241.0    |
| 130.90 | 532.0     |
| 132.00 | 736.0     |
| 133.90 | 1381.0    |
| 135.00 | 2778.0    |
| 135.80 | 1464.0    |
| 137.00 | 1547.0    |
| 137.80 | 453.0     |
| 139.90 | 1141.0    |
| 141.00 | 6176.0    |
| 142.00 | 1597.0    |
| 142.90 | 1086.0    |
| 145.10 | 467.0     |
| 146.00 | 1384.0    |
| 147.00 | 2953.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP86



|        |          |
|--------|----------|
| 148.00 | 7843.0   |
| 148.90 | 1834.0   |
| 150.90 | 539.0    |
| 153.10 | 1971.0   |
| 153.90 | 1238.0   |
| 155.00 | 3437.0   |
| 156.00 | 2453.0   |
| 156.70 | 319.0    |
| 157.20 | 353.0    |
| 157.90 | 1510.0   |
| 159.00 | 495.0    |
| 159.80 | 1491.0   |
| 161.00 | 2787.0   |
| 161.90 | 1342.0   |
| 164.90 | 2503.0   |
| 166.20 | 1692.0   |
| 167.00 | 10537.0  |
| 168.00 | 6399.0   |
| 169.10 | 1390.0   |
| 171.20 | 518.0    |
| 172.00 | 626.0    |
| 173.00 | 985.0    |
| 174.00 | 2220.0   |
| 175.00 | 4134.0   |
| 175.90 | 885.0    |
| 176.90 | 2322.0   |
| 178.00 | 821.0    |
| 179.00 | 7274.0   |
| 179.90 | 5183.0   |
| 181.00 | 2970.0   |
| 182.20 | 533.0    |
| 183.70 | 516.0    |
| 184.10 | 890.0    |
| 185.00 | 5040.0   |
| 186.00 | 27624.0  |
| 187.00 | 8877.0   |
| 187.90 | 1118.0   |
| 189.10 | 2687.0   |
| 190.00 | 521.0    |
| 191.00 | 1598.0   |
| 192.00 | 2390.0   |
| 193.00 | 3707.0   |
| 194.00 | 882.0    |
| 196.00 | 4812.0   |
| 197.00 | 779.0    |
| 198.00 | 177984.0 |
| 199.00 | 11696.0  |
| 199.90 | 675.0    |
| 200.20 | 695.0    |
| 201.20 | 603.0    |
| 203.00 | 1549.0   |
| 204.00 | 7900.0   |
| 205.00 | 13345.0  |
| 206.10 | 47600.0  |
| 207.00 | 10029.0  |
| 208.00 | 2576.0   |
| 209.00 | 797.0    |
| 209.90 | 1010.0   |
| 211.10 | 2246.0   |
| 212.10 | 447.0    |
| 215.00 | 1380.0   |
| 216.00 | 978.0    |
| 216.90 | 19472.0  |
| 218.10 | 2424.0   |
| 221.10 | 6356.0   |
| 221.80 | 1675.0   |
| 222.00 | 1692.0   |
| 223.00 | 3267.0   |
| 224.00 | 33960.0  |
| 225.00 | 9069.0   |
| 226.10 | 714.0    |
| 227.00 | 15356.0  |
| 228.00 | 2618.0   |
| 229.00 | 2763.0   |
| 230.10 | 464.0    |
| 230.80 | 892.0    |
| 233.00 | 692.0    |
| 233.90 | 831.0    |
| 235.00 | 1186.0   |
| 236.00 | 627.0    |
| 236.90 | 1161.0   |
| 239.10 | 404.0    |
| 239.80 | 531.0    |
| 241.00 | 888.0    |
| 242.10 | 1935.0   |
| 243.20 | 1313.0   |
| 244.00 | 25224.0  |
| 245.00 | 3691.0   |
| 245.90 | 5680.0   |
| 247.20 | 1513.0   |
| 249.00 | 1204.0   |
| 250.80 | 358.0    |
| 252.80 | 575.0    |
| 253.90 | 1483.0   |
| 255.00 | 137472.0 |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP86

|        |          |
|--------|----------|
| 256.00 | 17456.0  |
| 257.00 | 1406.0   |
| 258.00 | 10608.0  |
| 259.00 | 1132.0   |
| 260.20 | 439.0    |
| 260.90 | 433.0    |
| 265.10 | 2705.0   |
| 271.20 | 388.0    |
| 272.00 | 511.0    |
| 273.10 | 3087.0   |
| 274.00 | 9343.0   |
| 275.00 | 62536.0  |
| 276.00 | 7322.0   |
| 277.00 | 7150.0   |
| 278.00 | 837.0    |
| 281.10 | 749.0    |
| 282.90 | 809.0    |
| 283.90 | 735.0    |
| 285.20 | 683.0    |
| 289.20 | 473.0    |
| 291.90 | 346.0    |
| 293.10 | 1364.0   |
| 294.90 | 406.0    |
| 296.00 | 20784.0  |
| 297.00 | 3133.0   |
| 301.10 | 492.0    |
| 301.90 | 917.0    |
| 303.10 | 3356.0   |
| 304.00 | 638.0    |
| 314.00 | 550.0    |
| 315.00 | 3362.0   |
| 316.00 | 1154.0   |
| 321.10 | 559.0    |
| 323.00 | 5027.0   |
| 323.90 | 1162.0   |
| 327.00 | 1278.0   |
| 328.00 | 735.0    |
| 331.90 | 574.0    |
| 332.70 | 320.0    |
| 334.00 | 3580.0   |
| 335.20 | 673.0    |
| 341.10 | 1092.0   |
| 346.00 | 1050.0   |
| 352.00 | 1713.0   |
| 353.10 | 1051.0   |
| 354.00 | 2256.0   |
| 355.20 | 353.0    |
| 359.00 | 307.0    |
| 365.00 | 11109.0  |
| 366.00 | 1878.0   |
| 370.00 | 398.0    |
| 370.90 | 643.0    |
| 372.10 | 3562.0   |
| 373.10 | 639.0    |
| 383.00 | 1168.0   |
| 390.10 | 435.0    |
| 401.10 | 427.0    |
| 402.10 | 1250.0   |
| 403.10 | 1947.0   |
| 404.10 | 1081.0   |
| 404.80 | 334.0    |
| 421.10 | 1509.0   |
| 422.20 | 2529.0   |
| 423.10 | 10913.0  |
| 424.00 | 2261.0   |
| 425.00 | 300.0    |
| 441.10 | 29632.0  |
| 442.10 | 202624.0 |
| 443.10 | 37800.0  |
| 444.00 | 4405.0   |
| 445.10 | 375.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP86

| m/z    | Abundance |
|--------|-----------|
| 39.00  | 627.0     |
| 40.10  | 362.0     |
| 41.00  | 1535.0    |
| 41.90  | 630.0     |
| 43.10  | 2809.0    |
| 43.90  | 1330.0    |
| 44.80  | 1092.0    |
| 51.00  | 544.0     |
| 55.10  | 1075.0    |
| 57.10  | 875.0     |
| 60.10  | 793.0     |
| 67.10  | 317.0     |
| 67.90  | 638.0     |
| 69.00  | 533.0     |
| 70.20  | 471.0     |
| 72.90  | 1099.0    |
| 74.10  | 340.0     |
| 77.10  | 385.0     |
| 93.10  | 433.0     |
| 94.80  | 325.0     |
| 96.90  | 485.0     |
| 98.00  | 409.0     |
| 104.90 | 301.0     |
| 107.20 | 332.0     |
| 122.90 | 487.0     |
| 133.00 | 372.0     |
| 147.20 | 710.0     |
| 193.00 | 307.0     |
| 206.90 | 1553.0    |
| 208.20 | 446.0     |
| 209.10 | 332.0     |
| 221.00 | 568.0     |
| 280.90 | 733.0     |
| 296.20 | 476.0     |

Instrument :

BNA\_M

ClientSampleId :

DFTPP86