

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012921\
 Data File : BM028602.D
 Acq On : 29 Jan 2021 12:44
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040

Quant Time: Jan 29 17:52:06 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM012021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 29 16:19:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	27846	20.00	ng	0.00
21) Naphthalene-d8	10.745	136	111524	20.00	ng	# 0.00
39) Acenaphthene-d10	14.586	164	65700	20.00	ng	0.00
64) Phenanthrene-d10	17.321	188	131731	20.00	ng	# 0.00
76) Chrysene-d12	21.468	240	126977	20.00	ng	# 0.00
86) Perylene-d12	23.727	264	125773	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.481	112	130710	74.15	ng	0.00
7) Phenol-d6	7.122	99	182885	76.20	ng	0.00
23) Nitrobenzene-d5	9.110	82	168480	71.65	ng	0.00
42) 2,4,6-Tribromophenol	16.074	330	68677	78.89	ng	0.00
45) 2-Fluorobiphenyl	13.210	172	389697	74.42	ng	0.00
79) Terphenyl-d14	19.921	244	541862	77.06	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.264	88	24524	34.68	ng	# 72
3) Pyridine	3.699	79	72990	37.28	ng	# 57
4) n-Nitrosodimethylamine	3.605	42	34991	31.08	ng	# 65
6) Aniline	7.257	93	101070	38.86	ng	98
8) 2-Chlorophenol	7.504	128	74871	38.20	ng	93
9) Benzaldehyde	7.069	77	44988	36.02	ng	81
10) Phenol	7.146	94	85963	38.00	ng	84
11) bis(2-Chloroethyl)ether	7.357	93	69986	38.37	ng	94
12) 1,3-Dichlorobenzene	7.822	146	86975	37.60	ng	# 91
13) 1,4-Dichlorobenzene	7.969	146	88764	37.95	ng	96
14) 1,2-Dichlorobenzene	8.287	146	84094	37.48	ng	99
15) Benzyl Alcohol	8.187	79	63235	37.81	ng	# 84
16) 2,2'-oxybis(1-Chloropr...	8.475	45	114050	34.64	ng	72
17) 2-Methylphenol	8.398	107	60551	38.66	ng	# 88
18) Hexachloroethane	9.016	117	33086	37.58	ng	93
19) n-Nitroso-di-n-propyla...	8.751	70	55443	38.79	ng	# 69
20) 3+4-Methylphenols	8.734	107	83659	39.82	ng	89
22) Acetophenone	8.769	105	109142	37.75	ng	# 91
24) Nitrobenzene	9.151	77	81010	35.69	ng	# 85
25) Isophorone	9.681	82	142647	39.60	ng	94
26) 2-Nitrophenol	9.863	139	42209	41.09	ng	# 80
27) 2,4-Dimethylphenol	9.934	122	57968	38.81	ng	88
28) bis(2-Chloroethoxy)met...	10.157	93	97364	38.29	ng	99
29) 2,4-Dichlorophenol	10.410	162	70119	39.34	ng	100
30) 1,2,4-Trichlorobenzene	10.610	180	79110	37.37	ng	97
31) Naphthalene	10.798	128	236134	37.47	ng	99
32) Benzoic acid	10.098	122	54575	40.60	ng	# 83
33) 4-Chloroaniline	10.916	127	100628	38.44	ng	# 90
34) Hexachlorobutadiene	11.075	225	47270	37.42	ng	97
35) Caprolactam	11.716	113	22905	39.71	ng	# 68
36) 4-Chloro-3-methylphenol	12.057	107	72050	38.88	ng	93
37) 2-Methylnaphthalene	12.628	142	158247	36.43	ng	93
38) 1-Methylnaphthalene	12.628	142	158247	38.40	ng	99
40) 1,2,4,5-Tetrachloroben...	12.781	216	81068	37.41	ng	99
41) Hexachlorocyclopentadiene	12.751	237	40528	40.03	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012921\
 Data File : BM028602.D
 Acq On : 29 Jan 2021 12:44
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040

Quant Time: Jan 29 17:52:06 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM012021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 29 16:19:07 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	13.028	196	55260	38.02	ng	98
44) 2,4,5-Trichlorophenol	13.110	196	57344	38.12	ng	97
46) 1,1'-Biphenyl	13.422	154	208608	37.21	ng	99
47) 2-Chloronaphthalene	13.469	162	168722	36.82	ng	98
48) 2-Nitroaniline	13.675	65	52247	36.36	ng	88
49) Acenaphthylene	14.316	152	258252	37.98	ng	100
50) Dimethylphthalate	14.051	163	206104	38.16	ng	100
51) 2,6-Dinitrotoluene	14.175	165	44280	38.54	ng	86
52) Acenaphthene	14.651	154	158004	37.43	ng	99
53) 3-Nitroaniline	14.504	138	50405	38.91	ng	86
54) 2,4-Dinitrophenol	14.716	184	32062	46.85	ng	89
55) Dibenzofuran	14.986	168	241748	37.23	ng	98
56) 4-Nitrophenol	14.839	139	38310	37.56	ng	# 74
57) 2,4-Dinitrotoluene	14.963	165	61688	40.30	ng	# 88
58) Fluorene	15.633	166	198568	37.33	ng	98
59) 2,3,4,6-Tetrachlorophenol	15.216	232	51058	38.40	ng	95
60) Diethylphthalate	15.404	149	211491	38.57	ng	98
61) 4-Chlorophenyl-phenyle...	15.627	204	98903	37.75	ng	98
62) 4-Nitroaniline	15.663	138	54502	38.65	ng	# 79
63) Azobenzene	15.916	77	187414	36.41	ng	91
65) 4,6-Dinitro-2-methylph...	15.722	198	32724	39.98	ng	81
66) n-Nitrosodiphenylamine	15.845	169	171118	38.79	ng	99
67) 4-Bromophenyl-phenylether	16.516	248	61468	39.40	ng	95
68) Hexachlorobenzene	16.633	284	69837	38.52	ng	97
69) Atrazine	16.786	200	50084	38.30	ng	98
70) Pentachlorophenol	16.980	266	43234	43.76	ng	98
71) Phenanthrene	17.363	178	302228	37.31	ng	99
72) Anthracene	17.457	178	297407	38.01	ng	99
73) Carbazole	17.727	167	279354	37.64	ng	100
74) Di-n-butylphthalate	18.274	149	372917	41.77	ng	99
75) Fluoranthene	19.368	202	339412	37.30	ng	97
77) Benzidine	19.551	184	117961	41.31	ng	99
78) Pyrene	19.727	202	347177	38.13	ng	99
80) Butylbenzylphthalate	20.609	149	169961	42.34	ng	95
81) Benzo(a)anthracene	21.451	228	329091	37.53	ng	100
82) 3,3'-Dichlorobenzidine	21.386	252	115512	40.99	ng	# 98
83) Chrysene	21.503	228	315671	37.29	ng	100
84) Bis(2-ethylhexyl)phtha...	21.368	149	252696	44.52	ng	# 97
85) Di-n-octyl phthalate	22.250	149	430936	44.91	ng	# 92
87) Indeno(1,2,3-cd)pyrene	26.033	276	369105	39.39	ng	# 90
88) Benzo(b)fluoranthene	23.050	252	329947	38.64	ng	# 97
89) Benzo(k)fluoranthene	23.092	252	320030	38.39	ng	# 96
90) Benzo(a)pyrene	23.633	252	306667	38.76	ng	# 96
91) Dibenzo(a,h)anthracene	26.038	278	307260	39.63	ng	# 94
92) Benzo(g,h,i)perylene	26.733	276	297198	38.97	ng	# 91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012921\
 Data File : BM028602.D
 Acq On : 29 Jan 2021 12:44
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 Client Sample Id :
 SSTDCCC040


Quant Time: Jan 29 17:52:06 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM012021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 29 16:19:07 2021
 Response via : Initial Calibration








1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 7.934 min Scan# 807
 Delta R.T. -0.000 min
 Lab File: BM028602.D
 Acq: 29 Jan 2021 12:44

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040







Tgt Ion:	152	Resp:	27846
Ion	Ratio	Lower	Upper
152	100		
150	150.6	120.6	181.0
115	58.7	47.2	70.8



1,4-Dioxane
 Concen: 34.68 ng
 RT: 3.264 min Scan# 13
 Delta R.T. 0.000 min
 Lab File: BM028602.D
 Acq: 29 Jan 2021 12:44

Tgt Ion:	88	Resp:	24524
Ion	Ratio	Lower	Upper
88	100		
58	79.6	46.2	69.2#
43	35.9	19.0	28.6#









Pyridine
Concen: 37.28 ng
RT: 3.699 min Scan# 87
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:	79	Resp:	72990
Ion	Ratio	Lower	Upper
79	100		
52	78.2	35.8	53.8#
51	41.4	21.0	31.6#



n-Nitrosodimethylamine
Concen: 31.08 ng
RT: 3.605 min Scan# 71
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Tgt Ion:	42	Resp:	34991
Ion	Ratio	Lower	Upper
42	100		
74	112.9	127.8	191.6#
44	6.9	4.1	6.1#









2-Fluorophenol
 Concen: 74.15 ng
 RT: 5.481 min Scan# 390
 Delta R.T. 0.000 min
 Lab File: BM028602.D
 Acq: 29 Jan 2021 12:44




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040


Tgt Ion:	112	Resp:	130710
Ion	Ratio	Lower	Upper
112	100		
64	58.9	50.6	76.0
63	30.6	31.0	46.6#



Aniline
 Concen: 38.86 ng
 RT: 7.257 min Scan# 692
 Delta R.T. 0.000 min
 Lab File: BM028602.D
 Acq: 29 Jan 2021 12:44

Tgt Ion:	93	Resp:	101070
Ion	Ratio	Lower	Upper
93	100		
66	40.5	33.3	49.9
65	20.5	17.4	26.0









Phenol-d6
 Concen: 76.20 ng
 RT: 7.122 min Scan# 669
 Delta R.T. 0.000 min
 Lab File: BM028602.D
 Acq: 29 Jan 2021 12:44




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040


Tgt Ion:	99	Resp:	182885
Ion	Ratio	Lower	Upper
99	100		
42	19.4	13.5	20.3
71	32.9	39.4	59.0#



2-Chlorophenol
 Concen: 38.20 ng
 RT: 7.504 min Scan# 734
 Delta R.T. 0.000 min
 Lab File: BM028602.D
 Acq: 29 Jan 2021 12:44

Tgt Ion:	128	Resp:	74871
Ion	Ratio	Lower	Upper
128	100		
130	32.4	11.5	51.5
64	43.3	30.1	70.1









Benzaldehyde
 Concen: 36.02 ng
 RT: 7.069 min Scan# 660
 Delta R.T. 0.000 min
 Lab File: BM028602.D
 Acq: 29 Jan 2021 12:44




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040


Tgt Ion:	77	Resp:	44988
Ion	Ratio	Lower	Upper
77	100		
105	102.2	64.7	104.7
106	94.0	57.5	97.5



Phenol
 Concen: 38.00 ng
 RT: 7.146 min Scan# 673
 Delta R.T. 0.000 min
 Lab File: BM028602.D
 Acq: 29 Jan 2021 12:44

Tgt Ion:	94	Resp:	85963
Ion	Ratio	Lower	Upper
94	100		
65	28.8	16.4	56.4
66	37.9	30.2	70.2









bis(2-Chloroethyl)ether
 Concen: 38.37 ng
 RT: 7.357 min Scan# 709
 Delta R.T. 0.000 min
 Lab File: BM028602.D
 Acq: 29 Jan 2021 12:44




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040

Tgt Ion:	93	Resp:	69986
Ion	Ratio	Lower	Upper
93	100		
63	80.5	53.1	93.1
95	32.4	12.1	52.1




1,3-Dichlorobenzene
 Concen: 37.60 ng
 RT: 7.822 min Scan# 788
 Delta R.T. 0.000 min
 Lab File: BM028602.D
 Acq: 29 Jan 2021 12:44

Tgt Ion:	146	Resp:	86975
Ion	Ratio	Lower	Upper
146	100		
148	64.1	51.6	77.4
75	28.0	33.0	49.6#





Abundance Scan 830 (8.069 min): BM028463.D\data.ms (-822) (-)




1,4-Dichlorobenzene
Concen: 37.95 ng
RT: 7.969 min Scan# 813
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:	146	Resp:	88764
Ion	Ratio	Lower	Upper
146	100		
148	62.7	51.4	77.0
111	39.2	34.4	51.6



Abundance Scan 813 (7.969 min): BM028602.D\data.ms (-762) (-)





Abundance Scan 884 (8.387 min): BM028463.D\data.ms (-877) (-)




1,2-Dichlorobenzene
Concen: 37.48 ng
RT: 8.287 min Scan# 867
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44


Tgt Ion:	146	Resp:	84094
Ion	Ratio	Lower	Upper
146	100		
148	64.4	51.3	76.9
111	42.7	35.4	53.0



Abundance Scan 867 (8.287 min): BM028602.D\data.ms (-816) (-)






Abundance Scan 865 (8.275 min): BM028463.D\data.ms (-858) #15




Benzyl Alcohol
Concen: 37.81 ng
RT: 8.187 min Scan# 850
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 63235
Ion Ratio Lower Upper
79 100
108 80.1 49.0 73.4#
77 61.7 55.2 82.8






Abundance Scan 915 (8.569 min): BM028463.D\data.ms (-906) #16



2,2'-oxybis(1-Chloropropane)
Concen: 34.64 ng
RT: 8.475 min Scan# 899
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44


Tgt Ion: 45 Resp: 114050
Ion Ratio Lower Upper
45 100
77 13.0 9.5 49.5
79 10.2 1.8 41.8





Abundance Scan 900 (8.481 min): BM028463.D\data.ms (-892) #17

2-Methylphenol
Concen: 38.66 ng
RT: 8.398 min Scan# 886
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44


Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:	107	Resp:	60551
Ion	Ratio	Lower	Upper
107	100		
108	111.8	87.8	131.6
77	44.3	50.9	76.3#
79	44.8	46.6	69.8#




Abundance Scan 886 (8.398 min): BM028602.D\data.ms (-835) (-)





Abundance Scan 1009 (9.122 min): BM028463.D\data.ms (-1002) #18


Hexachloroethane
Concen: 37.58 ng
RT: 9.016 min Scan# 991
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Tgt Ion:	117	Resp:	33086
Ion	Ratio	Lower	Upper
117	100		
119	97.0	73.1	109.7
201	102.3	88.7	133.1



Abundance Scan 991 (9.016 min): BM028602.D\data.ms (-940) (-)









#19
 n-Nitroso-di-n-propylamine
 Concen: 38.79 ng
 RT: 8.751 min Scan# 946
 Delta R.T. 0.000 min
 Lab File: BM028602.D
 Acq: 29 Jan 2021 12:44




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040

Tgt Ion:	70	Resp:	55443
Ion	Ratio	Lower	Upper
70	100		
42	66.3	31.9	47.9#
101	9.8	7.0	10.4
130	24.4	13.4	20.0#



#20
 3+4-Methylphenols
 Concen: 39.82 ng
 RT: 8.734 min Scan# 943
 Delta R.T. 0.000 min
 Lab File: BM028602.D
 Acq: 29 Jan 2021 12:44


Tgt Ion:	107	Resp:	83659
Ion	Ratio	Lower	Upper
107	100		
108	88.5	64.5	104.5
77	31.9	24.2	64.2
79	28.3	17.8	57.8






Abundance Scan 1303 (10.851 min): BM028463.D\data.ms (-1234) (-)

Naphthalene-d8
Concen: 20.00 ng
RT: 10.745 min Scan# 1285
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion	Ratio	Lower	Upper
136	100		
137	10.6	8.6	12.8
54	8.7	4.6	7.0#
68	5.5	4.1	6.1






Abundance Scan 966 (8.869 min): BM028463.D\data.ms (-959) (-)

Acetophenone
Concen: 37.75 ng
RT: 8.769 min Scan# 949
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44






Tgt Ion	Ratio	Lower	Upper
105	100		
71	2.0	7.6	11.4#
51	28.7	18.6	28.0#
120	22.5	17.3	25.9




Abundance Scan 1024 (9.210 min): BM028463.D\data.ms (-1024)

Nitrobenzene-d5
Concen: 71.65 ng
RT: 9.110 min Scan# 1007
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040






Tgt Ion:	82	Resp:	168480
Ion	Ratio	Lower	Upper
82	100		
128	45.0	27.4	41.0#
54	53.0	31.4	47.0#




Abundance Scan 1031 (9.251 min): BM028463.D\data.ms (-1031)


Nitrobenzene
Concen: 35.69 ng
RT: 9.151 min Scan# 1014
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44



Tgt Ion:	77	Resp:	81010
Ion	Ratio	Lower	Upper
77	100		
123	48.3	29.1	43.7#
65	14.3	11.3	16.9






Abundance Scan 1120 (9.775 min): BM028463.D\data.ms (-11#25-)




Isophorone
Concen: 39.60 ng
RT: 9.681 min Scan# 1104
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:	82	Resp:	142647
Ion	Ratio	Lower	Upper
82	100		
95	7.4	6.8	10.2
138	19.4	13.1	19.7






Abundance Scan 1152 (9.863 min): BM028463.D\data.ms (-11#26-)



2-Nitrophenol
Concen: 41.09 ng
RT: 9.863 min Scan# 1135
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44


Tgt Ion:	139	Resp:	42209
Ion	Ratio	Lower	Upper
139	100		
109	27.5	33.8	50.8#
65	49.9	51.0	76.6#






Abundance Scan 1161 (10.016 min): BM028463.D\data.ms (-125) (-)

2,4-Dimethylphenol
Concen: 38.81 ng
RT: 9.934 min Scan# 1147
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:	122	Resp:	57968
Ion	Ratio	Lower	Upper
122	100		
107	113.5	106.6	159.8
121	61.2	48.8	73.2






Abundance Scan 1202 (10.257 min): BM028463.D\data.ms (-125) (-)

bis(2-Chloroethoxy)methane
Concen: 38.29 ng
RT: 10.157 min Scan# 1185
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44




Tgt Ion:	93	Resp:	97364
Ion	Ratio	Lower	Upper
93	100		
95	32.6	26.5	39.7
123	11.5	8.6	12.8





Abundance Scan 1242 (10.492 min): BM028463.D\data.ms (-1229) (-)

2,4-Dichlorophenol
Concen: 39.34 ng
RT: 10.410 min Scan# 1228
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44


Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:	162	Resp:	70119
Ion	Ratio	Lower	Upper
162	100		
164	63.5	43.2	83.2
98	35.8	15.7	55.7




Abundance Scan 1228 (10.410 min): BM028602.D\data.ms (-1177) (-)





Abundance Scan 1279 (10.710 min): BM028463.D\data.ms (-1230) (-)


1,2,4-Trichlorobenzene
Concen: 37.37 ng
RT: 10.610 min Scan# 1262
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44



Tgt Ion:	180	Resp:	79110
Ion	Ratio	Lower	Upper
180	100		
182	94.2	77.7	116.5
145	29.8	23.4	35.0




Abundance Scan 1262 (10.610 min): BM028602.D\data.ms (-1211) (-)





Abundance Scan 1312 (10.904 min): BM028463.D\data.ms (-1281) (-)

Naphthalene
Concen: 37.47 ng
RT: 10.798 min Scan# 1294
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44


Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:	128	Resp:	236134
Ion	Ratio	Lower	Upper
128	100		
129	11.1	9.0	13.4
127	12.8	10.6	15.8






Abundance Scan 1294 (10.798 min): BM028602.D\data.ms (-1243) (-)




Abundance Scan 1185 (10.157 min): BM028463.D\data.ms (-1182) (-)

Benzoic acid
Concen: 40.60 ng
RT: 10.098 min Scan# 1175
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44


Tgt Ion:	122	Resp:	54575
Ion	Ratio	Lower	Upper
122	100		
105	123.7	114.0	154.0
77	93.0	101.8	141.8#



Abundance Scan 1175 (10.098 min): BM028602.D\data.ms (-1124) (-)





Abundance Scan 1330 (11.010 min): BM028463.D\data.ms (-1263) (-)




4-Chloroaniline
Concen: 38.44 ng
RT: 10.916 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
127	100		
129	33.5	24.7	37.1
65	29.0	30.2	45.4#
92	18.2	17.8	26.8



Abundance Scan 1314 (10.916 min): BM028602.D\data.ms (-1263) (-)





Abundance Scan 1359 (11.180 min): BM028463.D\data.ms (-1263) (-)




Hexachlorobutadiene
Concen: 37.42 ng
RT: 11.075 min Scan# 1341
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Tgt Ion	Ratio	Lower	Upper
225	100		
223	61.3	51.0	76.4
227	65.7	50.6	75.8




Abundance Scan 1341 (11.075 min): BM028602.D\data.ms (-1290) (-)






Abundance Scan 1464 (11.798 min): BM028463.D\data.ms (-1435) (-)

Caprolactam
Concen: 39.71 ng
RT: 11.716 min Scan# 1450
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040






Tgt Ion:	113	Resp:	22905
Ion	Ratio	Lower	Upper
113	100		
55	169.7	103.2	143.2#
56	125.3	84.2	124.2#




Abundance Scan 1520 (12.127 min): BM028463.D\data.ms (-1532) (-)

4-Chloro-3-methylphenol
Concen: 38.88 ng
RT: 12.057 min Scan# 1508
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Tgt Ion:	107	Resp:	72050
Ion	Ratio	Lower	Upper
107	100		
144	26.9	20.1	30.1
142	82.4	60.3	90.5






Abundance Scan 1584 (12.504 min): BM028463.D\data.ms (-1535) (-)




2-Methylnaphthalene
Concen: 36.43 ng
RT: 12.628 min Scan# 1605
Delta R.T. 0.218 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:	142	Resp:	158247
Ion	Ratio	Lower	Upper
142	100		
141	93.4	70.3	105.5
115	35.4	32.2	48.4






Abundance Scan 1621 (12.722 min): BM028463.D\data.ms (-1535) (-)



1-Methylnaphthalene
Concen: 38.40 ng
RT: 12.628 min Scan# 1605
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44



Tgt Ion:	142	Resp:	158247
Ion	Ratio	Lower	Upper
142	100		
141	93.4	73.8	110.6
116	3.7	3.5	5.3





Abundance Scan 1953 (14.674 min): BM028463.D\data.ms (-1539) (-)

Acenaphthene-d10
Concen: 20.00 ng
RT: 14.586 min Scan# 1938
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040





Tgt Ion	Ratio	Lower	Upper
164	100		
162	101.5	80.4	120.6
160	43.9	35.9	53.9





Abundance Scan 1646 (12.869 min): BM028463.D\data.ms (-1549) (-)

1,2,4,5-Tetrachlorobenzene
Concen: 37.41 ng
RT: 12.781 min Scan# 1631
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44




Tgt Ion	Ratio	Lower	Upper
216	100		
214	78.1	62.1	93.1
179	20.4	17.2	25.8
108	18.8	14.9	22.3






Abundance Scan 1642 (12.845 min): BM028463.D\data.ms (-1642) (-)

Hexachlorocyclopentadiene
Concen: 40.03 ng
RT: 12.751 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion	Ratio	Lower	Upper
237	100		
235	63.0	43.1	83.1
272	13.7	0.0	33.0






Abundance Scan 2205 (16.157 min): BM028463.D\data.ms (-2205) (-)


2,4,6-Tribromophenol
Concen: 78.89 ng
RT: 16.074 min Scan# 2191
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44



Tgt Ion	Ratio	Lower	Upper
330	100		
332	97.6	77.8	116.8
141	34.9	28.4	42.6





Abundance Scan 1687 (13.110 min): BM028463.D\data.ms (-1640) (-)




2,4,6-Trichlorophenol
Concen: 38.02 ng
RT: 13.028 min Scan# 1673
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:	196	Resp:	55260
Ion	Ratio	Lower	Upper
196	100		
198	95.8	74.9	112.3
200	31.4	23.7	35.5



Abundance Scan 1673 (13.028 min): BM028602.D\data.ms (-1622) (-)





Abundance Scan 1699 (13.180 min): BM028463.D\data.ms (-1644) (-)




2,4,5-Trichlorophenol
Concen: 38.12 ng
RT: 13.110 min Scan# 1687
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Tgt Ion:	196	Resp:	57344
Ion	Ratio	Lower	Upper
196	100		
198	96.8	79.0	118.6
97	49.0	40.5	60.7
132	27.7	19.3	28.9





Abundance Scan 1687 (13.110 min): BM028602.D\data.ms (-1636) (-)





Abundance Scan 1719 (13.298 min): BM028463.D\data.ms (-1745) (-)

2-Fluorobiphenyl
Concen: 74.42 ng
RT: 13.210 min Scan# 1704
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040





Tgt Ion	Ratio	Lower	Upper
172	100		
171	34.8	27.8	41.8
170	22.4	18.6	27.8





Abundance Scan 1755 (13.510 min): BM028463.D\data.ms (-1746) (-)

1,1'-Biphenyl
Concen: 37.21 ng
RT: 13.422 min Scan# 1740
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44



Tgt Ion	Ratio	Lower	Upper
154	100		
153	40.6	20.0	60.0
76	12.2	0.0	32.0



Abundance Scan 1763 (13.557 min): BM028463.D\data.ms (-1745) (-)

2-Chloronaphthalene


Concen: 36.82 ng

RT: 13.469 min Scan# 1748

Delta R.T. 0.000 min

Lab File: BM028602.D

Acq: 29 Jan 2021 12:44





Tgt Ion: 162 Resp: 168722

Ion Ratio Lower Upper


162 100

127 38.5 29.6 44.4

164 32.5 26.2 39.4



Abundance Scan 1748 (13.469 min): BM028602.D\data.ms (-1697) (-)



Abundance Scan 1797 (13.757 min): BM028463.D\data.ms (-1746) (-)

2-Nitroaniline


Concen: 36.36 ng

RT: 13.675 min Scan# 1783

Delta R.T. 0.000 min

Lab File: BM028602.D

Acq: 29 Jan 2021 12:44





Tgt Ion: 65 Resp: 52247

Ion Ratio Lower Upper


65 100

92 68.2 51.5 77.3


138 107.4 72.7 109.1



Abundance Scan 1783 (13.675 min): BM028602.D\data.ms (-1732) (-)






Abundance Scan 1906 (14.398 min): BM028463.D\data.ms (-1849) (-)




Acenaphthylene
Concen: 37.98 ng
RT: 14.316 min Scan# 1892
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:	152	Resp:	258252
Ion Ratio	Lower	Upper	
152	100		
151	19.6	15.8	23.6
153	12.8	10.4	15.6






Abundance Scan 1860 (14.127 min): BM028463.D\data.ms (-1850) (-)



Dimethylphthalate
Concen: 38.16 ng
RT: 14.051 min Scan# 1847
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44


Tgt Ion:	163	Resp:	206104
Ion Ratio	Lower	Upper	
163	100		
194	4.1	3.4	5.2
164	10.1	8.1	12.1






Abundance Scan 1881 (14.251 min): BM028463.D\data.ms (-1851) (-)

2,6-Dinitrotoluene
Concen: 38.54 ng
RT: 14.175 min Scan# 1868
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:	165	Resp:	44280
Ion	Ratio	Lower	Upper
165	100		
63	54.5	51.1	76.7
89	54.4	53.3	79.9






Abundance Scan 1964 (14.739 min): BM028463.D\data.ms (-1852) (-)

Acenaphthene
Concen: 37.43 ng
RT: 14.651 min Scan# 1949
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44




Tgt Ion:	154	Resp:	158004
Ion	Ratio	Lower	Upper
154	100		
153	112.0	88.3	132.5
152	51.0	41.4	62.2





Abundance Scan 1937 (14.580 min): BM028463.D\data.ms (-1550) (-)

3-Nitroaniline
Concen: 38.91 ng
RT: 14.504 min Scan# 1924
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44


Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:	138	Resp:	50405
Ion	Ratio	Lower	Upper
138	100		
108	9.7	8.6	13.0
92	110.6	102.1	153.1




Abundance Scan 1924 (14.504 min): BM028602.D\data.ms (-1873) (-)





Abundance Scan 1972 (14.786 min): BM028463.D\data.ms (-1550) (-)


2,4-Dinitrophenol
Concen: 46.85 ng
RT: 14.716 min Scan# 1960
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44



Tgt Ion:	184	Resp:	32062
Ion	Ratio	Lower	Upper
184	100		
63	64.0	60.2	90.2
154	59.8	52.9	79.3




Abundance Scan 1960 (14.716 min): BM028602.D\data.ms (-1909) (-)






Abundance Scan 2020 (15.068 min): BM028463.D\data.ms (-1955) (-)

Dibenzofuran
Concen: 37.23 ng
RT: 14.986 min Scan# 2006
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:	168	Resp:	241748
Ion	Ratio	Lower	Upper
168	100		
139	36.5	30.2	45.4
169	13.1	10.6	16.0






Abundance Scan 1987 (14.874 min): BM028463.D\data.ms (-1955) (-)

4-Nitrophenol
Concen: 37.56 ng
RT: 14.839 min Scan# 1981
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44




Tgt Ion:	139	Resp:	38310
Ion	Ratio	Lower	Upper
139	100		
109	69.7	77.0	117.0#
65	96.0	103.4	143.4#





Abundance Scan 2014 (15.033 min): BM028463.D\data.ms (-2053) (-)

2,4-Dinitrotoluene
Concen: 40.30 ng
RT: 14.963 min Scan# 2002
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44


Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:	165	Resp:	61688
Ion	Ratio	Lower	Upper
165	100		
63	41.5	38.8	58.2
89	66.5	62.6	94.0
182	3.5	2.3	3.5#





Abundance Scan 2002 (14.963 min): BM028602.D\data.ms (-1951) (-)




Abundance Scan 2130 (15.715 min): BM028463.D\data.ms (-2153) (-)

Fluorene
Concen: 37.33 ng
RT: 15.633 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44


Tgt Ion:	166	Resp:	198568
Ion	Ratio	Lower	Upper
166	100		
165	96.8	76.2	114.2
167	13.5	10.7	16.1



Abundance Scan 2116 (15.633 min): BM028602.D\data.ms (-2065) (-)






Abundance Scan 2058 (15.292 min): BM028463.D\data.ms (-2058) (-)




2,3,4,6-Tetrachlorophenol
Concen: 38.40 ng
RT: 15.216 min Scan# 2045
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion	Ratio	Resp	Lower	Upper
232	100	51058		
131	45.3	32.2	48.2	
130	2.2	1.8	2.8	
166	30.8	24.5	36.7	






Abundance Scan 2091 (15.486 min): BM028463.D\data.ms (-2091) (-)



Diethylphthalate
Concen: 38.57 ng
RT: 15.404 min Scan# 2077
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44


Tgt Ion	Ratio	Resp	Lower	Upper
149	100	211491		
177	21.7	18.2	27.4	
150	12.3	10.2	15.4	





Abundance Scan 2128 (15.704 min): BM028463.D\data.ms (-2#11) (-)

4-Chlorophenyl-phenylether
Concen: 37.75 ng
RT: 15.627 min Scan# 2115
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44


Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:	204	Resp:	98903
Ion	Ratio	Lower	Upper
204	100		
206	32.1	26.1	39.1
141	56.3	43.7	65.5




Abundance Scan 2115 (15.627 min): BM028602.D\data.ms (-2064) (-)





Abundance Scan 2133 (15.733 min): BM028463.D\data.ms (-2#12) (-)


4-Nitroaniline
Concen: 38.65 ng
RT: 15.663 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44




Tgt Ion:	138	Resp:	54502
Ion	Ratio	Lower	Upper
138	100		
92	49.7	35.5	75.5
108	72.3	80.5	120.5#



Abundance Scan 2121 (15.663 min): BM028602.D\data.ms (-2070) (-)





Abundance Scan 2178 (15.998 min): BM028463.D\data.ms (-2113) (-)




Azobenzene
Concen: 36.41 ng
RT: 15.916 min Scan# 2164
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:	77	Resp:	187414
Ion	Ratio	Lower	Upper
77	100		
182	24.9	5.8	45.8
105	19.5	0.0	37.2
51	32.7	3.6	43.6



Abundance Scan 2164 (15.916 min): BM028602.D\data.ms (-2113) (-)





Abundance Scan 2417 (17.404 min): BM028463.D\data.ms (-2403) (-)




Phenanthrene-d10
Concen: 20.00 ng
RT: 17.321 min Scan# 2403
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Tgt Ion:	188	Resp:	131731
Ion	Ratio	Lower	Upper
188	100		
94	9.9	6.2	9.4#
80	11.0	7.5	11.3




Abundance Scan 2403 (17.321 min): BM028602.D\data.ms (-2352) (-)






Abundance Scan 2144 (15.798 min): BM028463.D\data.ms (-2155) (-)

4,6-Dinitro-2-methylphenol
Concen: 39.98 ng
RT: 15.722 min Scan# 2131
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:	198	Resp:	32724
Ion	Ratio	Lower	Upper
198	100		
51	51.7	13.0	53.0
105	43.2	18.7	58.7






Abundance Scan 2165 (15.921 min): BM028463.D\data.ms (-2166) (-)

n-Nitrosodiphenylamine
Concen: 38.79 ng
RT: 15.845 min Scan# 2152
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44




Tgt Ion:	169	Resp:	171118
Ion	Ratio	Lower	Upper
169	100		
168	66.3	53.8	80.8
167	35.3	29.0	43.4






Abundance Scan 2279 (16.592 min): BM028463.D\data.ms (-2279) (-)

4-Bromophenyl-phenylether
Concen: 39.40 ng
RT: 16.516 min Scan# 2266
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:	248	Resp:	61468
Ion	Ratio	Lower	Upper
248	100		
250	96.6	79.1	118.7
141	72.2	51.8	77.8






Abundance Scan 2299 (16.710 min): BM028463.D\data.ms (-2299) (-)

Hexachlorobenzene
Concen: 38.52 ng
RT: 16.633 min Scan# 2286
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44




Tgt Ion:	284	Resp:	69837
Ion	Ratio	Lower	Upper
284	100		
142	32.5	26.1	39.1
249	29.3	25.6	38.4





Abundance Scan 2324 (16.857 min): BM028463.D\data.ms (-2369) (-)

Atrazine
Concen: 38.30 ng
RT: 16.786 min Scan# 2312
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44


Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion	Ratio	Lower	Upper
200	100		
173	27.4	4.9	44.9
215	48.4	29.0	69.0




Abundance Scan 2312 (16.786 min): BM028602.D\data.ms (-2261) (-)





Abundance Scan 2357 (17.051 min): BM028463.D\data.ms (-2370) (-)


Pentachlorophenol
Concen: 43.76 ng
RT: 16.980 min Scan# 2345
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44




Tgt Ion	Ratio	Lower	Upper
266	100		
268	64.0	49.9	74.9
264	64.5	50.6	76.0



Abundance Scan 2345 (16.980 min): BM028602.D\data.ms (-2294) (-)






Abundance Scan 2424 (17.445 min): BM028463.D\data.ms (-2470) (-)




Phenanthrene
Concen: 37.31 ng
RT: 17.363 min Scan# 2410
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:	178	Resp:	302228
Ion	Ratio	Lower	Upper
178	100		
176	18.3	15.0	22.6
179	14.9	12.2	18.4






Abundance Scan 2439 (17.533 min): BM028463.D\data.ms (-2472) (-)



Anthracene
Concen: 38.01 ng
RT: 17.457 min Scan# 2426
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44


Tgt Ion:	178	Resp:	297407
Ion	Ratio	Lower	Upper
178	100		
176	17.8	15.0	22.4
179	15.4	12.5	18.7






Abundance Scan 2484 (17.798 min): BM028463.D\data.ms (-2473) (-)

Carbazole
Concen: 37.64 ng
RT: 17.727 min Scan# 2472
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion	Ratio	Lower	Upper
167	100		
166	21.0	16.9	25.3
139	12.5	9.9	14.9






Abundance Scan 2577 (18.345 min): BM028463.D\data.ms (-2574) (-)


Di-n-butylphthalate
Concen: 41.77 ng
RT: 18.274 min Scan# 2565
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44



Tgt Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.4	11.0
104	4.8	4.7	7.1






Abundance Scan 2762 (19.433 min): BM028463.D\data.ms (-2755) (-)




Fluoranthene
Concen: 37.30 ng
RT: 19.368 min Scan# 2751
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:	202	Resp:	339412
Ion Ratio	Lower	Upper	
202	100		
101	11.4	0.0	28.5
203	17.2	0.0	37.3







Abundance Scan 3119 (21.533 min): BM028463.D\data.ms (-3170) (-)



Chrysene-d12
Concen: 20.00 ng
RT: 21.468 min Scan# 3108
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Tgt Ion:	240	Resp:	126977
Ion Ratio	Lower	Upper	
240	100		
120	11.4	6.4	9.6#
236	24.0	20.1	30.1









Benzidine
Concen: 41.31 ng
RT: 19.551 min Scan# 2782
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:	184	Resp:	117961
Ion	Ratio	Lower	Upper
184	100		
185	13.5	11.0	16.6
183	11.2	9.7	14.5




Pyrene
Concen: 38.13 ng
RT: 19.727 min Scan# 2812
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Tgt Ion:	202	Resp:	347177
Ion	Ratio	Lower	Upper
202	100		
200	20.3	16.6	24.8
203	17.6	13.8	20.6





Abundance Scan 2857 (19.992 min): BM028463.D\data.ms (-2857) (-)




Terphenyl-d14
Concen: 77.06 ng
RT: 19.921 min Scan# 2845
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
244	100		
212	6.8	5.8	8.8
122	12.3	6.5	9.7#



Abundance Scan 2845 (19.921 min): BM028602.D\data.ms (-2794) (-)





Abundance Scan 2972 (20.668 min): BM028463.D\data.ms (-2966) (-)




Butylbenzylphthalate
Concen: 42.34 ng
RT: 20.609 min Scan# 2962
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44


Tgt Ion	Ratio	Lower	Upper
149	100		
91	68.8	58.1	87.1
206	21.1	18.7	28.1



Abundance Scan 2962 (20.609 min): BM028602.D\data.ms (-2911) (-)






Abundance Scan 3116 (21.515 min): BM028463.D\data.ms (-3116) (-)




Benzo(a)anthracene
Concen: 37.53 ng
RT: 21.451 min Scan# 3105
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:	228	Resp:	329091
Ion Ratio	Lower	Upper	
228	100		
226	26.4	21.1	31.7
229	19.8	15.7	23.5






Abundance Scan 3104 (21.445 min): BM028463.D\data.ms (-3104) (-)




3,3'-Dichlorobenzidine
Concen: 40.99 ng
RT: 21.386 min Scan# 3094
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Tgt Ion:	252	Resp:	115512
Ion Ratio	Lower	Upper	
252	100		
254	64.9	51.4	77.0
126	11.9	6.8	10.2#






Abundance Scan 3125 (21.568 min): BM028463.D\data.ms (-3125) (-)




Chrysene
Concen: 37.29 ng
RT: 21.503 min Scan# 3114
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:	228	Resp:	315671
Ion	Ratio	Lower	Upper
228	100		
226	29.2	23.4	35.0
229	19.1	15.7	23.5






Abundance Scan 3101 (21.427 min): BM028463.D\data.ms (-3101) (-)




Bis(2-ethylhexyl)phthalate
Concen: 44.52 ng
RT: 21.368 min Scan# 3091
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Tgt Ion:	149	Resp:	252696
Ion	Ratio	Lower	Upper
149	100		
167	26.6	22.7	34.1
279	4.0	4.0	6.0#






Abundance Scan 3252 (22.315 min): BM028463.D\data.ms (-3245) (-)




Di-n-octyl phthalate
Concen: 44.91 ng
RT: 22.250 min Scan# 3241
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:149 Resp: 430936
Ion Ratio Lower Upper
149 100
167 1.6 1.3 1.9
43 10.6 5.8 8.6#






Abundance Scan 3507 (23.815 min): BM028463.D\data.ms (-3486) (-)




Perylene-d12
Concen: 20.00 ng
RT: 23.727 min Scan# 3492
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Tgt Ion:264 Resp: 125773
Ion Ratio Lower Upper
264 100
260 22.4 19.0 28.4
265 21.5 17.7 26.5






Abundance Scan 3905 (26.156 min): BM028463.D\data.ms (-3887) (-)




Indeno(1,2,3-cd)pyrene
Concen: 39.39 ng
RT: 26.033 min Scan# 3884
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:	276	Resp:	369105
Ion	Ratio	Lower	Upper
276	100		
138	28.4	14.2	21.4#
277	25.2	19.8	29.8






Abundance Scan 3390 (23.127 min): BM028463.D\data.ms (-3886) (-)




Benzo(b)fluoranthene
Concen: 38.64 ng
RT: 23.050 min Scan# 3377
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Tgt Ion:	252	Resp:	329947
Ion	Ratio	Lower	Upper
252	100		
253	21.3	17.4	26.0
125	10.9	5.7	8.5#






Abundance Scan 3398 (23.174 min): BM028463.D\data.ms (-3398) (-)




Benzo(k)fluoranthene
Concen: 38.39 ng
RT: 23.092 min Scan# 3384
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.4	26.0
125	12.0	5.6	8.4#






Abundance Scan 3491 (23.721 min): BM028463.D\data.ms (-3491) (-)




Benzo(a)pyrene
Concen: 38.76 ng
RT: 23.633 min Scan# 3476
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.4	26.2
125	12.6	6.6	9.8#





Abundance Scan 3906 (26.162 min): BM028463.D\data.ms (-3893) (-)




Dibenzo(a,h)anthracene
Concen: 39.63 ng
RT: 26.038 min Scan# 3885
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:	278	Resp:	307260
Ion	Ratio	Lower	Upper
278	100		
139	17.8	9.0	13.6#
279	23.7	19.0	28.4



Abundance Scan 3885 (26.038 min): BM028602.D\data.ms (-3834) (-)





Abundance Scan 4027 (26.874 min): BM028463.D\data.ms (-4092) (-)



Benzo(g,h,i)perylene
Concen: 38.97 ng
RT: 26.733 min Scan# 4003
Delta R.T. 0.000 min
Lab File: BM028602.D
Acq: 29 Jan 2021 12:44

Tgt Ion:	276	Resp:	297198
Ion	Ratio	Lower	Upper
276	100		
277	23.5	18.6	28.0
138	24.6	12.2	18.2#



Abundance Scan 4003 (26.733 min): BM028602.D\data.ms (-3952) (-)

