


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM013015\
 Data File : BM000403.D
 Acq On : 30 Jan 2015 15:54
 Operator : TP/IZ
 Sample : DFTPP14
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP14

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-BM012915.M
 Title : SVOA CALIBRATION
 Last Update : Fri Jan 30 16:03:45 2015



AutoFind: Scans 2533, 2534, 2535; Background Corrected with Scan 2527

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.1	68928	PASS
68	69	0.00	2	1.8	1109	PASS
69	198	0.00	100	36.5	61245	PASS
70	69	0.00	2	0.3	184	PASS
127	198	10	80	46.1	77325	PASS
197	198	0.00	2	0.4	635	PASS
198	198	100	100	100.0	167672	PASS
199	198	5	9	6.4	10814	PASS
275	198	10	60	27.2	45658	PASS
365	198	1	100	4.5	7484	PASS
441	443	0.01	100	81.6	26010	PASS
442	198	50	100	98.5	165226	PASS
443	442	15	24	19.3	31861	PASS

m/z	Abundance
36.00	789.0
38.00	1222.0
39.10	5945.0
40.00	344064.0
44.00	3076.0
49.00	652.0
50.10	20520.0
51.10	86944.0
52.10	3645.0
53.00	498.0
55.00	953.0
56.00	2965.0
57.10	6545.0
61.00	1558.0
62.00	1684.0
63.00	3059.0
63.90	405.0
65.00	1582.0
67.00	362.0
68.10	983.0
69.00	74648.0
73.20	1725.0
74.10	7817.0
75.10	14018.0
76.10	3559.0
77.10	81280.0
78.00	4693.0
78.90	6046.0
80.10	4779.0
81.00	6460.0
82.00	1355.0
82.90	1657.0
85.00	1317.0
86.00	2782.0
87.10	575.0
87.90	393.0
89.10	330.0
91.10	1856.0
92.00	2419.0
93.00	10100.0
94.10	554.0
96.00	735.0
97.20	301.0
98.00	10219.0
99.00	7324.0
99.80	496.0
100.90	3925.0
103.00	1179.0
104.00	2723.0
104.90	2252.0
106.00	889.0
107.00	32216.0
108.10	3511.0
109.00	1263.0
110.00	46984.0
111.00	7021.0
112.00	756.0
113.00	618.0
116.10	1444.0
117.00	29192.0
118.00	2844.0
118.90	552.0
119.90	351.0
122.00	2032.0
122.90	2325.0
124.10	1365.0
125.10	1416.0
127.00	91008.0
128.00	6869.0
129.00	38880.0
130.00	3762.0
131.10	392.0
131.80	391.0
133.10	310.0
134.00	1305.0
135.00	4449.0
135.90	1811.0
137.00	1815.0
137.90	645.0
141.00	5627.0
142.00	1441.0
143.10	1077.0
144.20	501.0
145.00	401.0
145.90	1044.0
147.00	3541.0
148.00	7209.0
149.10	1105.0
151.20	943.0
152.20	1009.0
152.90	1602.0

Instrument :
BNA_M
ClientSampleId :
DFTPP14

153.90	1244.0
155.00	2911.0
156.10	4930.0
157.10	1303.0
158.10	1251.0
158.80	489.0
160.10	2162.0
161.00	2512.0
161.90	784.0
165.00	2224.0
165.90	1912.0
167.10	10713.0
168.00	5236.0
168.90	977.0
169.80	390.0
170.70	429.0
172.00	1380.0
173.00	1123.0
174.20	1843.0
175.00	4069.0
176.00	1425.0
176.90	1842.0
177.80	1294.0
179.00	8623.0
180.10	5566.0
181.00	3268.0
183.80	485.0
185.00	4582.0
186.00	29088.0
187.00	8029.0
188.00	1297.0
189.00	2406.0
189.90	515.0
191.00	1859.0
191.90	2317.0
193.00	3628.0
193.80	397.0
194.20	397.0
196.00	4758.0
197.00	845.0
198.00	184832.0
199.00	13788.0
200.00	1064.0
201.00	465.0
201.60	647.0
203.00	2369.0
204.10	7815.0
205.10	13486.0
206.10	54976.0
207.10	9179.0
208.00	2508.0
208.90	1095.0
210.00	1044.0
211.00	1911.0
212.10	330.0
214.90	755.0
215.90	972.0
217.00	15151.0
218.00	1486.0
218.80	316.0
219.70	352.0
220.50	897.0
221.20	4262.0
222.10	1586.0
223.00	3098.0
224.00	29728.0
225.10	7833.0
226.00	1206.0
227.10	15631.0
228.00	2489.0
229.00	1930.0
229.90	469.0
231.10	914.0
234.10	910.0
235.00	1443.0
236.00	874.0
237.10	1851.0
239.00	485.0
239.90	647.0
241.00	964.0
242.00	1427.0
242.90	1495.0
244.10	18184.0
245.00	2950.0
246.00	5752.0
246.90	1402.0
247.80	319.0
248.90	487.0
252.10	391.0
253.10	591.0
253.80	1016.0
255.10	106648.0
256.00	16920.0
256.90	1198.0
258.00	8770.0

Instrument :
BNA_M
ClientSampleId :
DFTPP14

258.90	1326.0
263.90	402.0
265.00	2569.0
266.10	438.0
271.80	490.0
273.00	3349.0
274.00	7906.0
275.00	49448.0
276.00	4926.0
277.10	4422.0
277.90	1042.0
282.10	621.0
282.80	578.0
284.10	355.0
284.90	875.0
285.90	320.0
288.90	335.0
293.10	960.0
296.00	12348.0
297.00	2913.0
303.10	1380.0
304.00	522.0
310.20	568.0
314.00	696.0
315.00	2186.0
316.20	751.0
317.00	307.0
321.20	592.0
323.10	4312.0
324.00	509.0
326.90	685.0
328.10	649.0
331.90	372.0
332.90	572.0
334.00	3211.0
334.90	999.0
340.90	412.0
346.00	871.0
347.10	415.0
352.00	1701.0
353.00	1070.0
354.10	1226.0
355.10	604.0
365.00	6038.0
365.90	911.0
371.00	458.0
372.00	2198.0
372.90	600.0
383.00	408.0
384.00	499.0
401.90	1140.0
403.10	1216.0
403.90	433.0
421.00	790.0
421.90	1399.0
423.10	7478.0
424.00	2912.0
424.80	386.0
441.00	19648.0
442.00	142016.0
443.00	30120.0
444.00	2803.0

Instrument :
BNA_M
ClientSampleId :
DFTPP14

m/z	Abundance
36.00	1138.0
37.00	455.0
38.20	1049.0
39.00	5370.0
40.00	348736.0
43.90	3909.0
45.00	331.0
48.90	461.0
50.10	18304.0
51.10	80408.0
52.10	3901.0
55.00	646.0
56.00	1911.0
57.00	5000.0
60.90	885.0
62.00	1260.0
63.00	2547.0
64.10	686.0
65.00	1838.0
68.10	1546.0
69.00	71416.0
70.10	554.0
73.10	1425.0
74.10	6577.0
75.00	13360.0
76.10	3877.0
77.10	77512.0
78.00	4319.0
79.00	6602.0
80.10	4742.0
81.00	5442.0
82.00	846.0
82.90	1436.0
85.00	1278.0
86.10	2103.0
87.00	1607.0
87.90	428.0
90.90	1562.0
92.10	2225.0
93.00	11955.0
94.00	553.0
96.00	537.0
98.00	8076.0
99.00	7444.0
100.10	579.0
101.00	3396.0
103.00	1336.0
104.10	2211.0
104.90	2985.0
105.90	953.0
107.00	31024.0
108.00	4024.0
108.90	701.0
110.00	45800.0
111.00	7734.0
112.00	1033.0
113.00	621.0
116.00	1917.0
117.00	26136.0
118.00	2169.0
119.00	548.0
122.00	1996.0
123.00	3555.0
124.10	1444.0
125.00	2256.0
126.20	741.0
127.00	89776.0
128.00	5992.0
129.00	41672.0
130.00	3398.0
131.00	595.0
132.90	497.0
133.90	1374.0
135.00	3383.0
136.10	1616.0
137.00	1956.0
137.90	581.0
139.90	711.0
140.90	5670.0
142.10	1757.0
143.00	1469.0
145.10	541.0
146.10	1017.0
147.00	4061.0
148.00	8614.0
149.00	1778.0
151.10	620.0
153.00	1878.0
154.20	1186.0
155.10	3336.0
156.10	6032.0

Instrument :
BNA_M
ClientSampleId :
DFTPP14

157.00	1030.0
158.10	906.0
158.80	723.0
159.90	1431.0
161.00	2392.0
161.90	803.0
165.00	1802.0
166.10	1582.0
167.00	10740.0
168.00	5639.0
168.90	1444.0
170.10	347.0
172.00	1307.0
172.90	1676.0
174.00	3305.0
175.10	4359.0
176.00	900.0
177.00	2820.0
178.10	889.0
179.00	9528.0
180.00	6645.0
181.00	2862.0
182.10	338.0
183.90	959.0
185.00	4335.0
186.00	33480.0
187.10	9931.0
188.10	704.0
189.00	2398.0
189.90	509.0
191.10	1098.0
192.00	2553.0
192.90	3119.0
194.20	594.0
195.00	338.0
196.00	4037.0
196.80	1060.0
198.00	198208.0
199.00	11921.0
200.00	1310.0
201.50	910.0
203.00	1842.0
204.10	10127.0
205.00	12981.0
206.10	57376.0
207.10	10974.0
208.10	3411.0
209.10	1305.0
209.90	960.0
211.00	2224.0
215.00	450.0
215.90	1207.0
217.00	16648.0
218.00	2157.0
221.10	4463.0
221.80	1972.0
223.00	4139.0
224.10	30672.0
225.10	8077.0
226.10	1395.0
227.00	15265.0
227.90	2217.0
229.10	2601.0
229.90	469.0
230.90	915.0
233.00	397.0
234.10	1232.0
235.10	854.0
236.20	794.0
237.10	2585.0
238.90	672.0
240.00	375.0
241.10	983.0
242.10	1446.0
243.00	1607.0
244.10	22392.0
245.00	3317.0
246.00	5461.0
247.10	1267.0
248.90	1000.0
250.90	383.0
252.10	513.0
253.10	638.0
254.10	1244.0
255.00	131072.0
256.10	20816.0
257.10	1397.0
258.00	8481.0
259.10	1122.0
260.00	687.0
261.00	343.0
265.00	3955.0
266.00	409.0
271.90	482.0
273.00	3261.0

Instrument :
BNA_M
ClientSampleId :
DFTPP14

274.00	9682.0
275.00	54272.0
276.00	6813.0
277.10	4491.0
277.90	1507.0
280.80	513.0
282.20	376.0
282.90	449.0
285.00	1063.0
289.10	647.0
290.10	340.0
291.90	370.0
293.10	1296.0
293.80	385.0
295.20	578.0
296.00	17424.0
297.00	2215.0
298.00	323.0
301.90	598.0
303.00	2209.0
304.20	834.0
307.90	683.0
314.00	1387.0
315.00	2324.0
316.00	936.0
317.00	321.0
321.10	685.0
321.90	313.0
323.00	5275.0
323.90	778.0
327.10	1240.0
327.80	532.0
331.90	620.0
333.00	653.0
334.00	3574.0
335.00	744.0
341.00	562.0
341.90	432.0
345.90	1601.0
352.00	1464.0
353.10	768.0
354.00	1864.0
355.00	516.0
365.00	9190.0
365.90	1739.0
371.00	552.0
372.10	3578.0
373.00	972.0
383.00	651.0
392.00	619.0
400.80	455.0
402.00	819.0
403.00	1903.0
403.90	552.0
421.00	1528.0
422.00	1624.0
423.00	10252.0
424.00	4046.0
424.90	309.0
441.00	32680.0
442.00	196544.0
443.00	37136.0
444.10	3800.0

Instrument :
BNA_M
ClientSampleId :
DFTPP14

m/z	Abundance
36.00	1184.0
38.00	659.0
39.10	2964.0
40.00	357056.0
44.00	2988.0
45.00	300.0
48.90	380.0
50.10	9354.0
51.10	39432.0
52.00	1578.0
54.90	438.0
56.00	1047.0
57.10	2292.0
60.90	465.0
61.90	612.0
63.10	2535.0
65.20	980.0
68.00	798.0
69.00	37672.0
72.90	1154.0
74.10	3967.0
75.10	6114.0
76.00	2372.0
77.10	39584.0
78.00	3230.0
79.00	2912.0
79.90	2368.0
81.10	3595.0
81.90	611.0
83.00	655.0
83.80	523.0
84.90	954.0
85.90	1310.0
86.90	755.0
91.20	702.0
92.10	890.0
93.00	7144.0
94.10	493.0
95.90	436.0
97.10	502.0
98.00	5376.0
99.00	3878.0
101.00	1578.0
102.90	528.0
103.90	1430.0
104.90	1112.0
106.20	345.0
107.00	14805.0
108.00	2292.0
109.00	783.0
110.00	23872.0
111.00	3611.0
112.10	618.0
115.90	883.0
117.00	15726.0
118.00	1245.0
122.10	786.0
123.00	1673.0
123.90	934.0
125.00	725.0
127.00	51192.0
128.00	3431.0
129.00	23968.0
129.90	1640.0
131.00	542.0
132.10	387.0
133.90	716.0
135.00	2409.0
136.00	964.0
137.10	1196.0
137.90	404.0
140.90	4108.0
142.10	982.0
142.90	567.0
146.00	427.0
147.00	1911.0
148.00	4105.0
149.00	776.0
150.00	425.0
151.00	611.0
152.90	1013.0
153.80	532.0
155.00	1682.0
156.10	2867.0
157.30	399.0
157.90	522.0
158.70	300.0
159.20	475.0
159.90	1315.0
161.00	2180.0
161.90	556.0

Instrument :
BNA_M
ClientSampleId :
DFTPP14

162.90	378.0
165.00	1166.0
166.00	971.0
167.00	6783.0
168.00	3253.0
168.90	437.0
172.10	862.0
173.00	800.0
174.00	1648.0
175.10	2523.0
176.00	462.0
177.00	2810.0
178.10	906.0
179.00	4712.0
180.10	3431.0
181.10	1380.0
181.90	370.0
184.10	377.0
185.00	3227.0
186.00	19248.0
187.10	5192.0
187.90	553.0
189.10	1563.0
191.00	916.0
192.00	2136.0
193.10	1488.0
194.20	394.0
195.00	548.0
196.00	2940.0
197.10	1232.0
198.00	119976.0
199.00	6735.0
199.90	1219.0
201.30	660.0
202.90	741.0
204.10	6170.0
205.00	9695.0
206.00	35136.0
207.10	6910.0
208.10	2467.0
208.90	494.0
209.20	520.0
209.80	940.0
211.10	1535.0
215.20	792.0
216.10	1211.0
217.00	12206.0
218.10	921.0
221.00	3395.0
222.00	689.0
222.90	2077.0
224.00	20600.0
225.10	5618.0
225.90	1258.0
227.00	9916.0
228.00	1301.0
229.00	2657.0
230.00	399.0
231.10	731.0
234.00	866.0
235.10	438.0
237.00	923.0
239.00	449.0
240.80	352.0
241.90	1248.0
243.00	1115.0
244.10	12933.0
245.00	2095.0
246.00	3613.0
247.20	854.0
248.90	538.0
250.80	340.0
253.10	443.0
254.00	765.0
255.00	83888.0
256.00	11493.0
257.10	664.0
258.00	5483.0
259.00	615.0
264.90	3071.0
266.10	333.0
273.00	2874.0
274.00	7356.0
275.00	33256.0
276.00	5163.0
277.10	3493.0
278.10	366.0
280.80	413.0
281.10	391.0
283.00	619.0
284.10	366.0
285.00	730.0
289.90	339.0
293.00	694.0
296.00	11071.0

Instrument :
BNA_M
ClientSampleId :
DFTPP14

296.90	1973.0
303.10	1939.0
304.00	406.0
314.10	887.0
314.90	1535.0
316.10	767.0
317.10	379.0
323.00	4261.0
324.00	903.0
326.90	1018.0
331.90	807.0
334.10	2059.0
340.80	666.0
345.90	806.0
352.00	1390.0
353.00	725.0
354.10	1110.0
354.90	768.0
365.00	7225.0
365.90	1191.0
372.00	2495.0
373.20	590.0
383.00	372.0
390.10	560.0
402.10	1062.0
403.00	892.0
404.10	389.0
420.90	692.0
422.00	1309.0
423.00	8796.0
424.10	2352.0
425.00	842.0
441.00	25704.0
442.00	157120.0
443.10	28328.0
443.90	2450.0

Instrument :
BNA_M
ClientSampleId :
DFTPP14

m/z	Abundance
36.00	1577.0
38.00	479.0
39.20	500.0
40.00	374976.0
44.00	3732.0
45.10	338.0
73.10	375.0
147.00	353.0
166.20	329.0
167.00	907.0
191.10	424.0
207.00	3132.0
207.90	610.0
208.20	596.0
209.00	572.0
280.90	519.0

Instrument :

BNA_M

ClientSampleId :

DFTPP14