

Data Path : Z:\HPCHEM1\BNA M\DATA\BM013016\
 Data File : BM004147.D
 Acq On : 29 Jan 2016 15:37
 Operator : SJ/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Jan 29 22:50:54 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM012216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 22 14:46:41 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.547	0.472	13.7	90	0.00
3	n-Nitrosodimethylamine	0.446	0.400	10.3	86	0.00
4 S	2-Fluorophenol	1.043	0.901	13.6	84	0.00
5 S	Phenol-d6	1.259	1.101	12.5	86	0.00
6	bis(2-Chloroethyl)ether	1.167	1.068	8.5	89	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
8 S	Nitrobenzene-d5	0.219	0.195	11.0	91	0.00
9	Nitrobenzene	0.250	0.221	11.6	91	0.00
10	Naphthalene	1.165	1.076	7.6	92	0.00
11	Hexachlorobutadiene	0.180	0.163	9.4	91	0.00
12	2-Methylnaphthalene	0.762	0.710	6.8	94	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.145	0.123	15.2	95	0.00
15 S	2-Fluorobiphenyl	1.532	1.381	9.9	95	0.00
16	Acenaphthylene	2.158	1.914	11.3	96	0.00
17	Acenaphthene	1.634	1.434	12.2	96	0.00
18	Fluorene	1.987	1.837	7.5	101	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
20	4-Bromophenyl-phenylether	0.163	0.146	10.4	95	0.00
21	Hexachlorobenzene	0.171	0.151	11.7	97	0.00
22	Pentachlorophenol	0.043	0.031	27.9#	101	0.00
23	Phenanthrene	1.074	0.962	10.4	101	0.00
24	Anthracene	1.142	0.994	13.0	97	0.00
25	Fluoranthene	1.355	1.193	12.0	99	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	85	0.00
27	Benzydine	0.241	0.230	4.6	92	0.02
28	Pvrene	1.423	1.468	-3.2	97	0.00
29 S	Terphenyl-d14	0.643	0.658	-2.3	98	0.00
30	Benzo(a)anthracene	1.518	1.340	11.7	84	0.00
31	3,3'-Dichlorobenzidine	0.338	0.313	7.4	96	0.00
32	Chrysene	1.372	1.651	-20.3	116	0.00
33	Bis(2-ethylhexyl)phthalate	0.595	0.537	9.7	109	0.00
34	Indeno(1,2,3-cd)pyrene	1.463	1.351	7.7	89	0.01
35 I	Perylene-d12	1.000	1.000	0.0	92	0.01
36	Benzo(b)fluoranthene	1.551	1.356	12.6	87	0.00
37	Benzo(k)fluoranthene	1.404	1.363	2.9	101	0.00
38 C	Benzo(a)pyrene	1.366	1.237	9.4	94	0.01
39	Dibenzo(a,h)anthracene	1.227	1.055	14.0	88	0.01
40	Benzo(a,h,i)perylene	1.339	1.151	14.0	89	0.01

Data Path : Z:\HPCHEM1\BNA M\DATA\BM013016\
Data File : BM004147.D
Acq On : 29 Jan 2016 15:37
Operator : SJ/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Jan 29 22:50:54 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM012216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 22 14:46:41 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				