

Data Path : Z:\HPCHEM1\BNA M\DATA\BM013016\
 Data File : BM004156.D
 Acq On : 29 Jan 2016 21:01
 Operator : SJ/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Jan 29 22:16:02 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM012216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 22 14:46:41 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	74	0.00
2	1,4-Dioxane	1.000	0.966	3.4	75	0.00
3	n-Nitrosodimethylamine	1.000	0.939	6.1	76	0.00
4 S	2-Fluorophenol	1.000	0.976	2.4	80	0.00
5 S	Phenol-d6	1.000	1.000	0.0	83	0.00
6	bis(2-Chloroethyl)ether	1.000	0.918	8.2	75	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	77	0.00
8 S	Nitrobenzene-d5	1.000	0.978	2.2	84	0.00
9	Nitrobenzene	1.000	0.953	4.7	83	0.00
10	Naphthalene	1.000	0.910	9.0	76	0.00
11	Hexachlorobutadiene	1.000	0.916	8.4	77	0.00
12	2-Methylnaphthalene	1.000	0.938	6.2	79	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	82	0.00
14 S	2,4,6-Tribromophenol	1.000	1.031	-3.1	98	0.00
15 S	2-Fluorobiphenyl	1.000	0.900	10.0	80	0.00
16	Acenaphthylene	1.000	0.919	8.1	83	0.00
17	Acenaphthene	1.000	0.873	12.7	80	0.00
18	Fluorene	1.000	0.939	6.1	86	0.00
19 I	Phenanthrene-d10	5.000	5.000	0.0	88	0.00
20	4-Bromophenyl-phenylether	1.000	0.894	10.6	81	0.00
21	Hexachlorobenzene	1.000	0.878	12.2	82	0.00
22	Pentachlorophenol	1.000	1.201	-20.1	143	0.00
23	Phenanthrene	1.000	0.882	11.8	85	0.00
24	Anthracene	1.000	0.922	7.8	87	0.00
25	Fluoranthene	1.000	0.965	3.5	92	0.00
26 I	Chrysene-d12	5.000	5.000	0.0	83	0.00
27	Benzydine	1.000	1.097	-9.7	94	0.02
28	Pvrene	1.000	0.997	0.3	90	0.00
29 S	Terphenyl-d14	1.000	0.995	0.5	91	0.00
30	Benzo(a)anthracene	1.000	0.890	11.0	83	0.00
31	3,3'-Dichlorobenzidine	1.000	1.350	-35.0#	126	0.00
32	Chrysene	1.000	1.169	-16.9	107	0.00
33	Bis(2-ethylhexyl)phthalate	1.000	1.503	-50.3#	151	0.00
34	Indeno(1,2,3-cd)pyrene	1.000	1.050	-5.0	99	0.01
35 I	Perylene-d12	5.000	5.000	0.0	96	0.00
36	Benzo(b)fluoranthene	1.000	0.875	12.5	90	0.00
37	Benzo(k)fluoranthene	1.000	0.984	1.6	107	0.00
38 C	Benzo(a)pyrene	1.000	0.949	5.1	102	0.00
39	Dibenzo(a,h)anthracene	1.000	0.930	7.0	99	0.00
40	Benzo(a,h,i)perylene	1.000	0.913	8.7	98	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM013016\
Data File : BM004156.D
Acq On : 29 Jan 2016 21:01
Operator : SJ/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Jan 29 22:16:02 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM012216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 22 14:46:41 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				